

INTERNATIONAL STANDARD

ISO
23539
CIE S 010/E

First edition
2005-08-01

Photometry — The CIE system of physical photometry

Photométrie — Le système CIE de photométrie physique



Reference number
ISO 23539:2005(E)
CIE S 010/E:2004

© ISO 2005

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

ISO 23539 was prepared as Standard CIE S 010/E by the International Commission on Illumination, which has been recognized by the ISO Council as an international standardizing body. It was adopted by ISO under a special procedure which requires approval by at least 75 % of the member bodies casting a vote, and is published as a joint ISO/CIE edition.

The International Commission on Illumination (abbreviated as CIE from its French title) is an organization devoted to international cooperation and exchange of information among its member countries on all matters relating to the science and art of lighting.

ISO 23539 was prepared by CIE Technical Committee 2-35 *CIE Standard for $V(\lambda)$ and $V'(\lambda)$* .



CIE S 010/E:2004

Standard

PHOTOMETRY - THE CIE SYSTEM OF PHYSICAL PHOTOMETRY

Photométrie – Le système CIE de photométrie physique

Photometrie - Das CIE-System der physikalischen Photometrie

CIE Standards are copyrighted and shall not be reproduced in any form, entirely or partly, without the explicit agreement of the CIE.

CIE Central Bureau, Vienna
Kegelgasse 27, A-1030 Vienna, Austria

CIE S 010/E:2004

UDC: 635.24

Descriptor:

Photometry

This document is a preview generated by EVS

© CIE 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from CIE Central Bureau at the address below.

CIE Central Bureau
Kegelgasse 27
A-1030 Vienna
Austria
Tel.: +43 1 714 3187 0
Fax: +43 1 714 3187 18
e-mail: ciecb@ping.at
Web: www.cie.co.at

FOREWORD

Standards produced by the Commission Internationale de l'Eclairage (CIE) are a concise documentation of data defining aspects of light and lighting, for which international harmony requires such unique definition. CIE Standards are therefore a primary source of internationally accepted and agreed data, which can be taken, essentially unaltered, into universal standard systems.

This International Standard has been prepared by CIE Technical Committee 2-35^{*)}, "CIE Standard for $V(\lambda)$ and $V'(\lambda)$ ", and was approved by the National Committees of the CIE.

TABLE OF CONTENTS

FOREWORD	vii
INTRODUCTION	1
1. SCOPE	1
2. PHOTOMETRIC QUANTITIES	1
2.1 Luminous flux	1
2.2 Other quantities	1
3. PHOTOMETRIC UNITS	2
3.1 Candela	2
3.2 Other units	2
4. PHOTOMETRIC STANDARDS	3
4.1 CIE Standard spectral luminous efficiency functions for photopic and scotopic vision	3
4.2 Maximum luminous efficacies for photopic and scotopic vision	3
4.3 Fundamental equations relating photometric and radiometric quantities	4
4.4 Measurement procedures	4
TABLES	5
Table 1. Definitive values of the spectral luminous efficiency function for photopic vision $V(\lambda)$	5
Table 2. Definitive values of the spectral luminous efficiency function for scotopic vision $V'(\lambda)$	10
ANNEX A (INFORMATIVE)	14
Vocabulary of related terms	14
ANNEX B (INFORMATIVE)	16
Background to the CIE system of physical photometry	16
B.1 Evolution of the photometric base unit	16
B.2 Spectral response of the human eye	16
B.3 Supplementary remarks	17
ANNEX C (INFORMATIVE)	18
Bibliography	18

© CIE 2004

^{*)} Chairman of this TC was K. D. Mielenz (US), members were: J. Bastie (FR), J. L. Gardner (AU), F. Hengstberger (ZA), J. R. Moore (GB), Y. Ohno (US), A. C. Parr (US), A. R. Robertson (CA), G. Sauter (DE), J. Schanda (HU).

PHOTOMETRY - THE CIE SYSTEM OF PHYSICAL PHOTOMETRY

INTRODUCTION

The visual brightness of a light source depends not only on the amount of radiation it emits but also on its spectral composition and on the visual response function of the observer viewing it. Because human visual response varies at different light levels and from person to person, precise photometry requires the definition of representative standard observers. The CIE system of physical photometry specifies procedures for the quantitative evaluation of optical radiation in terms of the spectral luminous efficiency functions of two such standard observers. One, $V(\lambda)$, represents photopic vision and the other, $V'(\lambda)$, scotopic vision. Used in conjunction with the SI photometric base unit, the candela, these functions constitute a system that enables the values of photometric quantities for all types of luminous source to be precisely determined, regardless of the spectral composition of the radiation emitted.

1. SCOPE

This international Standard specifies the characteristics of the system of physical photometry established by the CIE and accepted as the basis for the measurement of light. It defines the photometric quantities, units and standards that make up the CIE system of physical photometry and that have been officially accepted by the Comité International des Poids et Mesures (CIPM). They comprise:

- the definition of photometric quantities and units,
- the definition of CIE standard spectral luminous efficiency functions for photopic and scotopic vision,
- the definition of a CIE standard photometric observer that conforms to these functions,
- the definition of maximum luminous efficacies for photopic and scotopic vision.

An informative annex provides a vocabulary of related terms.

2. PHOTOMETRIC QUANTITIES

Photometric quantities are defined in the International Lighting Vocabulary (ILV) (CIE, 1987a).

2.1 Luminous flux

The fundamental physical quantity used in optical radiometry is the radiant flux or radiant power, Φ_e , measured in watts, which is emitted by a source of radiation, transmitted by a medium of propagation, or received at a surface. The corresponding photometric quantity is:

luminous flux (Φ_v) (see ILV 845-01-25)

quantity derived from radiant flux Φ_e by evaluating the radiation according to its action upon the CIE standard photometric observer

The procedure for deriving Φ_v from Φ_e is defined in 4.3, below.

2.2 Other quantities

The following are the photometric quantities that correspond to the most important radiometric quantities defined in the International Lighting Vocabulary.

luminous energy (also known as **quantity of light**) (Q_v) (see ILV 845-01-28)

time integral of the luminous flux Φ_v over a given duration Δt

$$Q_v = \int_{\Delta t} \Phi_v dt$$

luminous intensity (of a source in a given direction) (I_v) (see ILV 845-01-31)

quotient of the luminous flux $d\Phi_v$ leaving the source and propagated in the element of solid angle $d\Omega$ containing the given direction, by the element of solid angle