

ICS 13.030.01; 13.080.10

English Version

**Sludge, treated biowaste and soil - Determination of elements
using inductively coupled plasma mass spectrometry (ICP-MS)**

Boues, bio-déchets traités et sols - Détermination des
éléments en traces par spectrométrie de masse avec
plasma induit par haute fréquence (ICP-MS)

Schlamm, behandelter Bioabfall und Boden - Bestimmung
von Elementen mittels Massenspektrometrie mit induktiv
gekoppeltem Plasma (ICP-MS)

This Technical Specification (CEN/TS) was approved by CEN on 16 July 2012 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Principle.....	5
4 Interferences	6
4.1 General.....	6
4.2 Spectral interferences	6
4.2.1 Isobaric elemental interferences.....	6
4.2.2 Isobaric molecular and doubly-charged ion interferences	6
4.2.3 Non-spectral interferences	6
5 Reagents.....	7
6 Apparatus	9
7 Procedure	10
7.1 Test sample solution	10
7.2 Test solution.....	11
7.3 Instrument set-up.....	11
7.4 Calibration	11
7.4.1 Linear calibration function.....	11
7.4.2 Standard addition calibration	12
7.4.3 Determination of correction factors.....	12
7.4.4 Variable isotope ratio	12
7.5 Sample measurement.....	12
8 Calculation.....	13
9 Expression of results	13
10 Performance characteristics	13
10.1 Blank	13
10.2 Calibration check	14
10.3 Internal standard response.....	14
10.4 Interference	14
10.5 Recovery	14
10.6 Precision	14
11 Test report	14
Annex A (informative) Selected isotopes and spectral interferences for quadrupole ICP-MS instruments	15
Bibliography	16

Foreword

This document (CEN/TS 16171:2012) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This Technical Specification is applicable for several types of matrices as indicated in Table 1.

Table 1 — Matrices for which this Technical Specification is applicable

Matrix	Materials tested
Sludge	Municipal sludge
Biowaste	Compost
Soil	Sludge amended soils

WARNING — Persons using this Technical Specification should be familiar with usual laboratory practice. This Technical Specification does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this Technical Specification be carried out by suitably trained staff.

1 Scope

This Technical Specification specifies a method for the determination of the following elements in *aqua regia* or nitric acid digests of sludge, treated biowaste and soil:

Aluminium (Al), antimony (Sb), arsenic (As), barium (Ba), beryllium (Be), bismuth (Bi), boron (B), cadmium (Cd), calcium (Ca), cerium (Ce), cesium (Cs), chromium (Cr), cobalt (Co), copper (Cu), dysprosium (Dy), erbium (Er), europium (Eu), gadolinium (Gd), gallium (Ga), germanium (Ge), gold (Au), hafnium (Hf), holmium (Ho), indium (In), iridium (Ir), iron (Fe), lanthanum (La), lead (Pb), lithium (Li), lutetium (Lu), magnesium (Mg), manganese (Mn), mercury (Hg), molybdenum (Mo), neodymium (Nd), nickel (Ni), palladium (Pd), phosphorus (P), platinum (Pt), potassium (K), praseodymium (Pr), rhenium (Re), rhodium (Rh), rubidium (Rb), ruthenium (Ru), samarium (Sm), scandium (Sc), selenium (Se), silicon (Si), silver (Ag), sodium (Na), strontium (Sr), sulfur (S), tellurium (Te), terbium (Tb), thallium (Tl), thorium (Th), thulium (Tm), tin (Sn), titanium (Ti), tungsten (W), uranium (U), vanadium (V), ytterbium (Yb), yttrium (Y), zinc (Zn), and zirconium (Zr).

The working range depends on the matrix and the interferences encountered.

The limit of detection of the method is between 0,1 mg/kg dry matter and 2,0 mg/kg dry matter for most elements.

The limit of detection will be higher in cases where the determination is likely to be interfered (see Clause 4) or in case of memory effects (see e.g. 8.2 of EN ISO 17294-1:2006).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16173, *Sludge, treated biowaste and soil — Digestion of nitric acid soluble fractions of elements*

EN 16174, *Sludge, treated biowaste and soil — Digestion of aqua regia soluble fractions of elements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

EN ISO 17294-1:2006, *Water quality — Application of inductively coupled plasma mass spectrometry (ICP-MS) — Part 1: General guidelines (ISO 17294-1:2004)*

3 Principle

Digests of sludge, treated biowaste or soil with nitric acid or *aqua regia* (see EN 16173 and EN 16174) are analysed by ICP-MS to get a multi-elemental determination of analytes.

The method measures ions produced by a radio-frequency inductively coupled plasma. Analyte species originating in the digest solution are nebulised and the resulting aerosol is transported by argon gas into the plasma. The ions produced by the high temperatures of the plasma are entrained in the plasma gas and introduced, by means of an interface, into a mass spectrometer, sorted according to their mass-to-charge ratios and quantified with a detector (e.g. channel electron multiplier).