INTERNATIONAL STANDARD

ISO 10360-7

First edition 2011-06-01

Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) —

Part 7:

CMMs equipped with imaging probing systems

Spécification géométrique des produits (GPS) — Essais de réception et de vérification périodique des machines à mesurer tridimensionnelles (MMT) —

Partie 7: MMT équipées de systèmes de palpage imageurs



This document is a preview denerated by EUS **COPYRIGHT PROTECTED DOCUMENT**



© ISO 2011

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office Case postale 56 • CH-1211 Geneva 20 Tel. + 41 22 749 01 11 Fax + 41 22 749 09 47 E-mail copyright@iso.org Web www.iso.org

Published in Switzerland

Contents

Page

Forew	vord	v
Introductionvi		
1	Scope	
2	Normative references	1
3	Terms and efinitions	2
4	Symbols	7
5	Environmental and metrological requirements	8
5.1	Environmental conditions	8
5.2	Operating conditions Requirements for various configuration imaging probe CMMs General Length measurement errors	8
5.3	Requirements for various configuration imaging probe CMMs	8
5.3.1	General	8
5.3.2	Length measurement errors	9
5.3.3	Probing errors	10
5.3.4	Repeatability range of the length measurement error, R _B or R _U	10
5.3.5	Workpiece loading effects	10
6	Acceptance tests and reverification tests	11
6.1	Acceptance tests and reverification tests	11
6.2	Langth magazirament arrara	11
6.2.1	Constal	11
6.2.1	General Length measurement errors	۱۱
	Langth mass warment array E ar E	12
6.2.3	Length measurement error, E_{B} or E_{U}	13
6.2.4	Z length measurement error, E_{BZ} or E_{UZ}	14
6.2.5	At length measurement error, E_{RYV} or E_{HYV}	
6.2.6	Imaging probe length measurement error, E_{BV}	15
6.3	Squareness error, E_{SQ}	16
6.3.1	General	16
6.3.2	Squareness error, $E_{\rm SQ}$	16
6.3.3	Measuring positions	17
6.3.4	Measuring procedure	17
6.3.5	Derivation of test results	18
6.4	Repeatability range of the length measurement error, $R_{\rm B}$ or $R_{\rm U}$	19
6.5	Repeatability range of the length measurement error, $R_{\rm B}$ or $R_{\rm U}$. Probing performance ($P_{\rm F2D}$)	19
6.5.1	Principle	19
6.5.2	Measuring equipment	19
6.5.3	Procedure	19
6.5.4	Derivation of test results Probing error of the imaging probe, P_{FV2D}	21
6.6	Probing error of the imaging probe, P_{FV2D}	21
6.6.1	Principle	
6.6.2	Measuring equipment	
6.6.3	Procedure	
6.6.4	Derivation of test results	21
7	Compliance with specifications	22
7.1	Acceptance test	
7.1.1	Acceptance criteria	
7.1.2	Data rejection and repeated measurements	
7.2	Reverification test	
0	Applications	24
8	ADDITICATIONS	24

ISO 10360-7:2011(E)

8.1 Acceptance test	24
8.2 Reverification test	
8.3 Interim check	2!
9 Indication in product documentation and data sheets	20
Annex A (informative) Interim check	2
Annex B (normative) Artefacts that represent a calibrated test len	gth28
Annex C (informative) Alternative method for checking the square	eness error34
Annex D (normative) Mathematical adjustments to low CTE artefa	
Annex E (informative) Relation to the GPS matrix model	38
Bibliography	39
() *	

ation to this document is a previous generaled by this

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 10360-7 was prepared by Technical Committee ISO/TC 213, Dimensional and geometrical product specifications and verification.

ISO 10360 consists of the following parts, under the general title *Geometrical product specifications (GPS)* — *Acceptance and reverification tests for coordinate measuring machines (CMM)*:

- Part 1: Vocabulary
- Part 2: CMMs used for measuring linear dimension
- Part 3: CMMs with the axis of a rotary table as the fourth axis
- Part 4: CMMs used in scanning measuring mode
- Part 5: CMMs using single and multiple stylus contacting probing systems
- Part 6: Estimation of errors in computing of Gaussian associated tures
- Part 7: CMMs equipped with imaging probing systems
- Part 9: CMMs with multiple probing systems

The following part is under preparation:

Part 8: CMMs with optical distance sensors

© ISO 2011 – All rights reserved

Introduction

This part of ISO 10360 is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO/TR 14638). It influences chain link 5 of the chains of standards on size, distance, radius, angle, form, orientation, location, run-out and datums. For more detailed information on the relation of this part of ISO 10360 to other standards and the GPS matrix model, see Annex E.

The tests of this part of ISO 10360 have two technical objectives:

- a) to test the error of indication of a calibrated test length using an imaging probing system;
- b) to test the errors in the imaging probing system.

The benefits of these tests are that the measured result has a direct traceability to the unit length, the meter, and that it gives information on how the OMM will perform on similar length measurements.

The structure of this part of ISO 10360 parallels that of ISO 10360-2, which is for CMMs equipped with contact probing systems. The testing methodology between these two parts of ISO 10360 is intentionally similar. The differences that exist may be eliminated in future revisions of either this part of ISO 10360 or ISO 10360-2.

All the definitions in Clause 3 will appear in the reason of ISO 10360-1:2000.

Original Clause 3 will appear in the reason of ISO 10360-1:2000.

vi

Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) —

Part 7:

CMMs equipped with imaging probing systems

1 Scope

This part of ISO 10360 specifies the acceptance tests for verifying the performance of a coordinate measuring machine (CMM) used for measuring linear dimensions as stated by the manufacturer. It also specifies the reverification tests that enable the user to periodically reverify the performance of the CMM.

The acceptance and reverification tests oven in this part of ISO 10360 are applicable only to Cartesian CMMs using imaging probing systems of any type operating in the discrete-point probing mode.

This part of ISO 10360 does not explicitly apply to

- non-Cartesian CMMs; however, parties may apply this part of ISO 10360 to non-Cartesian CMMs by mutual agreement;
- CMMs using other types of optical probing; however, parties may apply this approach to other optical CMMs by mutual agreement;
- CMMs using contact probing systems (see ISO 10360-2 to contact probing systems).

This part of ISO 10360 specifies performance requirements that can be assigned by the manufacturer or the user of a CMM, the manner of execution of the acceptance and reprification tests to demonstrate the stated requirements, rules for proving conformance, and applications for which the acceptance and reverification tests can be used.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10360-1:2000, Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 1: Vocabulary

ISO 10360-2:2009, Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 2: CMMs used for measuring linear dimensions

ISO 14253-1:1998, Geometrical Product Specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or non-conformance with specifications

© ISO 2011 – All rights reserved

ISO 10360-7:2011(E)

ISO 14660-1:1999, Geometrical Product Specifications (GPS) — Geometrical features — Part 1: General terms and definitions

ISO/TS 23165:2006, Geometrical product specifications (GPS) — Guidelines for the evaluation of coordinate measuring machine (CMM) test uncertainty

ISO/IEC Guide 99, International vocabulary of metrology — Basic and general concepts and associated terms (VIM)

Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10360-1, ISO 10360-2, ISO 14253-1, ISO 14660-1, ISO/TS 2316 SO/IEC Guide 99 and the following apply.

3.1

imaging probing system

probing system which creates measurement points through the use of an imaging system

concerned with imaging probing systems that enable measurements in the This part of ISO 10360 is primarily lateral direction to the probing system axis.

A video or vision probing system is an imaging probing system. NOTE 2

3.2

imaging probe CMM

CMM equipped with an imaging probing system

3.3

field of view

FOV

area viewed by the imaging probing system

See Figure 1.

The measuring limits, or size, of the FOV are stated as the limits of the object space that is reproduced in the NOTE final image.

3.4

measuring window

region of interest in the FOV that is used in the determination of the measured

See Figure 1.

Configurations of measuring windows may vary widely between various imaging pro MMs and for different measuring applications on the same imaging probe CMM.

3.5

measuring plane (of the imaging probing system)

two-dimensional plane defined by the FOV of an imaging probing system

3.6

coefficient of thermal expansion

CTE

linear thermal expansion coefficient of a material at 20 °C