

Stationary source emissions - Determination of low
range mass concentration of dust - Part 2: Quality
assurance of automated measuring systems

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 13284-2:2017 sisaldab Euroopa standardi EN 13284-2:2017 ingliskeelset teksti.	This Estonian standard EVS-EN 13284-2:2017 consists of the English text of the European standard EN 13284-2:2017.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 01.11.2017.	Date of Availability of the European standard is 01.11.2017.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.040.40

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Stationary source emissions - Determination of low range
mass concentration of dust - Part 2: Quality assurance of
automated measuring systems**

Émissions de sources fixes - Détermination de faibles
concentrations en masse de poussières - Partie 2 :
Assurance qualité des systèmes de mesurage
automatisés

Emissionen aus stationären Quellen - Ermittlung der
Staubmassenkonzentration bei geringen
Staubkonzentrationen - Teil 2: Qualitätssicherung für
automatische Messeinrichtungen

This European Standard was approved by CEN on 11 September 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

European foreword.....	4
Introduction	5
1 Scope.....	6
2 Normative references.....	6
3 Terms and definitions	6
4 Symbols and abbreviations	6
4.1 Symbols.....	6
4.2 Abbreviations	7
5 Principle	8
6 Calibration and validation of the AMS (QAL2).....	8
6.1 General.....	8
6.2 Functional test.....	8
6.3 Parallel measurements with the SRM.....	8
6.3.1 General.....	8
6.3.2 Increasing the dust concentration	9
6.3.3 SRM measured values below the limit of quantification	9
6.3.4 Semi-volatiles	9
6.4 Data evaluation.....	10
6.4.1 Preparation of data.....	10
6.4.2 Selection of data points from automated SRM.....	10
6.4.3 Establishing the calibration function.....	10
6.5 Calibration function of the AMS and its validity	11
6.6 Calculation of variability	11
6.7 Test of variability	11
6.8 QAL2 report.....	11
7 Ongoing quality assurance during operation (QAL3).....	11
8 Annual Surveillance Test (AST)	11
9 Documentation.....	12
Annex A (informative) Example of calculation of the calibration function and of the variability test	13
A.1 General.....	13
A.2 Data evaluation.....	13
A.2.1 General.....	13
A.2.2 Calculation of measured values of the SRM at standard conditions.....	15
A.2.3 Calibration function	16
A.2.4 Calibrated AMS measured values	17
A.2.5 Conversion of data to standard conditions.....	18
A.2.6 Valid calibration range.....	18
A.2.7 Test of variability	19
Annex B (informative) Configuration of extractive dust measuring systems	22
B.1 General.....	22
B.2 Isokinetic sampling	22

B.3	Calibration function.....	23
Annex C	(informative) Number of parallel measurements for calibration	24
C.1	General	24
C.2	Errors in calibration at low dust concentration.....	24
Annex D	(informative) Extension of the calibration range by changing the dust concentration.....	25
D.1	General	25
D.2	Changing plant load	25
D.3	Changing operation of dust arrestment plant.....	25
D.4	Injection of dust.....	26
Annex E	(informative) Quadratic regression.....	28
Annex F	(informative) Configuration of AMS at mass concentration below the limit of quantification of the SRM.....	30
F.1	General	30
F.2	Techniques directly dealing with mass	30
F.3	Scattered-light measuring systems.....	30
F.4	Surrogate generators	30
F.5	Missing alternative calibration techniques.....	30
Annex G	(informative) Significant technical changes	32
Bibliography		33

European foreword

This document (EN 13284-2:2017) has been prepared by Technical Committee CEN/TC 264 “Air quality”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2018, and conflicting national standards shall be withdrawn at the latest by May 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 13284-2:2004.

This document is Part 2 of the EN 13284 series:

- EN 13284-1, *Stationary source emissions — Determination of low range mass concentration of dust — Part 1: Manual gravimetric method*;
- EN 13284-2, *Stationary source emissions — Determination of low range mass concentration of dust — Part 2: Quality assurance of automated measuring systems*.

Annex G provides details of significant technical changes between this European Standard and the previous edition.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This document describes the quality assurance procedures related to automated measuring systems (AMS) for the determination of dust in waste gas, in order to meet the uncertainty requirements on measured values given by regulations, e.g. EU Directives ([1]), national or other legislation.

This document is derived from EN 14181, which specifies general procedures for establishing quality assurance levels (QAL) for automated measuring systems (AMS) installed on industrial plants for the determination of the waste gas components and other waste gas parameters. It amends EN 14181 and provides guidance specific to dust measurements. It is only applicable in conjunction with EN 14181.

The calibration and validation of dust AMS is based on parallel measurements with the manual gravimetric standard reference method (SRM) described in EN 13284-1.

This document was primarily developed for the measurement of emissions from waste incinerators. From a technical point of view, it can be applied to other processes, for which measurement at an emission limit is required with defined uncertainty.

1 Scope

This European Standard specifies requirements for the calibration and validation (QAL2), the ongoing quality assurance during operation (QAL3) and the annual surveillance test (AST) of automated measuring systems (AMS) used for monitoring dust emissions from stationary sources to demonstrate compliance with emission limit values (ELV) below 50 mg/m³ at standard conditions. It specifically deals with measurements in wet gases and at low concentrations.

This document is derived from EN 14181 and is only applicable in conjunction with EN 14181.

This document is applicable by direct correlation with the standard reference method (SRM) described in EN 13284-1.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13284-1:2017, *Stationary source emissions — Determination of low range mass concentration of dust — Part 1: Manual gravimetric method*

EN 14181:2014, *Stationary source emissions — Quality assurance of automated measuring systems*

EN 15267-3, *Air quality — Certification of automated measuring systems — Part 3: Performance criteria and test procedures for automated measuring systems for monitoring emissions from stationary sources*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 13284-1:2017 and EN 14181:2014 apply.

4 Symbols and abbreviations

4.1 Symbols

For the purposes of this document, the following symbols apply.

a	intercept of the calibration function
$\hat{a}$	best estimate of a
b	slope of the calibration function
$\hat{b}$	best estimate of b
c	slope of the quadratic term of the calibration function
d	particle diameter
$\bar{D}$	average of differences D_i
e	error
E	emission limit value
h	absolute water vapour content (by volume)
i	counter