

Dentistry - Analysis of Fluoride Concentration in
Aqueous Solutions by use of Fluoride-Ion Selective
Electrode (ISO 19448:2018)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 19448:2018 sisaldab Euroopa standardi EN ISO 19448:2018 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 19448:2018 consists of the English text of the European standard EN ISO 19448:2018.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 31.01.2018.	Date of Availability of the European standard is 31.01.2018.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 71.100.70

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Analysis of Fluoride Concentration in Aqueous
Solutions by use of Fluoride-Ion Selective Electrode (ISO
19448:2018)

Médecine bucco-dentaire - Analyse de la concentration
en fluorure dans les solutions aqueuses à l'aide d'une
électrode sélective d'ions fluorures (ISO 19448:2018)

Zahnheilkunde - Analyse der Fluoridkonzentration in
wässrigen Lösungen mit einer fluoridionenselektiven
Elektrode (ISO 19448:2018)

This European Standard was approved by CEN on 23 December 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European Foreword

This document (EN ISO 19448:2018) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2018, and conflicting national standards shall be withdrawn at the latest by July 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 19448:2018 has been approved by CEN as EN ISO 19448:2018 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	1
5 Reagents and materials	2
6 Apparatus	2
7 Preparation	3
7.1 TISAB solution	3
7.2 TISAB blank solution	3
7.3 Fluoride standard solutions	3
7.4 Calibration curve	4
8 Procedure	5
8.1 Direct analysis technique	5
8.2 Standard addition technique	6
8.3 Dilution analysis technique	7
8.4 Multiple sample addition technique	7
8.5 Internal standards technique	7
9 Technical notes about the analytical techniques	7
10 Calculations	8
10.1 Direct analysis technique	8
10.2 Standard additions technique	8
10.3 Dilution analysis technique	9
10.4 Multiple sample addition technique	9
10.5 Internal standards technique	9
11 Test report	9
12 Trouble shooting	9
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral care products*.

Introduction

Fluoride is used in dentistry primarily for caries prevention. Fluoride is the active ingredient for caries prevention in many dental products such as dentifrices, gels, oral rinses, and fluoride releasing varnishes. The fluoride concentration in these products is an integral part of other standards where the product is intended to be analysed for fluoride content. This document provides methods based on the use of fluoride ion-selective electrode technology for the analysis of the total fluoride content in aqueous samples. The methods describe uses for fluoride ion-selective electrode technology and anticipate that the sample-specific preparations such as digestion, distillation, etc. have been described in the standards specific to product types.

Dentistry — Analysis of fluoride concentration in aqueous solutions by use of fluoride ion-selective electrode

1 Scope

This document specifies test methods for the quantification of fluoride concentrations in dental products including dentifrices, gels, oral rinses, fluoride releasing varnishes, and other fluoride containing products. The methods are based on fluoride ion-selective electrode technology for the analysis of fluoride in aqueous samples derived from dental products.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 78-2, *Chemistry — Layouts for standards — Part 2: Methods of chemical analysis*

ISO 835, *Laboratory glassware — Graduated pipettes*

ISO 1942, *Dentistry — Vocabulary*

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 78-2 and ISO 1942 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

4 Principle

This test method covers the determination of fluoride concentration in aqueous solutions derived from dental products. The following test methods are presupposed to use fluoride ion-selective electrodes (F-ISE).

The sensing element of the fluoride ion-selective electrode consists of a single crystal of europium (II)-doped lanthanum fluoride, LaF_3 . When the sensing element is in contact with a solution containing fluoride ions, an electrode potential develops across the sensing element. This potential, which depends on the activity of free fluoride ion in solution $\{\text{F}^-\}$, is measured against a constant reference electrode potential with a digital pH/mV meter or ISE (concentration) potentiometer. A potential (E) is developed which is proportional to the logarithm of the activity of the fluoride ion $\{\text{F}^-\}$ in solution following the Nernst Equation.

$$E = E_0 + S \cdot \log \{\text{F}^-\} \quad (1)$$