
**Coffee and coffee products —
Determination of acrylamide —
Methods using HPLC-MS/MS and GC-
MS after derivatization**

*Café et de ses dérivés — Dosage de l'acrylamide — Méthodes utilisant
CLHP-MS/MS et CG-MS après dérivation*

This document is a preview generated by EBS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2016, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

| | |
|---|-----------|
| Foreword | iv |
| Introduction | v |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Principle | 1 |
| 5 Reagents | 1 |
| 6 Apparatus | 3 |
| 7 Sampling | 4 |
| 8 Procedure | 4 |
| 8.1 General | 4 |
| 8.2 Preparation of the sample extract | 4 |
| 8.3 Clean-up of the extracts | 5 |
| 8.3.1 Carrez precipitation | 5 |
| 8.3.2 Solid phase extraction | 5 |
| 8.4 HPLC-MS/MS measurement | 5 |
| 8.4.1 High-performance liquid chromatography (HPLC) | 5 |
| 8.4.2 Identification and quantification by mass spectrometry (HPLC-MS/MS) | 6 |
| 8.5 Measurement with GC-MS | 6 |
| 8.5.1 Derivatization and sample preparation for gas chromatography | 6 |
| 8.5.2 Gas chromatography | 7 |
| 8.5.3 Identification and quantification by mass spectrometry | 7 |
| 9 Calibration | 7 |
| 9.1 General advice | 7 |
| 9.2 Determination of linearity and definition of the working range | 7 |
| 9.3 Calibration with internal standard solution | 7 |
| 9.4 Determination of the laboratory specific recovery | 8 |
| 10 Evaluation | 8 |
| 10.1 Criteria for identification | 8 |
| 10.2 Calculation and final results | 8 |
| 11 Precision data | 9 |
| 11.1 General | 9 |
| 11.2 Repeatability | 9 |
| 11.3 Reproducibility | 9 |
| 11.4 Recovery | 9 |
| 12 Measurement uncertainty | 9 |
| 13 Test report | 10 |
| Annex A (informative) Performance characteristics | 11 |
| Annex B (informative) Examples of absorber materials | 12 |
| Annex C (informative) Examples of columns and analysis conditions | 13 |
| Bibliography | 19 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 15, *Coffee*.

Introduction

When applying this document, all existing safety regulations have to be followed.

This document is a preview generated by EVS

Coffee and coffee products — Determination of acrylamide — Methods using HPLC-MS/MS and GC-MS after derivatization

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to take appropriate measures for ensuring the safety and health of the personnel prior to application of this document and to fulfil statutory requirements for this purpose.

1 Scope

This document specifies methods for the determination of acrylamide in coffee and coffee products by extraction with water, clean-up by solid-phase extraction and determination by HPLC-MS/MS and GC-MS. It was validated in a method validation study on roasted coffee, soluble coffee, coffee substitutes and coffee products with ranges from 53 µg/kg to 612,1 µg/kg.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

4 Principle

The coffee sample is extracted with water or, in the case of soluble products, dissolved in water. A clean-up by solid phase extraction is employed to remove interfering matrix compounds. Two alternative methods can be used for the determination: high-performance liquid chromatography with mass spectrometric detection (HPLC-MS/MS) or, after a bromination of the acrylamide, gas chromatography with mass spectrometric detection (GC-MS). In both cases, isotopic labelled internal standard solutions are used.

5 Reagents

WARNING — In view of health risks when working with acrylamide, appropriate preventive and protection measures shall be taken, such as using a fume cupboard, aspirating acrylamide-containing solutions only with a pipette, and avoiding skin and eye contact or inhalation of acrylamide-containing vapour.