Health informatics - Personal health device communication - Part 00103: Overview (ISO/IEEE 11073-00103:2015)



### **EESTI STANDARDI EESSÕNA**

### **NATIONAL FOREWORD**

|   | This Estonian standard EVS-EN ISO 11073-00103:2017 consists of the English text of the European standard EN ISO 11073-00103:2017.  |
|---|--|
| Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.  | This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation. |
| Euroopa standardimisorganisatsioonid on teinud<br>Euroopa standardi rahvuslikele liikmetele<br>kättesaadavaks 15.02.2017. | Date of Availability of the European standard is 15.02.2017.   |
| Standard on kättesaadav Eesti<br>Standardikeskusest.  | The standard is available from the Estonian Centre for Standardisation.  |

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile <u>standardiosakond@evs.ee</u>.

ICS 11.040.55, 35.240.80

### Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega: Koduleht <u>www.evs.ee</u>; telefon 605 5050; e-post <u>info@evs.ee</u>

### The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

# **EUROPEAN STANDARD**

## NORME EUROPÉENNE

**EUROPÄISCHE NORM** 

February 2017

ICS 11.040.55; 35.240.80

### **English Version**

### Health informatics - Personal health device communication - Part 00103: Overview (ISO/IEEE 11073-00103:2015)

Informatique de santé - Communication entre dispositifs de santé personnels - Partie 00103: Aperçu général (ISO/IEEE 11073-00103:2015)

Medizinische Informatik - Kommunikation von Geräten für die persönliche Gesundheit-Teil 00103: Überblick (ISO/IEEE 11073-00103:2015)

EN ISO 11073-00103

This European Standard was approved by CEN on 16 January 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

### **European foreword**

The text of ISO/IEEE 11073-00103:2015 has been prepared by Technical Committee ISO/TC 215 "Health informatics" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 11073-00103:2017 by Technical Committee CEN/TC 251 "Health informatics" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2017, and conflicting national standards shall be withdrawn at the latest by August 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 11073-00103:2015 has been approved by CEN as EN ISO 11073-00103:2017 without any modification.

### Introduction

This introduction is not part of IEEE Std 11073-00103-2012, Health informatics-Personal health device communication—Part 00103: Overview.

Within the context of the ISO/IEEE 11073 family of standards for device communication, this guide describes the landscape of transport-independent applications and information profiles for personal telehealth devices. These profiles define data exchange, data representation, and terminology for rsonal c
ie activity,
or mobile apph. communication between personal telehealth devices and compute engines (e.g., health appliances, set top boxes, cell phones, and personal computers). The guide provides a definition of personal telehealth devices as devices used for life activity, wellness monitoring, and/or health monitoring in domestic home, communal home, and/or mobile applications. Use cases relevant to these scenarios and environments are also presented.

### Contents

| 1. Overview  | 1    |
|--|------|
| 1.1 Scope  |      |
| 1.2 Purpose  |      |
| 1.3 The standards within 11073 standards applicable for the personal health devices (PHD) domain |      |
| 1.4 Audience   |      |
| 1.5 Document organization.   |      |
| 1.5 Document organization  | 5    |
| 2. Definitions, acronyms, and abbreviations  | 3    |
| 2.1 Definitions  |      |
| 2.2 Acronyms and abbreviations   |      |
| 2.2 retory his and decreated in  | •••• |
| 3. PHD environment overview  | 5    |
| 3.1 General  |      |
| 3.2 Topology of PHD systems  |      |
| 3.3 Use contexts.  |      |
| 3.4 Health and fitness   |      |
| 3.5 Independent living (aging independently)   |      |
|  |      |
| 3.6 Disease management   | . 13 |
| 5.7 Device examples  | . 20 |
| 4. Introduction into IEEE 11073 PHD standards (tutorial)   | 23   |
| 4.1 General description of the IEEE 11073 context  |      |
| 4.2 Domain information model   |      |
| 4.3 Nomenclature   |      |
| 4.4 Service model  |      |
| 4.5 Communication model  |      |
| 4.5 Communication model.   | . 33 |
| 5. Utilizing IEEE 11073 PHD standards in the development process                                 | 35   |
| 5.1 Example implementation: Introduction, how to use ISO/IEEE 11073-20601:2010(E) [B48]          |      |
| 5.2 Customer needs analysis, use knowledge from the IEEE 11073 PHD standards                     |      |
| 5.3 Risk management  |      |
| 5.4 Security   |      |
| 5.5 Quality of service (QoS)   |      |
| 5.6 Regulatory issues—what is a medical device?  | . 43 |
| 5.7 System development planning: Using available IEEE 11073 PHD conformant components may        | . 43 |
| speed things up  | 47   |
|  |      |
| 5.8 System requirements analysis: Drawing from the ISO/IEEE 11073-20601 blueprints               |      |
| 5.9 Software architectural design: Using the ISO/IEEE 11073-20601 building blocks                |      |
| 5.10 Software detailed design: Using the ISO/IEEE 11073-20601 engineering elements               | . 50 |
| 5.11 Software unit implementation and verification   | . 50 |
| 5.12 System integration and testing, validation  |      |
| 5.13 System release  |      |
| 5.14 Configuration management  |      |
| 5.15 Maintenance   | . 51 |
|  |      |
| 6 Conformance and interoperability   | 51   |

| nnex A (informative) Example use case regular blood pressure control, detailed description | 53 |
|--|----|
| nex B (informative) Example transaction profiles   | 55 |
| nex C (informative) Transport layer details  | 60 |
| nex D (informative) Bibliography   |    |
| xi<br>Copyright © 2012 IEEE. All rights reserved.  |    |

### Health informatics—Personal health device communication

### Part 00103: Overview

IMPORTANT NOTICE: IEEE Standards documents are not intended to ensure safety, health, or environmental protection, or ensure against interference with or from other devices or networks. Implementers of IEEE Standards documents are responsible for determining and complying with all appropriate safety, security, environmental, health, and interference protection practices and all applicable laws and regulations.

This IEEE document is made available for use subject to important notices and legal disclaimers. These notices and disclaimers appear in all publications containing this document and may be found under the heading "Important Notice" or "Important Notices and Disclaimers Concerning IEEE Documents." They can also be obtained on request from IEEE or viewed at <a href="http://standards.ieee.org/IPR/disclaimers.html">http://standards.ieee.org/IPR/disclaimers.html</a>.

### 1. Overview

### 1.1 Scope

Within the context of the ISO/IEEE 11073 family of standards for device communication, this guide describes the landscape of transport-independent applications and information profiles for personal telehealth devices. These profiles define data exchange, data representation, and terminology for communication between personal health devices and compute engines (e.g., health appliances, set top boxes, cell phones, and personal computers). The guide provides a definition of personal telehealth devices as devices used for life activity, wellness monitoring, and/or health monitoring in domestic home, communal home, and/or mobile applications as well as professional medical usage. Use cases relevant to these scenarios and environments are also presented.

### 1.2 Purpose

This guide sets a context for other personal telehealth standards in the ISO/IEEE 11073 framework of standards and describes the need for interoperability in personal telehealth environments. Interoperability is the key to growing the potential market for these devices and to enabling people to manage their own health independently.

#### IEEE Std 11073-00103-2012 Health informatics—Personal health device communication Part 00103: Overview

### 1.3 The standards within 11073 standards applicable for the personal health devices (PHD) domain

The IEEE 11073 series of standards date back to the 1990s. It was initially intended for connecting pointof-care medical devices in professional healthcare provider organizations. Examples of these devices are vital signs monitors, blood pressure monitors, and other "medical" devices, Initially, medical devices were in most cases used in healthcare organizations by medical experts. However, the use of medical devices at home increased over time. Additionally, fitness and health devices reached the market. The intended use of these devices is generally not by clinicians directly, but derived data may have clinical significance. The term PHD evolved for medical devices as well as for health and fitness devices used out of professional healthcare organizations, by users at home. Today, PHDs are commonly sold together with consumer electronics products. They are used in home and mobile environments. Most devices provide digital displays and local storage of readings. Because of their small size and limited power supply, many devices have low computational limits. Users increasingly find it cumbersome to read data from displays and to enter it manually into online forms. Any manual interference adds to the probability of error. Communicating the recorded data is therefore gaining importance, with the additional advantage of avoiding media breaches. This development was also reflected in standardization work.

The following standards have been developed within the IEEE 11073 PHD standards series so far.

— ISO/IEEE 11073-20601:2010(E) [B48] and its amendment IEEE Std 11073-20601a:2010 [B36]: The Optimized Exchange Protocol defines the core elements: the domain information model, the service model, and the communication model.<sup>1</sup>

A further series of "device specialization" standards then tailor the broad toolkit provided in ISO/IEEE 11073-20601:2010(E) [B48] and IEEE Std 11073-20601a:2010 [B36] to specific usages to meet the needs to the device type being specialized. The following device specializations are currently available: 6] 8] 229]

- ISO/IEEE 11073-10404:2010(E) [B43]
- IEEE Std 11073-10406<sup>™</sup>-2011 [B27]
- ISO/IEEE 11073-10407:2010(E) [B44]
- ISO/IEEE 11073-10408:2010(E) [B45]
- ISO/IEEE 11073-10415:2010(E) [B46]
- IEEE Std 11073-10417<sup>™</sup>-2011 [B28]
- IEEE Std 11073-10418<sup>™</sup>-2011 [B29]
- IEEE Std 11073-10420<sup>™</sup>-2010 [B30]
- IEEE Std 11073-10421<sup>™</sup>-2010 [B31]
- IEEE Std 11073-10441™-2008 [B32]
- IEEE Std 11073-10442<sup>™</sup>-2008 [B33]
- IEEE Std 11073-10471™-2008 [B34]
- IEEE Std 11073-10472<sup>™</sup>-2010 [B35]

Work is in progress to add further device specializations.

<sup>&</sup>lt;sup>1</sup> The numbers in brackets correspond to those of the bibliography in Annex D.

## IEEE Std 11073-00103-2012 Health informatics—Personal health device communication Part 00103: Overview

An attempt has been made to ensure that the device specialization documents together with the base standard are self-contained and complete. For example, ISO/IEEE 11073-10101:2004(E) [B41] provides an extensive list of terms for coding data elements in the domain information model. For convenience of use, the terms are repeated where they are used, so that the reader does not need to consult ISO/IEEE 11073-10101:2004(E). Equally, key concepts such as the information and communication models are reproduced.

### 1.4 Audience

This overview is intended for readers who are interested in standardization for interoperability in the PHD field. It targets readers, engineers, and nonengineers who plan to provide and contribute to personal healthcare services, to use or buy devices, as well as those who are interested in the manifold steps that are necessary for planning and implementation.

This guide might also be used by information technology (IT) experts, who are seeking an introductory level entry point into the multi-part IEEE 11073 PHD series of standards, and need to quickly identify which content is intended for which purpose, and where the details are described. It also explains the more user-oriented aspects so that engineers may learn more about how users see things. That is the main focus of the "environmental overview."

### 1.5 Document organization

Subsequent sections of this document are organized as follows.

- Clause 2 includes definitions, acronyms, and abbreviations. In general, these should be accessible by general technical and implementer audiences.
- Clause 3 describes user and device use characteristics in some detail, followed by a summary of device types; content is at a general technical level.
- Clause 4 is a "tutorial" of PHD technology, which is a set of structured models, generally following "top-down" organization, which are intended for technical audiences to gain a reasonably substantive understanding of standard architecture, although content should be accessible to general technical readers. It introduces the main technical concepts of the IEEE 11073 PHD standards.
- Clause 5 covers a number of topics concerning implementation, and along with Annex A through Annex C, is intended for in-depth technical understanding. It introduces how the IEEE 11073 PHD standards are used together with other specifications in implementation projects.
- Annex D is a detailed informative bibliography.

### 2. Definitions, acronyms, and abbreviations

### 2.1 Definitions

For the purposes of this document, the following terms and definitions apply. The *IEEE Standards Dictionary Online* should be consulted for terms not defined in this clause.<sup>2</sup>

agent: A node that collects and transmits personal health data to an associated manager.

<sup>&</sup>lt;sup>2</sup> The *IEEE Standards Dictionary Online* subscription is available at <a href="http://www.ieee.org/portal/innovate/products/standardsgandard