

---

# INTERNATIONAL STANDARD



# 1391 / III

---

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

---

## Paraformaldehyde for industrial use — Methods of test — Part III : Determination of iron content — 2,2'-Bipyridyl photometric method

*Paraformaldéhyde à usage industriel — Méthodes d'essai — Partie III : Dosage du fer — Méthode  
photométrique au bipyridyle-2,2'*

**First edition — 1976-12-15**

Corrected and reprinted 1977-01-06

---

UDC 661.727.1 : 620.1 : 543.8

Ref. No. ISO 1391/III-1976 (E)

**Descriptors :** paraformaldehyde, tests, chemical analysis, determination of content, ash, iron, impurities.

Price based on 2 pages

## FOREWORD

ISO (the International Organization for Standardization) is a worldwide federation of national standards institutes (ISO member bodies). The work of developing International Standards is carried out through ISO technical committees. Every member body interested in a subject for which a technical committee has been set up has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council.

Prior to 1972, the results of the work of the technical committees were published as ISO Recommendations; these documents are in the process of being transformed into International Standards. As part of this process, Technical Committee ISO/TC 47, *Chemistry*, has reviewed ISO Recommendation R 1391-1970 and found it technically suitable for transformation. The technical committee, however, divided the recommendation into four parts (ISO 1391, parts I to IV), which therefore replace ISO Recommendation R 1391-1970, to which they are technically identical.

ISO Recommendation R 1391 had been approved by the member bodies of the following countries :

Austria	Ireland	Romania
Belgium	Italy	South Africa, Rep. of
Brazil	Japan	Spain
Czechoslovakia	Korea, Rep. of	Sweden
France	Netherlands	Switzerland
Germany	New Zealand	Thailand
Hungary	Poland	Turkey
Iran	Portugal	United Kingdom

The member body of the following country had expressed disapproval of the Recommendation on technical grounds :

India

The member bodies of the following countries disapproved the transformation of the Recommendation into an International Standard :

France  
Netherlands

# Paraformaldehyde for industrial use — Methods of test — Part III : Determination of iron content — 2,2'-Bipyridyl photometric method

## 1 SCOPE AND FIELD OF APPLICATION

This part of ISO 1391 specifies a 2,2'-bipyridyl photometric method for the determination of the iron content of paraformaldehyde for industrial use.

This document should be read in conjunction with part I (see the annex).

## 2 PRINCIPLE

Conversion of any iron present in a test portion into the sulphate by hot sulphuric acid.

Oxidation of any iron present in the solution by hydrogen peroxide, followed by reduction, by hydroxylammonium chloride, of the trivalent iron. Formation of the coloured complex iron(II)-2,2'-bipyridyl in a buffered medium. Photometric measurement of the coloured complex at a wavelength of about 510 nm.

## 3 REAGENTS

During the analysis, use only reagents of recognized analytical grade and only distilled water or water of equivalent purity.

**3.1 Sulphuric acid**, approximately 5 N solution.

**3.2 Hydrogen peroxide**, 150 g/l solution.

**3.3 Hydroxylammonium chloride** ( $\text{NH}_2\text{OH}\cdot\text{HCl}$ ), 100 g/l solution.

**3.4 Ammonium acetate** ( $\text{CH}_3\text{COONH}_4$ ), 500 g/l solution.

**3.5 2,2'-Bipyridyl**, 5 g/l hydrochloric solution.

Dissolve 0,5 g of 2,2'-bipyridyl in 100 ml of 1 N hydrochloric acid solution.

**3.6 Iron**, standard solution corresponding to 0,100 g of Fe per litre.

Weigh, to the nearest 0,000 1 g, 0,702 2 g of ammonium iron(II) sulphate hexahydrate  $[(\text{NH}_4)_2\text{SO}_4\cdot\text{FeSO}_4\cdot 6\text{H}_2\text{O}]$  and dissolve in 50 ml of the sulphuric acid solution (3.1). Transfer quantitatively to a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 0,100 mg of Fe.

**3.7 Iron**, standard solution corresponding to 0,010 g of Fe per litre.

Transfer 100 ml of the standard iron solution (3.6) to a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 10  $\mu\text{g}$  of Fe.

Prepare this solution immediately before use.

## 4 APPARATUS

Ordinary laboratory apparatus and

**4.1 Spectrophotometer**, or

**4.2 Photoelectric absorptiometer**, fitted with filters allowing maximum transmission at about 510 nm.

## 5 PROCEDURE

### 5.1 Preparation of calibration graph

**5.1.1 Preparation of standard matching solutions** relating to measurements carried out with cells of 4 or 5 cm optical path length

Into a series of seven 400 ml beakers, place the volumes of the standard iron solution (3.7) shown in the following table :

Standard iron solution (3.7)	Corresponding mass of iron
ml	$\mu\text{g}$
0*	0
2,0	20
4,0	40
7,0	70
10,0	100
15,0	150
20,0	200

\* Blank test of reagents for calibration graph.

Add, to each beaker, in successive small portions, 10 ml of the hydrogen peroxide solution (3.2) and 10 ml of the sulphuric acid solution (3.1) and then heat on a sand bath in a fume cupboard until acid fumes are evolved.