

ELEKTRILISED MEDITSIINISEADMED. OSA 2-16:
ERINÕUDED HEMODIALÜÜSI, HEMODIAFILTRATSIOONI
JA HEMOFILTRATSIOONISEADMETE ESMASELE
OHUTUSELE JA OLULISTELE TOIMIMISNÄITAJATELE

Medical electrical equipment - Part 2-16: Particular
requirements for the basic safety and essential
performance of haemodialysis, haemodiafiltration and
haemofiltration equipment

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN IEC 60601-2-16:2019 sisaldab Euroopa standardi EN IEC 60601-2-16:2019 ingliskeelset teksti.	This Estonian standard EVS-EN IEC 60601-2-16:2019 consists of the English text of the European standard EN IEC 60601-2-16:2019.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 24.05.2019.	Date of Availability of the European standard is 24.05.2019.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.20, 11.040.25

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Medical electrical equipment - Part 2-16: Particular requirements
for the basic safety and essential performance of haemodialysis,
haemodiafiltration and haemofiltration equipment
(IEC 60601-2-16:2018)**

Appareils électromédicaux - Partie 2-16: Exigences
particulières pour la sécurité de base et les performances
essentielles des appareils d'hémodialyse,
d'hémodiafiltration et d'hémofiltration
(IEC 60601-2-16:2018)

Medizinische elektrische Geräte - Teil 2-16: Besondere
Festlegungen für die Sicherheit einschließlich der
wesentlichen Leistungsmerkmale von Hämodialyse-,
Hämodiafiltrations- und Hämofiltrationsgeräten
(IEC 60601-2-16:2018)

This European Standard was approved by CENELEC on 2018-05-25. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of document 62D/1557/FDIS, future edition 5 of IEC 60601-2-16, prepared by SC 62D "Electromedical equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 60601-2-16:2019.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2019-11-24
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2022-05-24

This document supersedes EN 60601-2-16:2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60601-2-16:2018 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-24	NOTE Harmonized as EN 60601-2-24
IEC 60601-2-39	NOTE Harmonized as EN 60601-2-39
IEC 80001-1:2010	NOTE Harmonized as EN 80001-1:2011 (not modified)
ISO 8637-2	NOTE Harmonized as EN ISO 8637-2
ISO 11197	NOTE Harmonized as EN ISO 11197
ISO 23500-1	NOTE Harmonized as EN ISO 23500-1
ISO 23500-2	NOTE Harmonized as EN ISO 23500-2
ISO 23500-3	NOTE Harmonized as EN ISO 23500-3
ISO 23500-4	NOTE Harmonized as EN ISO 23500-4
ISO 23500-5	NOTE Harmonized as EN ISO 23500-5
ISO 14971:2007	NOTE Harmonized as EN ISO 14971:2012 (not modified)

Annex ZA

(normative)

Normative references to international publications with their corresponding European publications

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 Where an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Replacement</i>				
IEC 60601-1-2	2014	Medical electrical equipment - Part 1-2:EN 60601-1-2 General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests		2015
IEC 60601-1-6	2010	Medical electrical equipment - Part 1-6:EN 60601-1-6 General requirements for basic safety and essential performance - Collateral standard: Usability		2010
+ A1	2013		+ A1	2015
IEC 60601-1-8	2006	Medical electrical equipment - Part 1-8:EN 60601-1-8 General requirements for basic safety and essential performance - Collateral Standard: General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems		2007
-	-		+ corrigendum Mar.	2010
+ A1	2012		+ A1	2013
-	-		+ AC	2014
-	-		+ A11	2017
<i>Addition</i>				
IEC 60601-1-10	2007	Medical electrical equipment - Part 1-10:EN 60601-1-10 General requirements for basic safety and essential performance - Collateral Standard: Requirements for the development of physiologic closed-loop controllers		2008
+ A1	2013		+ A1	2015

IEC 60601-1-11	2015	Medical electrical equipment - Part 1-11:EN 60601-1-11	2015
		General requirements for basic safety and essential performance - Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment	
IEC 61672-1	-	Electroacoustics - Sound level meters -EN 61672-1	-
		Part 1: Specifications	
ISO 3744	-	Acoustics - Determination of sound powerEN ISO 3744	-
		levels and sound energy levels of noise sources using sound pressure - Engineering methods for an essentially free field over a reflecting plane	

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1 Scope, object and related standards.....	8
201.2 Normative references	10
201.3 Terms and definitions	11
201.4 General requirements	14
201.5 General requirements for testing ME EQUIPMENT	17
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	17
201.7 ME EQUIPMENT identification, marking and documents	17
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	22
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 Protection against unwanted and excessive radiation HAZARDS	23
201.11 Protection against excessive temperatures and other HAZARDS	23
201.12 * Accuracy of controls and instruments and protection against hazardous outputs	24
201.13 HAZARDOUS SITUATIONS and fault conditions for ME EQUIPMENT.....	34
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	34
201.15 Construction of ME EQUIPMENT.....	35
201.16 * ME SYSTEMS	35
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS	36
202 Electromagnetic disturbances – Requirements and tests	36
208 General requirements, tests and guidance for ALARM SYSTEMS in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS	37
209 Requirements for environmentally conscious design	39
210 Requirements for the development of PHYSIOLOGIC CLOSED-LOOP CONTROLLERS	39
211 * Requirements for MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS used in the HOME HEALTHCARE ENVIRONMENT.....	39
Annexes	41
Annex G (normative) Protection against HAZARDS of ignition of flammable anaesthetic mixtures.....	42
Annex AA (informative) Particular guidance and rationale	43
Annex BB (informative) Examples of HAZARDS, foreseeable sequences of events, and HAZARDOUS SITUATIONS in HAEMODIALYSIS EQUIPMENT	62
Bibliography.....	71
Index of defined terms used in this particular standard.....	74
Figure 201.101 – Continuous air infusion test setup with example dimensions	30
Figure AA.1 – Example of a HAEMODIALYSIS ME SYSTEM	58
Table 201.101 – ESSENTIAL PERFORMANCE requirements	14

Table AA.1 – Example of ALARM CONDITION priorities according to 6.1.2 of IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, adapted for HAEMODIALYSIS EQUIPMENT needs.....	60
Table BB.1 – HAZARDOUS SITUATION list following ISO 14971:2007, Annex E	62

This document is a preview generated by EVS

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International standard IEC 60601-2-16 has been prepared by IEC subcommittee 62D: Electromedical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This fifth edition cancels and replaces the fourth edition of IEC 60601-2-16 published in 2012. This edition constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- a) update of references to IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, of references and requirements to IEC 60601-1-2:2014, of references to IEC 60601-1-6:2010 and IEC 60601-1-6:2010/AMD1:2013, of references and requirements to IEC 60601-1-8:2006 and IEC 60601-1-8:2006/AMD1:2012, of references to

IEC 60601-1-9:2007 and IEC 60601-1-9:2007/AMD1:2013, of references to IEC 60601-1-10:2007 and IEC 60601-1-10:2007/AMD1:2013 and of references to IEC 60601-1-11:2015;

- b) widening of the scope;
- c) editorial improvements;
- d) addition of requirements for anticoagulant delivery means;
- e) other few small technical changes.

The text of this particular standard is based on the following documents:

FDIS	Report on voting
62D/1557/FDIS	62D/1585/RVD

Full information on the voting for the approval of this particular standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this document, the following print types are used:

- requirements and definitions: roman type;
- *test specifications: italic type*;
- informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type;
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this document, the term

- "clause" means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- "subclause" means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this particular standard are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

NOTE The attention of users of this document is drawn to the fact that equipment manufacturers and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committees that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

INTRODUCTION

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT.

This document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-16: Particular requirements for the basic safety and essential performance of haemodialysis, haemodiafiltration and haemofiltration equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This part of IEC 60601 applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION EQUIPMENT, hereafter referred to as HAEMODIALYSIS EQUIPMENT.

This document does not take into consideration specific safety details of the DIALYSIS FLUID control system of HAEMODIALYSIS EQUIPMENT using regeneration of DIALYSIS FLUID or CENTRAL DELIVERY SYSTEMS for DIALYSIS FLUID. It does, however, take into consideration the specific safety requirements of such HAEMODIALYSIS EQUIPMENT concerning electrical safety and PATIENT safety.

This document specifies the minimum safety requirements for HAEMODIALYSIS EQUIPMENT. These HAEMODIALYSIS EQUIPMENT are intended for use either by medical staff or for use by the PATIENT or other trained personnel under medical supervision.

This document includes all ME EQUIPMENT that is intended to deliver a HAEMODIALYSIS, HAEMODIAFILTRATION and HAEMOFILTRATION treatment to a PATIENT, independent of the treatment duration and location.

If applicable, this document applies to the relevant parts of ME EQUIPMENT intended for other extracorporeal blood purification treatments.

The particular requirements in this document do not apply to:

- EXTRACORPOREAL CIRCUITS (see ISO 8637-2, [12]²);
- DIALYSERS (see ISO 8637-1, [11]);
- DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]);
- DIALYSIS WATER supply systems (see ISO 23500-2, [16]);
- CENTRAL DELIVERY SYSTEMS for DIALYSIS FLUID CONCENTRATES (see ISO 23500-4, [18]), described as systems for bulk mixing concentrate at a dialysis facility;
- equipment used to perform PERITONEAL DIALYSIS (see IEC 60601-2-39, [8]).

¹ The general standard is IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

² Numbers in square brackets refer to the Bibliography.