
**Water quality — Determination of
mercury — Method using atomic
absorption spectrometry (AAS) with and
without enrichment**

*Qualité de l'eau — Dosage du mercure — Méthode par spectrométrie
d'absorption atomique (SAA) avec et sans enrichissement*



This document is a preview generated by EVS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2012

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Principle	2
4 General interferences	2
5 Sample collection and pre-treatment for drinking, surface, ground and rain-water samples	3
6 Determination of mercury after tin(II) chloride reduction and enrichment by amalgamation	4
6.1 Reagents and standards	4
6.2 Apparatus and instrumentation	6
6.3 Instrumental set-up	8
6.4 Procedure	8
6.5 Calculation	9
7 Determination of mercury after tin(II) chloride reduction without enrichment	9
7.1 Reagents and standards	9
7.2 Apparatus and instrumentation	9
7.3 Sample collection and pre-treatment for drinking, surface, ground and rain-water samples	10
7.4 Sample collection and pre-treatment for waste-water samples	10
7.5 Instrumental set-up	10
7.6 Procedure	10
7.7 Calculation	11
8 Expression of results	11
9 Test report	11
Annex A (informative) Performance data	12
Bibliography	15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 12846 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

This first edition of ISO 12846 cancels and replaces the editions (ISO 5666:1999, ISO 16590:2000), which have been technically revised.

Introduction

In natural water sources, mercury compounds generally occur in very low concentrations of less than 0,1 µg/l. Higher concentrations may be found, for example, in industrial waste water. Both inorganic and organic compounds of mercury may be present. Mercury can also accumulate in sediments and sludges.

In order to fully decompose all of the mercury compounds in the presence of particles in the sample, an additional digestion procedure is necessary. This additional digestion can be omitted only if significant amounts of previous comparison data clearly demonstrate this.

For reliable measurements in the low-concentration range, the highest purity reagents, clean vessels, mercury-free air in the laboratory and a very stable measurement system are essential.

This International Standard is a state-of-the-art revision of existing standards for the determination of mercury by AAS with and without a pre-enrichment step combining the advantages of the existing methods with new developments and technique. The following methods are considered:

A) Methods without enrichment:

- ISO 5666:1999, *Water quality — Determination of mercury*;
- EN 1483:2007, *Water quality — Determination of mercury — Method using atomic absorption spectrometry*.

B) Methods with enrichment:

- ISO 16590:2000, *Water quality — Determination of mercury — Methods involving enrichment by amalgamation*;
- EN 12338:1998, *Water quality — Determination of mercury — Enrichment methods by amalgamation*.

Water quality — Determination of mercury — Method using atomic absorption spectrometry (AAS) with and without enrichment

WARNING – Potassium bromate, used in significant quantities for sample preservation in this International Standard, is carcinogenic and suitable safety precautions shall be taken. Precautions should be taken to detoxify any residual bromate by reduction to bromide before disposal.

Persons using this International Standard should be familiar with normal laboratory practice. This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted in accordance with this International Standard be carried out by suitably qualified staff. Mercury and mercury compounds are very toxic. Extreme caution should be exercised when handling samples and solutions which contain or may contain mercury.

1 Scope

This International Standard specifies two methods for the determination of mercury in drinking, surface, ground, rain and waste water after appropriate pre-digestion. For the first method (described in Clause 6), an enrichment step by amalgamation of the Hg on, for example, a gold/platinum adsorber is used. For the method given in Clause 7, the enrichment step is omitted.

The choice of method depends on the equipment available, the matrix and the concentration range of interest. Both methods are suitable for the determination of mercury in water. The method with enrichment (see Clause 6) commonly has a practical working range from 0,01 µg/l to 1 µg/l. The mean limit of quantification (LOQ) reported by the participants of the validation trial (see Annex A) was 0,008 µg/l. This information on the LOQ gives the user of this International Standard an orientation and does not replace the estimation of performance data based on laboratory-specific data. It has to be considered that it is possible to achieve lower LOQs with specific instrumentation (e.g. single mercury analysers).

The method without enrichment (in Clause 7) commonly has a practical working range starting at 0,05 µg/l. The LOQ reported by the participants of the validation trial (see Annex A) was 0,024 µg/l. It is up to the user, based on the specific application, to decide whether higher concentrations are determined by omitting the enrichment step and/or by diluting the sample(s). The sensitivity of both methods is dependent on the selected operating conditions.

Another possibility for the determination of extremely low Hg concentrations down to 0,002 µg/l without pre-concentration is the application of atomic fluorescence spectrometry (see ISO 17852). Specific atomic-absorption mercury analysers allow determinations down to 0,010 µg/l without pre-concentration.

In general, the determination of trace concentrations of Hg by AAS (or AFS) is dependent on clean operating conditions in the laboratory and on the use of high-purity chemicals with negligible low-Hg blanks.

NOTE This International Standard may be applied to industrial and municipal waste water after an additional digestion step performed under appropriate conditions and after suitable method validation (see 7.4). A potential sample stability issue (mercury loss) for anaerobic reducing industrial effluents has to be considered thoroughly.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes and sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Preservation and handling of water samples*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

ISO 8466-2, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 2: Calibration strategy for non-linear second order calibration functions*

3 Principle

Mono- or divalent mercury and organo-mercury compounds as well are converted to divalent mercury by oxidation with KBrO_3 -KBr and then reduced to the elemental form by tin(II) chloride in an acid medium.

Elemental mercury is then stripped from the solution with the aid of a stream of inert gas or mercury-free air.

Alternatively, samples may be preserved with dichromate/ HNO_3 and digested under appropriate conditions (e.g. using potassium permanganate/potassium peroxodisulfate). This technique is not part of this International Standard and has therefore to be validated appropriately by the user.

If the enrichment step is applied, the mercury (in the form of atomic vapour) is transported in a stream of inert gas with negligible mercury content to a quartz tube with a suitable heating and adsorbent (e.g. gold-platinum gauze) on which the mercury is adsorbed.

Other adsorbents based on the principle of amalgamation are allowed if the user demonstrates fitness for purpose.

The mercury is then released by rapid heating of the adsorbent (desorption at 600 °C as a minimum) and further transported in a stream of carrier gas to the absorption cell where the absorbance is measured at 253,7 nm in the radiation beam of an atomic absorption spectrometer. Concentrations are calculated using a calibration curve or using the standard addition method.

If the enrichment step is omitted, the mercury is transported directly into a cuvette. Absorbances are also measured at a wavelength of 253,7 nm.

4 General interferences

With mercury, there is a risk that exchange reactions, that is adsorption and desorption, will occur on the walls of the sampling and reaction vessels. The instructions in 6.2.5.2 should therefore be followed exactly.

Mercury vapour can diffuse through various plastics; this phenomenon needs to be taken into consideration in the choice of tubing material. Glass or special plastics tubing, e.g. FEP tubes [FEP = perfluoro(ethene-propene)], should be used. Silicone tubing, for example, is unsuitable.

Easily reducible elements, such as gold, silver and platinum, amalgamate with elemental mercury and may therefore cause suppression. Additionally, iodide is a significant interferent causing severe suppression of the mercury signal.

Volatile organic substances can absorb in the UV range and although automatic background correction is used to compensate for this, it still results in a degraded signal-to-noise ratio.

All solutions should be brought to the same temperature (<25 °C) before reduction and stripping of the mercury vapour. Water condensation on the cuvette windows has to be prevented, e.g. by heating the cuvette with, for example, an infrared lamp.

NOTE Tin(II) chloride causes extensive contamination of the apparatus with tin. Considerable interferences occur if tin has to be determined afterwards on the same system.