

Molecular in vitro diagnostic examinations -
Specifications for pre-examination processes for venous
whole blood - Part 2: Isolated genomic DNA (ISO
20186-2:2019)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 20186-2:2019 sisaldab Euroopa standardi EN ISO 20186-2:2019 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 20186-2:2019 consists of the English text of the European standard EN ISO 20186-2:2019.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 27.03.2019.	Date of Availability of the European standard is 27.03.2019.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.100.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

**Molecular in vitro diagnostic examinations - Specifications
for pre-examination processes for venous whole blood -
Part 2: Isolated genomic DNA (ISO 20186-2:2019)**

Analyses de diagnostic moléculaire in vitro -
Spécifications relatives aux processus préanalytiques
pour le sang total veineux - Partie 2: ADN génomique
extrait (ISO 20186-2:2019)

Molekularanalytische in-vitro-diagnostische Verfahren
- Spezifikationen für präanalytische Prozesse für
venöse Vollblutproben - Teil 2: Isolierte genomische
DNA (ISO 20186-2:2019)

This European Standard was approved by CEN on 2 February 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 20186-2:2019) has been prepared by Technical Committee ISO/TC 212 "Clinical laboratory testing and in vitro diagnostic test systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2019, and conflicting national standards shall be withdrawn at the latest by March 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16835-2:2015.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 20186-2:2019 has been approved by CEN as EN ISO 20186-2:2019 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 General considerations	5
5 Outside the laboratory	6
5.1 Specimen collection	6
5.1.1 Information about the specimen donor/patient	6
5.1.2 Selection of the venous whole blood collection tube by the laboratory	6
5.1.3 Venous whole blood specimen collection from the donor/patient and stabilization procedures	6
5.1.4 Information about the specimen and storage requirements at the blood collection facility	7
5.2 Transport requirements	8
6 Inside the laboratory	8
6.1 Specimen reception	8
6.2 Storage requirements	8
6.3 Isolation of the genomic DNA	10
6.3.1 General	10
6.3.2 Examination provider's instructions available	10
6.3.3 Examination provider's instructions not available	10
6.4 Quantity and quality assessment of isolated genomic DNA	11
6.5 Storage of isolated genomic DNA	11
6.5.1 General	11
6.5.2 Genomic DNA isolated with commercially available kits	11
6.5.3 Genomic DNA isolated with the laboratory's own protocols	12
Annex A (informative) Impact of pre-examination process steps on venous whole blood genomic DNA quality	13
Bibliography	19

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

A list of all parts in the ISO 20186 series can be found on the ISO website.

Introduction

Molecular in vitro diagnostics has enabled significant progress in medicine. Further progress is expected by new technologies analysing profiles of nucleic acids, proteins, and metabolites in human tissues and body fluids. However, the profiles of these molecules can change drastically during the pre-examination process, including the specimen collection, transport, storage and processing. Consequently, this makes the outcome from diagnostics or research unreliable or even impossible, because the subsequent examination might not determine the real situation in the patient but an artificial profile generated during the pre-examination processes.

Genomic DNA can fragment or degrade after blood collection. Therefore, special measures need to be taken to secure good quality specimens for genomic DNA examination. This is particularly relevant for examination test procedures requiring high molecular weight DNA (HMW DNA).

Standardization of the entire workflow from specimen collection to the genomic DNA examination is needed due to genomic DNA degradation and fragmentation after blood collection. Studies have been undertaken to determine the important influencing factors. This document draws upon such work to codify and standardize the steps for venous whole blood genomic DNA examination in what is referred to as the pre-examination phase.

In this document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission;
- “can” indicates a possibility or a capability.

Molecular in vitro diagnostic examinations — Specifications for pre-examination processes for venous whole blood —

Part 2: Isolated genomic DNA

1 Scope

This document gives guidelines on the handling, storage, processing and documentation of venous whole blood specimens intended for genomic DNA examination during the pre-examination phase before a molecular examination is performed. This document covers specimens collected in venous whole blood collection tubes.

This document is applicable to any molecular in vitro diagnostic examination performed by medical laboratories. It is also intended to be used by laboratory customers, in vitro diagnostics developers and manufacturers, biobanks, institutions and commercial organizations performing biomedical research, and regulatory authorities.

Different dedicated measures are taken for stabilizing blood cell free circulating DNA, which are not described in this document.

NOTE Circulating cell free DNA in blood is covered in ISO 20186-3.

Different dedicated measures are taken for collecting, stabilizing, transporting and storing capillary blood as well as for collecting and storing blood by paper based technologies or other technologies generating dried blood. These are not described in this document.

This document does not cover the isolation of specific blood cells and subsequent isolation of genomic DNA therefrom.

DNA in pathogens present in blood is not covered by this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15189:2012, *Medical laboratories — Requirements for quality and competence*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>