Geometrical product specifications (GPS) - General concepts and requirements for GPS measuring equipment (ISO 14978:2018)



EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 14978:2018 sisaldab Euroopa standardi EN ISO 14978:2018 ingliskeelset teksti.	'					
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.					
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.12.2018.	Date of Availability of the European standard is 19.12.2018.					
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.					

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile <u>standardiosakond@evs.ee</u>.

ICS 17.040.30

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega: Koduleht <u>www.evs.ee</u>; telefon 605 5050; e-post <u>info@evs.ee</u>

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

NORME EUROPÉENNE

EN ISO 14978

EUROPÄISCHE NORM

December 2018

ICS 17.040.30

Supersedes EN ISO 14978:2006

English Version

Geometrical product specifications (GPS) - General concepts and requirements for GPS measuring equipment (ISO 14978:2018)

Spécification géométrique des produits (GPS) -Concepts et exigences généraux pour les équipements de mesure GPS (ISO 14978:2018) Geometrische Produktspezifikation (GPS) - Allgemeine Begriffe und Anforderungen für GPS-Messeinrichtungen (ISO 14978:2018)

This European Standard was approved by CEN on 1 October 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 14978:2018) has been prepared by Technical Committee ISO/TC 213 "Dimensional and geometrical product specifications and verification" in collaboration with Technical Committee CEN/TC 290 "Dimensional and geometrical product specification and verification" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2019, and conflicting national standards shall be withdrawn at the latest by June 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 14978:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 14978:2018 has been approved by CEN as EN ISO 14978:2018 without any modification.

Introduction	Co	ntents	;		Page		
1 Scope	Fore	word			v		
2 Normative references 1 3 Terms and definitions 1 3.1 ISO/IEC Guide 99:2007 terms 2 3.2 Synonym terms to ISO/IEC Guide 99:2007 terms 2 3.3 ISO/IEC Guide 98-4:2012 terms 3 3.4 ISO/14253-5:2015 terms 3 3.5 Terms related to GPS measuring equipment 3 4 Abbreviations 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.1 Metrological characteristics for indicating measuring instruments 11 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 1 6.1.2 Standards for measuring equipment 1 6.1.2 Standards for measuri	Intr	oduction	l		vi		
2 Normative references 1 3 Terms and definitions 1 3.1 ISO/IEC Guide 99:2007 terms 2 3.2 Synonym terms to ISO/IEC Guide 99:2007 terms 2 3.3 ISO/IEC Guide 98-4:2012 terms 3 3.4 ISO/14253-5:2015 terms 3 3.5 Terms related to GPS measuring equipment 3 4 Abbreviations 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.1 Metrological characteristics for indicating measuring instruments 11 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 1 6.1.2 Standards for measuring equipment 1 6.1.2 Standards for measuri	1	Scope			1		
3.1 ISO/IEC Guide 99:2007 terms 2.3.1 ISO/IEC Guide 99:2007 terms 2.3.2 Synonym terms to ISO/IEC Guide 99:2007 terms 2.3.3 ISO/IEC Guide 98-4:2012 terms 3.3.4 ISO 14253-5:2015 terms 3.3.5 ISO 14353-5:2015 te	_						
3.1 ISO/IEC Guide 99:2007 terms 2 3.2 Synonym terms to ISO/IEC Guide 99:2007 terms 2 3.3 ISO/IEC Guide 98-4:2012 terms 3 3.4 ISO 14253-5:2015 terms 3 3.5 Terms related to GPS measuring equipment 10 5 10 Importance of design characteristics 10 5 1.1 Importance of design characteristics 10 5 1.2 Standards for measuring equipment 11 5 1.2 Standards for measuring equipment 11 5 1.3 Measuring equipment 11 5 1.4 Measuring equipment 11 5 1.2 Standards for measuring instruments 11 5 2 Design characteristics for indicating measuring instruments 11 5 3 Design characteristics for indicating measuring instruments 11 6 1.1 Importance of metrological characteristics 13 6 6 Metrological characteristics for material measures 12 6 1.1 Importance of metrological characteristics 13 6 6 1.2 Standards for measuring equipment 13 6 6 1.2 Standards for measuring equipment 13 6 6 1.5 Galibration and verification of measuring equipment 14 6 6 5 6 6 Measuring equipment 6 6 6 6 6 6 6 6 6							
3.2 Synonym terms to ISO/IEC Guide 99:2007 terms 2 3.3 1SO/IEC Guide 98-4:2012 terms 3 3.4 ISO 14253-5:2015 terms 3 3.5 Terms related to GPS measuring equipment 3 4 Abbreviations 10 5 Design characteristics 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Commerce 11 5.2 Design characteristics for indicating measuring instruments 11 5.2 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.2 Standards for measuring equipment 13	3						
3.3 ISO/IEC Guide 98-4:2012 terms 3 3.4 ISO 14253-5:2015 terms 3 3.5 Terms related to GPS measuring equipment 3 3 4 Abbreviations 10 5 Design characteristics 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics for material measures 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Internal use in a company 17 6.1 Indicating measuring instruments 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment (19 6.3.1 General 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 6.3.8 Tom of feature characteristics 20 6.3.9 Tom of feature characteristics 20 6.3.1 General 20 6.3.2 Specification of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 22			Synonym terms to ISO/IFC Guide 99:2007 terms				
3.4 ISO 14253-5:2015 terms 3.5 Terms related to GPS measuring equipment 3.6 Abbreviations 10			ISO/IEC Guide 98-4:2012 terms				
4 Abbreviations 10 5 Design characteristics 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Calibration and verification of measuring equipment 14 6.1.							
5 Design characteristics 10 5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and verification of metrological characteristics 13 6.1.4 Calibration and verification methods 15 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Internal use in a compan		3.5	Terms	related to GPS measuring equipment	3		
5.1 General 10 5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification of measuring equipment 14 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2	4	Abbre	viation	18	10		
5.1.1 Importance of design characteristics 10 5.1.2 Standards for measuring equipment 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.2 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Internal use in a company 17 6.2.1 Indicating measuring instruments 17 6.2.2 Scale interval — Resolution 18 6.2.1 General 17 6.2.2 Scale interval — Resol	5	Desig	Design characteristics				
5.1.2 Standards for measuring equipment — Commerce 11 5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification methods 15 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18		5.1	Genera				
5.1.3 Measuring equipment — Commerce 11 5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.6 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 </td <td></td> <td></td> <td>0.2.2</td> <td></td> <td></td>			0.2.2				
5.1.4 Measuring equipment — Internal use in a company 11 5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics							
5.2 Design characteristics for indicating measuring instruments 11 5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.2 Scale interval — Resolution 18 6.2.2 Error of indication 18 6.2.2 Error of indication 18 6.2.3 Error of indication 18 6.2.4 Error of indication 18 6.2.5							
5.3 Design characteristics for material measures 12 6 Metrological characteristics 13 6.1 General 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 Geome		5.2	_				
6 Metrological characteristics 13 6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.2 Standards for measuring equipment 13 6.1.2 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6							
6.1 General 13 6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 <td< td=""><td></td><td></td><td>U</td><td></td><td></td></td<>			U				
6.1.1 Importance of metrological characteristics 13 6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.5	6						
6.1.2 Standards for measuring equipment 13 6.1.3 Identification, definition and choice of metrological characteristics 13 6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.5		0.1					
6.1.3 Identification, definition and choice of metrological characteristics 6.1.4 Calibration and verification of measuring equipment 6.1.5 Calibration and verification methods 6.1.6 Measuring equipment — Commerce 7.6.1.7 Measuring equipment — Internal use in a company 7.6.2 Indicating measuring instruments 7.7 6.2.1 General 7.7 6.2.2 Scale interval — Resolution 7.8 6.2.3 Digital step 7.9 6.2.4 Error of indication 7.0 6.2.5 Temperature-related metrological characteristics 7.0 6.2.6 Characteristics related to measuring force 7.0 6.2.7 Geometry of contact element 7.0 6.2.8 Auxiliary equipment 7.0 6.2.8 Auxiliary equipment 7.0 6.2.9 General 7.0 6.3.1 General 7.0 6.3.2 Scale interval — Resolution of reading 7.0 6.3.3 Form of feature characteristics 7.0 6.3.4 Orientation of feature characteristics 7.0 6.3.5 Temperature-related metrological characteristics 7.0 Geometrical stability 7.0 6.3.7 Other possible metrological characteristics 7.1 General 7.2 Specification and presentation of metrological characteristics 7.1 General 7.2 Specification and presentation of metrological characteristics 7.1 General 7.2 Specification of metrological characteristics 7.1 General				Standards for measuring equipment.	13		
6.1.4 Calibration and verification of measuring equipment 14 6.1.5 Calibration and verification methods 15 6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.2.9 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characterist							
6.1.6 Measuring equipment — Commerce 17 6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21			6.1.4	Calibration and verification of measuring equipment	14		
6.1.7 Measuring equipment — Internal use in a company 17 6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21							
6.2 Indicating measuring instruments 17 6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 21 7.1 General 21 7.2 Specification and presentation of metrological characteristics <							
6.2.1 General 17 6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21 7.2 Specification and presentation of metrological characteristics 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22		6.0					
6.2.2 Scale interval — Resolution 18 6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2.1 General 21 7.2.1 General 21 7.2.2 Constant value MPE function 21		6.2		ting measuring instruments	1/		
6.2.3 Digital step 18 6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21 7.2 Specification and presentation of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
6.2.4 Error of indication 18 6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21 7.2 Specification and presentation of metrological characteristics 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
6.2.5 Temperature-related metrological characteristics 19 6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22				Error of indication	18		
6.2.6 Characteristics related to measuring force 19 6.2.7 Geometry of contact element 19 6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7.1 General 21 7.2 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22			6.2.5	Temperature-related metrological characteristics	19		
6.2.8 Auxiliary equipment 19 6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 20 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 21 7.2.3 Proportional value MPE function 22				Characteristics related to measuring force	19		
6.3 Material measures 20 6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
6.3.1 General 20 6.3.2 Scale interval — Resolution of reading 20 6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22		6.0					
6.3.2 Scale interval — Resolution of reading		6.3					
6.3.3 Form of feature characteristics 20 6.3.4 Orientation of feature characteristics 20 6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22				Scale interval — Pecalution of reading	20		
6.3.4Orientation of feature characteristics206.3.5Temperature-related metrological characteristics206.3.6Geometrical stability206.3.7Other possible metrological characteristics207Specification and presentation of metrological characteristics217.1General217.2Specification of metrological characteristics217.2.1General217.2.2Constant value MPE function217.2.3Proportional value MPE function22							
6.3.5 Temperature-related metrological characteristics 20 6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
6.3.6 Geometrical stability 20 6.3.7 Other possible metrological characteristics 20 7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
7 Specification and presentation of metrological characteristics 21 7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22			6.3.6	Geometrical stability	20		
7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22							
7.1 General 21 7.2 Specification of metrological characteristics 21 7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22	7	Specif	fication	and presentation of metrological characteristics	21		
7.2.1 General 21 7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22		7.1	7.1 General				
7.2.2 Constant value MPE function 21 7.2.3 Proportional value MPE function 22		7.2					
7.2.3 Proportional value MPE function							
1							

EVS-EN ISO 14978:2018

	7.3		itation of characteristic curves	
		7.3.1 7.3.2	General Presentation of characteristic curves – Reference point	
8	Calib		f metrological characteristics	
U	8.1	Manuf	acturer and supplier of measuring equipment	26
	8.2		f measuring equipment	
_	8.3		rement uncertainty	
9				
10			Is for specific measuring equipment	27
Annex			General minimum requirements and guidance for clauses in GPS r specific measuring equipment	28
Annex	x B (inf	formative	e) Data sheet for measuring equipment requirements	31
			Common design characteristics	
			e) Test uncertainty	
			e) Relation to the GPS matrix model	
				5
iv			© ISO 2018 – All r	ghts reserved

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee 213, *Dimensional and geometrical product specifications and verification*.

This second edition cancels and replaces the first edition (ISO 14978:2006), which has been technically revised. It also incorporates the Technical Corrigendum ISO 14978:2006/Cor. 1:2008.

The main changes compared to the previous edition are as follows:

- the terms and definitions have been updated relative to ISO/IEC Guide 99:2007;
- a number of design characteristics common in GPS measuring equipment have been added;
- an updated discussion of calibration and verification, including concepts from ISO 14253-5:2015, has been added.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

2/5

Introduction

This document is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO 14638). It influences chain links F and G for measuring equipment and calibration in the general GPS matrix model (see <u>Annex E</u>).

The ISO/GPS matrix model given in ISO 14638 gives an overview of the ISO/GPS system of which this document is a part. The fundamental rules of ISO/GPS given in ISO 8015 apply to this document and the default decision rules given in ISO 14253-1 apply to specifications made in accordance with this document, unless otherwise indicated; see ISO/TR 14253-6 for additional information on the selection of alternative decision rules.

For more detailed information of the relation of this document to other standards and the GPS matrix model, see Annex E.

This document contains guidance for writing the standards for specific GPS measuring equipment.

This document is intended to give the user a basic understanding of the use of ISO standards for GPS measuring equipment. This document presents and defines general concepts to be used in connection with GPS measuring equipment to avoid multiple repetitions in the ISO standards for specific GPS measuring equipment. This document is also intended as guidance for the manufacturer/supplier to evaluate and present specifications for characteristics for GPS measuring equipment.

str.
ad usin, This document is necessary when reading and using ISO standards for specific GPS measuring equipment.

Geometrical product specifications (GPS) — General concepts and requirements for GPS measuring equipment

1 Scope

This document specifies the general requirements, calibration, terms and definitions of characteristics of GPS measuring equipment, for example micrometers, callipers, gauge blocks and rotary axis form measuring instruments. This document forms the basis for standards defining and describing the design characteristics and metrological characteristics for measuring equipment and gives guidance for the development and content of standards for GPS measuring equipment.

This document is intended to ease the communication between manufacturer/supplier and customer/ user and to make the specification phase of GPS measuring equipment more accurate. This document is also intended as a tool to be used in companies in the process of defining and selecting relevant characteristics for measuring equipment.

This document includes terms which are frequently used in connection with the characterization of specific measuring equipment.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14253-1, Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for verifying conformity or nonconformity with specifications

ISO 14253-5:2015, Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 5: Uncertainty in verification testing of indicating measuring instruments

ISO/TR 14253-6, Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 6: Generalized decision rules for the acceptance and rejection of instruments and workpieces

ISO/IEC Guide 98-3, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 98-4:2012, Uncertainty of measurement — Part 4: Role of measurement uncertainty in conformity assessment

ISO/IEC Guide 99:2007, International vocabulary of metrology — Basic and general concepts and associated terms (VIM)

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO/IEC Guide 99:2007, ISO/IEC Guide 98-4:2012, ISO 14253-1, ISO 14253-5:2015, ISO/TR 14253-6 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at https://www.iso.org/obp
- IEC Electropedia: available at http://www.electropedia.org/