

Ophthalmic implants - Intraocular lenses - Part 7:
Clinical investigations of intraocular lenses for the
correction of aphakia (ISO 11979-7:2018)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 11979-7:2018 sisaldab Euroopa standardi EN ISO 11979-7:2018 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 11979-7:2018 consists of the English text of the European standard EN ISO 11979-7:2018.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 02.05.2018.	Date of Availability of the European standard is 02.05.2018.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.70

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Ophthalmic implants - Intraocular lenses - Part 7: Clinical investigations of intraocular lenses for the correction of aphakia (ISO 11979-7:2018)

Implants ophtalmiques - Lentilles intraoculaires -
Partie 7: Investigations cliniques de lentilles
intraoculaires pour la correction de l'aphakie (ISO
11979-7:2018)

Ophthalmische Implantate - Intraokularlinsen - Teil 7:
Klinische Prüfungen von Intraokularlinsen für die
Korrektion von Aphakie (ISO 11979-7:2018)

This European Standard was approved by CEN on 28 February 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 11979-7:2018) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2018, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11979-7:2014 and EN ISO 11979-9:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11979-7:2018 has been approved by CEN as EN ISO 11979-7:2018 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms, definitions and abbreviated terms	1
3.1 Terms and definitions	1
3.2 Abbreviated terms	1
4 Justification for a clinical investigation	2
5 Ethical considerations	2
6 General requirements	2
6.1 General	2
6.2 Design of a clinical investigation	2
6.2.1 Requirements for all types of IOLs	2
6.2.2 Additional requirements for toric IOLs (TIOLs)	3
6.2.3 Additional requirements for multifocal IOLs (MIOLs)	3
6.2.4 Additional requirements for accommodating IOLs (AIOLs)	3
6.2.5 Additional requirements for anterior chamber IOLs	3
6.3 Characteristics of clinical investigations	4
6.3.1 General	4
6.3.2 Characteristics to be studied for all types of IOLs	4
6.3.3 Additional characteristics to be studied for TIOLs	4
6.3.4 Additional characteristics to be studied for MIOLs	5
6.3.5 Additional characteristics to be studied for AIOLs	5
6.3.6 Additional characteristics applying to anterior chamber IOLs	5
6.3.7 Additional characteristics	5
6.4 Duration of the investigations	5
6.5 Enrolment	6
6.6 Bilateral implantation	6
6.7 Surgical technique	6
6.8 Examination and treatment of subjects	6
6.9 Adverse events reports	7
6.10 Inclusion and exclusion criteria	7
6.10.1 General	7
6.10.2 Additional inclusion criteria for TIOLs	7
6.10.3 Additional exclusion criteria for MIOLs	7
6.10.4 Additional exclusion criteria for anterior chamber IOLs	8
Annex A (informative) General elements in the clinical investigation of IOLs	9
Annex B (informative) Additional elements for the clinical investigation of TIOLs	14
Annex C (informative) Additional elements for the clinical investigation of multifocal IOLs	19
Annex D (informative) Additional elements in the clinical investigation of accommodating IOLs	22
Annex E (informative) Evaluation of post-operative adverse events and visual acuity rates	27
Annex F (informative) Clinical tests	31
Annex G (informative) Statistical methods and sample size calculations	37
Bibliography	42

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This fourth edition cancels and replaces the third edition (ISO 11979-7:2014). It also cancels and replaces the first edition of ISO 11979-9:2006 and its amendment ISO 11979-9:2006/Amd 1:2014.

The main changes compared to the previous edition are as follows:

- Integration of the multifocal intraocular lens document (ISO 11979-9:2006);
- Technical updates concerning the safety and efficacy of the intraocular lens subtypes monofocal, multifocal, toric and accommodating;
- Recommendations for the clinical investigations of novel lens models; and
- The separation of guidance for intraocular lenses used in cases of aphakia, and intraocular lens used for the correction of ametropia in phakic patients.

A list of all parts in the ISO 11979 series can be found on the ISO website.

Introduction

Intraocular lenses (IOLs) are used to correct residual refractive errors in subjects who have aphakia. Such residual refractive errors typically include sphere and astigmatism, but can also include accommodation. Different designs of IOLs can be used to correct for specific refractive errors. In the case where an IOL is designed to provide more than one type of refractive correction, that IOL will have to satisfy each of the separate requirements of those correction designs.

This document provides requirements and recommendations for intraocular lens investigations of new IOL models. In the case where an IOL model is a modification of a parent IOL model, a risk analysis can be used in order to determine the appropriate level of testing.

Ophthalmic implants — Intraocular lenses —

Part 7:

Clinical investigations of intraocular lenses for the correction of aphakia

1 Scope

This document specifies the particular requirements for the clinical investigations of intraocular lenses that are implanted in the eye in order to correct aphakia.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

ISO 11979-10:2018, *Ophthalmic implants — Intraocular lenses — Part 10: Clinical investigations of intraocular lenses for correction of ametropia in phakic eyes*

ISO 14155, *Clinical investigation of medical devices for human subjects — Good clinical practice*

ISO 14971, *Medical devices — Application of risk management to medical devices*

3 Terms, definitions and abbreviated terms

3.1 Terms and definitions

For the purposes of this document the terms and definitions given in ISO 11979-1 and ISO 14155 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.2 Abbreviated terms

UDVA	uncorrected distance visual acuity
UIVA	uncorrected intermediate visual acuity
UNVA	uncorrected near visual acuity
CDVA	corrected distance visual acuity
CIVA	corrected intermediate visual acuity