

Preparation and quality management of fluids for
haemodialysis and related therapies - Part 1: General
requirements (ISO 23500-1:2019)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 23500-1:2019 sisaldab Euroopa standardi EN ISO 23500-1:2019 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 23500-1:2019 consists of the English text of the European standard EN ISO 23500-1:2019.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 20.03.2019.	Date of Availability of the European standard is 20.03.2019.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.40

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Preparation and quality management of fluids for
haemodialysis and related therapies - Part 1: General
requirements (ISO 23500-1:2019)

Préparation et management de la qualité des liquides
d'hémodialyse et de thérapies annexes - Partie 1:
Exigences générales (ISO 23500-1:2019)

Leitfaden für die Vorbereitung und das
Qualitätsmanagement von Konzentraten für die
Hämodialyse und verwandte Therapien - Teil 1:
Allgemeine Anforderungen (ISO 23500-1:2019)

This European Standard was approved by CEN on 18 January 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 23500-1:2019) has been prepared by Technical Committee ISO/TC 150 "Implants for surgery" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2019, and conflicting national standards shall be withdrawn at the latest by September 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 23500:2015.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 23500-1:2019 has been approved by CEN as EN ISO 23500-1:2019 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
1.1 General	1
1.2 Inclusions	1
1.3 Exclusions	2
2 Normative references	2
3 Terms and definitions	2
4 Quality requirements	9
4.1 General	9
4.2 Dialysis water	9
4.2.1 General	9
4.2.2 Chemical contaminants in dialysis water	9
4.2.3 Organic Carbon, pesticides and other chemicals	11
4.2.4 Microbiological contaminants in dialysis water	11
4.3 Requirements for concentrate	12
4.3.1 Chemical and microbiological contaminants in concentrate	12
4.3.2 Water used to prepare concentrate	12
4.4 Requirements for dialysis fluid	12
4.4.1 General	12
4.4.2 Microbiological requirements for standard dialysis fluid	13
4.4.3 Microbiological requirements for ultrapure dialysis fluid	13
4.4.4 Microbiological requirements for online-prepared substitution fluid	13
4.5 Record retention	13
5 Critical aspects of system design	14
5.1 General	14
5.2 Technical aspects	14
5.3 Microbiological aspects	15
5.4 Environmental impact	16
6 Validation of system performance	16
6.1 General	16
6.2 Validation plan	17
6.3 Installation and operational qualification	17
6.4 Performance qualification	18
6.5 Routine surveillance and revalidation	18
7 Quality management	19
7.1 General	19
7.2 Surveillance of fluid quality	19
7.2.1 Surveillance of dialysis water quality	19
7.2.2 Surveillance of concentrate quality	20
7.2.3 Surveillance of dialysis fluid quality	20
7.3 Surveillance of water treatment equipment	20
7.3.1 General	20
7.3.2 Surveillance of sediment filters	20
7.3.3 Surveillance of cartridge filters	21
7.3.4 Surveillance of softeners	21
7.3.5 Surveillance of carbon media	21
7.3.6 Surveillance of chemical injection systems	22
7.3.7 Surveillance of reverse osmosis	22
7.3.8 Surveillance of deionization	24
7.3.9 Surveillance of endotoxin-retentive filters	24
7.4 Surveillance of dialysis water storage and distribution	24

7.4.1	Surveillance of water storage tanks	24
7.4.2	Surveillance of the water distribution systems	24
7.4.3	Surveillance of bacterial control devices	25
7.5	Surveillance of concentrate preparation	25
7.5.1	Surveillance of mixing systems	25
7.5.2	Surveillance of additives	26
7.6	Surveillance of concentrate distribution	26
7.7	Surveillance of dialysis fluid proportioning	26
8	Strategies for microbiological control	26
8.1	General	26
8.2	Disinfection	27
8.2.1	General	27
8.2.2	Microbiological aspects of fluid system design	27
8.2.3	Disinfection frequency	28
8.3	Microbiological surveillance methods	29
8.3.1	General	29
8.3.2	Sample collection	29
8.3.3	Heterotrophic plate count	30
8.3.4	Bacterial endotoxin test	32
8.3.5	Determination of yeast and mould	32
9	Location of and access to the water treatment system	32
10	Personnel	32
Annex A	(informative) Rationale for the development and provisions of this document	33
Annex B	(informative) Equipment	38
Annex C	(informative) Surveillance guidelines for water treatment equipment, distribution systems, and dialysis fluid	56
Annex D	(informative) Strategies for microbiological control	61
Annex E	(informative) Validation	68
Annex F	(informative) Special considerations for home haemodialysis	71
Annex G	(informative) Special considerations for acute haemodialysis	77
Bibliography		82

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 150, *Implants for surgery*, Subcommittee SC 2, *Cardiovascular implants and extracorporeal systems*.

This first edition cancels and replaces ISO 23500:2014, which has been technically revised. The main changes compared to the previous edition are as follows:

- The document forms part of a revised and renumbered series dealing with the preparation and quality management of fluids for haemodialysis and related therapies. The series comprise ISO 23500-1 (previously ISO 23500), ISO 23500-2, (previously ISO 26722), ISO 23500-3, (previously ISO 13959), ISO 23500-4, (previously ISO 13958), and ISO 23500-5, (previously ISO 11663).

A list of all parts in the ISO 23500 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document is the base standard for a number of other standards dealing with water treatment and the production of dialysis fluid (ISO 23500 series).

The objective of the ISO 23500 series is to provide users with guidance for handling water and concentrates and for the production and quality oversight of dialysis fluid used for haemodialysis. The need for such guidance is based on the critical role of dialysis fluid quality in providing safe and effective haemodialysis, and the recognition that day-to-day dialysis fluid quality is under the control of the healthcare professionals who deliver dialysis therapy.

[Annex A](#) provides further information on the rationale for the development and provisions of this document.

The equipment used in the various stages of dialysis fluid preparation is generally obtained from specialized vendors. Dialysis practitioners are generally responsible for maintaining that equipment following its installation. Therefore, this document provides guidance on quality oversight and maintenance of the equipment to ensure that dialysis fluid quality is acceptable at all times. At various places throughout this International Standard, the user is advised to follow the manufacturer's instructions regarding the operation and maintenance of equipment. In those instances in which the equipment is not obtained from a specialized vendor, it is the responsibility of the user to validate the performance of the equipment in the haemodialysis setting and to ensure that appropriate operating and maintenance manuals are available.

[Annex B](#) to this document provides further information on the system components that are used for water treatment, concentrate, and dialysis fluid preparation at a dialysis facility. These descriptions are intended to provide the user with a basis for understanding why certain equipment might be required and how it should be configured; they are not intended as detailed design standards. Requirements for water treatment equipment are provided in ISO 23500-2.

Increasingly, self-contained, integrated systems designed and validated to produce water and dialysis fluid are becoming available and used clinically. This document applies to systems assembled from individual components. Consequently, some of the requirements in ISO 23500-1 and ISO 23500-2 might not apply to integrated systems, however such systems are required to comply with the requirements of ISO 23500-3, ISO 23500-4, and ISO 23500-5. In order to ensure conformity when using such systems, adherence to the manufacturer's instructions regarding the operation, testing, and maintenance of such systems is required to ensure that the system is being operated under the validated conditions.

This document reflects the conscientious efforts of healthcare professionals, patients, and medical device manufacturers to develop recommendations for handling water and concentrates and for the production and surveillance of dialysis fluid for haemodialysis and protecting haemodialysis patients from adverse effects arising from known chemical and microbial contaminants that might be found in improperly prepared dialysis fluid. [Annexes F](#) and [G](#) provide further information in respect of special considerations for home and acute haemodialysis. The standard together with its constituent parts is directed towards the healthcare professionals involved in the management or routine care of haemodialysis patients and responsible for the quality of dialysis fluid. However, the physician in charge of dialysis has the ultimate responsibility for ensuring that the dialysis fluid is correctly formulated and meets the requirements of all applicable quality standards.

The provisions contained in this document might not be applicable in all circumstances and they are not intended for regulatory application.

Preparation and quality management of fluids for haemodialysis and related therapies —

Part 1: General requirements

1 Scope

1.1 General

This document is the base standard for a number of other standards dealing with water treatment equipment, water, dialysis water, concentrates, and dialysis fluid (ISO 23500 series) and provides dialysis practitioners with guidance on the preparation of dialysis fluid for haemodialysis and related therapies and substitution fluid for use in online therapies, such as haemodiafiltration and haemofiltration. As such, this document functions as a recommended practice.

This document does not address clinical issues that might be associated with inappropriate usage of the water, dialysis water, concentrates, or dialysis fluid. Healthcare professionals involved in the provision of treatment for kidney failure should make the final decision regarding the applications with which these fluids are used, for example, haemodialysis, haemodiafiltration, high-flux haemodialysis, and the reprocessing of dialysers, and need to be aware of the issues that the use of inappropriate fluid quality raises in each of the therapies.

The concepts incorporated in this document should not be considered inflexible or static. The recommendations presented here should be reviewed periodically in order to assimilate increased understanding of the role of dialysis fluid purity in patient outcomes and technological developments.

1.2 Inclusions

This document addresses the user's responsibility for dialysis fluid once the equipment used in its preparation has been delivered and installed.

For the purposes of this document, dialysis fluid includes:

- a) dialysis water (see [3.17](#) for definition) used for the preparation of dialysis fluid and substitution fluid,
- b) dialysis water used for the preparation of concentrates at the user's facility,
- c) concentrates,
- d) the final dialysis fluid and substitution fluid.

The scope of this document includes

- a) the quality management of equipment used to treat and distribute water used for the preparation of dialysis fluid and substitution fluid, from the point at which municipal water enters the dialysis facility to the point at which the final dialysis fluid enters the dialyser or the point at which substitution fluid is infused,
- b) equipment used to prepare concentrate from powder or other highly concentrated media at a dialysis facility, and
- c) preparation of the final dialysis fluid or substitution fluid from dialysis water and concentrates.

NOTE Because water used to prepare dialysis fluid can also be used to reprocess dialysers not marked intended for single use, this aspect of water use is also covered by this document.

1.3 Exclusions

This document does not apply to sorbent-based dialysis fluid regeneration systems that regenerate and recirculate small volumes of dialysis fluid, systems for continuous renal replacement therapy that use pre-packaged solutions, and systems and solutions for peritoneal dialysis.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For undated references, the latest edition of the referenced document (including any amendments) applies. For dated references, only the edition cited applies.

ISO 23500-3, *Preparation and quality management of fluids for haemodialysis and related therapies — Part 3: Water for haemodialysis and related therapies*

ISO 23500-4, *Preparation and quality management of fluids for haemodialysis and related therapies — Part 4: Concentrates for haemodialysis and related therapies*

ISO 23500-5, *Preparation and quality management of fluids for haemodialysis and related therapies — Part 5: Quality of dialysis fluid for haemodialysis and related therapies*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <http://www.iso.org/obp>
- IEC Electropedia: available at <http://www.electropedia.org/>

3.1
acetate concentrate
concentrated solution of salts containing acetate, which, when diluted with dialysis water, yields bicarbonate-free dialysis fluid for use in dialysis

Note 1 to entry: Acetate concentrate can contain glucose.

Note 2 to entry: Sodium acetate is used to provide buffer in place of sodium bicarbonate.

Note 3 to entry: Acetate concentrate is used as a single concentrate.

3.2
acid concentrate
A-concentrate
acidified concentrated mixture of salts that, when diluted with dialysis water and bicarbonate concentrate, yields dialysis fluid for use in dialysis

Note 1 to entry: The term “acid” refers to the small amount of acid (for example, acetic acid or citric acid) that is included in the concentrate.

Note 2 to entry: Acid concentrate can contain glucose.

Note 3 to entry: Acid concentrate can be in the form of a liquid, a dry powder, other highly concentrated media, or some combination of these forms.