
**Ophthalmic implants — Intraocular
lenses — Guidance on assessment of
the need for clinical investigation of
intraocular lens design modifications**

*Implants ophtalmiques — Lentilles intraoculaires — Directives
relatives à l'évaluation de la nécessité d'investigation clinique pour les
modifications de conception des lentilles intraoculaires*



This document is a preview generated by ERS



COPYRIGHT PROTECTED DOCUMENT

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Ch. de Blandonnet 8 • CP 401
CH-1214 Vernier, Geneva, Switzerland
Tel. +41 22 749 01 11
Fax +41 22 749 09 47
copyright@iso.org
www.iso.org

Contents

Page

Foreword	iv
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Modifications to parent models	1
4.1 General	1
4.2 Modification levels	2
4.2.1 General	2
4.2.2 Level A modifications	2
4.2.3 Level B modifications	2
4.2.4 Level C modifications	2
4.2.5 Clinical investigation with multiple IOL models	3
5 Considerations for the assignment of modification level	3
5.1 General	3
5.2 Risk assessment	3
5.3 Special considerations	3
5.3.1 Phakic lenses	3
5.3.2 Anterior chamber lenses	3
5.3.3 Posterior chamber lenses intended for implantation in the sulcus	4
6 Modifications of optical design features	4
6.1 Optical design changes	4
6.2 Multifocal lenses (MIOL)	4
6.3 Toric lenses (TIOL)	4
6.4 Accommodating lenses (AIOL)	4
7 Modifications to the mechanical design	5
7.1 General	5
7.2 Mechanical analysis	5
8 Modifications to material	5
8.1 Interchanging IOL materials	5
8.2 New materials	5
Annex A (informative) Examples of modifications to a parent IOL model	6
Annex B (informative) Mechanical data analysis	11
Bibliography	22

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, SC 7 *Ophthalmic optics and instruments*.

This second edition cancels and replaces the first edition (ISO/TR 22979:2006), which has been technically revised.

Ophthalmic implants — Intraocular lenses — Guidance on assessment of the need for clinical investigation of intraocular lens design modifications

1 Scope

This document provides guidance on the application of all parts of the ISO 11979 series of International Standards for intraocular lenses (IOLs).^[1-9] It addresses factors to be considered in the risk management process of modifications to anterior and posterior chamber IOLs in accordance with ISO 14971.^[11] It also suggests methods of data analysis and interpretation that can be used to determine the need for a clinical investigation and its design.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11979-1 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

NOTE The terms listed are related to [Annex B](#).

3.1

open-loop IOL

IOL model which contains two loops, each loop having one end attached to the body of the IOL and the other end free

3.2

closed-loop IOL

IOL model, which contains two loops, each loop having both ends attached to the body of the optic

3.3

hybrid open-loop/closed-loop IOL

IOL model which contains two loops, with one loop having one end attached to the body of the IOL and the other end free, and the other loop having both ends attached to the body of the IOL

4 Modifications to parent models

4.1 General

IOLs, that are modifications of a parent IOL, have different requirements for clinical investigations depending on the risk associated with the modifications and depending on their location in the eye.