

TECHNICAL REPORT



**Medical electrical equipment –
Part 4-1: Guidance and interpretation – Medical electrical equipment and medical
electrical systems employing a degree of autonomy**



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2017 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland

Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00
info@iec.ch
www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

IEC Catalogue - webstore.iec.ch/catalogue

The stand-alone application for consulting the entire bibliographical information on IEC International Standards, Technical Specifications, Technical Reports and other documents. Available for PC, Mac OS, Android Tablets and iPad.

IEC publications search - www.iec.ch/searchpub

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - webstore.iec.ch/justpublished

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and also once a month by email.

Electropedia - www.electropedia.org

The world's leading online dictionary of electronic and electrical terms containing 20 000 terms and definitions in English and French, with equivalent terms in 16 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

IEC Glossary - std.iec.ch/glossary

65 000 electrotechnical terminology entries in English and French extracted from the Terms and Definitions clause of IEC publications issued since 2002. Some entries have been collected from earlier publications of IEC TC 37, 77, 86 and CISPR.

IEC Customer Service Centre - webstore.iec.ch/csc

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: csc@iec.ch.

TECHNICAL REPORT



**Medical electrical equipment –
Part 4-1: Guidance and interpretation – Medical electrical equipment and medical
electrical systems employing a degree of autonomy**

INTERNATIONAL
ELECTROTECHNICAL
COMMISSION

ICS 11.040.01

ISBN 978-2-8322-4329-9

Warning! Make sure that you obtained this publication from an authorized distributor.

CONTENTS

FOREWORD.....	5
INTRODUCTION.....	7
1 Scope.....	9
2 Normative references	9
3 Terms and definitions	10
4 DEGREE OF AUTONOMY (DOA).....	17
4.1 Introduction to DEGREE OF AUTONOMY.....	17
4.2 Methodology to determine DEGREE OF AUTONOMY.....	17
4.3 Relationship between DOA and RISK	18
5 PROCESS STANDARDS supporting DOA.....	18
5.1 General.....	18
5.2 RISK MANAGEMENT PROCESS	19
5.2.1 Defining INTENDED USE	19
5.2.2 INTENDED USE and characteristics related to SAFETY	19
5.3 RISK CONTROL.....	20
5.3.1 General	20
5.3.2 RISK CONTROL hierarchy	21
5.4 USABILITY engineering considerations for MEE or MES having a higher DOA.....	22
5.4.1 General	22
5.4.2 OPERATOR situation awareness.....	22
5.4.3 OPERATOR reaction time.....	23
5.4.4 OPERATOR sensory input and response	23
5.4.5 Detectability by OPERATOR of malfunction or errors of MEE or MES with a higher DOA	23
5.5 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) and software development LIFE CYCLE (IEC 62304)	23
5.6 Application of RISK MANAGEMENT for IT-networks incorporating medical devices	24
6 BASIC SAFETY and ESSENTIAL PERFORMANCE related to DOA	25
6.1 GENERAL	25
6.2 BASIC SAFETY related to DOA	25
6.3 ESSENTIAL PERFORMANCE related to DOA.....	26
Annex A (informative) Rationale for defining the AUTOMATIC, AUTONOMY and DOA framework and the distinction between a MEDICAL ROBOT and other MEE or MES.....	28
A.1 General.....	28
A.2 Existing definitions and limitations	28
A.3 New approaches	29
A.4 Definition of MONITOR – GENERATE – SELECT – EXECUTE	30
A.5 Approaches to define ROBOT and MEDICAL ROBOT.....	31
A.6 Conclusions	31
Annex B (informative) DOA and relevant terms used in MEE standards.....	32
B.1 General.....	32
B.2 Procedure	32
B.3 Results	32
B.3.1 Summary	32
B.3.2 Tables	33
Annex C (informative) Exemplar methods for classifying DEGREE OF AUTONOMY	42

C.1	Descriptive method	42
C.2	Binary method.....	43
C.3	Weighted method.....	44
Annex D (informative)	Examples of introducing DOA to MEE/MES	50
D.1	General.....	50
D.2	Example 1 – Lower extremity exoskeleton.....	50
D.2.1	Description of the medical procedures	50
D.2.2	DOA classification method.....	50
D.2.3	Effect of DOA on the RISK MANAGEMENT PROCESS.....	52
D.3	Example 2 – Orthopaedic MEE/MES/MEDICAL ROBOT for reshaping bone.....	54
D.3.1	Description of the medical procedures	54
D.3.2	DOA classification method.....	54
D.3.3	Effect of DOA on the RISK MANAGEMENT PROCESS.....	55
D.3.4	Summary and conclusions	55
D.4	Example 3 – Instrument exchange on robotically-assisted surgical equipment	55
D.4.1	Description of the medical procedures	55
D.4.2	DOA classification method.....	56
D.4.3	Effect of DOA on the RISK MANAGEMENT PROCESS.....	56
D.4.4	Summary and conclusions	57
D.5	Example 4 – Master–slave robotically-assisted surgical equipment	57
D.5.1	Description of the medical procedures	57
D.5.2	DOA classification method.....	58
D.5.3	Effect of DOA on RISK MANAGEMENT PROCESS.....	58
D.5.4	Summary and conclusions	58
D.6	Example 5 – Image-guided radiotherapy equipment	58
D.6.1	Description of the medical procedures	58
D.6.2	DOA classification method.....	59
D.6.3	RISK ANALYSIS for each level of DOA	61
D.6.4	Effect of DOA on the RISK MANAGEMENT PROCESS.....	61
D.6.5	Summary and conclusions	61
D.7	Example 6 – Automated external defibrillator (AED).....	62
D.7.1	Description of the medical procedures	62
D.7.2	DOA classification method.....	63
D.7.3	Effect of DOA on the RISK MANAGEMENT PROCESS.....	64
D.7.4	Summary and conclusions	64
Annex E (informative)	PATIENT SAFETY characteristics to be taken into account during RISK MANAGEMENT for MEE or MES employing DOA.....	65
E.1	Types of PATIENTS	65
E.2	Additional attention for child (PATIENT) SAFETY	65
E.3	PATIENT abilities and variability of physiological signals	66
E.3.1	ISO/IEC Guide 71	66
E.3.2	Changing need and abilities of PATIENTS	66
E.3.3	PATIENT’S sensory abilities.....	66
E.3.4	PATIENT’S PHYSICAL ABILITIES	67
E.3.5	PATIENT’S COGNITIVE ABILITIES.....	67
E.3.6	PATIENT ALLERGIES	67
ANNEX F (informative)	PHYSIOLOGIC CLOSED-LOOP CONTROL SYSTEM AND DOA	69
Annex G (informative)	Examples of distributed ESSENTIAL PERFORMANCE	72
Bibliography.....		75

Figure 1 – Basic model of interoperability of MEE in an MES (Order of execution: 1 to 3)	25
Figure A.1 – ALFUS approach applied to MEE or MES applications	30
Figure C.1 – Application of weighted method to the “MONITOR” TASK	45
Figure C.2 – Application of weighted method to “GENERATE OPTIONS”	46
Figure C.3 – Application of weighted method to “SELECT OPTION” TASK	47
Figure C.4 – Application of weighted method to the “EXECUTE” TASK	48
Figure F.1 – Functional diagram indicating typical components of a PHYSIOLOGIC CLOSED-LOOP CONTROL SYSTEM (PCLCS) utilizing a PCLC	69
Figure F.2 – Examples of introducing DOA into the MONITORING TASK via PCLCS	70
Figure F.3 – Examples of introducing DOA into the GENERATING TASK via PCLCS	70
Figure F.4 – Examples of introducing DOA into the SELECTION TASK via PCLCS	70
Figure F.5 – Examples of introducing DOA into the EXECUTION TASK via PCLCS	71
Table 1 – Examples of ESSENTIAL PERFORMANCE of MEE or MES with a DOA	27
Table B.1 – List of terms that indicate the use of AUTONOMY	33
Table B.2 – List of reviewed standards – sorted by standard number (1 of 4)	34
Table B.3 – List of identified inconsistencies in reviewed standards (1 of 2)	40
Table C.1 – Descriptive classification of DOA	43
Table C.2 – Binary classification of DOA	44
Table D.1 – Example 1 – Effect of DOA on the RISK MANAGEMENT PROCESS	52
Table D.2 – Example 1 – Physical and cognitive capability of individual and CLINICAL FUNCTION needed	52
Table D.3 – Example 1 – Sub-function TASK example	53
Table D.4 – Example 2 – Effect of DOA on the RISK MANAGEMENT PROCESS	55
Table D.5 – Example 3 – Comparison of instrument exchange design implementations	57
Table D.6 – Example 3 – Effect of DOA on the RISK MANAGEMENT PROCESS	57
Table D.7 – Example 4 – Effect of DOA on the RISK MANAGEMENT PROCESS	58
Table D.8 – Example 5 – Descriptive classification of DOA for IGRT MEE	60
Table D.9 – Example 5 – Binary classification of DOA for IGRT MEE	60
Table D.10 – Example 5 – Effect of DOA on the RISK MANAGEMENT PROCESS	62
Table D.11 – Example 6 – Descriptive method classification of DOA in external defibrillators	63
Table D.12 – Example 6 – Effect of DOA on the RISK MANAGEMENT PROCESS	64
Table G.1 – Examples of distributed ESSENTIAL PERFORMANCE (1 of 3)	72

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 4-1: Guidance and interpretation –
Medical electrical equipment and medical
electrical systems employing a degree of autonomy****FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

The main task of IEC technical committees is to prepare International Standards. However, a technical committee may propose the publication of a technical report when it has collected data of a different kind from that which is normally published as an International Standard, for example "state of the art".

IEC 60601-4-1, which is a technical report, has been prepared by a Joint Working Group of IEC subcommittee 62A: Common aspects of electrical equipment used in medical practice, of IEC technical committee 62: Electrical equipment in medical practice, and ISO technical committee 299: Robotics.

The text of this technical report is based on the following documents:

Enquiry draft	Report on voting
62A/1099/DTR	62A/1129A/RVDTR

Full information on the voting for the approval of this technical report can be found in the report on voting indicated in the above table.

This document has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this technical report, the following print types are used:

- recommendations and definitions: roman type.
- *test instructions: italic type.*
- informative material appearing outside of tables, such as notes, examples and references: in smaller type.
Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THIS TECHNICAL REPORT OR AS NOTED: SMALL CAPITALS.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

A bilingual version of this publication may be issued at a later date.

IMPORTANT – The 'colour inside' logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.

INTRODUCTION

This Technical Report is the result of work that began in ISO/TC 184/SC 2/WG 7 in October 2006 on personal care ROBOTS, to address an emerging type of ROBOT that was used outside of an industrial environment. That group was working on a new standard, ISO 13482, which was published as an International Standard (IS) in 2014. While initially focused on non-medical applications, WG 7 recognized that work was likely to be needed on medical devices utilizing robotic technology. In September 2009, ISO/TC 184/SC 2 established a WG 7, *Study Group (SG) on Medical care robots*, comprised of experts from Canada, France, Germany, Japan, Korea, Romania, Switzerland, UK and USA.

The work of ISO/TC 184/SC 2/WG 7¹ SG cumulated in a proposal to form a Joint Working Group with IEC/SC 62A to develop general requirements and guidance related to the SAFETY of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that utilize robotic technology. The work would include medical applications (including aids for the disabled) covering invasive and non-invasive procedures such as surgery, rehabilitation therapy, imaging and other ROBOTS for medical diagnosis and treatment. The proposal was approved, resulting in the formation of Joint Working Group (JWG) 9 (*Medical electrical equipment and systems using robotic technology*) and the first meeting was held in Los Angeles in June 2011.

JWG 9 examined the definition of a ROBOT from ISO 8373:2012 (which was later modified to a “programmed actuated mechanism with a DEGREE OF AUTONOMY (DOA), moving within its environment, to perform intended TASKS”) and AUTONOMY (the “ability to perform intended TASKS based on current state and sensing, without human intervention”). It was recognized by JWG 9 that these definitions could need further refinement to establish the appropriate boundaries for future standardisation work. AUTONOMY and DEGREE OF AUTONOMY (DOA) were felt to be key ingredients in distinguishing a “MEDICAL ROBOT” from other types of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS.

However, JWG 9 came to realize that there are currently standardized MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that exhibit a DOA. Therefore, DOA by itself is not a unique characteristic of a MEDICAL ROBOT. This can be stated more clearly as follows:

- not all MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that exhibit a DOA are MEDICAL ROBOTS; but
- all MEDICAL ROBOTS exhibit a DOA.

Hence a MEDICAL ROBOT can be a MEDICAL ELECTRICAL EQUIPMENT or part of a MEDICAL ELECTRICAL SYSTEM, but not all MEDICAL ELECTRICAL EQUIPMENT are MEDICAL ROBOTS.

NOTE The majority of existing MEDICAL ELECTRICAL EQUIPMENT are not considered as MEDICAL ROBOTS.

The MANUFACTURER states clearly the type of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEM through the INTENDED USE of their product. For this INTENDED USE, a definition of MEDICAL ROBOT would be helpful to have a common understanding if this MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM can be tagged as a MEDICAL ROBOT equipment or MEDICAL ROBOT system. The definition of MEDICAL ROBOT is therefore helpful to distinguish if the MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEM is a MEDICAL ROBOT and the INTENDED USE as indicated by the MANUFACTURER. This distinction is clarified in Annex A.

DOA is normally considered for adoption into MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS for the following reasons:

- DOA could give benefits to CLINICAL FUNCTION outcomes;

1) ISO TC 184/SC 2 was reformed to ISO TC 299 in January 2016.

- DOA could give economic value to MEDICAL ELECTRICAL EQUIPMENT;
- DOA could improve the consistency of medical procedures;
- DOA could handle more complex data;
- DOA could lead to faster reaction times;
- DOA could optimise medical procedure times or duration;
- DOA could make it easier to integrate MEDICAL ELECTRICAL SYSTEMS;
- DOA could decrease the overall level of RISK; and
- DOA could change the role of an OPERATOR to a more supervisory than active (hands on) function.

In order to progress the work of JWG 9, it was agreed to focus on the IEC 60601-1 standard family and see how specific clauses could be extended to cover the additional DOA issues in a possible new Technical Report once fully developed. JWG 9 looked at existing MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS that had characteristics of a ROBOT based on the definition, and investigated the suitability of the existing standards to address the HAZARDS associated with their use. As a result of this investigation, it was acknowledged that IEC 60601 (all parts), ISO 14971, IEC 62366-1 and IEC 62304 provide appropriate general requirements and guidance on how to address the HAZARDS; however, emerging functionality associated with increased DOA on MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, whether a ROBOT or not, could result in situations where BASIC SAFETY and ESSENTIAL PERFORMANCE are considered again by the MANUFACTURER.

Current MEDICAL ELECTRICAL EQUIPMENT standards do not fully address higher DOA modes of operation, and this document is intended to provide guidance for MANUFACTURERS and others in this field on how DOA could be introduced into MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS. Incorporation of higher levels of AUTONOMY in MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS is still new and rapidly evolving, and at the time of writing this document does not lend itself to general standardization.

The importance of understanding DOA can be illustrated by examining its effects in other industries. The airline industry is one example in which increasing DOA has often been implemented as a RISK CONTROL measure. However, there are numerous examples in the airline industry in which increased DOA was found to have been a major contributor to a fatal accident [70].² To avoid similar mistakes in the field of MEDICAL ELECTRICAL EQUIPMENT and MEDICAL ELECTRICAL SYSTEMS, MANUFACTURERS learn from these other fields and not only characterize DOA but also understand its potential for unintentionally increasing RISK.

It is important to point out that this IEC document is an informative document as are all IEC Technical Reports (ISO/IEC Guide 2 [61]). The concept and approach stated in the STATE OF THE ART are not intended to be addressed through this informative document. This document is not used as a normative requirement as per the claimed STATE OF THE ART by any country or community. This document is an **informative** document, which is intended to provide existing and future designers of MEDICAL ELECTRICAL EQUIPMENT or MEDICAL ELECTRICAL SYSTEMS some guidance and direction concerning the adoption of DOA. This document is not applicable as a base for a testing procedure or writing a test protocol template.

² Numbers in square brackets refer to the bibliography.

MEDICAL ELECTRICAL EQUIPMENT –

Part 4-1: Guidance and interpretation – Medical electrical equipment and medical electrical systems employing a degree of autonomy

1 Scope

This Part of IEC 60601 is intended to help a MANUFACTURER through the key decisions and steps to be taken to perform a detailed RISK MANAGEMENT and USABILITY ENGINEERING PROCESSES for MEDICAL ELECTRICAL EQUIPMENT or a MEDICAL ELECTRICAL SYSTEM, hereafter referred to as MEE or MES, employing a DEGREE OF AUTONOMY (DOA).

This document provides a definition of DOA of MEE or MES and a MEDICAL ROBOT, and also provides guidance on:

- methodologies to perform the RISK MANAGEMENT PROCESS and USABILITY ENGINEERING for an MEE or MES with a DOA;
- considerations of BASIC SAFETY and ESSENTIAL PERFORMANCE for an MEE and MES with a DOA; and
- identifying the use of DOA, and similar concepts in existing ISO/IEC standards dealing with MEE or MES with the goal to facilitate alignment of standards by consistent use of the concept of DOA; and
- distinguishing between MEDICAL ROBOTS, and other MEE and MES.

Unless specified otherwise, this document considers MEE and MES together.

The MANUFACTURER of an MEE or MES with a DOA is expected to design and manufacture an MEE or MES that fulfils its INTENDED USE and does not have unacceptable RISK throughout its LIFE-CYCLE.

This document provides guidance to help the MANUFACTURER in complying with the requirements of IEC 60601-1:2005 and IEC 60601-1:2005/AMD1:2012 for MEE and MES with DOA. The document is also intended as guidance for future standard writers.

There are no prerequisites to this document.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

IEC 60601-1:2005/AMD1:2012³

³ There exists a consolidated edition 3.1, including IEC 60601-1:2005 and its Amendment 1:2012.