

Dentistry - Soft lining materials for removable dentures
- Part 1: Materials for short-term use (ISO
10139-1:2018)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 10139-1:2018 sisaldab Euroopa standardi EN ISO 10139-1:2018 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 10139-1:2018 consists of the English text of the European standard EN ISO 10139-1:2018.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 02.05.2018.	Date of Availability of the European standard is 02.05.2018.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.060.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English Version

Dentistry - Soft lining materials for removable dentures -
Part 1: Materials for short-term use (ISO 10139-1:2018)

Médecine bucco-dentaire - Produits souples pour
intradoss de prothèses dentaires amovibles - Partie 1:
Produits pour usage à court terme (ISO 10139-1:2018)

Zahnheilkunde - Weichbleibende
Unterfütterungswerkstoffe für Prothesen - Teil 1:
Werkstoffe für kurzzeitige Anwendungen (ISO 10139-
1:2018)

This European Standard was approved by CEN on 23 April 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 10139-1:2018) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2018, and conflicting national standards shall be withdrawn at the latest by November 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10139-1:2005.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 10139-1:2018 has been approved by CEN as EN ISO 10139-1:2018 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification	2
4.1 Types	2
4.2 Classes	2
5 Requirements	2
5.1 Shore A0 hardness	2
5.1.1 Shore A0 hardness at 2 h	2
5.1.2 Shore A0 hardness at 7 d	2
5.2 Consistency	2
5.3 Detail reproduction	3
6 Sampling	3
7 Test methods	3
7.1 Ambient conditions for testing	3
7.2 Shore A0 hardness	3
7.2.1 Apparatus	3
7.2.2 Procedure	3
7.3 Consistency test	5
7.3.1 Test conditions	5
7.3.2 Apparatus	5
7.3.3 Test procedure	5
7.4 Detail reproduction test	5
7.4.1 General	5
7.4.2 Apparatus	5
7.4.3 Procedure	6
8 Requirement for packaging, marking and instructions supplied by the manufacturer	8
8.1 Packaging	8
8.2 Marking and manufacturer's instructions for use	8
Bibliography	10

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 2, *Prosthodontic materials*.

This third edition cancels and replaces the second edition (ISO 10139-1:2005), which has been technically revised. It also incorporates the Technical Corrigendum ISO 10139-1:2005/Cor.1:2006.

A list of all parts in the ISO 10139 series can be found on the ISO website.

Introduction

Clinically, short-term denture-lining materials are used commonly as tissue conditioners and as temporary soft lining materials. Furthermore, some materials are also indicated for functional impression taking. Therefore, the tests are designed to cover the more common usages.

It is recognized that the short-term material, when used as a tissue conditioner, is commonly changed every few days with the aim of returning the mucosa to a healthy condition as quickly as possible. As a temporary soft lining, the material is commonly placed in immediate dentures and in dentures that need to be modified as part of implant treatment. Therefore the specification has been so designed to necessitate that a material exhibit the required properties over a 7 d period. It is of course recognized that there are a number of clinical situations where it is appropriate to retain the soft lining in the denture for periods longer than 7 d. It is also recognized that manufacturers may wish to provide more than one set of times, temperatures, proportions and procedures to mix or prepare the material properly in order that the material can satisfy the requirements of more than one type or class.

In an attempt to establish some degree of harmony with the procedures used to evaluate related dental materials, the detail reproduction test has been adopted for materials also used for functional impression taking (ISO 4823). As well, in this revision of the standard, the Shore A0 hardness test has replaced the depth of penetration test, and the consistency test has been reintroduced as a replacement of the elastic recovery test due to complexity of this method.

Specific qualitative and quantitative test methods for demonstrating freedom from unacceptable biological hazards are not included in this document, but it is recommended that, for the assessment of possible biological hazards, reference should be made to ISO 10993-1 and ISO 7405.

Dentistry — Soft lining materials for removable dentures —

Part 1: Materials for short-term use

1 Scope

This document specifies requirements for the physical properties, test methods, packaging, marking and manufacturer's instructions for soft denture lining materials suitable for short-term use, including functional impression taking using existing removable prosthesis.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 4823:2015, *Dentistry — Elastomeric impression materials*

ISO 7619-1, *Rubber, vulcanized or thermoplastic — Determination of indentation hardness — Part 1: Durometer method (Shore hardness)*

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

3.1

soft denture lining material

soft resilient material bonded to the fitting surface of a denture to reduce trauma to the supporting tissues

Note 1 to entry: A soft lining material can be used as a tissue conditioning material when placed in the fitting surface of a denture and intended to be in contact with the denture-supporting mucosa, commonly for a period of up to 7 d, with the aim of assisting its return to a healthy condition.

3.2

short-term use

use normally for a continuous period of between 60 minutes and 30 days

3.3

functional impression taking

use of a *soft denture lining material* (3.1) to take a functional impression using existing removable prosthesis