
International Standard



4943

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Steel and cast iron — Determination of copper content — Flame atomic absorption spectrometric method

Aciers et fontes — Dosage du cuivre — Méthode par spectrométrie d'absorption atomique dans la flamme

First edition — 1985-12-15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 4943 was prepared by Technical Committee ISO/TC 17, *Steel*.

Users should note that all International Standards undergo revision from time to time and that any reference made herein to any other International Standard implies its latest edition, unless otherwise stated.

Steel and cast iron — Determination of copper content — Flame atomic absorption spectrometric method

1 Scope and field of application

This International Standard specifies a method for the determination of copper in steel and cast iron by means of flame atomic absorption spectrometry.

The method is applicable to copper contents in the range 0,004 to 0,5 % (*m/m*).

2 Reference

ISO/R 377, *Selection and preparation of samples and test pieces for wrought steel*.

3 Principle

Dissolution of a test portion in a mixture of hydrochloric, nitric and perchloric acids.

Spraying of the solution into an air-acetylene flame. Spectrometric measurement of the atomic absorption of the 324,7 nm spectral line emitted by a copper hollow cathode lamp.

4 Reagents

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and having a very low copper content, and only distilled water or water of equivalent purity.

If possible, use only freshly prepared distilled or de-ionized water.

4.1 Iron of high purity,
copper content < 0,000 5 % (*m/m*).

4.2 Hydrochloric acid-nitric acid mixture.

Mix three parts by volume of hydrochloric acid (ρ about 1,19 g/ml), one part by volume of nitric acid (ρ about 1,40 g/ml), and two parts by volume of water.

Prepare this mixture immediately before use.

4.3 Hydrochloric acid-nitric acid-perchloric acid mixture.

Mix 20 ml of hydrochloric acid (ρ about 1,19 g/ml) with 55 ml of nitric acid (ρ about 1,40 g/ml) and 75 ml of perchloric acid (ρ about 1,54 g/ml).

NOTE — Perchloric acid (ρ about 1,67 g/ml) may also be used. 100 ml of perchloric acid (ρ about 1,54 g/ml) is equivalent to 79 ml of perchloric acid (ρ about 1,67 g/ml).

4.4 Copper, standard solution.

4.4.1 Stock solution, corresponding to 1 g of Cu per litre.

Weigh, to the nearest 0,000 1 g, 1,000 0 g of high purity copper [$> 99,95$ % (*m/m*) Cu].

Transfer to a 400 ml beaker and dissolve in 25 ml of nitric acid (ρ about 1,40 g/ml, diluted 1 + 4). Cover with a watch-glass. When dissolution is complete, evaporate on a water-bath until the onset of crystallization. Dissolve the residue in water, cool, transfer to a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

4.4.2 Standard solution, corresponding to 20 mg of Cu per litre.

Transfer 20,0 ml of the stock solution (4.4.1) into a 1 000 ml one-mark volumetric flask, dilute to the mark and mix.

1 ml of this standard solution contains 20 μ g of Cu.

Prepare this standard solution immediately before use.

5 Apparatus

Ordinary laboratory apparatus, and

5.1 Atomic absorption spectrometer.

A copper hollow cathode lamp; supplies of air and acetylene sufficiently pure to give a steady clear fuel-lean flame, free from water and oil, and free from copper.