

Environmental testing -- Part 3-11: Supporting documentation and guidance - Calculation of the uncertainty of conditions in climatic test chambers

Environmental testing -- Part 3-11: Supporting
documentation and guidance - Calculation of the
uncertainty of conditions in climatic test chambers

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 60068-3-11:2007 sisaldab Euroopa standardi EN 60068-3-11:2007 ingliskeelset teksti. Käesolev dokument on jõustatud 27.08.2007 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes. Standard on kättesaadav Eesti standardiorganisatsioonist.	This Estonian standard EVS-EN 60068-3-11:2007 consists of the English text of the European standard EN 60068-3-11:2007. This document is endorsed on 27.08.2007 with the notification being published in the official publication of the Estonian national standardisation organisation. The standard is available from Estonian standardisation organisation.
--	---

Käsitlusala: This part of IEC 60068 demonstrates how to estimate the uncertainty of steady-state temperature and humidity conditions in temperature and humidity chambers. Since this is inextricably linked to the methods of measurement, these are also described. This standard is equally applicable to all environmental enclosures, including rooms or laboratories. The methods used apply both to temperature chambers and combined temperature and humidity chambers. This standard is meant to help everyone using climatic test chambers. Those already familiar with uncertainty of measurement will find it useful for guidance on typical sources of uncertainty and how they should be quantified and combined. It is also intended to assist the first-time or occasional user who has little or no knowledge of the subject.	Scope: This part of IEC 60068 demonstrates how to estimate the uncertainty of steady-state temperature and humidity conditions in temperature and humidity chambers. Since this is inextricably linked to the methods of measurement, these are also described. This standard is equally applicable to all environmental enclosures, including rooms or laboratories. The methods used apply both to temperature chambers and combined temperature and humidity chambers. This standard is meant to help everyone using climatic test chambers. Those already familiar with uncertainty of measurement will find it useful for guidance on typical sources of uncertainty and how they should be quantified and combined. It is also intended to assist the first-time or occasional user who has little or no knowledge of the subject.
---	---

ICS 19.040

Võtmesõnad:

**Environmental testing -
Part 3-11: Supporting documentation and guidance -
Calculation of uncertainty of conditions in climatic test chambers
(IEC 60068-3-11:2007)**

Essais d'environnement -
Partie 3-11: Documentation
d'accompagnement et guide -
Calcul de l'incertitude des conditions
en chambres d'essais climatiques
(CEI 60068-3-11:2007)

Umgebungseinflüsse -
Teil 3-11: Unterstützende Dokumentation
und Leitfaden -
Berechnung der Messunsicherheit
von Umgebungsbedingungen
in Klimaprüfkammern
(IEC 60068-3-11:2007)

This European Standard was approved by CENELEC on 2007-06-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 104/409/FDIS, future edition 1 of IEC 60068-3-11, prepared by IEC TC 104, Environmental conditions, classification and methods of test, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60068-3-11 on 2007-06-01.

The following dates were fixed:

- | | | |
|--|-------|------------|
| – latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2008-03-01 |
| – latest date by which the national standards conflicting with the EN have to be withdrawn | (dow) | 2010-06-01 |

Annex ZA has been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60068-3-11:2007 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60068-3-5	- ¹⁾	Environmental testing - Part 3-5: Supporting documentation and guidance - Confirmation of the performance of temperature chambers	EN 60068-3-5	2002 ²⁾
IEC 60068-3-6	- ¹⁾	Environmental testing - Part 3-6: Supporting documentation and guidance - Confirmation of the performance of temperature/humidity chambers	EN 60068-3-6	2002 ²⁾
ISO 3534-1	2006	Statistics - Vocabulary and symbols - Part 1: General statistical terms and terms used in probability	-	-
ISO 3534-2	2006	Statistics - Vocabulary and symbols - Part 2: Applied statistics	-	-
ISO Guide 99	1996	International vocabulary of basic and general terms in metrology	-	-
ISO/IEC Guide 98	1995	Guide to the expression of uncertainty in measurement	-	-

¹⁾ Undated reference.

²⁾ Valid edition at date of issue.

INTERNATIONAL
STANDARD

IEC
CEI

NORME
INTERNATIONALE

60068-3-11

First edition
Première édition
2007-05

Environmental testing –

**Part 3-11:
Supporting documentation and guidance –
Calculation of uncertainty of conditions
in climatic test chambers**

Essais d'environnement –

**Partie 3-11:
Documentation d'accompagnement et guide –
Calcul de l'incertitude des conditions
en chambres d'essais climatiques**



Reference number
Numéro de référence
IEC/CEI 60068-3-11:2007



THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2007 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office
3, rue de Varembe
CH-1211 Geneva 20
Switzerland
Email: inmail@iec.ch
Web: www.iec.ch

About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

- Catalogue of IEC publications: www.iec.ch/searchpub

The IEC on-line Catalogue enables you to search by a variety of criteria (reference number, text, technical committee,...). It also gives information on projects, withdrawn and replaced publications.

- IEC Just Published: www.iec.ch/online_news/justpub

Stay up to date on all new IEC publications. Just Published details twice a month all new publications released. Available on-line and also by email.

- Customer Service Centre: www.iec.ch/webstore/custserv

If you wish to give us your feedback on this publication or need further assistance, please visit the Customer Service Centre FAQ or contact us:

Email: csc@iec.ch
Tel.: +41 22 919 02 11
Fax: +41 22 919 03 00

A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

- Catalogue des publications de la CEI: www.iec.ch/searchpub/cur_fut-f.htm

Le Catalogue en-ligne de la CEI vous permet d'effectuer des recherches en utilisant différents critères (numéro de référence, texte, comité d'études,...). Il donne aussi des informations sur les projets et les publications retirées ou remplacées.

- Just Published CEI: www.iec.ch/online_news/justpub

Restez informé sur les nouvelles publications de la CEI. Just Published détaille deux fois par mois les nouvelles publications parues. Disponible en-ligne et aussi par email.

- Service Clients: www.iec.ch/webstore/custserv/custserv_entry-f.htm

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions, visitez le FAQ du Service clients ou contactez-nous:

Email: csc@iec.ch
Tél.: +41 22 919 02 11
Fax: +41 22 919 03 00

**INTERNATIONAL
STANDARD**

**IEC
CEI**

**NORME
INTERNATIONALE**

60068-3-11

First edition
Première édition
2007-05

Environmental testing –

**Part 3-11:
Supporting documentation and guidance –
Calculation of uncertainty of conditions
in climatic test chambers**

Essais d'environnement –

**Partie 3-11:
Documentation d'accompagnement et guide –
Calcul de l'incertitude des conditions
en chambres d'essais climatiques**



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE
CODE PRIX

V

*For price, see current catalogue
Pour prix, voir catalogue en vigueur*

CONTENTS

FOREWORD.....	4
INTRODUCTION.....	6
1 Scope.....	7
2 Normative references	7
3 Terms and definitions	8
4 Concept of uncertainty.....	11
4.1 Uncertainty, error and “true value”.....	11
4.2 Statements of uncertainty	12
4.3 Combining uncertainties	13
5 Tolerance	13
6 Humidity and temperature measurement	13
7 Methods for determining climatic test chamber uncertainties	14
7.1 Empty chamber	16
7.2 Typical load.....	16
7.3 Measurement of conditions in the chamber during the test	17
7.4 Conditions to measure.....	17
7.5 Measurements required.....	18
7.6 Sources of uncertainty.....	19
7.7 Essential contributions of uncertainty	20
8 Estimation of uncertainty components and their combination	24
9 Overall uncertainty of temperature measurement.....	24
9.1 General.....	24
9.2 Further considerations.....	26
10 Overall uncertainty of relative humidity measurement.....	26
10.1 Uncertainty of temperature measurement at each sensor point.....	27
10.2 Uncertainty of the relative humidity measurement.....	27
11 Anomalous data and presentation of results	30
11.1 Average case analysis.....	30
11.2 Worst case analysis	30
Annex A (informative) Measurement data sets – Loaded chamber	32
Bibliography.....	34

Figure 1 – Approaches to calibration method and uncertainty calculation	15
Figure 2 – Illustration of the fluctuation of a temperature sensor	23
Table 1 – Combination of temperature uncertainties	24
Table 2 – Combination of temperature uncertainties at each point	27
Table 3 – Combination of humidity uncertainties	28
Table A.1 – Typical temperature measurement data set and it's analysis and refs	32
Table A.2 – Humidity measurements analysis based on Table A.1 temperatures.....	33

INTERNATIONAL ELECTROTECHNICAL COMMISSION

ENVIRONMENTAL TESTING –

**Part 3-11: Supporting documentation and guidance –
Calculation of uncertainty of conditions in climatic test chambers**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with an IEC Publication.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60068-3-11 has been prepared by IEC technical committee 104: Environmental conditions, classification and methods of test.

The text of this standard is based on the following documents:

FDIS	Report on voting
104/409/FDIS	104/415/RVD

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

A list of all parts in the IEC 60068 series, under the general title *Environmental testing* can be found on the IEC website.

The committee has decided that the contents of this publication will remain unchanged until the maintenance result date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

INTRODUCTION

This part of IEC 60068 provides guidance for analysing uncertainties of temperature and humidity in climatic test chambers. It has been written for technicians, engineers and managers in environmental testing, and for anyone who needs to understand the results of environmental tests.

The performance of climatic test chambers is a key concern in environmental test engineering. To comply with any test specification, the performance of the chamber needs to be characterized to decide whether the generated conditions fall within the specified limits. This characterization can be a difficult task, and the analysis of uncertainties in chamber performance is often surrounded by confusion. This publication is intended to ease that process.

In what follows, the concept of uncertainty of measurement is introduced first and then the significance of tolerance discussed. Aspects of humidity and temperature measurement are considered, followed by methods for determining and combining uncertainties. The cases of both calibrating an empty chamber and of measuring conditions in a loaded chamber are considered. Finally, detailed guidance and worked examples are given for analysing results to give estimates of uncertainty in the measured performance.

ENVIRONMENTAL TESTING –

Part 3-11: Supporting documentation and guidance – Calculation of uncertainty of conditions in climatic test chambers

1 Scope

This part of IEC 60068 demonstrates how to estimate the uncertainty of steady-state temperature and humidity conditions in temperature and humidity chambers. Since this is inextricably linked to the methods of measurement, these are also described.

This standard is equally applicable to all environmental enclosures, including rooms or laboratories. The methods used apply both to temperature chambers and combined temperature and humidity chambers.

This standard is meant to help everyone using climatic test chambers. Those already familiar with uncertainty of measurement will find it useful for guidance on typical sources of uncertainty and how they should be quantified and combined. It is also intended to assist the first-time or occasional user who has little or no knowledge of the subject.

To discuss uncertainty, it is important first to understand what is being measured or characterized. The calibration or characterization of the performance of a chamber is concerned with the humidity and temperature of the air in the chamber, as experienced by the item under test, at a given set point. This should not be confused with characterizing or calibrating the chamber sensor, which is a separate matter.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068-3-5: *Environmental testing – Part 3-5: Supporting documentation and guidance – Confirmation of the performance of temperature chambers*

IEC 60068-3-6: *Environmental testing – Part 3-6: Supporting documentation and guidance – Confirmation of the performance of temperature/humidity chambers*

ISO 3534-1:2006, *Statistics – Vocabulary and symbols – Part 1: General statistical terms and terms used in probability*

ISO 3534-2:2006, *Statistics – Vocabulary and symbols – Part 2: Applied statistics*

International Vocabulary of basic and general standard terms in metrology. ISO, Geneva, Switzerland 1993 (ISBN 92-67-10175-1) – VIM

Guide to the expression of uncertainty in measurement. ISO, Geneva, Switzerland 1993.
(ISBN 92-67-10188-9) – GUM

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

calibration authority

laboratory or other organization that performs calibrations and is itself accredited by the appropriate national accreditation body

3.2

climatic test chamber

enclosure

chamber or enclosed space where the internal temperature or temperature and humidity can be controlled within specified limits

3.3

combined standard uncertainty

standard uncertainty of the result of a measurement when that result is obtained from the values of a number of other quantities, equal to the positive square root of a sum of terms, the terms being the variances or covariances of these other quantities weighted according to how the measurement result varies with changes in these quantities

See also GUM.

3.4

correction

value added algebraically to the result of a measurement to compensate determinable systematic error

See also VIM.

3.5

confidence level

value of probability associated with a confidence interval

NOTE The confidence level is the likelihood that the “true value” lies within the stated range of uncertainty usually expressed as a percentage, e.g. 95 %.

See also ISO 3534-1.

3.6

coverage factor

numerical factor to multiply the combined standard uncertainty to obtain an expanded uncertainty

NOTE A coverage factor of $k=2$ corresponds to a confidence level of approximately 95 % if normally distributed and if the number of degrees of freedom is sufficiently large.

See also GUM.