

---

---

**Ophthalmic optics — Visual acuity  
testing — Standard and clinical  
optotypes and their presentation**

*Optique ophtalmique — Mesure de l'acuité visuelle — Optotype  
normalisé et optotypes cliniques et leur présentation*



This document is a preview generated by EBS



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2017, Published in Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Ch. de Blandonnet 8 • CP 401  
CH-1214 Vernier, Geneva, Switzerland  
Tel. +41 22 749 01 11  
Fax +41 22 749 09 47  
copyright@iso.org  
www.iso.org

# Contents

Page

|   |           |
|---|-----------|
| <b>Foreword</b> .....   | <b>iv</b> |
| <b>1 Scope</b> .....  | <b>1</b>  |
| <b>2 Normative references</b> .....                                       | <b>1</b>  |
| <b>3 Terms and definitions</b> .....                                      | <b>1</b>  |
| <b>4 Requirements</b> .....   | <b>2</b>  |
| 4.1 Specifications of the standard optotype.....                          | 2         |
| 4.2 Visual acuity grades and standard optotype grades.....                | 3         |
| 4.3 Test area and spacing between standard optotypes.....                 | 4         |
| 4.4 Positions of the optotype.....  | 4         |
| 4.5 Quality of presentation.....  | 4         |
| 4.6 Luminance.....  | 5         |
| <b>5 Test methods</b> .....   | <b>6</b>  |
| 5.1 Viewing distance for distance visual acuity testing.....              | 6         |
| 5.2 Criteria for determination and assignment of visual acuity grade..... | 6         |
| <b>6 Test report</b> .....  | <b>6</b>  |
| <b>Annex A (informative) Optotypes for clinical investigations</b> .....  | <b>7</b>  |
| <b>Bibliography</b> .....   | <b>10</b> |

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This third edition cancels and replaces the second edition (ISO 8596:2009), which has been technically revised.

The main changes compared to the previous edition are as follows:

- a) restructuring of technical content into the [Clauses 4](#) and [6](#) has been applied;
- b) terms and definitions with the terms standard optotype, clinical optotype, visual acuity, and the systems decimal visual acuity, Snellen fraction, LogMAR acuity, and visual acuity grade have been added;
- c) Snellen fraction values in [Table 1](#) have been added;
- d) [Figure 2](#) has been added;
- e) [Annex A](#) has been added.

# Ophthalmic optics — Visual acuity testing — Standard and clinical optotypes and their presentation

## 1 Scope

This document specifies a range of Landolt ring optotypes and describes a method for measuring distance visual acuity under photopic conditions for the purposes of certification or licensing.

This document is neither intended as a standard for clinical measurements nor for the certification of blindness or partial sight.

Other optotypes used for clinical investigations are described in [Annex A](#) for information.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3:1973, *Preferred numbers — Series of preferred numbers*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

### 3.1

#### **standard optotype**

Landolt ring

Note 1 to entry: The Landolt ring is specified in [Table 1](#) and [Figure 1](#).

### 3.2

#### **clinical optotype**

optotype other than the *standard optotype* ([3.1](#)) used for measuring *visual acuity* ([3.3](#))

Note 1 to entry: This definition does not exclude the standard optotype from being used for the same purposes as a clinical optotype.

Note 2 to entry: Since clinical optotypes can differ greatly in legibility, it is crucial to refer to the standard optotype whenever the comparability of the results is important. ISO/TR 19498 provides a method for correlation of clinical optotypes to the standard optotype.

### 3.3

#### **visual acuity**

number characterizing the ability of the visual system to recognize optotypes

Note 1 to entry: Currently, three different scaling systems are used to describe the visual acuity of a patient. These are decimal visual acuity, Snellen fraction, LogMAR acuity. See [Table 1](#).