

Elektrilised meditsiiniseadmed. Osa 2-24: Erinõuded infusioonpumpade ja kontrolleri ohutusele

Medical electrical equipment – Part 2-24: Particular requirements for the safety of infusion pumps and controllers

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 60601-2-24:2001 sisaldab Euroopa standardi EN 60601-2-24:1998 ingliskeelset teksti. Käesolev dokument on jõustatud 19.06.2001 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes. Standard on kättesaadav Eesti standardiorganisatsioonist.	This Estonian standard EVS-EN 60601-2-24:2001 consists of the English text of the European standard EN 60601-2-24:1998. This document is endorsed on 19.06.2001 with the notification being published in the official publication of the Estonian national standardisation organisation. The standard is available from Estonian standardisation organisation.
--	---

ICS 11.040.20

Standardite reprodutseerimis- ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute Estonian Standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: +372 605 5050; E-mail: info@evs.ee

Descriptors: Medical electrical equipment, infusion pump, infusion controller, safety requirements, protection against electric shock, protection against mechanical hazard, radiation protection, fire protection, environmental conditions

English version

Medical electrical equipment
Part 2: Particular requirements for the safety of
infusion pumps and controllers
(IEC 60601-2-24:1998)

Appareils électromédicaux
Partie 2: Règles particulières de sécurité
des pompes et régulateurs de perfusion
(CEI 60601-2-24:1998)

Medizinische elektrische Geräte
Teil 2: Besondere Festlegungen für
die Sicherheit von Infusionspumpen
und Steuergeräten
(IEC 60601-2-24:1998)

This European Standard was approved by CENELEC on 1998-04-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

The text of document 62D/250/FDIS, future edition 1 of IEC 60601-2-24, prepared by SC 62D, Electromedical equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-24 on 1998-04-01.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 1999-01-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2001-01-01

CLC/TC 62 notes to the text of IEC 60601-2-24:

- 1) Where the phrase "ISO class III water for medical use" is used it should be replaced by "water, Grade III, according to ISO 3696".
- 2) In figure 104b the syringe used in the test procedure needs to comply with ISO 7886-2.

Endorsement notice

The text of the International Standard IEC 60601-2-24:1998 was approved by CENELEC as a European Standard without any modification.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies (including amendments).

NOTE: When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Addition to annex ZA of EN 60601-1:1990/A2:1995:				
IEC 60529	1989	Degrees of protection provided by enclosures (IP Code)	EN 60529	1991
IEC 60601-1	1988	Medical electrical equipment Part 1: General requirements for safety	EN 60601-1	1990
A1	1991		+ corr. July A1	1994 1993
A2	1995		+ corr. July A2 ¹⁾ A13	1994 1995 1996
IEC 60601-1-2	1993	Medical electrical equipment Part 1: General requirements for safety 2. Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2	1993
IEC 60651	1979	Sound level meters	EN 60651	1994
A1	1993		A1	1994
IEC 60804	1985	Integrating-averaging sound level meters	EN 60804	1994
+ A1	1989		A2	1994
A2	1993			
ISO 3696	1987	Water for analytical laboratory use Specification and test methods	EN ISO 3696	1995
ISO 3744	1994	Acoustics - Determination of sound power levels of noise sources using sound pressure - Engineering method in an essentially free field over a reflecting plane	EN ISO 3744	1995
ISO 7864	1993	Sterile hypodermic needles for single use	EN ISO 7864	1995

1) A2 includes corrigendum June 1995 to IEC 60601-1:1988/A2.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
ISO 7886-2	1996	Sterile hypodermic syringes for single use Part 2: Syringes for use with power-driven syringe pumps	-	-
ISO 8536-4	1987	Infusion equipment for medical use Part 4: Infusion sets for single use	-	-

Annex ZB (informative)

Other international publications mentioned in this standard with the references of the relevant European publications

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Addition to annex ZB of EN 60601-1:1990/A2:1995:				
IEC 60521	1988	Class 0,5, 1 and 2 alternating-current watthour meters	EN 60521	1995
IEC 60801-1	1984	Electromagnetic compatibility for industrial-process measurement and control equipment Part 1 : General introduction	HD 481.1 S1	1987
IEC 60801-2	1991	Part 2: Electrostatic discharge requirements	EN 60801-2	1993
IEC 61000-4-3 (mod)	1995	Electromagnetic compatibility (EMC) Part 4: Testing and measurement techniques Section 3: Radiated, radio-frequency, electromagnetic field immunity test	EN 61000-4-3	1996
IEC 61000-4-4	1995	Section 4: Electrical fast transient/burst immunity test	EN 61000-4-4	1995

INTERNATIONAL STANDARD

IEC
60601-2-24

First edition
1998-02

Medical electrical equipment –

Part 2-24:

**Particular requirements for the safety of infusion
pumps and controllers**

Appareils électromédicaux –

Partie 2-24:

*Règles particulières de sécurité des pompes et régulateurs
de perfusion*



Reference number
IEC 60601-2-24:1998(E)

Numbering

As from 1 January 1997 all IEC publications are issued with a designation in the 60000 series.

Consolidated publications

Consolidated versions of some IEC publications including amendments are available. For example, edition numbers 1.0, 1.1 and 1.2 refer, respectively, to the base publication, the base publication incorporating amendment 1 and the base publication incorporating amendments 1 and 2.

Validity of this publication

The technical content of IEC publications is kept under constant review by the IEC, thus ensuring that the content reflects current technology.

Information relating to the date of the reconfirmation of the publication is available in the IEC catalogue.

Information on the revision work, the issue of revised editions and amendments may be obtained from IEC National Committees and from the following IEC sources:

- **IEC Bulletin**
- **IEC Yearbook**
On-line access*
- **Catalogue of IEC publications**
Published yearly with regular updates
(On-line access)*

Terminology, graphical and letter symbols

For general terminology, readers are referred to IEC 60050: *International Electrotechnical Vocabulary* (IEV).

For graphical symbols, and letter symbols and signs approved by the IEC for general use, readers are referred to publications IEC 60027: *Letter symbols to be used in electrical technology*, IEC 60417: *Graphical symbols for use on equipment. Index, survey and compilation of the single sheets* and IEC 60617: *Graphical symbols for diagrams*.

IEC publications prepared by the same technical committee

The attention of readers is drawn to the end pages of this publication which list the IEC publications issued by the technical committee which has prepared the present publication.

* See web site address on title page.

INTERNATIONAL STANDARD

IEC
60601-2-24

First edition
1998-02

Medical electrical equipment –

Part 2-24:

Particular requirements for the safety of infusion pumps and controllers

Appareils électromédicaux –

Partie 2-24:

Règles particulières de sécurité des pompes et régulateurs de perfusion

© IEC 1998 Copyright - all rights reserved — Droits de reproduction réservés

Aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'éditeur.

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Electrotechnical Commission
Telefax: +41 22 919 0300

3, rue de Varembé, Geneva, Switzerland
e-mail: inmail@iec.ch IEC web site: <http://www.iec.ch>



Commission Electrotechnique Internationale
International Electrotechnical Commission
Международная Электротехническая Комиссия

PRICE CODE **XB**

For price, see current catalogue

CONTENTS

	Page
FOREWORD	4
INTRODUCTION	5
Clause	
SECTION ONE – GENERAL	
1 Scope and object	6
2 Terminology and definitions	7
3 General requirements	10
5 Classification	11
6 Identification, marking and documents	11
SECTION TWO – ENVIRONMENTAL CONDITIONS	
10 Environmental conditions	13
SECTION THREE – PROTECTION AGAINST ELECTRIC SHOCK HAZARDS	
14 Requirements related to classification	14
17 Separation	14
19 Continuous LEAKAGE CURRENTS and PATIENT AUXILIARY CURRENTS	14
SECTION FOUR – PROTECTION AGAINST MECHANICAL HAZARDS	
21 Mechanical strength	16
SECTION FIVE – PROTECTION AGAINST HAZARDS FROM UNWANTED OR EXCESSIVE RADIATION	
36 Electromagnetic compatibility	17
SECTION SIX – PROTECTION AGAINST HAZARDS OF IGNITION OF FLAMMABLE ANAESTHETIC MIXTURES	
SECTION SEVEN – PROTECTION AGAINST EXCESSIVE TEMPERATURES AND OTHER SAFETY HAZARDS	
44 Overflow, spillage, leakage, humidity, ingress of liquids, cleaning, sterilization, disinfection and compatibility	19
47 Electrostatic charges	20
49 Interruption of the power supply	20

Clause		Page
	SECTION EIGHT – ACCURACY OF OPERATING DATA AND PROTECTION AGAINST HAZARDOUS OUTPUT	
50	Accuracy of operating data.....	21
51	Protection against hazardous output	38
	SECTION NINE – ABNORMAL OPERATION AND FAULT CONDITIONS: ENVIRONMENTAL TESTS	
	SECTION TEN – CONSTRUCTIONAL REQUIREMENTS	
54	General	42
56	Components and general assembly.....	44
	Annexes	
L	References – Publications mentioned in this standard	45
AA	General guidance and rationale.....	47

INTERNATIONAL ELECTROTECHNICAL COMMISSION

MEDICAL ELECTRICAL EQUIPMENT –**Part 2-24: Particular requirements for the safety
of infusion pumps and controllers**

FOREWORD

- 1) The IEC (International Electrotechnical Commission) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of the IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, the IEC publishes International Standards. Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. The IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of the IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested National Committees.
- 3) The documents produced have the form of recommendations for international use and are published in the form of standards, technical reports or guides and they are accepted by the National Committees in that sense.
- 4) In order to promote international unification, IEC National Committees undertake to apply IEC International Standards transparently to the maximum extent possible in their national and regional standards. Any divergence between the IEC Standard and the corresponding national or regional standard shall be clearly indicated in the latter.
- 5) The IEC provides no marking procedure to indicate its approval and cannot be rendered responsible for any equipment declared to be in conformity with one of its standards.
- 6) Attention is drawn to the possibility that some of the elements of this International Standard may be the subject of patent rights. The IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60601-2-24 has been prepared by subcommittee 62D: Electro-medical equipment, of IEC technical committee 62: Electrical equipment in medical practice.

The text of this Particular Standard is based on the following documents:

FDIS	Report on voting
62D/250/FDIS	62D/268/RVD

Full information on the voting for the approval of this Particular Standard can be found in the report on voting indicated in the above table.

Annex L is an integral part of this standard.

Annex AA is for information only.

In this Particular Standard the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type;
- notes, explanations, advice, introductions, general statements, exceptions and references: in smaller type;
- *test specifications: in italic type;*
- TERMS DEFINED IN CLAUSE 2 OF THE GENERAL STANDARD IEC 60601-1 OR THIS PARTICULAR STANDARD: SMALL CAPITALS.

A bilingual version of this standard may be issued at a later date.

INTRODUCTION

This Particular Standard deals with the safety of INFUSION PUMPS and CONTROLLERS. The relationship between this Particular Standard, IEC 60601-1 (including amendments 1 and 2), and the Collateral Standards is explained in 1.3.

The safe use of infusion pumps and controllers is primarily the responsibility of the OPERATOR. It is also recognized that OPERATORS should be trained in the operation of MEDICAL ELECTRICAL EQUIPMENT and that safe use of the EQUIPMENT can only be achieved if it is operated in accordance with the manufacturer's instructions for use. The minimum specified safety requirements are considered to provide a practical degree of safety in operation. It is the responsibility of the manufacturer to ensure that the requirements of this Particular Standard are reliably implemented. This Particular Standard has been developed in accordance with these principles.

Safe use can be ensured only if the associated disposable parts, especially lines and syringes are consistent with the system. ISO 7886-2:1996, *Sterile hypodermic syringes for single use – Part 2: Syringes for use with power-driven syringe pumps* should be taken into account.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-24: Particular requirements for the safety of infusion pumps and controllers

SECTION ONE – GENERAL

The clauses and subclauses of this section of the General Standard and of this section of the Collateral Standard IEC 60601-1-2 apply, except as follows:

1 Scope and object

This clause of the General Standard and this clause of the Collateral Standard IEC 60601-1-2 apply, except as follows:

1.1* Scope

Addition:

This Particular Standard specifies the requirement for INFUSION PUMPS, INFUSION CONTROLLERS, SYRINGE PUMPS and PUMPS FOR AMBULATORY USE, as defined in 2.101 to 2.110. These devices are intended for use by medical staff and home PATIENTS as prescribed and medically indicated. These particular requirements do not apply to devices:

- 1) specifically intended for diagnostic or similar use (e.g. angiography or other pumps permanently controlled or supervised by the OPERATOR),
- 2) enteral infusion,
- 3) extracorporeal circulation of blood,
- 4) implantable or disposable devices,
- 5) EQUIPMENT specifically intended for diagnostic use within urodynamics (measurement of pressure-volume relationship of the urinary bladder when filled through a catheter with water);
- 6) EQUIPMENT specifically intended for diagnostic use within male impotence testing (measurement of amount of liquid infused, necessary to maintain a preset pressure level for maintaining penile erection: cavernosometry, cavernosography).

1.3 Particular standards

Addition:

This Particular Standard refers to IEC 60601-1:1988, *Medical electrical equipment – Part 1: General requirements for safety* as amended by its amendment 1 (1991) and amendment 2 (1995) and to the Collateral Standard IEC 60601-1-2:1993, *Medical electrical equipment – Part 1: General requirements for safety – 2. Collateral Standard: Electromagnetic compatibility – Requirements and tests*.

For brevity, Part 1 is referred to in this Particular Standard either as the General Standard or as the General Requirement(s) and IEC 60601-1-2 as the Collateral Standard.

The numbering of sections, clauses and subclauses of this Particular Standard corresponds to that of the General Standard. The changes to the text of the General Standard are specified by the use of the following words: