

Elektrilised meditsiiniseadmed. Osa 2-4: Erinõuded südamedefibrillaatorite esmasele ohutusele ja olulistele toimimisinäitajatele

Medical electrical equipment - Part 2-4: Particular requirements for basic safety and essential performance of cardiac defibrillators

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 60601-2-4:2011 sisaldab Euroopa standardi EN 60601-2-4:2011 ingliskeelset teksti.	This Estonian standard EVS-EN 60601-2-4:2011 consists of the English text of the European standard EN 60601-2-4:2011.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.08.2011.	Date of Availability of the European standard is 19.08.2011.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.10

Võtmesõnad: electrical medic, electrical medical e, electromedicine, equipment safety, measuring instruments, medical equipment, medical sciences, medicine, monitors, safety, safety engineering, safety requirements, specification (approval), specifications, testing conditions,

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Aru 10, 10317 Tallinn, Eesti; www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:
Aru 10, 10317 Tallinn, Estonia; www.evs.ee; phone 605 5050; e-mail info@evs.ee

English version

**Medical electrical equipment -
Part 2-4: Particular requirements for the basic safety and essential
performance of cardiac defibrillators
(IEC 60601-2-4:2010)**

Appareils électromédicaux -
Partie 2-4: Exigences particulières pour la
sécurité de base et les performances
essentielle des défibrillateurs cardiaques
(CEI 60601-2-4:2010)

Medizinische elektrische Geräte -
Teil 2-4: Besondere Festlegungen für die
Sicherheit einschließlich der wesentlichen
Leistungsmerkmale von Defibrillatoren
(IEC 60601-2-4:2010)

This European Standard was approved by CENELEC on 2011-01-12. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62D/857/FDIS, future edition 3 of IEC 60601-2-4, prepared by SC 62D, Electromedical equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-4 on 2011-01-12.

This European Standard supersedes EN 60601-2-4:2003.

EN 60601-2-4:2011 constitutes a technical revision, revised to structurally align it with EN 60601-1:2006 and to implement the decision of IEC SC 62A that the clause numbering structure of particular standards written to EN 60601-1:2006 would adhere to the form specified in ISO/IEC Directives, Part 2:2004. The aim of this third edition is to bring this particular standard up to date with reference to the third edition of the general standard through reformatting and technical changes.

The principle technical changes are as follows:

- 201.8.8.3, test 4: added additional test options;
- Figure 201.105: provided example of stainless steel plates. Added note for 10 Hz generator or shockable rhythm generator;
- Figure 201.101: Changed orientation of the lower diode at the oscilloscope connection;
- 202.6.1, .2, .4: "Additions" and "Replacements" corrected to be as originally intended;
- 201.101.1: Clarified preconditioning of a non-rechargeable battery;
- 201.3.207: Clarified definition of DUMMY COMPONENT;
- 201.15.4.101: In paragraph b), added reduced flex requirements for sterilizable internal paddles with specified limit on sterilization cycles;
- 201.15.4.3.103: Added an option for devices having non-changeable pre-programmed energy-setting sequences;
- 201.102.3.1, 2: Changed from specified defibrillation cycles to use of pre-programmed defibrillation sequence;
- 202.6.2.2.1: Changed ESD discharge sequence to match EN 60601-1-2:2007.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- | | | |
|--|-------|------------|
| <ul style="list-style-type: none"> – latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2012-02-12 |
| <ul style="list-style-type: none"> – latest date by which the national standards conflicting with the EN have to be withdrawn | (dow) | 2014-01-12 |

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.

– TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive MDD (93/42/EEC). See Annex ZZ.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-4:2010 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

[2] IEC 60601-2-27 NOTE Harmonized as EN 60601-2-27.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Annex ZA of EN 60601-1:2006 applies except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Replace IEC 60601-1-2 by:				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr. March	2007 2010
Add:				
IEC 61000-4-2	-	Electromagnetic compatibility (EMC) - Part 4-2: Testing and measurement techniques - Electrostatic discharge immunity test	EN 61000-4-2	-
ISO 15223-1	2007	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements	-	-

Annex ZZ
(informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

This document is a preview generated by EVS

CONTENTS

FOREWORD.....	4
201.1 Scope, object and related standards.....	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	9
201.4 General requirements	11
201.5 General requirements for testing of ME EQUIPMENT.....	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents.....	13
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	17
201.9 Protection against MECAHNMICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	23
201.10 Protection against unwanted and excessive radiation HAZARDS.....	23
201.11 Protection against excessive temperatures and other HAZARDS.....	23
201.12 * Accuracy of controls and instruments and protection against hazardous outputs.....	25
201.13 HAZARDOUS SITUATIONS and fault conditions.....	27
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	27
201.15 Construction of ME EQUIPMENT.....	27
201.16 ME SYSTEMS	32
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS.....	32
201.101 * Charging time	32
201.102 Internal electrical power source.....	35
201.103 * Endurance.....	36
201.104 * Synchronizer.....	37
201.105 * Recovery of the MONITOR and/or ECG input after defibrillation.....	37
201.106 * Disturbance to the MONITOR from charging or internal discharging	41
201.107 * Requirements for RHYTHM RECOGNITION DETECTOR.....	42
201.108 DEFIBRILLATOR ELECTRODES.....	43
201.109 * External pacing (U.S.).....	45
202 * Electromagnetic compatibility – Requirements and tests.....	49
Annexes.....	52
Annex C (informative) Guide to marking and labelling requirements for ME EQUIPMENT and ME SYSTEMS	53
Annex AA (informative) Particular guidance and rationale.....	55
Annex BB (informative) Mapping between the elements of the second edition of IEC 60601-2-4 and IEC 60601-2-4:2010	68
Bibliography.....	73
Index of defined terms used in this particular standard	74
Figure 201.101 – Dynamic test for limitation of energy from different parts of the ME EQUIPMENT	18
Figure 201.102 – Allowed current versus applied test voltage	22
Figure 201.103 – Examples of cord anchorages that require testing.....	31
Figure 201.104 – Test apparatus for flexible cords and their anchorages.....	32

Figure 201.105 – Arrangement for test of recovery after defibrillation.....	39
Figure 201.106 – Arrangement of monitoring electrodes on sponge	40
Figure 201.107 – Arrangement for recovery test after defibrillation	40
Figure 201.108 – Arrangement for test of disturbance from charging and internal discharging.....	42
Figure 201.109 – Test circuit for offset instability/internal noise determination	49
Figure 201.110 – Test circuit for DEFIBRILLATOR overload test of pacing output circuitry.....	49
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	12
Table 201.102 – Rhythm recognition detector categories.....	42
Table 201.C.101 – Marking on the outside of a CARDIAC DEFIBRILLATOR or its parts	53
Table 201.C.102 – Marking of controls and instruments of a CARDIAC DEFIBRILLATOR.....	53
Table 201.C.103 – ACCOMPANYING DOCUMENTS, general.....	53
Table 201.C.104 – ACCOMPANYING DOCUMENTS, instructions for use	54
Table 201.C.105 – ACCOMPANYING DOCUMENTS, technical description.....	54
Table BB.1 – Mapping between the elements of the second edition of IEC 60601-2-4 and IEC 60601-2-4:2010	68

document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of CARDIAC DEFIBRILLATORS, hereafter referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

HAZARDS inherent in the intended physiological function of ME EQUIPMENT or ME SYSTEMS within the scope of this standard are not covered by specific requirements in this standard except in 7.2.13 and 8.4.1 of the general standard.

NOTE See also 4.2 of the general standard.

This particular standard does not apply to implantable defibrillators, remote control DEFIBRILLATORS, external transcutaneous pacemakers, or separate stand-alone cardiac monitors (which are standardized by IEC 60601-2-27 [2]²). Cardiac monitors which use separate ECG monitoring electrodes are not within the scope of this standard unless they are used as the sole basis for AED rhythm recognition detection or beat detection for synchronized cardioversion.

Defibrillation waveform technology is evolving rapidly. Published studies indicate that the effectiveness of waveforms varies. The choice of a particular waveform including waveshape, delivered energy, efficacy, and safety has been specifically excluded from the scope of this standard.

However, due to the critical importance of the therapeutic waveform, comments have been added to the rationale which addresses considerations in waveform selection.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for cardiac defibrillators as defined in 201.3.202.

¹ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

² Numbers in square brackets refer to the bibliography.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 applies as modified in Clause 202. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general

standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

NOTE Informative references are listed in the bibliography beginning on page 73.

Clause 2 of the general standard applies, except as follows:

Amendment:

IEC 60601-1-2:2007, *Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

Addition:

IEC 61000-4-2, *Electromagnetic compatibility (EMC) – Part 4-2: Testing and measurement techniques – Electrostatic discharge immunity test*

ISO 15223-1:2007, *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements*

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005 apply, except as follows:

Addition:

NOTE An index of defined terms is found beginning on page 74.

201.3.201

AUTOMATED EXTERNAL DEFIBRILLATOR AED

DEFIBRILLATOR that, once activated by the OPERATOR, analyses the ECG obtained from electrodes placed on the patient's skin identifies shockable cardiac rhythms, and automatically operates the DEFIBRILLATOR when a shockable rhythm is detected, hereinafter referred to as an AED

NOTE AEDs may provide varying levels of automation and be referred to by various terms. A semi-automatic DEFIBRILLATOR requires manual shock activation. A fully automatic DEFIBRILLATOR will provide shock without OPERATOR intervention.

201.3.202

CARDIAC DEFIBRILLATOR

MEDICAL ELECTRICAL EQUIPMENT intended to normalize the rhythm of the heart by an electrical pulse via electrodes applied either to the PATIENT's skin with external electrodes or to the exposed heart with internal electrodes

NOTE 1 A CARDIAC DEFIBRILLATOR can be referred to in this standard as a DEFIBRILLATOR or as ME EQUIPMENT.

NOTE 2 Such ME EQUIPMENT may also include other monitoring or therapeutic functions.

201.3.203

CHARGING CIRCUIT

circuit within the DEFIBRILLATOR intended for charging the ENERGY STORAGE DEVICE. This circuit includes all parts conductively connected to the ENERGY STORAGE DEVICE during the charging period