

**Elektrilised meditsiiniseadmed. Osa 2-43: Erinõuded
invasiivprotseduuride röntgenseadmete esmasele
ohutusele ja olulistele toimimisnäitajatele**

Medical electrical equipment - Part 2-43: Particular
requirements for basic safety and essential performance of X
ray equipment for interventional procedures

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 60601-2-43:2010 sisaldab Euroopa standardi EN 60601-2-43:2010 ingliskeelset teksti.

Standard on kinnitatud Eesti Standardikeskuse 31.08.2010 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 11.06.2010.

Standard on kättesaadav Eesti standardiorganisatsioonist.

This Estonian standard EVS-EN 60601-2-43:2010 consists of the English text of the European standard EN 60601-2-43:2010.

This standard is ratified with the order of Estonian Centre for Standardisation dated 31.08.2010 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

Date of Availability of the European standard text 11.06.2010.

The standard is available from Estonian standardisation organisation.

ICS 11.040.50, 37.040.25

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute Estonian Standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: +372 605 5050; E-mail: info@evs.ee

**Medical electrical equipment -
Part 2-43: Particular requirements for basic safety and essential
performance of X-ray equipment for interventional procedures
(IEC 60601-2-43:2010)**

Appareils électromédicaux -
Partie 2-43: Exigences particulières
pour la sécurité de base
et les performances essentielles
des appareils à rayonnement X
lors d'interventions
(CEI 60601-2-43:2010)

Medizinische elektrische Geräte -
Teil 2-43: Besondere Festlegungen
für die Sicherheit
von Röntgeneinrichtungen
für interventionelle Verfahren
(IEC 60601-2-43:2010)

This European Standard was approved by CENELEC on 2010-06-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62B/779/FDIS, future edition 2 of IEC 60601-2-43, prepared by SC 62B, Diagnostic imaging equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-43 on 2010-06-01.

This European Standard supersedes EN 60601-2-43:2000 and partially supersedes EN 60601-2-54:2009.

This particular standard has been revised to provide a complete set of safety requirements for X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, based on EN 60601-1:2006 and relevant collaterals. EN 60601-2-43:2010 is extended to become a system standard for X-RAY EQUIPMENT designed for the use during interventional procedures using X-ray imaging, whether of prolonged or normal duration.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

The following dates were fixed:

- latest date by which the EN has to be implemented
at national level by publication of an identical
national standard or by endorsement (dop) 2011-03-01
- latest date by which the national standards conflicting
with the EN have to be withdrawn (dow) 2013-06-01

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive 93/42/EEC. See Annex ZZ.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-43:2010 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

[2] IEC 60601-2-44 NOTE Harmonized as EN 60601-2-44.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
<i>Amendment:</i>				
IEC 60601-1-2	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr. March 2010	2007
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3 + corr. March 2010	2008
<i>Addition:</i>				
IEC 60580	-	Medical electrical equipment - Dose area product meters	EN 60580	-
IEC 60601-2-54	2009	Medical electrical equipment - Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy	EN 60601-2-54	2009
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 62220-1	2003	Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1: Determination of the detective quantum efficiency	EN 62220-1	2004

Annex ZZ (informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directives concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

This document is a preview generated by EVS

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references	9
201.3 Terms and definitions	9
201.4 General requirements.....	10
201.5 General requirements for testing of ME EQUIPMENT.....	11
201.6 Classification of ME EQUIPMENT and ME SYSTEMS	12
201.7 ME EQUIPMENT identification, marking and documents.....	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS.....	16
201.10 Protection against unwanted and excessive radiation HAZARDS.....	17
201.11 Protection against excessive temperatures and other HAZARDS.....	17
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	19
201.13 HAZARDOUS SITUATIONS and fault conditions.....	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	22
201.15 Construction of ME EQUIPMENT	22
201.16 ME SYSTEMS	23
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	23
202 Electromagnetic compatibility – Requirements and tests	23
203 Radiation protection in diagnostic X-ray equipment.....	24
Annexes	34
Annex AA (informative) Particular guidance and rationale.....	35
Annex BB (normative) Distribution maps of STRAY RADIATION.....	43
Annex CC (informative) Mapping between this Edition 2 of IEC 60601-2-43 and Edition 1	47
Bibliography.....	49
Index of defined terms used in this particular standard.....	51
Figure BB.1 – Example of isokerma map at 100 cm height in lateral configuration	45
Figure BB.2 – Example of isokerma map at 100 cm height in vertical configuration.....	46
Table 201.101 – Additional list of potential ESSENTIAL PERFORMANCE to be considered by MANUFACTURER in the RISK MANAGEMENT analysis.....	10
Table 201.102 – Other subclauses requiring statements in ACCOMPANYING DOCUMENTS	15
Table AA.1 – Examples of prolonged RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects of IRRADIATION are possible	35
Table AA.2 – Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES for which deterministic effects are unlikely	36

INTRODUCTION

X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES may subject PATIENTS and OPERATORS to higher levels of RADIATION than those which normally prevail during diagnostic X-ray imaging procedures. One consequence for the PATIENT may be the occurrence of deterministic injury when procedures involve the delivery of substantial amounts of RADIATION to localized areas. Another consequence can be an increased RISK of stochastic effects, such as cancer. These health concerns apply also to the OPERATOR. In addition, for this particular type of equipment, there is a need for availability of critical functions with minimal periods of loss.

Interventional procedures of the type envisaged are well established in clinical fields such as:

- invasive cardiology;
- interventional RADIOLOGY;
- interventional neuroradiology.

These procedures also include many newly developing and emerging applications in a wide range of medical and surgical specialities.

NOTE Attention is drawn to the existence of legislation in some countries concerning RADIOLOGICAL PROTECTION, which may not align with the provisions of this standard.

This document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 * Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, hereafter referred to as INTERVENTIONAL X-RAY EQUIPMENT. Its scope excludes, in particular:

- equipment for RADIOTHERAPY;
- equipment for COMPUTED TOMOGRAPHY;
- ACCESSORIES intended to be introduced into the PATIENT;
- mammographic X-RAY EQUIPMENT;
- dental X-RAY EQUIPMENT.

NOTE 1 Examples of RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, for which the use of INTERVENTIONAL X-RAY EQUIPMENT complying with this standard is recommended, are given in Annex AA.

NOTE 2 Specific requirements for magnetic navigation devices, and for the use of INTERVENTIONAL X-RAY EQUIPMENT in an operating room environment were not considered in this particular standard; therefore no specific requirements have been developed for these devices or uses. In any case, such devices or uses remain under the general clause requirements.

NOTE 3 INTERVENTIONAL X-RAY EQUIPMENT when used in cross-sectional imaging mode (sometimes described as CT-like mode or cone-beam CT) is covered by this particular standard and not by IEC 60601-2-44 [2]²⁾. Additional requirements for operation in CT-like mode or cone-beam CT were not considered in the present standard.

INTERVENTIONAL X-RAY EQUIPMENT declared by the MANUFACTURER to be suitable for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, which does not include a PATIENT SUPPORT as part of the system, is exempt from the PATIENT SUPPORT provisions of this standard.

If a clause or subclause is specifically intended to be applicable to INTERVENTIONAL X-RAY EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to INTERVENTIONAL X-RAY EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE 4 See also 4.2 of the general standard.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.

²⁾ Figures in square brackets refer to the Bibliography.

201.1.2 Object

Replacement:

The object of this particular standard is:

- to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for the design and manufacture of X-RAY EQUIPMENT for RADIOSCOPICALLY GUIDED INTERVENTIONAL PROCEDURES, as defined in 201.3.203.
- to specify information which is to be provided with such INTERVENTIONAL X-RAY EQUIPMENT for the assistance of the RESPONSIBLE ORGANIZATION and OPERATOR in managing the RADIATION RISK and equipment failure RISK arising from these procedures which could affect PATIENTS or staff.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clause 202 and Clause 203 respectively. IEC 60601-1-8 and IEC 60601-1-10 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.