

ELEKTRILISED MEDITSIINISEADMED. OSA 2-45:
ERINÕUDED MAMMOGRAAFILISTE RÖNTGENSEADMETE
JA MAMMOGRAAFILISTE STEREOTAKTILISTE SEADISTE
ESMASELE OHUTUSELE JA OLULISTELE
TOIMIMISNÄITAJATELE

Medical electrical equipment - Part 2-45: Particular
requirements for the basic safety and essential
performance of mammographic X-ray equipment and
mammographic stereotactic devices

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 60601-2-45:2011 sisaldab Euroopa standardi EN 60601-2-45:2011 ingliskeelset teksti.	This Estonian standard EVS-EN 60601-2-45:2011 consists of the English text of the European standard EN 60601-2-45:2011.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 18.03.2011.	Date of Availability of the European standard is 18.03.2011.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.50

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

English version

**Medical electrical equipment -
Part 2-45: Particular requirements for the basic safety and essential
performance of mammographic X-ray equipment and mammographic
stereotactic devices
(IEC 60601-2-45:2011)**

Appareils électromédicaux -
Partie 2-45: Exigences particulières pour
la sécurité de base et les performances
essentielles des appareils de
mammographie à rayonnement X et des
appareils mammographiques
stéréotaxiques
(CEI 60601-2-45:2011)

Medizinische elektrische Geräte -
Teil 2-45: Besondere Festlegungen für die
Sicherheit einschließlich der wesentlichen
Leistungsmerkmale von Röntgen-
Mammographiegeräten und
mammographischen Stereotaxie-
Einrichtungen
(IEC 60601-2-45:2011)

This European Standard was approved by CENELEC on 2011-03-17. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Management Centre: Avenue Marnix 17, B - 1000 Brussels

Foreword

The text of document 62B/817/FDIS, future edition 3 of IEC 60601-2-45, prepared by SC 62B, Diagnostic imaging equipment, of IEC TC 62, Electrical equipment in medical practice, was submitted to the IEC-CENELEC parallel vote and was approved by CENELEC as EN 60601-2-45 on 2011-03-17.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN and CENELEC shall not be held responsible for identifying any or all such patent rights.

This European Standard supersedes EN 60601-2-45:2001.

EN 60601-2-45:2011 has been aligned to EN 60601-1:2006 and to EN 60601-1-3:2008 + corrigendum March 2010. Further modifications have been made with respect to the current technology of MAMMOGRAPHIC X-RAY EQUIPMENT.

The following dates were fixed:

- | | | |
|--|-------|------------|
| – latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement | (dop) | 2011-12-17 |
| – latest date by which the national standards conflicting with the EN have to be withdrawn | (dow) | 2014-03-17 |

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- Test specifications: italic type.
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g., Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g., 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or”, so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex AA.

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and covers essential requirements of EC Directive MDD (93/423/EEC). See Annex ZZ.

Annexes ZA and ZZ have been added by CENELEC.

Endorsement notice

The text of the International Standard IEC 60601-2-45:2011 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60601-2-7:1998	NOTE Harmonized as EN 60601-2-7:1998 (not modified).
IEC 60601-2-28:2010	NOTE Harmonized as EN 60601-2-28:2010 (not modified).
IEC 60601-2-32:1994	NOTE Harmonized as EN 60601-2-32:1994 (not modified).
IEC 60664-1:2007	NOTE Harmonized as EN 60664-1:2007 (not modified).
ISO 4090:2001	NOTE Harmonized as EN ISO 4090:2004 (not modified).
ISO 12052	NOTE Harmonized as EN ISO 12052 (not modified).

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE When an international publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

Annex ZA of EN 60601-1:2006 applies, except as follows:

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
Replace IEC 60601-1-2 and IEC 60601-1-3 by:				
IEC 60601-1-2 (mod)	2007	Medical electrical equipment - Part 1-2: General requirements for basic safety and essential performance - Collateral standard: Electromagnetic compatibility - Requirements and tests	EN 60601-1-2 + corr. March	2007 2010
IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3 + corr. March	2008 2010
Add:				
IEC 60336	2005	Medical electrical equipment - X-ray tube assemblies for medical diagnosis - Characteristics of focal spots	EN 60336	2005
IEC 60613	2010	Electrical and loading characteristics of X-ray tube assemblies for medical diagnosis	EN 60613	2010
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 61223-3-2	2007	Evaluation and routine testing in medical imaging departments - Part 3-2: Acceptance tests - Imaging performance of mammographic X-ray equipment	EN 61223-3-2	2008
IEC 62220-1-2	2007	Medical electrical equipment - Characteristics of digital X-ray imaging devices - Part 1-2: Determination of the detective quantum efficiency - Detectors use in mammography	EN 62220-1-2	2007
ISO 9236-3	1999	Photography - Sensitometry of screen/film systems for medical radiography - Part 3: Determination of sensitometric curve shape, speed and average gradient for mammography	-	-

Annex ZZ (informative)

Coverage of Essential Requirements of EC Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association and within its scope the standard covers all relevant essential requirements as given in Annex I of the EC Directive 93/42/EEC.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EC Directives may be applicable to the products falling within the scope of this standard.

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	6
201.1 Scope, object and related standards	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements	11
201.5 General requirements for testing of ME EQUIPMENT	12
201.6 Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7 ME EQUIPMENT identification, marking and documents	12
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	16
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	19
201.10 Protection against unwanted and excessive radiation HAZARDS	21
201.11 Protection against excessive temperatures and other HAZARDS	22
201.12 Accuracy of controls and instruments and protection against hazardous outputs	22
201.13 Hazardous situations and fault conditions	22
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	22
201.15 Construction of ME EQUIPMENT.....	22
201.16 ME SYSTEMS	22
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	23
202 Electromagnetic compatibility – Requirements and tests	23
203 Radiation protection in diagnostic X-ray equipment.....	23
Annex AA (informative) Particular guidance and rationale.....	46
Bibliography.....	48
Index of defined terms used in this particular standard.....	49
Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements	11
Table 203.101 – Minimum values of TOTAL FILTRATION and factors for determining the minimum AIR KERMA RATE	38

INTRODUCTION

The third edition of this particular standard has been prepared to provide a complete set of safety requirements for MAMMOGRAPHIC X-RAY EQUIPMENT, based on IEC 60601-1:2005 (3rd edition) and its collaterals. This particular standard addresses the system level of MAMMOGRAPHIC X-RAY EQUIPMENT, which consists of a combination of an X-RAY GENERATOR, associated equipment and ACCESSORIES. Components functions are addressed as far as necessary.

The minimum safety requirements specified in this particular standard are considered to provide for a practical degree of safety in the operation of MAMMOGRAPHIC X-RAY EQUIPMENT.

Like the previous edition of this Part 2-45, the present third edition includes requirements on HIGH-VOLTAGE GENERATORS for mammography.

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-45: Particular requirements for the basic safety and essential performance of mammographic X-ray equipment and mammographic stereotactic devices

201.1 Scope, object and related standards

Clause 1 of the general standard¹⁾ applies, except as follows:

201.1.1 Scope

Replacement:

This international standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of MAMMOGRAPHIC X-RAY EQUIPMENT and MAMMOGRAPHIC STEREOTACTIC DEVICES, hereafter also referred to as ME EQUIPMENT.

NOTE 1 This includes MAMMOGRAPHIC X-RAY EQUIPMENT using integrated digital X-RAY IMAGE RECEPTORS or integrated storage phosphor subsystems.

Excluded from the scope of this document are:

- reconstructive tomography modes of operation;
- diagnostic consoles;
- picture archiving and communication systems (PACS);
- non-integrated storage phosphor readers;
- hard copy cameras;
- films, screens and cassettes;
- computer aided detection (CAD);
- devices for performing core biopsy and other biopsy instruments;
- modes of operation intended to demonstrate local contrast medium uptake (contrast enhanced digital mammography);

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE 2 IEC 60601-2-7:1998 and IEC 60601-2-32 are not part of the 3rd edition scheme for MAMMOGRAPHIC X-RAY EQUIPMENT and MAMMOGRAPHIC STEREOTACTIC DEVICES.

201.1.2 Object

Replacement:

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for MAMMOGRAPHIC X-RAY EQUIPMENT and MAMMOGRAPHIC STEREOTACTIC DEVICES, to ensure safety, to specify methods for demonstrating compliance with those requirements and to provide guidance for RISK MANAGEMENT.

¹⁾ The general standard is IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*.