

---

# International Standard



# 8288

---

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

---

## **Water quality — Determination of cobalt, nickel, copper, zinc, cadmium and lead — Flame atomic absorption spectrometric methods**

*Qualité de l'eau — Dosage du cobalt, nickel, cuivre, zinc, cadmium et plomb — Méthodes par spectrométrie d'absorption atomique avec flamme*

**First edition — 1986-03-15**

---

**UDC 543.3 : 543.422 : 546.47/.48 : 546.56 : 546.74/.76**

**Ref. No. ISO 8288-1986 (E)**

**Descriptors:** water, quality, chemical analysis, determination of content, cobalt, nickel, copper, zinc, cadmium, lead, atomic absorption method.

Price based on 11 pages

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work.

Draft International Standards adopted by the technical committees are circulated to the member bodies for approval before their acceptance as International Standards by the ISO Council. They are approved in accordance with ISO procedures requiring at least 75 % approval by the member bodies voting.

International Standard ISO 8288 was prepared by Technical Committee ISO/TC 147, *Water quality*.

Users should note that all International Standards undergo revision from time to time and that any reference made herein to any other International Standard implies its latest edition, unless otherwise stated.

# Water quality — Determination of cobalt, nickel, copper, zinc, cadmium and lead — Flame atomic absorption spectrometric methods

## 1 Scope

This International Standard specifies three methods for the determination of cobalt, nickel, copper, zinc, cadmium and lead in water by flame atomic absorption spectrometry:

**Section one:** method A, for direct determination by flame atomic absorption spectrometry;

**Section two:** method B, for determination by flame atomic absorption spectrometry after chelation (APDC) and extraction (MIBK);

**Section three:** method C, for determination by flame atomic absorption spectrometry after chelation (HMA, HMDC) and extraction (DIPK-xylene).

## 2 Field of application

**2.1** Method A is particularly applicable when concentrations of elements to be analysed are relatively high and when there are no interferences.

When the samples are of a complex or unknown nature or when they contain high concentrations of dissolved solids (brines or brackish waters) method A is not applicable and either method B or C should be selected.

The concentrations of elements which can be determined by method A may vary according to the characteristics of the atomic absorption spectrometric apparatus used but are generally in the ranges indicated in table 1.

Table 1

Element to be determined	Range of determination (mg/l)
Cobalt	0,1 to 10
Nickel	0,1 to 10
Copper	0,05 to 6
Zinc	0,05 to 2
Cadmium	0,02 to 2
Lead	0,2 to 10

If the concentrations are greater than the upper limits indicated in the table, the sample may be diluted before analysis.

**2.2** Methods B and C are applicable when concentrations of elements to be analysed in the sample (or dilution of the sample) are greater than 0,5 µg/l.

### 2.2.1 Method B

The concentrations of the elements which can be determined by method B may vary according to the characteristics of the atomic absorption spectrometer used but are generally in the ranges indicated in table 2.

Table 2

Element to be determined	Range of determination (µg/l)
Cobalt	1 to 200
Nickel	1 to 200
Copper	1 to 200
Zinc	0,5 to 50
Cadmium	0,5 to 50
Lead	5 to 200

### 2.2.2 Method C

With a ratio of test portion to extraction solution of 20 to 1 by volume as indicated in 21.2, the concentrations of elements which can be determined by method C vary as indicated in table 3.

Table 3

Element to be determined	Range of determination (µg/l)
Cobalt	0,5 to 100
Nickel	0,5 to 100
Copper	0,5 to 100
Zinc	0,2 to 50
Cadmium	0,2 to 50
Lead	2 to 200