

Foodstuffs - Determination of elements and their chemical species - Determination of aluminium by inductively coupled plasma mass spectrometry (ICP-MS)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 17264:2019 sisaldab Euroopa standardi EN 17264:2019 ingliskeelset teksti.	This Estonian standard EVS-EN 17264:2019 consists of the English text of the European standard EN 17264:2019.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 11.09.2019.	Date of Availability of the European standard is 11.09.2019.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 67.050

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:
Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

ICS 67.050

English Version

Foodstuffs - Determination of elements and their chemical species - Determination of aluminium by inductively coupled plasma mass spectrometry (ICP-MS)

Produits alimentaires - Dosage des éléments et de leurs espèces chimiques - Dosage de l'aluminium par spectrométrie de masse avec plasma à couplage inductif (ICP-MS)

Lebensmittel - Bestimmung von Elementen und ihren Verbindungen - Bestimmung von Aluminium mittels Massenspektrometrie mit induktiv gekoppeltem Plasma (ICP-MS)

This European Standard was approved by CEN on 28 July 2019.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword.....	3
1 Scope	3
2 Normative references	4
3 Terms and definitions	4
4 Principle	4
5 Reagents	4
6 Apparatus	6
7 Procedure	7
7.1 Digestion	7
7.2 Mass spectrometry with inductively coupled plasma	8
8 Evaluation	10
8.1 Calculation	10
8.2 Precision	10
8.3 Repeatability	10
8.4 Reproducibility	11
9 Test report	11
Annex A (informative) Results of interlaboratory study	12
Annex B (normative) Explanation notes	14
Bibliography	16

European foreword

This document (EN 17264:2019) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2020, and conflicting national standards shall be withdrawn at the latest by March 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This document specifies a method for the determination of aluminium in food by inductively coupled plasma mass spectrometry (ICP-MS) after pressure digestion. This method was validated for infant formula, wheat noodle, cheese, liver, beetroot and cocoa powder at mass fractions in the range of 1 mg/kg to 200 mg/kg. At concentrations above 200 mg/kg, digestion temperatures higher than 220 °C can be necessary to recover the aluminium as completely as possible.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13804, *Foodstuffs - Determination of elements and their chemical species - General considerations and specific requirements*

EN 13805, *Foodstuffs - Determination of trace elements - Pressure digestion*

EN ISO 3696, *Water for analytical laboratory use - Specification and test methods (ISO 3696)*

3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

4 Principle

Aluminium is determined quantitatively by ICP-MS after digestion of the sample with nitric acid (with addition of water in case of foods with low water content) according to the pressure digestion process described in EN 13805 but without the use of hydrofluoric acid. The digestion conditions are chosen in such a way that even for samples with aluminium compounds of low solubility (e.g. silicates, oxides) low findings are kept to a minimum.

5 Reagents

The mass concentration of aluminium shall be low enough in the reagents and water not to affect the results.

All reagents shall be of analytical grade, i.e. pro analysi, p.a. or similar unless otherwise specified.

Use water conforming to grade 2 of EN ISO 3696.

5.1 Nitric acid, mass fraction w = at least 65 %, density = 1,4 g/ml.

5.2 Aluminium stock solution, with a certified mass concentration ρ = 1 000 mg/l.

5.3 Rhodium stock solution, ρ = 1 000 mg/l as internal standard.

The internal standard shall be free from aluminium impurities and shall be present in the sample in negligibly small amounts only.