

Geometrical Product Specifications (GPS) - Acceptance and reverification tests for coordinate measuring machines (CMM) - Part 5: CMMs using single and multiple stylus contacting probing systems

This document is a preview generated by EVS

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN ISO 10360-5:2010 sisaldab Euroopa standardi EN ISO 10360-5:2010 ingliskeelset teksti.

Standard on kinnitatud Eesti Standardikeskuse 31.10.2010 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 15.09.2010.

Standard on kättesaadav Eesti standardiorganisatsioonist.

This Estonian standard EVS-EN ISO 10360-5:2010 consists of the English text of the European standard EN ISO 10360-5:2010.

This standard is ratified with the order of Estonian Centre for Standardisation dated 31.10.2010 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

Date of Availability of the European standard text 15.09.2010.

The standard is available from Estonian standardisation organisation.

ICS 17.040.30

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

English Version

Geometrical product specifications (GPS) - Acceptance and
reverification tests for coordinate measuring machines (CMM) -
Part 5: CMMs using single and multiple stylus contacting probing
systems (ISO 10360-5:2010)

Spécification géométrique des produits (GPS) - Essais de
réception et de vérification périodique des machines à
mesurer tridimensionnelles (MMT) - Partie 5: MMT utilisant
des systèmes de palpation à stylet simple ou à stylets
multiples (ISO 10360-5:2010)

Geometrische Produktspezifikation (GPS) -
Annahmeprüfung und Bestätigungsprüfung für
Koordinatenmessgeräte (KMG) - Teil 5: Prüfung der
Antastabweichungen von KMG mit berührendem
Messkopfsystem (ISO 10360-5:2010)

This European Standard was approved by CEN on 12 June 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

The text of ISO 10360-5:2010 has been prepared by Technical Committee ISO/TC 213 “Dimensional and geometrical product specifications and verification” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 10360-5:2010 by Technical Committee CEN/TC 290 “Dimensional and geometrical product specification and verification” the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2011, and conflicting national standards shall be withdrawn at the latest by March 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10360-5:2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 10360-5:2010 has been approved by CEN as a EN ISO 10360-5:2010 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	2
3 Terms and definitions	2
4 Symbols	6
5 Requirements for metrological characteristics	7
5.1 Single-stylus probing error	7
5.2 Single-stylus probing configuration	8
5.3 Multi-stylus probing errors and values	8
5.4 Multi-stylus probing configurations	9
5.5 Styli	9
5.6 Environmental conditions	9
5.7 Operating conditions	9
6 Acceptance tests and reverification tests	10
6.1 General	10
6.2 Single-stylus probing configuration	10
6.2.1 Application	10
6.2.2 Principle	10
6.2.3 Measuring equipment	10
6.2.4 Procedure	11
6.2.5 Derivation of test results	12
6.3 Fixed multi-probe and multi-stylus probing systems	12
6.3.1 Principle	12
6.3.2 Measuring equipment	13
6.3.3 Procedure	14
6.3.4 Data analysis	15
6.4 Articulating probing systems	16
6.4.1 Principle	16
6.4.2 Measuring equipment	17
6.4.3 Procedure	17
6.4.4 Data analysis	19
7 Compliance with specification	19
7.1 Acceptance tests	19
7.2 Reverification tests	20
8 Applications	20
8.1 Acceptance tests	20
8.2 Reverification tests	20
8.3 Interim checks	20
9 Indication in product documentation and data sheets	21
Annex A (informative) Symbols and subscripts	23
Annex B (informative) Checking the probing system prior to the ISO 10360-2 test	24
Annex C (informative) Interpretation of multi-stylus test results	25
Annex D (normative) Maximum permissible error/limit figures	27

Annex E (informative) Relation to the GPS matrix model	28
Bibliography	30

This document is a preview generated by EVS

Introduction

This part of ISO 10360 is a geometrical product specification (GPS) standard and is to be regarded as a general GPS standard (see ISO/TR 14638). It influences chain link 5 of the chains of standards of size, distance, radius, angle, form, orientation, location, run-out and datums.

For more detailed information on the relation of this part of ISO 10360 to other standards and the GPS matrix model, see Annex E.

The acceptance and reverification tests described in this part of ISO 10360 are applicable to coordinate measuring machines (CMMs) that use contacting probes, with or without multiple styli or multiple articulated-probe positions, when measuring a workpiece.

Experience has shown that the multi-stylus errors calculated using this part of ISO 10360 are significant and, at times, the dominant errors in the CMM. Owing to the virtually infinite variety of modern CMM probing system configurations, the tests specified by this part of ISO 10360 have been limited to providing a testing format only. The tests are intended to provide information on the ability of a CMM to measure a feature or features, using a contacting probe and, when relevant, using multiple styli, multiple probes or multiple articulated-probe positions.

The situations to which they are applicable include

- single-stylus probing systems,
- multiple styli connected to the CMM probe (e.g. a star),
- installations using an articulating probing system (motorized or manual) that can be prequalified,
- installations using a repeatable probe-changing system,
- installations using a repeatable stylus-changing system, and
- multi-probe installations.

It is believed that the procedures given in this part of ISO 10360 will be helpful in identifying CMM system uncertainty components for specific measurement tasks, and that the user will be able to reduce errors by removing contributing elements such as long probe extensions and styli, then retesting the new configuration set.

The tests in this part of ISO 10360 are sensitive to many errors attributable to both the CMM and the probing system, and are to be performed in addition to the length-measuring tests given in ISO 10360-2.

The primary objective is to determine the practical performance of the complete CMM and probing system. Therefore, the tests are designed to reveal measuring errors which are likely to occur when such a combined system is used on real workpieces, e.g. errors generated by the interaction between large probe-tip-offset lengths and uncorrected CMM rotation errors. The errors found here differ from those found in the E_L tests in ISO 10360-2:2009, 6.5, because with multiple styli the net CMM travel may be very different from the measured length. See Annex C for more information.

Geometrical product specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) —

Part 5: CMMs using single and multiple stylus contacting probing systems

1 Scope

This part of ISO 10360 specifies acceptance and periodic reverification tests of CMM performance with contacting probing systems and is only applicable to CMMs using

- any type of contacting probing system,
- a discrete point probing mode, and
- spherical or hemispherical stylus tip(s).

It complements ISO 10360-7, which is the module for CMMs with video probing systems, and ISO 10360-2, which is universal, i.e. not probe-type specific.

NOTE It is the CMM probing performance tests which are specified by the maximum permissible errors (MPEs), due to the impracticality of isolating the performance of the probing system from that of the CMM, even on a small artefact such as a test sphere.

This part of ISO 10360 applies to CMMs supplied with any of the following:

- a) single-stylus probing system;
- b) multi-stylus probing systems with fixed multiple styli attached to a single probe (e.g. “star” stylus);
- c) multiple probing systems such as those with a stylus for each of their probes;
- d) systems with articulating probing;
- e) stylus and probe changing systems;
- f) manual (non-driven) CMMs.

This part of ISO 10360 is not applicable to non-contacting probing systems, which require different testing procedures.

The terms “multi-stylus size error”, etc., should strictly be written “combined CMM and multi-stylus probing-system size error”, etc. For convenience, the wording has been truncated.

If it is desired to isolate the probing-system performance as far as is practical, the influence of the CMM can be minimized. See Annex C for more information.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10360-1:2000, *Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 1: Vocabulary*

ISO 10360-2:2009, *Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 2: CMMs used for measuring linear dimensions*

ISO 14253-1, *Geometrical Product Specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformance or non-conformance with specifications*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10360-1, ISO 14253-1, ISO/IEC Guide 99 and the following apply.

NOTE This clause contains eight definitions (3.6 to 3.13) which supersede fourteen similar definitions in Clause 9 of ISO 10360-1:2000. Some of these revised definitions are required to avoid ambiguities which would otherwise have been introduced with this edition of ISO 10360-5. Others effectively supersede identical definitions in ISO 10360-1, because the symbols used have been revised and expanded for clarification. The superseded definitions are 9.3, 9.4 and 9.15 to 9.26.

3.1 inferred probing-system qualification

probing-system qualification method where the parameters for each probing system attached to an articulation system are inferred by interpolation, extrapolation, or other relevant model, for significantly different angular position(s) from parameters acquired by **empirical probing-system qualification** (3.3) at a few angular positions

3.2 angular positioning device qualification

establishment of the parameters of the angular positioning device in an articulating probing system necessary for subsequent **inferred probing-system qualification** (3.1)

3.3 empirical probing-system qualification

probing-system qualification method where the parameters for each probing system attached to an articulation system must be acquired by measurement of the reference sphere at each angular position used

3.4 effective stylus tip diameter

diameter used for the tip correction vector, for compensating measured feature size, etc.

NOTE 1 For the position of the tip correction vector, see ISO 10360-1:2000, Figure 4.

NOTE 2 The effective stylus tip diameter may be a parameter established by a probing-system qualification.

3.5 probing-system pre-qualification

probing-system qualification which is separated from subsequent measurement by probe or stylus change(s), and/or articulating probing-system re-orientation(s)