

EVALVEERIMINE JA KORRALINE KATSETAMINE  
MEDITSIINIPIILTDIAGNOSTIKA OSAKONDADES. OSA  
3-8: HEAKSKIIDU- JA PÜSIVUSKATSED. RADIOGRAAFIA  
JA FLUOROSKOOPIA RÖNTGENSEADMETE  
PILDISTUSNÄITAJAD

Evaluation and routine testing in medical imaging  
departments - Part 3-8: Acceptance and constancy  
tests - Imaging performance of X-ray equipment for  
radiography and radioscopy

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>See Eesti standard EVS-EN IEC 61223-3-8:2024 sisaldab Euroopa standardi EN IEC 61223-3-8:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 17.05.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN IEC 61223-3-8:2024 consists of the English text of the European standard EN IEC 61223-3-8:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 17.05.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 11.040.50

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

ICS 11.040.50

English Version

Evaluation and routine testing in medical imaging departments -  
Part 3-8: Acceptance and constancy tests - Imaging  
performance of X-ray equipment for radiography and radioscopy  
(IEC 61223-3-8:2024)

Essais d'évaluation et de routine dans les services  
d'imagerie médicale - Partie 3-8: Essais d'acceptation et de  
constance - Performance d'imagerie des appareils à  
rayonnement X pour la radiographie et la radioscopie  
(IEC 61223-3-8:2024)

Bewertung und routinemäßige Prüfung in Abteilungen für  
medizinische Bildgebung - Teil 3-8: Abnahme- und  
Konstanzprüfungen - Leistungsmerkmale zur Bildgebung  
von Röntgeneinrichtungen für Radiographie und  
Radioskopie  
(IEC 61223-3-8:2024)

This European Standard was approved by CENELEC on 2024-04-30. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.



European Committee for Electrotechnical Standardization  
Comité Européen de Normalisation Electrotechnique  
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

## European foreword

The text of document 62B/1347/FDIS, future edition 1 of IEC 61223-3-8, prepared by SC 62B "Diagnostic imaging equipment" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN IEC 61223-3-8:2024.

The following dates are fixed:

- latest date by which the document has to be implemented at national (dop) 2025-01-30 level by publication of an identical national standard or by endorsement
- latest date by which the national standards conflicting with the (dow) 2027-04-30 document have to be withdrawn

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national committee. A complete listing of these bodies can be found on the CENELEC website.

## Endorsement notice

The text of the International Standard IEC 61223-3-8:2024 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standard indicated:

IEC 60627:2013	NOTE	Approved as EN 60627:2015 (not modified)
IEC 61223-3-6:2020	NOTE	Approved as EN IEC 61223-3-6:2020 (not modified)
IEC 62220-1-1:2015	NOTE	Approved as EN 62220-1-1:2015 (not modified)
IEC 62563-1:2009	NOTE	Approved as EN 62563-1:2010 (not modified)
IEC 62563-1:2009/A1:2016	NOTE	Approved as EN 62563-1:2010/A1:2016 (not modified)
IEC 62563-1:2009/AMD2:2021	NOTE	Approved as EN 62563-1:2010/A2:2021 (not modified)
IEC 62563-2:2021	NOTE	Approved as EN IEC 62563-2:2021 (not modified)

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Evaluation and routine testing in medical imaging departments –  
Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray  
equipment for radiography and radioscopy**

**Essais d'évaluation et de routine dans les services d'imagerie médicale –  
Partie 3-8: Essais d'acceptation et de constance – Performance d'imagerie des  
appareils à rayonnement X pour la radiographie et la radioscopie**



**THIS PUBLICATION IS COPYRIGHT PROTECTED**  
**Copyright © 2024 IEC, Geneva, Switzerland**

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester. If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de l'IEC ou du Comité national de l'IEC du pays du demandeur. Si vous avez des questions sur le copyright de l'IEC ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de l'IEC de votre pays de résidence.

IEC Secretariat  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

#### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

#### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigendum or an amendment might have been published.

#### IEC publications search - [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

The advanced search enables to find IEC publications by a variety of criteria (reference number, text, technical committee, ...). It also gives information on projects, replaced and withdrawn publications.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available online and once a month by email.

#### IEC Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Discover our powerful search engine and read freely all the publications previews, graphical symbols and the glossary. With a subscription you will always have access to up to date content tailored to your needs.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary on electrotechnology, containing more than 22 500 terminological entries in English and French, with equivalent terms in 25 additional languages. Also known as the International Electrotechnical Vocabulary (IEV) online.

---

#### A propos de l'IEC

La Commission Electrotechnique Internationale (IEC) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

#### A propos des publications IEC

Le contenu technique des publications IEC est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Recherche de publications IEC -

#### [webstore.iec.ch/advsearchform](http://webstore.iec.ch/advsearchform)

La recherche avancée permet de trouver des publications IEC en utilisant différents critères (numéro de référence, texte, comité d'études, ...). Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

#### IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications IEC. Just Published détaille les nouvelles publications parues. Disponible en ligne et une fois par mois par email.

#### Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [sales@iec.ch](mailto:sales@iec.ch).

#### IEC Products & Services Portal - [products.iec.ch](http://products.iec.ch)

Découvrez notre puissant moteur de recherche et consultez gratuitement tous les aperçus des publications, symboles graphiques et le glossaire. Avec un abonnement, vous aurez toujours accès à un contenu à jour adapté à vos besoins.

#### Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire d'électrotechnologie en ligne au monde, avec plus de 22 500 articles terminologiques en anglais et en français, ainsi que les termes équivalents dans 25 langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (IEV) en ligne.

# INTERNATIONAL STANDARD

## NORME INTERNATIONALE



**Evaluation and routine testing in medical imaging departments –  
Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray  
equipment for radiography and radioscopy**

**Essais d'évaluation et de routine dans les services d'imagerie médicale –  
Partie 3-8: Essais d'acceptation et de constance – Performance d'imagerie des  
appareils à rayonnement X pour la radiographie et la radioscopie**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-8565-7

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	8
INTRODUCTION.....	10
1 Scope and object.....	11
2 Normative references .....	11
3 Terms, definitions, symbols and abbreviated terms.....	12
3.1 Terms and definitions.....	12
3.2 Symbols and abbreviated terms .....	21
4 General aspects of ACCEPTANCE TESTING.....	21
4.1 General.....	21
4.2 Levels of requirements.....	22
4.2.1 General .....	22
4.2.2 Local regulatory requirements .....	22
4.2.3 Contractual requirements.....	22
4.2.4 MANUFACTURER specifications .....	22
4.3 General conditions in test procedures .....	22
4.4 Documents and data for the tests.....	22
4.5 Test conditions .....	23
4.6 Scope of tests.....	25
4.7 Test equipment including PHANTOMS and TEST DEVICES.....	26
4.7.1 General .....	26
4.7.2 Analysis software.....	26
4.7.3 Viewing conditions.....	26
4.7.4 High-voltage measuring instrument.....	26
4.7.5 DOSIMETER.....	26
4.7.6 PHANTOMS, TEST DEVICES and their application .....	26
4.8 Evaluating the test results.....	29
5 Specific aspects of CONSTANCY TESTING .....	29
5.1 General.....	29
5.2 Establishment of BASELINE VALUES.....	29
5.3 CONSTANCY TESTING and SIMPLE CHECKS .....	30
5.4 Frequency of CONSTANCY TESTS.....	30
6 Summary of tests for X-RAY EQUIPMENT .....	30
7 Requirements and test methods .....	32
7.1 General.....	32
7.2 Functional properties .....	33
7.2.1 Test requirements.....	33
7.2.2 Test method for ACCEPTANCE TEST.....	33
7.2.3 Test method for CONSTANCY TEST .....	33
7.3 IMAGE DISPLAY DEVICE performance.....	33
7.3.1 Test requirements.....	33
7.3.2 Test method for ACCEPTANCE TEST.....	33
7.3.3 Test method for CONSTANCY TEST .....	34
7.4 X-RAY TUBE VOLTAGE.....	34
7.4.1 Test requirements.....	34
7.4.2 Test method for ACCEPTANCE TEST.....	34
7.4.3 Test method for CONSTANCY TEST .....	35

7.5	X-RAY TUBE CURRENT or CURRENT TIME PRODUCT .....	35
7.5.1	Test requirements.....	35
7.5.2	Test method for ACCEPTANCE TEST.....	35
7.5.3	Test method for CONSTANCY TEST .....	35
7.6	HALF-VALUE LAYER of the X-RAY SOURCE ASSEMBLY .....	35
7.6.1	Test requirements.....	35
7.6.2	Test method for ACCEPTANCE TEST.....	36
7.6.3	Test method for CONSTANCY TEST .....	36
7.7	* QUALITY EQUIVALENT FILTRATION of ADDED FILTERS .....	36
7.7.1	Test requirements.....	36
7.7.2	* Test method for ACCEPTANCE TEST .....	36
7.7.3	Test method for CONSTANCY TEST .....	36
7.8	* ATTENUATION EQUIVALENT of materials between the PATIENT and the X-RAY IMAGE RECEPTOR .....	36
7.8.1	Test requirements.....	36
7.8.2	* Test method for ACCEPTANCE TEST .....	37
7.8.3	Test method for CONSTANCY TEST .....	37
7.9	FOCAL SPOT of the X-RAY TUBE .....	37
7.9.1	Test requirements.....	37
7.9.2	Test method for ACCEPTANCE TEST.....	37
7.9.3	Test method for CONSTANCY TEST .....	37
7.10	Assessing geometrical properties of the X-RAY EQUIPMENT .....	37
7.10.1	Test requirements.....	37
7.10.2	Test method for ACCEPTANCE TEST.....	37
7.10.3	Test method for CONSTANCY TEST .....	37
7.11	AUTOMATIC CONTROL SYSTEM – Performance.....	38
7.11.1	Test requirements.....	38
7.11.2	* Test method for ACCEPTANCE TEST .....	38
7.11.3	* Test method for CONSTANCY TEST.....	38
7.12	AUTOMATIC CONTROL SYSTEM – Reproducibility .....	38
7.12.1	Test requirements.....	38
7.12.2	Test method for ACCEPTANCE TEST.....	38
7.12.3	Test method for CONSTANCY TEST .....	38
7.13	* AUTOMATIC CONTROL SYSTEM – NOMINAL SHORTEST IRRADIATION TIME.....	38
7.13.1	Test requirements.....	38
7.13.2	* Test method for ACCEPTANCE TEST .....	38
7.13.3	Test method for CONSTANCY TEST .....	39
7.14	Correspondence between the X-RAY FIELD and EFFECTIVE IMAGE RECEPTION AREA .....	39
7.14.1	Test requirements.....	39
7.14.2	Test method for ACCEPTANCE TEST.....	40
7.14.3	Test method for CONSTANCY TEST .....	41
7.15	Accuracy of indication of the LIGHT FIELD-INDICATOR.....	41
7.15.1	Test requirements.....	41
7.15.2	Test method for ACCEPTANCE TEST.....	42
7.15.3	Test method for CONSTANCY TEST .....	42
7.16	Accuracy of marked and written indications of the X-RAY FIELD size .....	43
7.16.1	Test requirements.....	43
7.16.2	Test method for ACCEPTANCE TEST.....	43

7.16.3	Test method for CONSTANCY TEST .....	43
7.17	AIR KERMA RATE at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR for RADIOSCOPY.....	43
7.17.1	Test requirements.....	43
7.17.2	Test method for ACCEPTANCE TEST.....	44
7.17.3	Test method for CONSTANCY TEST .....	44
7.18	REFERENCE AIR KERMA RATE for RADIOSCOPY .....	44
7.18.1	Test requirements.....	44
7.18.2	Test method for ACCEPTANCE TEST.....	44
7.18.3	Test method for CONSTANCY TEST .....	45
7.19	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR for RADIOGRAPHY.....	46
7.19.1	Test requirements.....	46
7.19.2	Test method for ACCEPTANCE TEST.....	46
7.19.3	Test method for CONSTANCY TEST .....	46
7.20	REFERENCE AIR KERMA for SERIAL RADIOGRAPHY .....	46
7.20.1	Test requirements.....	46
7.20.2	Test method for ACCEPTANCE TEST.....	46
7.20.3	Test method for CONSTANCY TEST .....	46
7.21	Limitation of the REFERENCE AIR KERMA RATE in RADIOSCOPY.....	46
7.21.1	Test requirements.....	46
7.21.2	Test method for ACCEPTANCE TEST.....	46
7.21.3	Test method for CONSTANCY TEST .....	47
7.22	SPATIAL RESOLUTION for RADIOSCOPY and RADIOGRAPHY .....	47
7.22.1	Test requirements.....	47
7.22.2	Test method for ACCEPTANCE TEST.....	47
7.22.3	Test method for CONSTANCY TEST .....	49
7.23	LOW CONTRAST RESOLUTION for RADIOSCOPY and RADIOGRAPHY.....	49
7.23.1	Test requirements.....	49
7.23.2	Test method for ACCEPTANCE TEST.....	50
7.23.3	Test method for CONSTANCY TEST .....	50
7.24	DYNAMIC RANGE for RADIOSCOPY and RADIOGRAPHY .....	50
7.24.1	Test requirements.....	50
7.24.2	Test method for ACCEPTANCE TEST.....	51
7.24.3	Test method for CONSTANCY TEST .....	51
7.25	DOSE AREA PRODUCT and REFERENCE AIR KERMA (RATE) indicator.....	51
7.25.1	Test requirements.....	51
7.25.2	Test method for ACCEPTANCE TEST.....	51
7.25.3	Test method for CONSTANCY TEST .....	51
7.26	Back-up timer and safety measures.....	51
7.26.1	Test requirements.....	51
7.26.2	Test method for ACCEPTANCE TEST.....	51
7.26.3	Test method for CONSTANCY TEST .....	52
7.27	ARTEFACTS .....	52
7.27.1	Test requirements.....	52
7.27.2	Test method for ACCEPTANCE TEST.....	52
7.27.3	Test method for CONSTANCY TEST .....	53
7.28	Accuracy of image measuring functions .....	53
7.28.1	Test requirements.....	53

7.28.2	Test method for ACCEPTANCE TEST.....	53
7.28.3	Test method for CONSTANCY TEST .....	53
7.29	DYNAMIC RANGE for SUBTRACTION IMAGING.....	53
7.29.1	Test requirements.....	53
7.29.2	Conditions for testing the SUBTRACTION IMAGING mode.....	53
7.29.3	Test method for ACCEPTANCE TEST.....	53
7.29.4	Test method for CONSTANCY TEST .....	54
7.30	Contrast sensitivity for SUBTRACTION IMAGING.....	54
7.30.1	Test requirements.....	54
7.30.2	Test method for ACCEPTANCE TEST.....	54
7.30.3	Test method for CONSTANCY TEST .....	54
7.31	SPATIAL RESOLUTION for SUBTRACTION IMAGING.....	54
7.31.1	Test requirements.....	54
7.31.2	Test method for ACCEPTANCE TEST.....	55
7.31.3	Test method for CONSTANCY TEST .....	55
7.32	ARTEFACTS in SUBTRACTION IMAGING.....	55
7.32.1	Test requirements.....	55
7.32.2	Test method for ACCEPTANCE TEST.....	55
7.32.3	Test method for CONSTANCY TEST .....	55
7.33	X-RAY IMAGE RECEPTOR EXPOSURE INDEX .....	55
7.33.1	Test requirements.....	55
7.33.2	Test method for ACCEPTANCE TEST.....	56
7.33.3	Test method for CONSTANCY TEST .....	56
7.34	LEAKAGE RADIATION in the LOADING STATE .....	57
7.34.1	Test requirements.....	57
7.34.2	Test method for ACCEPTANCE TEST.....	57
7.34.3	Test method for CONSTANCY TEST .....	57
Annex A	(informative) General guidance and rationale.....	58
A.1	Aim and background information.....	58
A.2	Medical radiological equipment lifecycle testing .....	58
A.2.1	General .....	58
A.2.2	The "lifecycle".....	58
A.3	Rationale for particular clauses and subclauses.....	61
Annex B	(informative) Use of REFERENCE AIR KERMA and AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR .....	64
B.1	General.....	64
B.2	Issues with AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR as a measure of performance.....	64
B.3	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR as an ACCEPTANCE TEST or CONSTANCY TEST .....	64
B.4	REFERENCE AIR KERMA as a measure of performance .....	65
B.5	REFERENCE AIR KERMA as an ACCEPTANCE TEST .....	65
Annex C	(informative) Assessing clinical protocol performance .....	66
C.1	General.....	66
C.2	Discussion .....	66
Annex D	(informative) Examples of TEST DEVICES for image quality evaluation .....	67
D.1	General.....	67
D.2	TEST DEVICE according to DIN 6868-150.....	67
D.3	Leeds X-ray test objects for RADIOGRAPHY/RADIOSCOPY systems .....	67

D.4	LOW CONTRAST RESOLUTION TEST DEVICE of the University of Alabama at Birmingham, USA.....	67
Annex E	(informative) Examples of DSA PHANTOMS .....	69
E.1	General.....	69
E.2	Generic DSA PHANTOM.....	69
E.3	DSA PHANTOM with test step for compensation .....	71
Annex F	(informative) QUALITY CONTROL in cone beam computed tomography (CBCT) .....	73
F.1	General.....	73
F.2	Image quality parameters.....	73
F.2.1	Summary of tests for CBCT .....	73
F.2.2	Uniformity .....	73
F.2.3	Geometrical precision .....	74
F.2.4	Voxel density values .....	75
F.2.5	Noise .....	75
F.2.6	LOW CONTRAST RESOLUTION .....	75
F.2.7	SPATIAL RESOLUTION .....	76
F.3	Tests of RADIATION output.....	76
F.3.1	General .....	76
F.3.2	Cumulative DOSE AREA PRODUCT.....	77
F.3.3	AIR KERMA at the ENTRANCE PLANE of the X-RAY IMAGE RECEPTOR .....	78
F.3.4	In-phantom dosimetry indexes – CTDI metrics .....	78
F.3.5	Cumulative AIR KERMA .....	79
F.4	Image quality PHANTOMS.....	79
Annex G	(informative) Guidance for MANUFACTURER supplied QUALITY CONTROL manual.....	80
G.1	General.....	80
G.2	Generic considerations for a QUALITY CONTROL manual.....	80
G.3	Prerequisites for guidance for MANUFACTURER supplied QUALITY CONTROL manual.....	80
G.3.1	General .....	80
G.3.2	General aspects .....	81
G.3.3	Objective of test .....	81
G.3.4	Test methods, test equipment and test objects .....	81
G.3.5	Frequency of test.....	81
G.3.6	Acceptance and constancy criteria.....	81
G.4	Example of a QUALITY CONTROL manual for a DOSE AREA PRODUCT verification.....	82
Annex H	(informative) Uncertainty aspects of testing on X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY .....	84
H.1	General.....	84
H.2	Measurements performed in a clinical facility .....	85
H.3	Measurements made by placing a probe in the X-RAY BEAM.....	85
H.3.1	General .....	85
H.3.2	Direct setting of the X-ray equipment.....	86
H.3.3	Indirect setting of the X-ray equipment .....	86
Annex I	(informative) Examples of SIMPLE CHECKS .....	87
I.1	General.....	87
I.2	Examples of visual inspection .....	87
I.3	Examples of indicated values of parameters .....	87
Annex J	(informative) Examples of system changes and relevant ACCEPTANCE TESTS .....	89
Bibliography	.....	90

Index of defined terms .....	93
Figure 1 – Indicative measuring arrangement for RADIOGRAPHY and RADIOSCOPY equipment for AIR KERMA measurements.....	24
Figure 2 – Indicative measuring arrangement for RADIOGRAPHY and RADIOSCOPY equipment to test geometry and resolutions .....	25
Figure 3 – Example of a LINE PAIR RESOLUTION TEST DEVICE .....	28
Figure 4 – MODULATION TRANSFER FUNCTION TEST DEVICE .....	28
Figure 5 – Discrepancies in covering the IMAGE RECEPTION AREA .....	41
Figure 6 – Discrepancies in visual indication of the X-RAY FIELD .....	42
Figure A.1 – Lifecycle of X-ray imaging equipment.....	59
Figure E.1 – Generic DSA PHANTOM .....	70
Figure E.2 – DSA PHANTOM with test step for compensation .....	72
Figure F.1 – Position of the ROIs for the calculation of the uniformity .....	74
Figure F.2 – CTDI PHANTOM with the positions of the DOSIMETER .....	78
Figure G.1 – Measurement set-up for DAP measurement.....	82
Figure H.1 – Decision points with respect to MEASUREMENT UNCERTAINTY.....	85
Table 1 – QUALITY CONTROL tests and recommended frequencies.....	31
Table E.1 – Summary of properties of the generic DSA PHANTOM. ....	69
Table E.2 – Summary of properties of the DSA PHANTOM with test step for compensation.....	71
Table F.1 – Summary of image quality tests for CBCT including suggested frequencies .....	73
Table F.2 – Summary of RADIATION output tests for CBCT including suggested frequency.....	77
Table J.1 – Overview of relevant ACCEPTANCE TESTS related to system changes.....	89

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

---

**EVALUATION AND ROUTINE TESTING  
IN MEDICAL IMAGING DEPARTMENTS –**
**Part 3-8: Acceptance and constancy tests –  
Imaging performance of X-ray equipment for radiography and radioscopy**
**FOREWORD**

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) IEC draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). IEC takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, IEC had not received notice of (a) patent(s), which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at <https://patents.iec.ch>. IEC shall not be held responsible for identifying any or all such patent rights.

IEC 61223-3-8 has been prepared by subcommittee 62B: Medical imaging equipment, software, and systems, of IEC technical committee 62: Medical equipment, software, and systems. It is an International Standard.

The text of this document is based on the following documents:

Draft	Report on voting
62B/1347/FDIS	62B/1351/RVD

Full information on the voting for its approval can be found in the report on voting indicated in the above table.

The language used for the development of this International Standard is English.

This document was drafted in accordance with ISO/IEC Directives, Part 2, and developed in accordance with ISO/IEC Directives, Part 1 and ISO/IEC Directives, IEC Supplement, available at [www.iec.ch/members\\_experts/refdocs](http://www.iec.ch/members_experts/refdocs). The main document types developed by IEC are described in greater detail at [www.iec.ch/publications](http://www.iec.ch/publications).

In this document, the following print types are used:

- requirements, compliance with which can be tested, and definitions: in roman type.
- explanations, advice, notes, general statements, exceptions and references: in smaller type.
- *test specifications: in italic type.*
- TERMS USED THROUGHOUT THIS DOCUMENT THAT HAVE BEEN LISTED IN THE INDEX OF DEFINED TERMS AND DEFINED IN CLAUSE 3, OR IN OTHER STANDARDS: IN SMALL CAPITALS.

References to clauses within this document are preceded by the term "Clause" followed by the clause number. References to subclauses within this document are by number only.

In this document, the conjunctive "or" is used as an "inclusive or" so a statement is true if any combination of the conditions is true.

The verbal forms used in this document conform to usage described in Clause 7 of the ISO/IEC Directives, Part 2. For the purposes of this document, the auxiliary verb:

- "shall" means that compliance with a requirement or a test is mandatory for compliance with this document;
- "should" means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this document;
- "may" is used to describe a permissible way to achieve compliance with a requirement or test.

An asterisk (\*) as the first character of a title or at the beginning of a paragraph or table title indicates that there is guidance or rationale related to that item in Annex A.

A list of all parts of the IEC 61223 series, published under the general title *Evaluation and routine testing in medical imaging departments*, can be found on the IEC website.

The committee has decided that the contents of this document will remain unchanged until the stability date indicated on the IEC website under [webstore.iec.ch](http://webstore.iec.ch) in the data related to the specific document. At this date, the document will be

- reconfirmed,
- withdrawn, or
- revised.

NOTE The attention of the users of this document is drawn to the fact that equipment MANUFACTURERS and testing organizations might need a transitional period following publication of a new, amended or revised IEC publication in order to develop contractual specifications in accordance with the new test procedures and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The "colour inside" logo on the cover page of this document indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this document using a colour printer.**

## INTRODUCTION

The IEC 61223 series gives methods for ACCEPTANCE TESTS and CONSTANCY TESTS for diagnostic X-RAY EQUIPMENT.

IEC 60601-2-54:2022 and IEC 60601-2-43:2022 state that the MANUFACTURER provides ACCOMPANYING DOCUMENTS with instructions for MANUFACTURER-recommended QUALITY CONTROL PROCEDURES and tests to be performed on the X-RAY EQUIPMENT by the RESPONSIBLE ORGANIZATION. In this document, these instructions are referred to as a QUALITY CONTROL manual.

This part of IEC 61223 provides guidance on the content to be considered for inclusion in the QUALITY CONTROL manual for X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY, including INTERVENTIONAL X-RAY EQUIPMENT. The QUALITY CONTROL manual is intended to be used independently by the RESPONSIBLE ORGANIZATION to ensure safe and effective equipment performance during its EXPECTED SERVICE LIFE, without the need to consult IEC standards.

This document provides parameters to be evaluated, examples of test methods, recommended minimum frequencies of evaluation and tools intended to be used by the RESPONSIBLE ORGANIZATION on installed X-RAY EQUIPMENT within the scope of this document. For the parameters applicable to the specific X-RAY EQUIPMENT to be tested, specific acceptance criteria are supplied by MANUFACTURERS, which can be superseded with contractual criteria with the RESPONSIBLE ORGANIZATION and further superseded with locally applicable regulatory criteria.

A major purpose of this document is that of facilitating technical communication between stakeholders in the areas of ACCEPTANCE and CONSTANCY TESTING. The three major stakeholders responsible for assuring the safety and efficacy of X-RAY EQUIPMENT are the MANUFACTURER, the RESPONSIBLE ORGANIZATION and the regulatory authorities.

Generally, equipment installed in accordance with the MANUFACTURER'S QUALITY CONTROL PROCESS will comply with applicable IEC standards as well as meeting both local regulatory requirements and the MANUFACTURER'S general specifications and contractual specifications by the RESPONSIBLE ORGANIZATION.

The performance of installed equipment that is tested is assessed by the RESPONSIBLE ORGANIZATION and regulators using a wide variety of PROCESSES and tools. Acceptability criteria and PROCESSES can differ from those of equipment MANUFACTURERS.

## EVALUATION AND ROUTINE TESTING IN MEDICAL IMAGING DEPARTMENTS –

### Part 3-8: Acceptance and constancy tests – Imaging performance of X-ray equipment for radiography and radioscopy

#### 1 Scope and object

This part of IEC 61223 applies to evaluation of the imaging performance and related QUALITY CONTROL parameters of X-RAY EQUIPMENT for RADIOGRAPHY and RADIOSCOPY that conform to IEC 60601-2-54:2022 or IEC 60601-2-43:2022.

NOTE Cone-beam CT is a MODE OF OPERATION in INTERVENTIONAL X-RAY EQUIPMENT. This document discusses such MODE OF OPERATION in the informative Annex F.

This document applies to the evaluation of the imaging performance of the entire imaging chain from image acquisition, image processing and image display.

This document applies to ACCEPTANCE TESTS and CONSTANCY TESTS, which are part of the QUALITY ASSURANCE PROGRAM in medical imaging departments and are intended to be performed by or under the responsibility of the RESPONSIBLE ORGANIZATION. A detailed discussion of the position of these tests within the medical radiological equipment lifecycle is provided in Clause A.2. The methods included rely mainly on non-invasive measurements that use appropriate test equipment and are performed after the installation is completed in accordance with the MANUFACTURER'S installation instructions.

IEC 60601-2-54:2022 and IEC 60601-2-43:2022 require information to be provided to the RESPONSIBLE ORGANIZATION with respect to QUALITY CONTROL. This document provides guidance to MANUFACTURERS regarding the ACCEPTANCE and CONSTANCY TESTS for the X-RAY EQUIPMENT in a MANUFACTURER supplied QUALITY CONTROL manual. Annex G provides guidance for such a manual.

#### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60336:2020, *Medical electrical equipment – X-ray tube assemblies for medical diagnosis – Focal spot dimensions and related characteristics*

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*  
IEC 60601-1:2005/AMD1:2012  
IEC 60601-1:2005/AMD2:2020

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral standard: Radiation protection in diagnostic X-ray equipment*  
IEC 60601-1-3:2008/AMD1:2013  
IEC 60601-1-3:2008/AMD2:2021

IEC 60601-2-43:2022, *Medical electrical equipment – Part 2-43: Particular requirements for the basic safety and essential performance of X-ray equipment for interventional procedures*

IEC 60601-2-54:2022, *Medical electrical equipment – Part 2-54: Particular requirements for the basic safety and essential performance of X-ray equipment for radiography and radioscopy*

IEC 61267:2005, *Medical diagnostic X-ray equipment – Radiation conditions for use in the determination of characteristics*

IEC 61674:2012, *Medical electrical equipment – Dosimeters with ionization chambers and/or semi-conductor detectors as used in X-ray diagnostic imaging*

IEC 61676:2023, *Medical electrical equipment – Dosimetric instruments used for non-invasive measurement of X-ray tube voltage in diagnostic radiology*

IEC 62494-1:2008, *Medical electrical equipment – Exposure index of digital X-ray imaging systems – Part 1: Definitions and requirements for general radiography*

### **3 Terms, definitions, symbols and abbreviated terms**

#### **3.1 Terms and definitions**

For the purposes of this document, the terms and definitions in IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020, IEC 60601-1-3:2008, IEC 60601-1-3:2008/AMD1:2013, IEC 60601-1-3:2008/AMD2:2021, IEC 60601-2-54:2022, IEC 60601-2-43:2022 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- IEC Electropedia: available at <https://www.electropedia.org/>
- IEC Glossary; available at <http://std.iec.ch/glossary>
- ISO Online browsing platform: available at <https://www.iso.org/obp>

##### **3.1.1**

###### **ACCEPTANCE TEST**

###### **ACCEPTANCE TESTING**

test carried out after new equipment has been installed, or major modifications have been made to existing equipment, in order to verify compliance with specifications

Note 1 to entry: The specifications could include contractual specifications, requirements enforced by legislation, the MANUFACTURER's specifications or requirements from standards, for example in the IEC 60601 series.

[SOURCE: IEC 61223-1:1993, 3.2.4, modified – Words "contractual specifications" replaced with "specifications" and Note 1 to entry added.]

##### **3.1.2**

###### **ACTUAL FOCAL SPOT**

area on the surface of the anode that intercepts the beam of accelerated particles

[SOURCE: IEC 60336:2020, 3.1, modified – Word "target" replaced with "anode" and Note 1 to entry removed.]