

Ophthalmic implants - Intraocular lenses - Part 2:
Optical properties and test methods (ISO
11979-2:2024)

This document is a preview generated by EVS

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>See Eesti standard EVS-EN ISO 11979-2:2024 sisaldab Euroopa standardi EN ISO 11979-2:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 13.11.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 11979-2:2024 consists of the English text of the European standard EN ISO 11979-2:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 13.11.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.70

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 11979-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2024

ICS 11.040.70

Supersedes EN ISO 11979-2:2014

English Version

Ophthalmic implants - Intraocular lenses - Part 2: Optical properties and test methods (ISO 11979-2:2024)

Implants ophtalmiques - Lentilles intraoculaires -
Partie 2: Propriétés optiques et méthodes d'essai (ISO
11979-2:2024)

Ophthalmische Implantate - Intraokularlinsen - Teil 2:
Optische Eigenschaften und Prüfverfahren (ISO 11979-
2:2024)

This European Standard was approved by CEN on 1 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 11979-2:2024) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2025, and conflicting national standards shall be withdrawn at the latest by May 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11979-2:2014.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 11979-2:2024 has been approved by CEN as EN ISO 11979-2:2024 without any modification.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	1
4.1 General.....	1
4.2 Dioptric power.....	2
4.2.1 General.....	2
4.2.2 Dioptric power for toric IOL (TIOL).....	2
4.2.3 Dioptric power for simultaneous vision IOL (SVIOL).....	2
4.2.4 Dioptric power for accommodating IOL (AIOL).....	3
4.3 Imaging quality.....	3
4.3.1 General.....	3
4.3.2 Monofocal IOL.....	4
4.3.3 Toric IOL (TIOL).....	4
4.3.4 Simultaneous vision IOL (SVIOL).....	4
4.3.5 Accommodating IOL (AIOL).....	4
4.3.6 Combination of optical principles.....	4
4.3.7 Exceptions.....	4
4.4 Optical characterization.....	5
4.5 Spectral transmittance.....	5
4.5.1 Measurement of spectral transmittance.....	5
4.5.2 Cut-off wavelength.....	5
Annex A (normative) Measurement of dioptric power	6
Annex B (normative) Measurement of MTF	14
Annex C (normative) Optical characterization	18
Bibliography	21

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This third edition cancels and replaces the second edition (ISO 11979-2:2014), which has been technically revised.

The main changes are as follows:

- A new category of simultaneous vision IOLs (SVIOL) is introduced for non-accommodating lenses that provide simultaneous vision at multiple distances. It includes multifocal IOLs (MIOL), extended depth of focus IOLs (EDF), and full visual range IOLs (FVR).
- Dioptric power, imaging quality, and characterization clauses and annexes were modified to include requirements for SVIOLs.
- Respective units of mm^{-1} and degree^{-1} were adopted for linear and angular spatial frequencies per ISO 9334.
- The resolution efficiency and associated annex have been removed from this document due to advancements in optical designs and the availability of modulation transfer function (MTF) imaging quality measurement methods.
- A new [Annex C](#) with associated requirements for all IOL categories has been added.
- Clarified description of UV cut-off wavelength.
- New references were added to the Bibliography.

A list of all parts in the ISO 11979 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

This document initially addressed monofocal IOLs and now includes the optical requirements and test methods for monofocal, toric, simultaneous vision, and accommodating IOLs. This document generally provides specific test methods and requirements connected to the optical function of intraocular lenses. In some cases, test methods do not have specified requirements, including:

- the spectral transmittance test that provides information related to UV transmission and potential exposure situations, e.g. when using laser light sources for diagnosis and treatment;
- optical characterization testing that informs potential optical design risks and guides potential clinical investigation design.

The specified dioptric power and imaging quality limits result from the analysis of extensive interlaboratory testing of the original spherical monofocal IOLs. Based on these studies, the respective dioptric power repeatability and reproducibility were about 0,5 % and 1 %, respectively, of the dioptric power as described in Reference [1]. Additionally, for IOLs in the 10 D to 30 D range, the respective expected imaging quality repeatability and reproducibility were 0,09 and 0,16 modulation transfer function values as described in Reference [2]. For other non-monofocal IOL designs, manufacturers should utilize model-specific repeatability and reproducibility precision limits to establish reliable final release criteria.

During the interlaboratory testing, some problems were encountered with measuring dioptric power, as described in Reference [1]. Specifically, the accuracy in determining dioptric power has an error that is not negligible in relation to the half dioptre steps in which intraocular lenses are commonly labelled. The dioptric power tolerances take this fact into account. Hence the limits set may lead to some overlap into the next labelled power, especially for high dioptre lenses. Reference [1] further discusses this subject.

Historically, imaging quality was tested using either

- a) Air Force target-based resolution efficiency, or
- b) MTF using a minimal spherical aberration model eye, or
- c) a manufacturer-defined spherical aberration model eye using modulation transfer function (MTF) testing.

Since the test method with Air Force target-based resolution efficiency is not optimal for quantifying image contrast, and better methods using MTF measurements have become mainstream in the industry, Air Force target-based resolution efficiency is not included in this revision as a reference method. The model eye with manufacturer-defined spherical aberration includes the option of having a model eye with minimal spherical aberration. Therefore, the original model eye with minimal spherical aberration is removed from this document. For lenses that have already been approved using the measurements in the previous edition, it is not necessary to retest these lens models with the method in this document.

[Annex B](#) describes a test method used to establish quality criteria for IOLs. The quality criteria assure consistent IOL optical quality. This document also includes a new normative optical characterization text (see [Annex C](#)), that is meant to provide preclinical assessments to inform of risks and benefits associated with the optical design and guide the design of the potential clinical investigation. The additional optical characterization is required only for lens models to be approved after publication of this document.

Ophthalmic implants — Intraocular lenses —

Part 2: Optical properties and test methods

1 Scope

This document specifies requirements and test methods for certain optical properties of intraocular lenses (IOLs) with monofocal, toric, simultaneous vision, and/or accommodative optics. The generic descriptor 'IOL' used throughout this document also includes phakic intraocular lenses (PIOL).

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes the requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 9334, *Optics and photonics — Optical transfer function — Definitions and mathematical relationships*

ISO 9335, *Optics and photonics — Optical transfer function — Principles and procedures of measurement*

ISO 11979-1, *Ophthalmic implants — Intraocular lenses — Part 1: Vocabulary*

ISO 11979-4, *Ophthalmic implants — Intraocular lenses — Part 4: Labelling and information*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 11979-1 and ISO 9334 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

4 Requirements

4.1 General

The manufacturer shall assure that the entire range of available powers meets the specifications herein. All optical properties apply at in situ conditions, either by being measured at simulated in situ conditions, or being measured at other conditions and then corrected to in situ conditions.

For IOLs where the optic is intended to be deformed during implantation, it shall be demonstrated that dioptric power and imaging quality are retained at in situ or equivalent conditions following surgical manipulation and recovery. See ISO 11979-3^[3] for more details.

The test methods described in this document are reference methods. Alternative methods that produce equivalent results to those obtained with the reference methods may be used if the manufacturer can demonstrate that the IOLs meet the minimum dioptric power and imaging quality requirements.