

TAIMEDE BIIOSTIMULAATORID. SPETSIIFILISTE  
AINETE MÄÄRAMINE. OSA 2: CD, PB, NI, AS, CR, CU JA ZN  
ÜLDSISALDUSE MÄÄRAMINE

Plant biostimulants - Determination of specific  
elements - Part 2: Determination of total content of Cd,  
Pb, Ni, As, Cr, Cu and Zn

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>See Eesti standard EVS-EN 17701-2:2024 sisaldab Euroopa standardi EN 17701-2:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 27.11.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN 17701-2:2024 consists of the English text of the European standard EN 17701-2:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 27.11.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 65.080

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

EUROPEAN STANDARD

EN 17701-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2024

ICS 65.080

Supersedes CEN/TS 17701-2:2022

English Version

Plant biostimulants - Determination of specific elements -  
Part 2: Determination of total content of Cd, Pb, Ni, As, Cr,  
Cu and Zn

Biostimulants des végétaux - Dosage des éléments  
spécifiques - Partie 2 : Détermination de la teneur  
totale en Cd, Pb, Ni, As, Cr, Cu et Zn

Pflanzen-Biostimulanzien - Bestimmung spezifischer  
Elemente - Teil 2: Bestimmung des Gesamtgehaltes an  
Cd, Pb, Ni, As, Cr, Cu und Zn

This European Standard was approved by CEN on 26 August 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
<b>European foreword</b> .....	<b>3</b>
<b>Introduction</b> .....	<b>4</b>
<b>1 Scope</b> .....	<b>5</b>
<b>2 Normative references</b> .....	<b>5</b>
<b>3 Terms and definitions</b> .....	<b>5</b>
<b>4 Principle</b> .....	<b>6</b>
<b>5 Interferences</b> .....	<b>6</b>
<b>5.1 General</b> .....	<b>6</b>
<b>5.2 Spectral interferences</b> .....	<b>6</b>
<b>5.3 Transport interferences</b> .....	<b>7</b>
<b>5.4 Excitation interferences</b> .....	<b>7</b>
<b>5.5 Chemical interferences</b> .....	<b>7</b>
<b>5.6 Memory interferences</b> .....	<b>7</b>
<b>6 Reagents</b> .....	<b>7</b>
<b>7 Apparatus</b> .....	<b>8</b>
<b>8 Procedure</b> .....	<b>8</b>
<b>8.1 Preparation of test sample solution and blank test solution</b> .....	<b>8</b>
<b>8.2 Preparation of the calibration solutions</b> .....	<b>9</b>
<b>8.3 Measurement</b> .....	<b>9</b>
<b>8.3.1 Instrument conditions</b> .....	<b>9</b>
<b>8.3.2 Optimization of the instrument conditions</b> .....	<b>9</b>
<b>8.3.3 Interferences</b> .....	<b>9</b>
<b>8.3.4 Measurement</b> .....	<b>10</b>
<b>8.3.5 Matrix effects</b> .....	<b>10</b>
<b>8.3.6 Spiking</b> .....	<b>10</b>
<b>9 Calculation and expression of the results</b> .....	<b>12</b>
<b>10 Test report</b> .....	<b>12</b>
<b>Annex A (informative) Results of the inter-laboratory study</b> .....	<b>13</b>
<b>A.1 Inter-laboratory tests</b> .....	<b>13</b>
<b>A.2 Statistical results for the determination of the total contents of the elements</b> .....	<b>13</b>
<b>Annex ZA (informative) Relationship of this European Standard and the essential requirements of Regulation (EU) 2019/1009 making available on the market of EU fertilising products aimed to be covered</b> .....	<b>21</b>
<b>Bibliography</b> .....	<b>22</b>

## European foreword

This document (EN 17701-2:2024) has been prepared by Technical Committee CEN/TC 455 “Plant biostimulants”, the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2025, and conflicting national standards shall be withdrawn at the latest by May 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 17701-2:2022.

EN 17701-2:2024 includes the following significant technical changes with respect to CEN/TS 17701-2:2022:

- scope text regarding blends has been changed;
- Annex A “Results of the inter-laboratory study” has been added;
- Annex ZA has been added.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

EN 17701 series, *Plant biostimulants — Determination of specific elements*, consists of the following parts:

- *Part 1: Digestion by aqua regia for subsequent determination of elements;*
- *Part 2: Determination of total content of Cd, Pb, Ni, As, Cr, Cu and Zn;*
- *Part 3: Determination of mercury.*

Any feedback and questions on this document should be directed to the users’ national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## Introduction

The European Committee for Standardization (CEN) was requested by the European Commission (EC) to draft European Standards or European Standardization deliverables to support the implementation of Regulation (EU) 2019/1009 of 5 June 2019 [1] laying down rules on the making available on the market of EU fertilising products (“FPR” or “Fertilising Products Regulation”).

This standardization request, presented as SR M/564 and relevant amendments, also contributes to the Communication on “Innovating for Sustainable Growth: A Bio economy for Europe”. The interest in plant biostimulants has increased significantly in Europe as a valuable tool to use in agriculture. Standardization was identified as having an important role in order to promote the use of biostimulants. The work of CEN/TC 455 seeks to improve the reliability of the supply chain, thereby improving the confidence of farmers, industry, and consumers in biostimulants, and will promote and support commercialisation of the European biostimulant industry

This document concerns the analytical measurement step. Aqua regia digestion followed by inductively coupled plasma atomic emission spectrometry (ICP-AES) is widely used for determination of many elements, for example, a multi-matrix standard for aqua regia extraction of soils, sludges and biowaste [2] and soil [3]. A similar procedure was applied for determination of aqua regia extractable contents of arsenic, mercury, cadmium, chromium, nickel and lead in fertilizers and liming materials [4, 5] and for determination of micronutrients in mineral fertilizers [6]. Inductively coupled plasma atomic emission spectrometry (ICP-AES) is nowadays widely used and a well-established method in many laboratories.

The inter-laboratory study reflects the final statistical characteristics of the method for the determination of the specific elements in aqua regia digests including both, the digestion and the measurement steps.

**WARNING** — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety issues, if any, associated with its use. It is the responsibility of the user to establish appropriate health and safety practices and to ensure compliance with any national regulatory conditions.

**IMPORTANT** — It is absolutely essential that tests conducted according to this document are carried out by suitably trained staff.

## 1 Scope

This document specifies a method for the determination of total contents of arsenic (As), cadmium (Cd), copper (Cu), chromium (Cr), lead (Pb), nickel (Ni) and zinc (Zn) in aqua regia plant biostimulant digests using inductively coupled plasma-atomic emission spectrometry (ICP-AES).

This document is applicable to the blends of fertilizing products where a blend is a mix of at least two of the following component EU fertilising products categories: Fertilizers, Liming Materials, Soil Improvers, Growing Media, Inhibitors, Plant Biostimulants, and where the following category Plant Biostimulants is the highest % in the blend by mass or volume, or in the case of liquid form by dry mass. If Plant Biostimulants is not the highest % in the blend, the European Standard for the highest % of the blend applies. In case a blend of fertilizing products is composed of components in equal quantity or in case the component EU fertilising products used for the blend have identical formulations<sup>1</sup>, the user decides which standard to apply.

This method is applicable to aqua regia digests prepared according to EN 17701-1:2024. The method can be used for the determination of other elements, provided the user has verified the applicability.

NOTE Alternatively, inductively coupled plasma mass spectrometry (ICP-MS) can be used for the determination of the elements in the aqua regia digests if the user proves that the method gives the same results.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 17701-1:2024, *Plant biostimulants — Determination of specific elements — Part 1: Digestion by aqua regia for subsequent determination of elements*

EN 17704:2024, *Plant biostimulants — Determination of dry matter*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp/>
- IEC Electropedia: available at <https://www.electropedia.org/>

### 3.1

#### **analyte**

parameter to be determined

### 3.2

#### **blank calibration solution**

solution prepared in the same way as the calibration solution but leaving out the analytes

### 3.3

#### **blank test solution**

solution prepared in the same way as the test sample solution but omitting the test portion

---

<sup>1</sup> An example of such a blend is a product with two claimed functions consisting of a non-microbial plant biostimulant and an organic fertilizer composed of 1kg/kg of plant biostimulant from seaweed.