

Ambient air - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>See Eesti standard EVS-EN 14212:2024 sisaldab Euroopa standardi EN 14212:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 18.12.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN 14212:2024 consists of the English text of the European standard EN 14212:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 18.12.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 13.040.20

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

English Version

## Ambient air - Standard method for the measurement of the concentration of sulphur dioxide by ultraviolet fluorescence

Air ambiant - Méthode normalisée pour le mesurage de la concentration en dioxyde de soufre par fluorescence U.V.

Außenluft - Messverfahren zur Bestimmung der Konzentration von Schwefeldioxid mit Ultraviolett-Fluoreszenz

This European Standard was approved by CEN on 11 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

<b>Contents</b>	<b>Page</b>
European foreword.....	8
<b>1</b> Scope.....	<b>9</b>
<b>2</b> Normative references.....	<b>10</b>
<b>3</b> Terms and definitions.....	<b>11</b>
<b>4</b> Abbreviated terms.....	<b>16</b>
<b>5</b> Principle.....	<b>17</b>
5.1 General.....	17
5.2 Measuring principle.....	17
5.3 Type testing.....	18
5.4 Field operation and quality control.....	18
<b>6</b> Sampling.....	<b>18</b>
6.1 General.....	18
6.2 Sampling location.....	19
6.3 Sampling system.....	19
6.4 Control and regulation of sample flow rate.....	20
6.5 Sampling pump for the manifold.....	20
<b>7</b> Analyser equipment.....	<b>21</b>
7.1 General.....	21
7.2 Selective traps for interfering agents.....	21
7.3 Optical assembly.....	21
7.4 Pressure measurement.....	21
7.5 Flow rate indicator.....	22
7.6 Sampling pump for the analyser.....	22
7.7 Internal sulfur dioxide span source.....	22
7.8 Particle filter.....	22
<b>8</b> Type testing of ultraviolet fluorescence sulfur dioxide analysers.....	<b>22</b>
8.1 General.....	22
8.2 Relevant performance characteristics and criteria.....	23
Table 1 — Relevant performance characteristics and criteria.....	24
8.3 Design change.....	26
8.4 Procedures for determination of the performance characteristics during the laboratory test.....	26
Table 2 — Set points and stability of test parameters.....	27
Table 3 — Methods for preparation of test gases.....	27
Table 4a — Specification for purity of test gas.....	28
Table 4b — Specification for purity of zero gas for interferents testing [9].....	28
Table 4c — Specification for purity of zero gas for other tests [9].....	28
Figure 1 — Diagram illustrating the response time.....	30
Figure 2 — Concentration variations for the averaging effect test.....	37
8.5 Determination of the performance characteristics during the field test.....	38
8.6 Type testing and uncertainty calculation.....	42
<b>9</b> Field operation and ongoing quality control.....	<b>42</b>

9.1	General .....	42
9.2	Suitability evaluation .....	43
	<b>Table 5 — Site-specific conditions to be evaluated .....</b>	<b>43</b>
9.3	Initial installation .....	44
9.4	Ongoing quality assurance/quality control .....	46
	<b>Table 6 — Required frequency of calibration, checks and maintenance .....</b>	<b>47</b>
9.5	Calibration of the analyser .....	49
	<b>Figure 3 — Flow scheme for performance of evaluation of effects of violation of performance requirements and possibilities for data correction .....</b>	<b>52</b>
9.6	Checks .....	53
9.7	Maintenance .....	56
9.8	Data handling and data reports .....	57
9.9	Measurement uncertainty .....	57
10	Expression of results .....	58
11	Test reports and documentation .....	58
11.1	Type testing .....	58
11.2	Field operation .....	59
	<b>Annex A (normative) Test of lack of fit .....</b>	<b>61</b>
A.1	Establishment of the regression line .....	61
A.2	Calculation of the residuals of the averages .....	61
	<b>Annex B (informative) Sampling equipment .....</b>	<b>63</b>
	<b>Figure B.1 — Sampling layout with a main sampling manifold .....</b>	<b>63</b>
	<b>Figure B.2 — Sampling layout with individual lines .....</b>	<b>64</b>
	<b>Annex C (informative) Ultraviolet fluorescence analyser .....</b>	<b>65</b>
	<b>Figure C.1 — Schematic diagram of a UV fluorescence sulfur dioxide analyser .....</b>	<b>65</b>
	<b>Annex D (informative) Manifold testing .....</b>	<b>66</b>
D.1	Procedure for applying test gas .....	66
	<b>Figure D.1 — Schematic diagram of manifold testing equipment .....</b>	<b>67</b>
D.2	Procedure for the cross test .....	67
D.2.1	General .....	67
D.2.2	Initial stage .....	67
D.2.3	Stage 1 .....	68
D.2.4	Stage 2 .....	68
D.2.5	Data Processing .....	68
D.2.6	Evaluation .....	68
	<b>Annex E (normative) Type testing .....</b>	<b>70</b>
E.1	Type testing and uncertainty calculation .....	70
E.1.1	Type testing .....	70
E.1.2	Uncertainty calculation .....	70
E.2	Type testing Requirement a) .....	70

<b>Table E.1 — Relevant performance characteristics and criteria .....</b>	<b>71</b>
<b>E.3 Type testing Requirement b) .....</b>	<b>72</b>
<b>E.3.1 General .....</b>	<b>72</b>
<b>Table E.2 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory tests .....</b>	<b>73</b>
<b>Table E.3 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory tests .....</b>	<b>73</b>
<b>E.3.2 Calculation of standard uncertainties .....</b>	<b>75</b>
<b>E.3.2.1 General .....</b>	<b>75</b>
<b>E.3.2.2 Repeatability at zero .....</b>	<b>75</b>
<b>E.3.2.3 Repeatability at the hourly limit value of sulfur dioxide .....</b>	<b>75</b>
<b>E.3.2.4 Lack of fit .....</b>	<b>76</b>
<b>E.3.2.5 Influence quantities .....</b>	<b>76</b>
<b>E.3.2.5.1 General .....</b>	<b>76</b>
<b>E.3.2.5.2 Sample gas pressure .....</b>	<b>77</b>
<b>E.3.2.5.3 Sample gas temperature .....</b>	<b>77</b>
<b>E.3.2.5.4 Surrounding temperature .....</b>	<b>78</b>
<b>E.3.2.5.5 Electrical voltage .....</b>	<b>78</b>
<b>E.3.2.5.6 Water vapour .....</b>	<b>79</b>
<b>E.3.2.5.7 Other interferents .....</b>	<b>79</b>
<b>E.3.2.5.8 Averaging effect .....</b>	<b>80</b>
<b>E.3.2.5.9 Calibration gas .....</b>	<b>81</b>
<b>E.3.2.5.10 Difference sample/calibration port .....</b>	<b>81</b>
<b>E.3.3 Example calculation .....</b>	<b>82</b>
<b>Table E.4 — Example calculation .....</b>	<b>82</b>
<b>E.4 Type testing Requirement c) .....</b>	<b>84</b>
<b>Table E.5 — Relevant performance characteristics and criteria .....</b>	<b>84</b>
<b>E.5 Type testing Requirement d) .....</b>	<b>84</b>
<b>E.5.1 General .....</b>	<b>84</b>
<b>Table E.6 — Standard uncertainties to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests .....</b>	<b>85</b>
<b>Table E.7 — Standard uncertainty of the calibration gas to be incorporated in the calculation of the expanded uncertainty after the laboratory and field tests .....</b>	<b>85</b>
<b>E.5.2 Combined standard uncertainty .....</b>	<b>86</b>
<b>E.5.3 Absolute expanded uncertainty .....</b>	<b>86</b>
<b>E.5.4 Relative expanded uncertainty .....</b>	<b>87</b>
<b>E.5.5 Calculation of standard uncertainties .....</b>	<b>87</b>
<b>E.5.6 Example calculation .....</b>	<b>89</b>

<b>Table E.8 —Example calculation .....</b>	<b>89</b>
<b>Annex F (informative) Calculation of uncertainty in field operation at the hourly limit value ...</b>	<b>91</b>
<b>F.1 General .....</b>	<b>91</b>
<b>F.2 Combined standard uncertainty .....</b>	<b>91</b>
<b>F.3 Standard uncertainties.....</b>	<b>92</b>
<b>F.3.1 General .....</b>	<b>92</b>
<b>F.3.1.1 General .....</b>	<b>92</b>
<b>F.3.1.2 Repeatability at zero .....</b>	<b>92</b>
<b>F.3.1.3 Repeatability at the hourly limit value .....</b>	<b>92</b>
<b>F.3.1.4 Lack of fit.....</b>	<b>92</b>
<b>F.3.2 Influence quantities.....</b>	<b>93</b>
<b>F.3.2.1 General .....</b>	<b>93</b>
<b>F.3.2.2 Sample gas pressure.....</b>	<b>93</b>
<b>F.3.2.3 Sample gas temperature .....</b>	<b>93</b>
<b>F.3.2.4 Surrounding temperature .....</b>	<b>94</b>
<b>F.3.2.5 Electrical voltage.....</b>	<b>95</b>
<b>F.3.3 Interferents.....</b>	<b>95</b>
<b>F.3.3.1 Water vapour .....</b>	<b>95</b>
<b>F.3.3.2 Other interferents .....</b>	<b>96</b>
<b>F.3.4 Averaging effect.....</b>	<b>96</b>
<b>F.3.5 Reproducibility under field conditions .....</b>	<b>96</b>
<b>F.3.6 Long-term drift at zero .....</b>	<b>97</b>
<b>F.3.7 Long-term drift at level of the hourly limit value .....</b>	<b>97</b>
<b>F.3.8 Zero gas .....</b>	<b>97</b>
<b>F.3.9 Calibration gas .....</b>	<b>97</b>
<b>F.3.10 Difference sample/calibration port.....</b>	<b>97</b>
<b>F.4 Example calculation.....</b>	<b>98</b>
<b>Table F.1 —Example calculation .....</b>	<b>98</b>
<b>Annex G (informative) Calculation of uncertainty in field operation at the daily limit value ...</b>	<b>100</b>
<b>G.1 General .....</b>	<b>100</b>
<b>G.2 Combined standard uncertainty .....</b>	<b>100</b>
<b>G.3 Standard uncertainties.....</b>	<b>101</b>
<b>G.3.1 General .....</b>	<b>101</b>
<b>G.3.1.1 General .....</b>	<b>101</b>
<b>G.3.1.2 Repeatability at zero .....</b>	<b>101</b>
<b>G.3.1.3 Repeatability at the daily limit value .....</b>	<b>102</b>

G.3.1.4 Lack of fit.....	102
G.3.2 Influence quantities .....	102
G.3.2.1 General.....	102
G.3.2.2 Sample gas pressure .....	103
G.3.2.3 Sample gas temperature.....	103
G.3.2.4 Surrounding temperature.....	104
G.3.2.5 Electrical voltage.....	104
G.3.3 Interferents .....	105
G.3.3.1 General.....	105
G.3.3.2 Water vapour.....	105
G.3.3.3 Other interferents.....	106
G.3.4 Averaging effect.....	107
G.3.5 Zero gas.....	107
G.3.6 Calibration gas .....	107
G.3.7 Difference sample/calibration port.....	107
G.3.8 Reproducibility under field conditions.....	108
G.3.9 Long-term drift at zero.....	108
G.3.10 Long-term drift at level of the daily limit value .....	108
G.4 Example calculation .....	109
Table G.1 —Example calculation .....	109
Annex H (informative) Calculation of uncertainty in field operation at the annual critical level.....	111
H.1 General.....	111
H.2 Combined standard uncertainty.....	111
H.3 H.3 Standard uncertainties.....	113
H.3.1 General.....	113
H.3.2 Repeatability at zero.....	113
H.3.3 Repeatability at the annual limit value .....	113
H.3.4 Lack of fit.....	114
H.3.5 Influence quantities .....	114
H.3.5.1 General.....	114
H.3.5.2 Sample gas pressure .....	114
H.3.5.3 Sample gas temperature.....	115
H.3.5.4 Surrounding temperature.....	115
H.3.5.5 Electrical voltage .....	116
H.3.5.6 Interferents .....	117
H.3.5.7 Water vapour.....	117

<b>H.3.5.8 Other interferents .....</b>	<b>117</b>
<b>H.3.6 Averaging effect.....</b>	<b>118</b>
<b>H.3.7 Zero gas .....</b>	<b>119</b>
<b>H.3.8 Calibration gas .....</b>	<b>119</b>
<b>H.3.9 Difference sample/calibration port.....</b>	<b>119</b>
<b>H.3.10 Reproducibility under field conditions .....</b>	<b>120</b>
<b>H.3.11 Long term drift at zero.....</b>	<b>120</b>
<b>H.3.12 Long term drift at level of the annual critical level .....</b>	<b>121</b>
<b>Annex I (informative) Test stand for the test point “sensitivity coefficient of sample gas pressure” .....</b>	<b>122</b>
<b>Figure I.1 — Test stand for the test point “sensitivity coefficient of sample gas pressure” .....</b>	<b>122</b>
<b>Annex J (informative) Significant technical changes .....</b>	<b>123</b>
<b>Table J.1 —Details of significant technical changes between this document and the previous edition.....</b>	<b>123</b>
<b>Bibliography .....</b>	<b>124</b>

## European foreword

This document (EN 14212:2024) has been prepared by Technical Committee CEN/TC 264 "Air quality", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 14212:2012.

The technical modifications in comparison with the previous edition are listed in Annex J of this document.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## 1 Scope

This document specifies a continuous measurement method for the determination of the concentration of sulfur dioxide present in ambient air based on the ultraviolet fluorescence measuring principle. This document describes the performance characteristics and sets the relevant minimum criteria required to select an appropriate ultraviolet fluorescence analyser by means of type testing. It also includes the evaluation of the suitability of an analyser for use in a specific fixed site so as to meet the data quality requirements (see Annex I of Directive 2008/50/EC [1] for additional information) and requirements during sampling, calibration and quality assurance for use.

The method is applicable to the determination of the mass concentration of sulfur dioxide present in ambient air up to 1000  $\mu\text{g}/\text{m}^3$ . This concentration range represents the certification range for sulfur dioxide for type testing.

NOTE 1 It is possible to use other ranges depending on the levels present in ambient air.

NOTE 2 Exemplar uncertainty budget calculations are given in Annexes E to H referring to Directive 2008/50/EC [1]. In the event that the Limit Values are updated in future iterations of Directive 2008/50/EC [1], the user can use these new values to calculate measurement uncertainties.

The method covers the determination of ambient air concentrations of sulfur dioxide in zones classified as rural areas, urban-background areas, traffic-oriented locations and locations influenced by industrial sources.

The results are expressed in  $\mu\text{g}/\text{m}^3$  (at 20 °C and 101,3 kPa).

NOTE 3 1 000  $\mu\text{g}/\text{m}^3$  of  $\text{SO}_2$  corresponds to 376 nmol/mol of  $\text{SO}_2$ .

This document contains information for different groups of users.

Clause 5 to Clause 7 and Annex C and Annex D contain general information about the principles of sulfur dioxide measurement by ultraviolet fluorescence analyser and sampling equipment.

Clause 8 and Annex E are specifically directed towards test houses and laboratories that perform type testing of sulfur dioxide analysers. These sections contain information about:

- type testing conditions, test procedures and test requirements;
- analyser performance requirements;
- evaluation of the type testing results;
- evaluation of the associated uncertainty of the measurement performed by the sulfur dioxide analyser based on the type testing results.

Clause 9 to Clause 11 and Annex F, Annex G and Annex H are directed towards monitoring networks performing the practical measurements of sulfur dioxide in ambient air. These sections contain information about:

- initial installation of the analyser in the monitoring network and acceptance testing;
- ongoing quality assurance/quality control;
- calculation and reporting of measurement results;
- evaluation of the uncertainty of the measurement results under practical monitoring conditions.

This document represents an evolution of earlier editions (EN 14212:2005 and EN 14212:2012).

NOTE 4 Analysers type tested prior to the publication of this document can still be used for regulated monitoring purposes. As newer versions of analysers tested under this document become available, discontinue the use of older reference analysers.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15267-1, *Air quality - Assessment of air quality monitoring equipment - Part 1: General principles of certification*

EN 15267-2, *Air quality - Assessment of air quality monitoring equipment - Part 2: Initial assessment of the manufacturer's quality management system and post certification surveillance for the manufacturing process*

EN ISO 6142, *Gas analysis — Preparation of calibration gas mixtures — Gravimetric method (ISO 6142)*

EN ISO 6143, *Gas analysis - Comparison methods for determining and checking the composition of calibration gas mixtures (ISO 6143)*

EN ISO 6144, *Gas analysis - Preparation of calibration gas mixtures - Static volumetric method (ISO 6144)*

EN ISO 6145-6, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 6: Critical flow orifices (ISO 6145-6)*

EN ISO 6145-7, *Gas analysis - Preparation of calibration gas mixtures using dynamic methods - Part 7: Thermal mass-flow controllers (ISO 6145-7)*

EN ISO 6145-10, *Gas analysis - Preparation of calibration gas mixtures using dynamic volumetric methods - Part 10: Permeation method (ISO 6145-10)*

EN ISO 14956, *Air quality - Evaluation of the suitability of a measurement procedure by comparison with a required measurement uncertainty (ISO 14956)*

EN ISO 19229, *Gas analysis - Purity analysis and the treatment of purity data*