



**International  
Standard**

**ISO 13402**

**Surgical and dental hand  
instruments — Determination of  
resistance against autoclaving,  
corrosion and thermal exposure**

*Instruments chirurgicaux et dentaires à main — Détermination  
de la résistance au passage à l'autoclave, à la corrosion et à  
l'exposition à la chaleur*

**Second edition  
2025-03**

This document is a preview generated by EMS



**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2025

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 Test methods</b> .....	<b>1</b>
4.1 General.....	1
4.2 Sampling plan.....	1
4.3 Overview of test methods and applicability.....	2
<b>5 Test report</b> .....	<b>2</b>
<b>Annex A (normative) Boiling test in deionized water</b> .....	<b>4</b>
<b>Annex B (normative) Boiling test in 0,9 % NaCl solution</b> .....	<b>6</b>
<b>Annex C (normative) Copper sulfate test for steels with equal or greater than 16 % chromium</b> .....	<b>8</b>
<b>Annex D (normative) Copper sulfate test for martensitic steels with less than 16 % chromium</b> .....	<b>10</b>
<b>Annex E (informative) Test in 0,3 % sodium chloride solution of austenitic steels</b> .....	<b>12</b>
<b>Annex F (informative) Test with citric acid solution for austenitic steels</b> .....	<b>14</b>
<b>Annex G (informative) Thermal test</b> .....	<b>16</b>
<b>Annex H (informative) Autoclave test</b> .....	<b>17</b>
<b>Bibliography</b> .....	<b>18</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 170, *Surgical instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 55, *Dentistry*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 13402:1995), which has been technically revised.

The main changes are as follows:

- restructuring of the document;
- update of [Clause 2](#);
- addition of [Clause 3](#);
- addition of [Clause 4](#) including a table with an overview of test methods;
- addition of [Clause 5](#);
- addition of [Annexes A](#) to [H](#) including test methods.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

The procedures described in this document form a harmonized series of tests that can be referred to, individually or in combination, in other separate product standards. The requirements for such tests are defined and stated within the body of the product standard along with the number of cycles for each test procedure. The tests apply to dental and surgical instruments and are already standardized in relevant product standards (e.g. ISO 7151, ISO 9173-1). However, the test procedures as stated in the product standards differ in minor details. An alignment and a compilation were established.

The most important test methods for dental and surgical instruments have been compiled in this document.

This document is a preview generated by EVS



# Surgical and dental hand instruments — Determination of resistance against autoclaving, corrosion and thermal exposure

## 1 Scope

This document describes test methods to determine the resistance of stainless steel surgical and dental hand instruments against autoclaving, corrosion and thermal exposure.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 7554-3, *Surgical instruments — Terms, measuring methods and tests — Part 3: Test methods*<sup>1)</sup>

EN 13018, *Non-destructive testing — Visual testing — General principles*

## 3 Terms and definitions

No terms and definitions are listed in this document.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

## 4 Test methods

### 4.1 General

All tests are type tests.

### 4.2 Sampling plan

- a) Each instrument manufacturer shall define an appropriate sampling plan for ensuring instrument safety and quality. The sampling shall be done according to ISO 7554-3.
- b) Rejected lots are often re-passivated and re-tested, typically using an increased sample size of instruments. If the lot is rejected a second time, the cause of the repeat failure should be evaluated before proceeding.

---

1) Under preparation. Stage at the time of publication: ISO/DIS 7664-3.