

Measurement of radioactivity in the environment - Soil
- Part 7: In situ measurement of gamma-emitting
radionuclides (ISO 18589-7:2025)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>See Eesti standard EVS-EN ISO 18589-7:2025 sisaldab Euroopa standardi EN ISO 18589-7:2025 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 24.12.2025.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 18589-7:2025 consists of the English text of the European standard EN ISO 18589-7:2025.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 24.12.2025.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.080.99, 17.240

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 18589-7

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2025

ICS 13.080.99; 17.240

Supersedes EN ISO 18589-7:2016

English Version

**Measurement of radioactivity in the environment - Soil -
Part 7: In situ measurement of gamma-emitting
radionuclides (ISO 18589-7:2025)**

Mesurage de la radioactivité dans l'environnement -
Sol - Partie 7: Mesurage in situ des radionucléides
émetteurs gamma (ISO 18589-7:2025)

Ermittlung der Radioaktivität in der Umwelt -
Erdboden - Teil 7: In-situ-Messung von
Gammastrahlung emittierenden Radionukliden (ISO
18589-7:2025)

This European Standard was approved by CEN on 15 December 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 18589-7:2025) has been prepared by Technical Committee ISO/TC 85 "Nuclear energy, nuclear technologies, and radiological protection" in collaboration with Technical Committee CEN/TC 430 "Nuclear energy, nuclear technologies, and radiological protection" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2026, and conflicting national standards shall be withdrawn at the latest by June 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18589-7:2016.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 18589-7:2025 has been approved by CEN as EN ISO 18589-7:2025 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Symbols	3
5 Principles	6
5.1 Measurement method.....	6
5.2 Uncertainties of the measurement method.....	7
6 Equipment	7
6.1 Portable in situ spectrometry system.....	7
6.2 Detector system.....	7
6.2.1 General.....	7
6.2.2 Field-of-view.....	8
6.2.3 Special requirements.....	8
6.3 Pulse processing electronics.....	8
6.3.1 Components.....	8
6.3.2 Special requirements.....	9
6.3.3 Requirements for the evaluation program.....	9
6.4 Assembly jig for a detector system.....	10
6.5 Collimated detector.....	10
6.5.1 Construction.....	10
6.5.2 Collimator parameter.....	11
7 Procedure	12
7.1 Calibration.....	12
7.2 Method of combined calibrations.....	13
7.2.1 General.....	13
7.2.2 Intrinsic efficiency.....	14
7.2.3 Geometry factor.....	14
7.2.4 Angular correction factor.....	15
7.2.5 Numerical calibration procedure.....	16
8 Quality assurance and quality control program	17
8.1 General.....	17
8.2 Influencing variables.....	17
8.3 Instrument verification.....	17
8.4 Method verification.....	17
8.5 Quality control program.....	18
8.5.1 General.....	18
8.5.2 Description of periodical quality checks.....	18
8.5.3 Measurement verification.....	18
8.5.4 Qualification.....	18
8.5.5 Documentation of quality controls.....	19
8.6 Standard operating procedure.....	19
9 Expression of results	19
9.1 Calculation of activity per unit of surface area or unit of mass.....	19
9.2 Calculation of the characteristic limits and the best estimate of the measurand as well as its standard uncertainty.....	19
9.2.1 General.....	19
9.2.2 Standard uncertainty.....	20
9.2.3 Decision threshold and detection limit.....	20
9.2.4 Limits of coverage interval and best estimate of the measurand.....	21

9.3	Calculation of the radionuclide specific ambient dose rate.....	21
10	Test report.....	23
Annex A	(informative) Influence of radionuclides in air on the result of surface or mass activity measured by in situ gamma spectrometry.....	24
Annex B	(informative) Influence quantities.....	25
Annex C	(informative) Characteristics of germanium detectors.....	28
Annex D	(informative) Field-of-view of an in situ gamma spectrometer as a function of the photon energy for different radionuclide distributions in soil.....	30
Annex E	(informative) Methods for calculating geometry factors and angular correction factors.....	34
Annex F	(informative) Example for calculation of the characteristic limits as well as the best estimate of the measurand and its standard uncertainty.....	42
Annex G	(informative) Conversion factors for surface or mass activity to air kerma rate and ambient dose equivalent rate for different radionuclide distribution in soil.....	46
Annex H	(informative) Mass attenuation factors for soil and attenuation factors for air as a function of photon energy and deviation of $G(E,V)$ for different soil compositions.....	52
Bibliography	54

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee, TC 85, *Nuclear energy, nuclear technologies, and radiological protection*, Subcommittee SC 2, *Radiological protection*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 430, *Nuclear energy, nuclear technologies, and radiological protection*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 18589-7:2013), of which it constitutes a minor revision.

The main changes are as follows:

- [B.10](#): correction of the information related to the activity concentration of ^{40}K ;
- [E.2](#) and [E.6](#): correction of [Formulae \(E.5\)](#) and [\(E.11\)](#);
- [F.4](#): correction of β , according to the numerical values of the example;
- [F.6](#): modify $\beta = 50 \text{ g}\cdot\text{cm}^{-2}$ into $\beta = 50 \text{ kg}\cdot\text{m}^{-2}$;
- [G.3](#), Footnote 1 of [Table G.3](#): modify $1 \text{ g}\cdot\text{cm}^{-2} = 10 \text{ kg}\cdot\text{cm}^{-2}$ into $1 \text{ g}\cdot\text{cm}^{-2} = 10 \text{ kg}\cdot\text{m}^{-2}$.

A list of all parts in the ISO 18589 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

In situ gamma spectrometry is a rapid and accurate technique to assess the activity concentration of gamma-emitting radionuclides present in the top soil layer or deposited onto the soil surface. This method is also used to assess the dose rates of individual radionuclides.

In situ gamma spectrometry is a direct physical measurement of radioactivity that does not need any soil samples, thus reducing the time and cost of laboratory analysis of large number of soil samples.

The quantitative analysis of the recorded line spectra requires a suitable area for the measurement. Furthermore, it is required to know the physicochemical properties of the soil and the vertical distribution in the soil to assess the activity of the radionuclides.

Measurement of radioactivity in the environment — Soil —

Part 7: In situ measurement of gamma-emitting radionuclides

1 Scope

This document specifies the identification of radionuclides and the measurement of their activity in soil using in situ gamma spectrometry with portable systems equipped with germanium or scintillation detectors.

This document is suitable to rapidly assess the activity of artificial and natural radionuclides deposited on or present in soil layers of large areas of a site under investigation.

This document can be used in connection with radionuclide measurements of soil samples in the laboratory (see ISO 18589-3) in the following cases:

- routine surveillance of the impact of radioactivity released from nuclear installations or of the evolution of radioactivity in the region;
- investigations of accident and incident situations;
- planning and surveillance of remedial action;
- decommissioning of installations or the clearance of materials.

It can also be used for the identification of airborne artificial radionuclides, when assessing the exposure levels inside buildings or during waste disposal operations.

Following a nuclear accident, in situ gamma spectrometry is a powerful method for rapid evaluation of the gamma activity deposited onto the soil surface as well as the surficial contamination of flat objects.

NOTE The method described in this document is not suitable when the spatial distribution of the radionuclides in the environment is not precisely known (influence quantities, unknown distribution in soil) or in situations with very high photon flux. However, the use of small volume detectors with suitable electronics allows measurements to be performed under high photon flux.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11929 (all parts), *Determination of the characteristic limits (decision threshold, detection limit and limits of the coverage interval) for measurements of ionizing radiation — Fundamentals and application*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

IEC 61275, *Radiation protection instrumentation — Measurement of discrete radionuclides in the environment — In situ photon spectrometry system using a germanium detector*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.