

TSEMENDI KATSETAMINE. OSA 2: TSEMENDI
KEEMILINE ANALÜÜS

Methods of testing cement - Part 2: Chemical analysis
of cement

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>See Eesti standard EVS-EN 196-2:2026 sisaldab Euroopa standardi EN 196-2:2025 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 19.11.2025.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN 196-2:2026 consists of the English text of the European standard EN 196-2:2025.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 19.11.2025.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 91.100.10

<p>Standardite ja standardilaadsete dokumentide reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele</p> <p>Eesti standardid ja standardilaadsed dokumendid on Eesti Standardimis- ja Akrediteerimiskeskuse intellektuaalomand ning neid kasutatakse litsentsi alusel dokumentide kasutuslepingu tingimuste kohaselt.</p> <p>Ilma Eesti Standardimis- ja Akrediteerimiskeskuse eelneva kirjaliku loata on keelatud standardite ja standardilaadsete dokumentide täielik või osaline reprodutseerimine, levitamine, muutmine või kasutamine mis tahes kujul ja viisil - sealhulgas kopeerimise, skaneerimise, salvestamise või jagamise teel digiplatvormidel (k.a masinõppe ja tehisintellekti rakendustes). Loata kasutamine väljaspool litsentsi tingimusi käsitletakse õigusrikkumisenä.</p> <p>Kui Teil on küsimusi standardite ja standardilaadsete dokumentide autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Veebileht www.evs.ee; telefon +372 6055050; e-post info@evs.ee</p> <p>The right to reproduce and distribute standards and standard-like documents belongs to the Estonian Centre for Standardisation and Accreditation</p> <p>Estonian standards and standard-like documents are the intellectual property of the Estonian Centre for Standardisation and Accreditation and are made available under license in accordance with the terms and conditions of the document use agreement.</p> <p>Without the prior written permission of the Estonian Centre for Standardisation and Accreditation, the full or partial reproduction, distribution, modification, or use of standards and standard-like documents in any form or by any means - including photocopying, scanning, storing, or sharing via digital platforms (incl. in machine learning and artificial intelligence applications) - is strictly prohibited. Any unauthorized use beyond the scope of the granted license is prohibited and may result in legal action.</p> <p>If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee</p>
--

EUROPEAN STANDARD

EN 196-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2025

ICS 91.100.10

Supersedes EN 196-2:2013

English Version

Methods of testing cement - Part 2: Chemical analysis of cement

Méthodes d'essais des ciments - Partie 2: Analyse chimique des ciments

Prüfverfahren für Zement - Teil 2: Chemische Analyse von Zement

This European Standard was approved by CEN on 22 September 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

Contents	Page
European foreword	3
Introduction	5
1 Scope	6
2 Normative references	6
3 Terms and definitions	6
4 General requirements for testing	7
4.1 Number of tests	7
4.2 Repeatability and reproducibility	7
4.3 Expression of masses, volumes, factors and results	7
5 Analysis by wet chemistry	8
5.1 General	8
5.2 Reagents	9
5.3 Apparatus	21
5.4 Analysis procedure	27
5.5 Determination of major elements	35
6 Chemical analysis by X-ray fluorescence	60
6.1 Reagents and reference materials	60
6.2 Apparatus	62
6.3 Flux	63
6.4 Determination of loss on ignition and the change in mass on fusion of the cement ..	64
6.5 Factoring test results and correcting total analyses for presence of sulfides and halides	66
6.6 Preparation of fused beads and pressed pellets	68
6.7 Calibration and validation	70
6.8 Calculation and expression of results	80
6.9 Performance criteria (repeatability, accuracy and reproducibility limits)	81
7 Chemical analysis by inductively coupled plasma optical emission spectroscopy (ICP-OES)	81
7.1 Determination of SO₃ - alternative method	81
Annex A (informative) Examples of fluxes	84
Annex B (informative) Sources of certified reference materials	85
Annex C (informative) Examples of calibration standards and monitor beads and pellets ...	86
Annex D (informative) Guidance on preparation of standard calibration solutions	87
D.1 Stock solutions	87
D.2 Blank calibration solution	87
Bibliography	88

European foreword

This document (EN 196-2:2025) has been prepared by Technical Committee CEN/TC 51 “Cement and building limes”, the secretariat of which is held by NBN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2026, and conflicting national standards shall be withdrawn at the latest by May 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 196-2:2013.

EN 196-2:2025 includes the following significant technical changes with respect to EN 196-2:2013:

- redefinition of the introduction;
- introduction of concept;
- determination of reactive silica;
- introduction of the concept of determination of total carbonate content instead of determination of CO₂;
- determination of chloride by potentiometric titration as alternative method;
- determination of total carbonate content by gas volumetric method as alternative method;
- determination of total carbonate content by infrared detection system (method A and B) as alternative method;
- determination of SO₃ by inductively coupled plasma optical emission spectroscopy as alternative method.

This European Standard series, under the general title *Methods of testing cement*, comprises the following parts:

- *Part 1: Determination of strength;*
- *Part 2: Chemical analysis of cement;*
- *Part 3: Determination of setting times and soundness;*
- *Part 5: Pozzolanicity test for pozzolanic cement;*
- *Part 6: Determination of fineness;*
- *Part 7: Methods of taking and preparing samples of cement;*
- *Part 8: Heat of hydration — Solution method;*
- *Part 9: Heat of hydration — Semi-adiabatic method;*

- *Part 10: Determination of the water-soluble chromium (VI) content of cement;*
- *Part 11: Heat of hydration — Isothermal conduction calorimetry method.*

NOTE Another document, CEN/TR 196-4 Methods of testing cement — Part 4: Quantitative determination of constituents, has been published as a CEN Technical Report.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Introduction

In this document, the elemental chemical analysis of a clinker or cement, as of any natural or artificial mineral material (volcanic pozzolan, blast furnace slag, fly ash, etc.) is expressed in a conventional manner in the form of the mass proportions of the common oxides corresponding to the most frequent and stable degree of oxidation (Si: SiO_2 ; Al: Al_2O_3 , etc.). This expression of the chemical analysis does not indicate how these elements are combined into minerals such as silicates, oxides, carbonates, sulphates or sulphides.

This means that the presence of SiO_2 in a chemical analysis sheet does not necessarily imply the presence of crystalline silica e.g. quartz or cristobalite, but most likely silicates. Moreover, even if quartz is detected in the material, this does not necessarily imply an inhalation hazard if the particle size is larger than the inhalable fraction. Similarly, the presence of TiO_2 in a chemical analysis report does not necessarily imply the presence of the mineralogical phase(s) corresponding to this formula (rutile, anatase, brookite), although the presence of the mineralogical phase(s) does not necessarily imply an inhalation hazard if the particle size is larger than the inhalable fraction. In the same way, determination of total carbonate content (TCC) can be done by dosing CO_2 induced by acid attack, which does not mean, in any case, that the material contains CO_2 .

The only way to identify mineralogical phases in a powdered mineral material is by X-ray diffraction, which is the only analytical technique that is sensitive to the crystalline character of minerals. For an accurate quantification of inhalable fractions, it is advisable to be necessary to perform a quantitative particle size selection (e.g. by aerosolization technique) in order to eliminate coarse fractions that could reduce the quantitative character of the X-ray diffraction analysis.

Standardization to 100 made it possible to know if no constituent element had been left out. Hence the expression "percentage by mass of oxides" in the chemical analysis.

1 Scope

This document specifies the methods for the chemical analysis of cement.

This document describes the reference methods and, in certain cases, an alternative method which can be considered to be equivalent. In the case of a dispute, only the reference methods are used.

An alternative performance-based method using X-ray fluorescence (XRF) is described for SiO₂, Al₂O₃, Fe₂O₃, CaO, MgO, SO₃, K₂O, Na₂O, TiO₂, P₂O₅, Mn₂O₃, SrO, Cl and Br. This method is based on beads of fused sample and analytical validation using certified reference materials, together with performance criteria. A method based on pressed pellets of un-fused sample can be considered as equivalent, providing that the analytical performance satisfies the same criteria.

An alternative performance-based method using inductively coupled plasma optical emission spectroscopy (ICP-OES) is described for SO₃.

When correctly calibrated according to the specified procedures and reference materials, XRF and ICP-OES provides methods equivalent to the reference methods but has not been validated for use yet as a reference procedure for conformity and dispute purposes. They can be applied to other relevant elements when adequate calibrations have been established.

Any other methods can be used provided they are calibrated, either against the reference methods or against internationally accepted reference materials, in order to demonstrate their equivalence.

This document describes methods which apply principally to cements, but which can also be applied to their constituent materials. They can also be applied to other materials, the standards for which call up these methods. Standard specifications state which methods are to be used.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 196-7, *Methods of testing cement — Part 7: Methods of taking and preparing samples of cement*

EN 13656, *Soil, treated biowaste, sludge and waste — Digestion with a hydrochloric (HCl), nitric (HNO₃) and tetrafluoroboric (HBF₄) or hydrofluoric (HF) acid mixture for subsequent determination of elements*

EN ISO 385, *Laboratory glassware — Burettes (ISO 385)*

EN ISO 835, *Laboratory glassware — Graduated pipettes (ISO 835)*

ISO 33401, *Reference materials — Contents of certificates, labels and accompanying documentation*

ISO Guide 30, *Reference materials — Selected terms and definitions*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp/>
- IEC Electropedia: available at <https://www.electropedia.org/>