

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Medical electrical equipment –  
Part 2-44: Particular requirements for the basic safety and essential performance  
of X-ray equipment for computed tomography**

**Appareils électromédicaux –  
Partie 2-44: Exigences particulières pour la sécurité de base et les performances  
essentielles des équipements à rayonnement X de tomодensitométrie**



## THIS PUBLICATION IS COPYRIGHT PROTECTED

Copyright © 2012 IEC, Geneva, Switzerland

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either IEC or IEC's member National Committee in the country of the requester.

If you have any questions about IEC copyright or have an enquiry about obtaining additional rights to this publication, please contact the address below or your local IEC member National Committee for further information.

Droits de reproduction réservés. Sauf indication contraire, aucune partie de cette publication ne peut être reproduite ni utilisée sous quelque forme que ce soit et par aucun procédé, électronique ou mécanique, y compris la photocopie et les microfilms, sans l'accord écrit de la CEI ou du Comité national de la CEI du pays du demandeur.

Si vous avez des questions sur le copyright de la CEI ou si vous désirez obtenir des droits supplémentaires sur cette publication, utilisez les coordonnées ci-après ou contactez le Comité national de la CEI de votre pays de résidence.

IEC Central Office  
3, rue de Varembe  
CH-1211 Geneva 20  
Switzerland

Tel.: +41 22 919 02 11  
Fax: +41 22 919 03 00  
[info@iec.ch](mailto:info@iec.ch)  
[www.iec.ch](http://www.iec.ch)

### About the IEC

The International Electrotechnical Commission (IEC) is the leading global organization that prepares and publishes International Standards for all electrical, electronic and related technologies.

### About IEC publications

The technical content of IEC publications is kept under constant review by the IEC. Please make sure that you have the latest edition, a corrigenda or an amendment might have been published.

#### Useful links:

IEC publications search - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

The advanced search enables you to find IEC publications by a variety of criteria (reference number, text, technical committee,...).

It also gives information on projects, replaced and withdrawn publications.

IEC Just Published - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Stay up to date on all new IEC publications. Just Published details all new publications released. Available on-line and also once a month by email.

Electropedia - [www.electropedia.org](http://www.electropedia.org)

The world's leading online dictionary of electronic and electrical terms containing more than 30 000 terms and definitions in English and French, with equivalent terms in additional languages. Also known as the International Electrotechnical Vocabulary (IEV) on-line.

Customer Service Centre - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

If you wish to give us your feedback on this publication or need further assistance, please contact the Customer Service Centre: [csc@iec.ch](mailto:csc@iec.ch).

---

### A propos de la CEI

La Commission Electrotechnique Internationale (CEI) est la première organisation mondiale qui élabore et publie des Normes internationales pour tout ce qui a trait à l'électricité, à l'électronique et aux technologies apparentées.

### A propos des publications CEI

Le contenu technique des publications de la CEI est constamment revu. Veuillez vous assurer que vous possédez l'édition la plus récente, un corrigendum ou amendement peut avoir été publié.

#### Liens utiles:

Recherche de publications CEI - [www.iec.ch/searchpub](http://www.iec.ch/searchpub)

La recherche avancée vous permet de trouver des publications CEI en utilisant différents critères (numéro de référence, texte, comité d'études,...).

Elle donne aussi des informations sur les projets et les publications remplacées ou retirées.

Just Published CEI - [webstore.iec.ch/justpublished](http://webstore.iec.ch/justpublished)

Restez informé sur les nouvelles publications de la CEI. Just Published détaille les nouvelles publications parues. Disponible en ligne et aussi une fois par mois par email.

Electropedia - [www.electropedia.org](http://www.electropedia.org)

Le premier dictionnaire en ligne au monde de termes électroniques et électriques. Il contient plus de 30 000 termes et définitions en anglais et en français, ainsi que les termes équivalents dans les langues additionnelles. Egalement appelé Vocabulaire Electrotechnique International (VEI) en ligne.

Service Clients - [webstore.iec.ch/csc](http://webstore.iec.ch/csc)

Si vous désirez nous donner des commentaires sur cette publication ou si vous avez des questions contactez-nous: [csc@iec.ch](mailto:csc@iec.ch).

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE



**Medical electrical equipment –  
Part 2-44: Particular requirements for the basic safety and essential performance  
of X-ray equipment for computed tomography**

**Appareils électromédicaux –  
Partie 2-44: Exigences particulières pour la sécurité de base et les performances  
essentielles des équipements à rayonnement X de tomographie**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

ICS 11.040.50

ISBN 978-2-8322-0346-0

**Warning! Make sure that you obtained this publication from an authorized distributor.  
Attention! Veuillez vous assurer que vous avez obtenu cette publication via un distributeur agréé.**

## CONTENTS

FOREWORD.....	3
<b>Introduction to Amendment 1 .....</b>	<b>5</b>
201.1 Scope, object and related standards.....	6
201.2 Normative references .....	8
201.3 Terms and definitions .....	9
201.4 General requirements.....	17
201.5 General requirements for testing of ME EQUIPMENT.....	18
201.6 Classification of ME EQUIPMENT and ME SYSTEMS .....	18
201.7 ME EQUIPMENT identification, marking and documents.....	19
201.8 Protection against electrical HAZARDS from ME EQUIPMENT.....	21
201.9 Protection against mechanical HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	24
201.10 Protection against unwanted and excessive RADIATION HAZARDS .....	27
201.11 Protection against excessive temperatures and other HAZARDS.....	28
201.12 Accuracy of controls and instruments and protection against hazardous outputs.....	28
201.13 Hazardous situations and fault conditions.....	29
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS) .....	29
201.15 Construction of ME EQUIPMENT .....	29
201.16 ME SYSTEMS.....	29
201.17 ELECTROMAGNETIC COMPATIBILITY of ME EQUIPMENT and ME SYSTEMS.....	29
<b>202 Electromagnetic compatibility – Requirements and tests .....</b>	<b>35</b>
203 General requirements for RADIATION protection in diagnostic X-ray equipment .....	35
Annexes .....	52
Annex AA (informative) Choosing LOADING FACTORS for tests .....	52
Annex BB (informative) Estimating $CTDI_{vol}$ for scan projection RADIOGRAPHY (SPR).....	53
<b>Annex CC (informative) The <math>CTDI_{100}</math> concept in IEC 60601-2-44: Relationship between <math>CTDI_{100}</math> and <math>CTDI_{\infty}</math> .....</b>	<b>54</b>
<b>Annex DD (informative) Measuring <math>CTDI_{free\ air}</math> .....</b>	<b>58</b>
Bibliography.....	60
Index of defined terms used in this particular standard.....	62
Figure 201.101 – Coordinate system.....	11
<b>Figure 201.102 – Illustration of <math>N \times T</math>, <math>R</math> and <math>(N \times T) + R</math> .....</b>	<b>14</b>
<b>Figure 201.103 – Vertical alignment of the PATIENT SUPPORT .....</b>	<b>30</b>
<b>Figure 201.104 – Z-axis alignment of the PATIENT SUPPORT in the horizontal plane .....</b>	<b>32</b>
Figure 203.101 – Zone of extra-focal RADIATION .....	41
Figure 203.102 – Minimum dimensions for STRAY RADIATION measurement.....	43
<b>Figure CC.1 – <math>CTDI_w</math> versus beam width along <math>z</math>.....</b>	<b>55</b>
Table 203.101 – Test pattern for $CTDI_{free\ air}$ .....	48
<b>Table CC.1 – Ratios of <math>CTDI</math> by phantom length .....</b>	<b>57</b>

## INTERNATIONAL ELECTROTECHNICAL COMMISSION

## MEDICAL ELECTRICAL EQUIPMENT –

**Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography**

## FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as “IEC Publication(s)”). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

**This consolidated version of IEC 60601-2-44 consists of the third edition (2009) [documents 62B/727/FDIS and 62B/734/RVD], its amendment 1 (2012) [documents 62B/879/FDIS and 62B/890/RVD] and its corrigendum of May 2010. It bears the edition number 3.1.**

**The technical content is therefore identical to the base edition and its amendment and has been prepared for user convenience. A vertical line in the margin shows where the base publication has been modified by amendment 1. Additions and deletions are displayed in red, with deletions being struck through.**

International standard IEC 60601-2-44 has been prepared by subcommittee 62B: Diagnostic imaging equipment, of IEC technical committee 62: Electrical equipment in medical practice.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

In this standard, the following print types are used:

- Requirements and definitions: roman type.
- *Test specifications: italic type.*
- Informative material appearing outside of tables, such as notes, examples and references: in smaller type. Normative text of tables is also in a smaller type.
- TERMS DEFINED IN CLAUSE 3 OF THE GENERAL STANDARD, IN THIS PARTICULAR STANDARD OR AS NOTED: SMALL CAPITALS.

In referring to the structure of this standard, the term

- “clause” means one of the seventeen numbered divisions within the table of contents, inclusive of all subdivisions (e.g. Clause 7 includes subclauses 7.1, 7.2, etc.);
- “subclause” means a numbered subdivision of a clause (e.g. 7.1, 7.2 and 7.2.1 are all subclauses of Clause 7).

References to clauses within this standard are preceded by the term “Clause” followed by the clause number. References to subclauses within this particular standard are by number only.

In this standard, the conjunctive “or” is used as an “inclusive or” so a statement is true if any combination of the conditions is true.

The verbal forms used in this standard conform to usage described in Annex H of the ISO/IEC Directives, Part 2. For the purposes of this standard, the auxiliary verb:

- “shall” means that compliance with a requirement or a test is mandatory for compliance with this standard;
- “should” means that compliance with a requirement or a test is recommended but is not mandatory for compliance with this standard;
- “may” is used to describe a permissible way to achieve compliance with a requirement or test.

A list of all parts of the IEC 60601 series, published under the general title *Medical electrical equipment*, can be found on the IEC website.

The committee has decided that the contents of the base publication and its amendments will remain unchanged until the stability date indicated on the IEC web site under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

NOTE The attention of National Committees is drawn to the fact that equipment MANUFACTURERS and testing organizations may need a transitional period following publication of a new, amended or revised IEC publication in which to make products in accordance with the new requirements and to equip themselves for conducting new or revised tests. It is the recommendation of the committee that the content of this publication be adopted for implementation nationally not earlier than 3 years from the date of publication.

**IMPORTANT – The “colour inside” logo on the cover page of this publication indicates that it contains colours which are considered to be useful for the correct understanding of its contents. Users should therefore print this publication using a colour printer.**

## Introduction to Amendment 1

The main topic addressed in this amendment is an extended concept of *CTDI* to accommodate CT SCANNERS with very large z-coverage. The other principal subject areas include:

- 1) a dose-check feature associated with a pre-scanning alert if expected values of dose indices exceed user-configurable DOSE NOTIFICATION VALUES or DOSE ALERT VALUES and
- 2) requirements covering the use of CT data in radiotherapy treatment planning (RTP).

The CT dose metric in use has been based on the  $CTDI_{100}$ , i.e. measurement of dose in PHANTOMS and limited integration of scattered radiation, and it is used in many countries' legislation to define "dose reference values" (also called "diagnostic reference levels") for CT examinations. Many people use these indices,  $CTDI_{vol}$  and *DLP*, to derive estimates for effective dose via conversion factors.  $CTDI_{100}$  is also part of CT acceptance and constancy testing. The introduction of a new dose index would change all CT SCANNERS' *CTDI* values. Therefore the intention is to stay with the  $CTDI_{100}$ , i.e. the integration of primary radiation and scatter over 100 mm, but adapt the way of measuring and reporting the dose index to incorporate large collimations and to rate all collimations the same way, i.e. to reflect approximately the same percentage of  $CTDI_{\infty}$  for all collimations.

As defined in the amendment,  $CTDI_{100}$  is to be measured only for collimations up to 40 mm with the current equipment, i.e. the PMMA PHANTOMS and a 100-mm chamber, or other suitable methods that use a RADIATION DETECTOR. For these collimations there is no significant change of the ratio  $CTDI_{100} / CTDI_{\infty}$  according to published data. For larger collimations at the same CT CONDITIONS OF OPERATION, the z-efficiency may be different and must be evaluated in the dose measurement. This can be accomplished by the measurement of dose 'free air'. Based on these considerations  $CTDI_{100}$  and the  $CTDI_{free\ air}$  have been refined. Both types of measurement are combined now to determine the *CTDI* values for larger collimations and they are explained in detail in informative Annexes CC and DD.

Some additional requirements and refinements related to dose have been added:  $CTDI_{vol}$  and *DLP* are defined for a new type of scan mode ('shuttle mode'). In body CT EXAMINATION it is clarified that the  $CTDI_{vol}$  and *DLP* always be reported for the 32-cm diameter PHANTOM. In the amendment it is now required that CT SCANNERS support user-configurable DOSE NOTIFICATION VALUES and DOSE ALERT VALUES.

A new subject area in this Amendment 1 covers requirements for CT SCANNERS providing images for radiotherapy treatment planning. With this amendment begins the implementation of this important CT application into the CT safety standard with a set of requirements that is considered to be safety relevant. It mainly covers scanner hardware adjustments, accuracy of CT image data, and the conversion of HU to electron and mass density.

## MEDICAL ELECTRICAL EQUIPMENT –

### Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography

#### 201.1 Scope, object and related standards

Clause 1 of the general standard<sup>1)</sup> applies, except as follows:

##### 201.1.1 Scope

*Replacement:*

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of CT SCANNERS, hereafter also referred to as ME EQUIPMENT.

If a clause or subclause is specifically intended to be applicable to ME EQUIPMENT only, or to ME SYSTEMS only, the title and content of that clause or subclause will say so. If that is not the case, the clause or subclause applies both to ME EQUIPMENT and to ME SYSTEMS, as relevant.

NOTE 1 See also 4.2 of the general standard.

The scope of this document is limited to CT SCANNERS intended to be used for both head and body characterised by an ENCLOSURE of the X-ray source(s) and imaging detector(s) in a common protective cover in the shape of a toroid. It includes safety requirements for the X-RAY GENERATORS used in CT SCANNERS, including those where HIGH-VOLTAGE GENERATORS are integrated with an X-RAY TUBE ASSEMBLY.

NOTE 2 Requirements for X-RAY GENERATORS and for ASSOCIATED EQUIPMENT, which were previously specified in IEC 60601-2-7 and IEC 60601-2-32, have been included in either IEC 60601-1:2005 (Ed3) or this edition of IEC 60601-2-44. Therefore IEC 60601-2-7 and IEC 60601-2-32 are not part of the 3<sup>rd</sup> edition scheme for COMPUTED TOMOGRAPHY.

**The scope of this International Standard excludes RADIOTHERAPY SIMULATORS and systems where the image is created by a source other than an X-RAY TUBE.**

##### 201.1.2 Object

*Replacement:*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for CT SCANNERS as defined in 201.3.201, to ensure safety, and to specify methods for demonstrating compliance with those requirements, for CT SCANNERS.

NOTE 1 Requirements for reproducibility, linearity, constancy and accuracy are given because of their relationship to the quality and quantity of the IONIZING RADIATION produced and are confined to those considered necessary for safety.

NOTE 2 Both the levels for compliance and the tests prescribed to determine compliance reflect the fact that the safety of HIGH-VOLTAGE GENERATORS is not sensitive to small differences in levels of performance. The combinations of LOADING FACTORS specified for the tests are therefore limited in number but chosen from experience as being appropriate in most cases. It is considered important to standardize the choice of combinations of LOADING FACTORS so that comparison can be made between tests performed in different places on different occasions. However, combinations other than those specified could be of equal technical validity.

<sup>1)</sup> IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance.*

NOTE 3 The safety philosophy on which this standard is based is described in the introduction to the general standard and in IEC TR 60513.

NOTE 4 Concerning RADIOLOGICAL PROTECTION, it is assumed that MANUFACTURERS and RESPONSIBLE ORGANIZATIONS accept the general principles of justification, optimisation, and application of dose limits of the International Commission on Radiological Protection as stated in ICRP 103, 2007, paragraph 203, [12]<sup>2)</sup> namely:

(a) "The principle of justification: Any decision that alters the RADIATION exposure situation should do more good than harm."

(b) "The principle of optimisation of protection: The likelihood of incurring exposures, the number of people exposed, and the magnitude of their individual doses should all be kept as low as reasonably achievable, taking into account economic and societal factors."

(c) "The principle of application of dose limits: The total dose to any individual from regulated sources in planned exposure situations other than medical exposure of PATIENTS should not exceed the appropriate limits recommended by the Commission."

(d) "Application of dose limits for the PATIENT dose might be to the PATIENT'S detriment. Therefore dose limits should not be applied to medical exposures. However, considerations should be given to the use of dose constraints or investigation levels for some common diagnostic procedures. This concept, now renamed as diagnostic reference levels, has been introduced in a large number of countries."

NOTE 5 It is recognized that many of the judgements necessary to follow the ICRP general principles have to be made by the RESPONSIBLE ORGANIZATIONS and not by the MANUFACTURER of the ME EQUIPMENT.

### 201.1.3 Collateral standards

*Addition:*

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

~~IEC 60601-1-3 applies as modified in Clause 203. IEC 60601-1-8, IEC 60601-1-9 and IEC 60601-1-10<sup>3)</sup> do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.~~

IEC 60601-1-2 and IEC 60601-1-3 apply as modified in Clauses 202 and 203. IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10<sup>4)</sup>, IEC 60601-1-11<sup>5)</sup> and IEC 60601-1-12<sup>6)</sup> do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

For collateral standards published after this particular standard, MANUFACTURERS need to determine the applicability in accordance with the RISK MANAGEMENT PROCESS.

2) Figures in square brackets refer to the Bibliography.

~~3) IEC 60601-1-10, Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers~~

~~4) IEC 60601-1-10, Medical electrical equipment – Part 1-10: General requirements for basic safety and essential performance – Collateral Standard: Requirements for the development of physiologic closed-loop controllers~~

~~5) IEC 60601-1-11, Medical electrical equipment – Part 1-11: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems used in the home healthcare environment~~

~~6) IEC 60601-1-12, Medical electrical equipment – Part 1-12: General requirements for basic safety and essential performance – Collateral Standard: Requirements for medical electrical equipment and medical electrical systems intended to be used in the emergency medical services environment~~

#### **201.1.4 Particular standards**

##### *Replacement:*

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document numbers.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x" where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding section, clause or subclause in this particular standard, the section, clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any parts of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

#### **201.2 Normative references**

NOTE Informative references are listed in the bibliography beginning on page 60.

Clause 2 of the general standard applies, except as follows:

*Replacement:*

*IEC 60601-1-2:2007, Medical electrical equipment – Part 1-2: General requirements for basic safety and essential performance – Collateral standard: Electromagnetic compatibility – Requirements and tests*

*IEC 60601-1-3:2008, Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment*

*Addition:*

*IEC 60336 Medical electric equipment – X-Ray Tube assemblies for medical diagnosis – Characteristics of focal spots*

*IEC 60601-1:2005, Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

*IEC 61223-3-5, Evaluation and routine testing in medical imaging departments – Part 3-5: Acceptance tests – Imaging performance of computed tomography X-ray equipment*

*ISO 12052, Health informatics – Digital imaging and communication in medicine (DICOM) including workflow and data management*

### **201.3 Terms and definitions**

For the purposes of this document, the terms and definitions given in IEC 60601-1:2005, IEC 60601-1-3:2008 and IEC 60788:2004 apply, except as follows:

NOTE 101 An index of defined terms is to be found at the end of this document.

NOTE 102 In accordance with the definitions in IEC 60601-1-3, in this standard unless otherwise indicated:

- values of X-RAY TUBE VOLTAGE refer to peak values, transients being disregarded;
- values of X-RAY TUBE CURRENT refer to average values.

*Addition:*

#### **201.3.201**

##### **CT SCANNER**

X-RAY EQUIPMENT intended to generate cross-sectional images of the body by computer reconstruction of X-ray transmission data obtained at different angles, which may include signal analysis and display equipment, PATIENT SUPPORT, support parts and ACCESSORIES

NOTE 1 The scope of this document is limited to CT SCANNERS intended to be used for both head and body characterised by an ENCLOSURE of the X-ray source(s) and imaging detector(s) in a common protective cover in the shape of a toroid.

NOTE 2 Secondary imaging processing is not included in the scope of this standard.

#### **201.3.202**

##### **CT CONDITIONS OF OPERATION**

all selectable parameters governing the operation of a CT SCANNER

NOTE 1 Examples of such conditions include NOMINAL TOMOGRAPHIC SECTION THICKNESS, CT PITCH FACTOR, FILTRATION, peak X-RAY TUBE VOLTAGE and either X-RAY TUBE CURRENT and LOADING TIME or CURRENT TIME PRODUCT.

NOTE 2 Some CT CONDITIONS OF OPERATION may vary during the exposure.