

**MÕÕTMISMEETODITE JA TULEMUSTE
MÕÕTETÄPSUS (TÕELINE VÄÄRTUS JA TÄPSUS)
Osa 6: Täpsusväärtuste kasutus praktikas**

Accuracy (trueness and precision) of measurement
methods and results

Part 6: Use in practice of accuracy values

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

Käesolev Eesti standard EVS-ISO 5725-6:2002 "Mõõtmismeetodite ja tulemuste mõõtetäpsus (tõeline väärtus ja täpsus). Osa 6: Täpsusväärtuste kasutus praktikas" sisaldab rahvusvahelise standardi ISO 5725-6:1994 "Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values" identset ingliskeelset teksti.

Standard EVS-ISO 5725-6:2002 on kinnitatud Eesti Standardikeskuse 03.05.2002 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Standard on kättesaadav Eesti Standardikeskusest.

This Estonian Standard EVS-ISO 5725-6:2002 consists of the identical English text of the International Standard ISO 5725-6:1994 "Accuracy (trueness and precision) of measurement methods and results - Part 6: Use in practice of accuracy values".

This standard is ratified with the order of Estonian Centre for Standardisation dated 03.05.2002 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

The standard is available from Estonian Centre for Standardisation.

Käsitlusala

1.1 Käesoleva ISO 5725 osa eesmärk on anda juhised, millistel viisidel saab täpsusandmeid kasutada erinevates praktilistes olukordades:

a) esitab korduste piirangu, reprodutseeritavuse ja teiste limiitide arvutamise standardmeetodit, mida kasutatakse standardmõõtmise meetodi abil saadud katsetulemuste uurimisel;

b) tagab, et on võimalik kontrollida korratavuse või reprodutseeritavuse katsetel saadud tulemuste aktsepteeritavust;

c) kirjeldab, kuidas hinnata tulemuste stabiilsust laboris pikema aja vältel, mis omakorda võimaldab "kvaliteedikontrolli" selles laboris;

d) kirjeldab, kuidas hinnata, kas teatud labor on võimeline kasutama teatud standardmõõtemetodit rahuldaval viisil;

e) kirjeldab, kuidas võrrelda alternatiivseid mõõtmismeetodeid.

1.2 Käesolev ISO 5725 osa puutub ainult mõõtemetoditesse, mis annavad mõõtmisi pideval skaalal ning annavad katse tulemuseks ühe numbrilise väärtuse, kuigi see võib olla mitmete vaatluste põhjal tehtud arvutuse tulemus.

1.3 Eeldatakse, et tõesuse ja kordustäpsuse hinnangud käesoleva meetodi kohta on omandatud vastavuses ISO 5725 osadega 1 kuni 5.

1.4 Kogu lisateave selle rakenduse valdkonnast antakse iga üksiku rakenduse alguses.

Scope

1.1 The purpose of this part of ISO 5725 is to give some indications of the way in which accuracy data can be used in various practical situations by:

a) giving a standard method of calculating the repeatability limit, the reproducibility limit and other limits to be used in examining the test results obtained by a standard measurement method;

b) providing a way of checking the acceptability of test results obtained under repeatability or reproducibility conditions;

c) describing how to assess the stability of results within a laboratory over a period of time, and thus providing a method of "quality control" of the operations within that laboratory;

d) describing how to assess whether a given laboratory is able to use a given standard measurement method in a satisfactory way;

e) describing how to compare alternative measurement methods.

1.2 This part of ISO 5725 is concerned exclusively with measurement methods which yield measurements on a continuous scale and give a single numerical figure as the result, although this single figure may be the outcome of a calculation from a set of observations.

1.3 It is assumed that the estimates of trueness and precision for the method have been obtained in accordance with parts 1 to 5 of ISO 5725.

1.4 Any additional information regarding the field of application will be given at the beginning of each particular application.

This document is a preview generated by EVS

ICS 03.120.30 Statistiliste meetodite rakendamine

Standardite reprodutseerimis- ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

This document is a preview generated by EVS

Contents

	Page
1 Scope	1
2 Normative references	1
3 Definitions	2
4 Determination of limits	2
4.1 Repeatability and reproducibility limits	2
4.2 Comparisons based on more than two values	3
5 Methods for checking the acceptability of test results and determining the final quoted result	4
5.1 General	4
5.2 Methods for checking the acceptability of test results obtained under repeatability conditions	4
5.3 Methods for checking the acceptability of test results obtained under both repeatability and reproducibility conditions	11
6 Method for checking the stability of test results within a laboratory	13
6.1 Background	13
6.2 Methods for checking stability	13
7 Use of repeatability and reproducibility standard deviations in assessing laboratories	25
7.1 Assessment method	25
7.2 Evaluation of the use of a measurement method by a laboratory not previously assessed	26
7.3 Continued assessment of previously approved laboratories	29
8 Comparison of alternative measurement methods	32
8.1 Origin of alternative measurement methods	32
8.2 Purpose of comparing measurement methods	32
8.3 Method B is a candidate for an alternative standard method ("Standardization experiment" not defined)	33

© ISO 1994

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

8.4	Accuracy experiment	33
8.5	Method B is a candidate for a routine method	38
Annex		
A	Symbols and abbreviations used in ISO 5725	40

This document is a preview generated by EVS

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 5725-6 was prepared by Technical Committee ISO/TC 69, *Applications of statistical methods*, Subcommittee SC 6, *Measurement methods and results*.

ISO 5725 consists of the following parts, under the general title *Accuracy (trueness and precision) of measurement methods and results*:

- *Part 1: General principles and definitions*
- *Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- *Part 3: Intermediate measures of the precision of a standard measurement method*
- *Part 4: Basic methods for the determination of the trueness of a standard measurement method*
- *Part 5: Alternative methods for the determination of the precision of a standard measurement method*
- *Part 6: Use in practice of accuracy values*

Parts 1 to 6 of ISO 5725 together cancel and replace ISO 5725:1986, which has been extended to cover trueness (in addition to precision) and intermediate precision conditions (in addition to repeatability and reproducibility conditions).

Annex A forms an integral part of this part of ISO 5725.

Introduction

0.1 ISO 5725 uses two terms "trueness" and "precision" to describe the accuracy of a measurement method. "Trueness" refers to the closeness of agreement between the arithmetic mean of a large number of test results and the true or accepted reference value. "Precision" refers to the closeness of agreement between test results.

0.2 The need to consider "precision" arises because tests performed on presumably identical materials in presumably identical circumstances do not, in general, yield identical results. This is attributed to unavoidable random errors inherent in every measurement procedure; the factors that influence the outcome of a measurement cannot all be completely controlled. In the practical interpretation of measurement data, this variability has to be taken into account. For instance, the difference between a test result and some specified value may be within the scope of unavoidable random errors, in which case a real deviation from such a specified value has not been established. Similarly, comparing test results from two batches of material will not indicate a fundamental quality difference if the difference between them can be attributed to the inherent variation in the measurement procedure.

0.3 Parts 1 to 5 of ISO 5725 discuss the background to, and given methods for, the assessment of the precision (in terms of the repeatability standard deviation and the reproducibility standard deviation) and the trueness (in terms of the various components of bias) of measurements produced by a standard measurement method. Such assessment would, however, be pointless if there were no practical uses to which the results could be put.

0.4 Given that the accuracy of a measurement method has been established, this part of ISO 5725 applies that knowledge in practical situations in such a way as to facilitate commercial transactions and to monitor and improve the operational performance of laboratories.

Accuracy (trueness and precision) of measurement methods and results —

Part 6:

Use in practice of accuracy values

1 Scope

1.1 The purpose of this part of ISO 5725 is to give some indications of the way in which accuracy data can be used in various practical situations by:

- a) giving a standard method of calculating the repeatability limit, the reproducibility limit and other limits to be used in examining the test results obtained by a standard measurement method;
- b) providing a way of checking the acceptability of test results obtained under repeatability or reproducibility conditions;
- c) describing how to assess the stability of results within a laboratory over a period of time, and thus providing a method of "quality control" of the operations within that laboratory;
- d) describing how to assess whether a given laboratory is able to use a given standard measurement method in a satisfactory way;
- e) describing how to compare alternative measurement methods.

1.2 This part of ISO 5725 is concerned exclusively with measurement methods which yield measurements on a continuous scale and give a single numerical figure as the result, although this single figure may be the outcome of a calculation from a set of observations.

1.3 It is assumed that the estimates of trueness and precision for the method have been obtained in accordance with parts 1 to 5 of ISO 5725.

1.4 Any additional information regarding the field of application will be given at the beginning of each particular application.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 5725. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 5725 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 3534-1:1993, *Statistics — Vocabulary and symbols — Part 1: Probability and general statistical terms*.

ISO 5725-1:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*.

ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*.

ISO 5725-3:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 3: Intermediate measures of the precision of a standard measurement method*.

ISO 5725-4:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 4: Basic methods for the determination of the trueness of a standard measurement method*.

ISO 8258:1991, *Shewhart control charts*.

ISO Guide 33:1989, *Uses of certified reference materials*.

ISO Guide 35:1989, *Certification of reference materials — General and statistical principles*.

ISO/IEC Guide 25:1990, *General requirements for the competence of calibration and testing laboratories*.

3 Definitions

For the purposes of this part of ISO 5725, the definitions given in ISO 3534-1 and ISO 5725-1 apply.

The symbols used in ISO 5725 are given in annex A.

4 Determination of limits

4.1 Repeatability and reproducibility limits

4.1.1 In ISO 5725-2, attention has been focussed on estimating the standard deviations associated with operations under repeatability or reproducibility conditions. However, normal laboratory practice requires examination of the difference(s) observed between two (or more) test results, and for this purpose some measure akin to a critical difference is required, rather than a standard deviation.

4.1.2 When a quantity is based on sums or differences of n independent estimates each having a standard deviation σ , then that resultant quantity will have a standard deviation $\sigma\sqrt{n}$. The reproducibility limit (R) or repeatability limit (r) are for differences between two test results, so the associated standard deviation is $\sigma\sqrt{2}$. In normal statistical practice, for examining the difference between these two values the critical difference used is f times this standard deviation, i.e. $f\sigma\sqrt{2}$. The value of f (the critical range factor) depends on the probability level to be associated with the critical difference and on the shape of the underlying distribution. For the reproducibility and repeatability limits, the probability level is specified as 95 %, and throughout the analysis in ISO 5725 the assumption is made that the underlying distribution is approximately normal. For a normal distribution at 95 % probability level, f is 1,96 and $f\sqrt{2}$ then is 2,77. As the purpose of this part of ISO 5725 is to give some simple "rule of thumb" to be applied by non-statisticians when examining the results of tests, it seems reasonable to use a rounded value of 2,8 instead of $f\sqrt{2}$.

4.1.3 As has been stated, the process of estimating precision leads to estimates of the true standard deviations while the true standard deviations remain unknown. Therefore in statistical practice they should be denoted by s rather than σ . However, if the procedures given in ISO 5725-1 and ISO 5725-2 are followed, these estimates will be based on an appreciable number of test results, and will give the best information we are likely to have of the true values of the standard deviations. In other applications that follow, for estimates of these standard deviations based on more limited data, the symbol s (estimate of a standard deviation) is used. Therefore it seems best to use the symbol σ to denote the values obtained from a full precision experiment, and treat these as true standard deviations with which other estimates (s) will be compared.

4.1.4 In view of 4.1.1 to 4.1.3, when examining two single test results obtained under repeatability or reproducibility conditions, the comparison shall be made with the repeatability limit

$$r = 2,8\sigma_r$$

or the reproducibility limit

$$R = 2,8\sigma_R$$