

PAIKSETE SAASTEALLIKATE HEITED
Vääveldioksiidi massikontsentratsiooni määramine
Automaatmõõtemetodite suutlikkusnäitajad

Stationary source emissions

Determination of the mass concentration of sulfur dioxide

Performance characteristics of automated methods

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-ISO 7935:2006 "Paiksete saasteallikate heited. Vääveldioksiidi massikontsentratsiooni määramine. Automaatmõõtemetodite suutlikkusnäitajad" sisaldab rahvusvahelise standardi ISO 7935:1992 "Stationary source emissions - Determination of the mass concentration of sulfur dioxide - Performance characteristics of automated methods" identset ingliskeelset teksti.</p>	<p>This Estonian Standard EVS-ISO 7935:2006 consists of the identical English text of the International Standard ISO 7935:1992 "Stationary source emissions - Determination of the mass concentration of sulfur dioxide - Performance characteristics of automated methods".</p>
<p>Standardi avaldamise korraldas Eesti Standardikeskus.</p>	<p>Estonian standard is published by the Estonian Centre for Standardisation.</p>
<p>Standard EVS-ISO 7935:2006 on kinnitatud Eesti Standardikeskuse 22.12.2006 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teataja 2007. aasta jaanuarikuu numbris.</p>	<p>This standard is ratified with the order of Estonian Centre for Standardisation dated 22.12.2006 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.</p>
<p>Standard on kättesaadav Eesti Standardikeskusest.</p>	<p>The standard is available from Estonian Centre for Standardisation.</p>

Käsitlusala

Käesolev standard kehtestab paiksete allikate heidetes vääveldioksiidi massikontsentratsiooni pidevaks mõõtmiseks mõeldud automaatmõõtesüsteemide (AMS) suutlikkusnäitajate kõik väärtused.

Märkus 1. Kui AMS-i suutlikkusnäitajad on loetletud vastavalt tabelile 1, tagab see AMS-i usaldusväärsuse ja rahuldavad pidevmõõtmise tulemused.

Tabelis 1 loetletud andmed on mõõtemetodi suutlikkusnäitajad, mis hõlmavad kõiki etappe proovivõtust andmete registreerimise ja vajadusel säilitamiseni.

Standardit kohaldatakse gaasi väljavõtuga (ekstraktiivsete) ja väljavõtuta (*in situ*) automaatsetele vääveldioksiidi mõõtemetoditele. Mõlema meetoditüübi puhul eeldab standard null- ja kalibreerimisgaasi kasutamise võimalust ning võrreldavate proovide olemasolu. AMS-i saab kalibreerida kalibreerimisgaaside, standardis ISO 7934 kirjeldatud käsitsi teostatava meetodi või teisel määramispõhimõttel töötava, selle standardi kohaselt taadeldud AMS-i abil. Üldsuutlikkus (vt jaotis 3.7) määratakse standardi ISO 7934 põhjal või teisel määramispõhimõttel töötava, selle standardi kohaselt taadeldud AMS-i abil. Praegu kehtivad need näitajad vahemikus 0 g/m³ kuni 0,1 g/m³ ja 0 g/m³ 8 g/m³ (täpsemalt vt tabel 2).

Märkus 2. Ehkki täpseid katseandmeid anda ei saa, kehtivad nõuded ja katsepõhimõtted ka *in situ* süsteemidele.

ICS 13.040.40 Püsiallikate heitmed

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse poolt antud kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

Right to reproduce and distribute belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without permission in writing from Estonian Centre for Standardisation.

If you have any questions about standards copyright, please contact Estonian Centre for Standardisation:
Aru str 10 Tallinn 10317 Estonia; www.evs.ee; Phone: 605 5050; E-mail: info@evs.ee

Contents

	Page
1 Scope	1
2 Normative references	2
3 Definitions	2
4 Description of the automated measuring systems	3
5 Numerical values of performance characteristics and their applicability	4
 Annexes	
A Determination of main performance characteristics of automated measuring systems	6
B Additional performance characteristics	10
C Bibliography	11

This document is a preview generated by EVS

© ISO 1992

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 7935 was prepared by Technical Committee ISO/TC 146 *Air quality*, Sub-Committee SC 1, *Stationary source emissions*.

Annex A forms an integral part of this International Standard. Annexes B and C are for information only.

This document is a preview generated by EVS

Introduction

Sulfur dioxide can arise in considerable quantities from combustion of fossil fuels used for energy generation, industrial activities processing sulfur or sulfur containing material, and from combustion of sulfur containing waste. The waste gas from these processes, containing sulfur dioxide, is usually discharged into the ambient atmosphere, via a duct or a chimney.

For evaluating the mass concentration of sulfur dioxide present in the waste gas of stationary source emissions, a number of highly developed methods of integrated sampling and subsequent determination by chemical analysis and automated measuring systems are available. Considerable experience exists on their application under plant conditions. One of these methods is standardized as ISO 7934.

ISO 7934 is used for example in comparative measurements, where the automated measuring methods are involved. The automated technique is capable of continuous measurement of the mass concentration of sulfur dioxide.

For methods where performance characteristics are given, the values of performance characteristics are used to decide whether a method is suitable for a given measuring task (see ISO 6879:1983, clause 1). Values of the main performance characteristics of automated measuring systems, capable of determining the mass concentration of sulfur dioxide present in waste gas stationary emission sources, are given in clause 5.

Additional performance characteristics are given in informative annex B.

The procedure for evaluating the values of the performance characteristics listed in clause 5, is described in normative annex A.

Stationary source emissions — Determination of the mass concentration of sulfur dioxide — Performance characteristics of automated measuring methods

1 Scope

This International Standard specifies a complete set of values of performance characteristics for automated measuring systems for the continuous measurement of the mass concentrations of sulfur dioxide in stationary source emissions.

NOTE 1 If the performance characteristics of an automated measuring system are listed according to table 1, this ensures that the automated measuring system is reliable and gives satisfactory continuous results.

The set of data listed in table 1 refers to the performance characteristics of measurement methods, including all steps from sampling to recording and, if necessary, storage of data.

This International Standard is applicable to extractive and non-extractive automated sulfur dioxide measuring methods. For both methods it implies the applicability of zero and calibration gas and the availability of comparable samples. The automated measuring system can be calibrated with calibration gases, by applying the manual method described in ISO 7934, or by applying an automated measuring system previously verified according to this International Standard using a different principle of detection. The value of the integral performance (3.7) is determined by using ISO 7934 or an automated measuring system verified according to this International Standard with a different principle of detection. At present, the range over which this specification applies is between 0 g/m^3 to $0,1 \text{ g/m}^3$ and 0 g/m^3 to 8 g/m^3 (see table 2 for details).

NOTE 2 Although it is impossible to give precise testing details, the requirements and testing principles are also applicable to non-extractive systems.

Table 1 — Main performance characteristics

Performance characteristics	Numerical value	Test methods (see annex A)
Detection limit	2 % ¹⁾	A.4.2.1.1
Effect of interfering substances	$\pm 2 \%$ ^{1) 2)}	A.4.2.1.2
Response time	$\leq 200 \text{ s}$ ³⁾	A.4.2.1.3
Integral performance (s_A)	$\pm 2,5 \%$ ^{1) 4)}	A.4.2.2

1) Related to the upper limit of measurement.
 2) The main interfering substances in the flue gas from combustion plants are CO_2 , CO , NO , H_2O and, in smaller concentrations, NO_2 and NH_3 . If the water vapour is not removed from the flue gas of coal and waste fired incinerators, HCl and H_2S may also interfere. In special cases there may be other interfering substances (e.g. cyanide).
 3) Assuming an integration time of 30 min.
 4) See 3.7.

The facilities at which the values of the performance characteristics given in table 1 have been verified according to this International Standard in the appropriate ranges are listed in table 2.

Table 2 — Facilities and measuring ranges

Facility	Measuring range g/m ³ of SO ₂ 1)
Furnaces for hard coal	0 - 1 to 0 - 8
Furnaces for hard coal with stack gas desulfuration plant	0 to 0,1
Furnaces for brown coal	0 - 0,1 to 0 - 3,0
Furnaces for heavy fuel oil	0 - 0,1 to 0 - 5,0
Refuse incinerator	0 - 0,4 to 0 - 1,0
Coke oven	0 to 1
Calcar with heavy fuel oil	0 to 5
Sulfuric acid recovery plant	0 to 1

1) Related to 101,3 kPa, 273 K and dry gas.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 6879:1983, *Air quality — Performance characteristics and related concepts for air quality measuring methods*.

ISO 7934:1989, *Stationary source emissions — Determination of the mass concentration of sulfur dioxide — Hydrogen peroxide/barium perchlorate/Thorin method*.

3 Definitions

For the purposes of this International Standard, the following definitions apply.

3.1 automated measuring system (AMS): A complete system that may be attached to a chimney to continuously measure and record the mass concentration of sulfur dioxide passing through the chimney.

3.2 analyser: Analytical part in an extractive AMS.

3.3 verified AMS: AMS previously verified in ISO 7935.

3.4 calibration gas: A gas of known and reliable composition that may be used to check the response of an AMS.

3.5 comparative measurements: Measurements that are performed in the same chimney in the same sampling plane for the same period of time.

3.6 manual method: The test method defined in ISO 7934 for the manual sampling and analysis of stationary source emissions containing sulfur dioxide.

3.7 integral performance, s_A : The integral performance is a measure of the working accuracy of the AMS. It is calculated according to the formula for standard deviations.

The integral performance is derived from the difference in the pairs of measured values of sulfur dioxide by the AMS under investigation, and by an ISO manual method or a verified AMS of different measuring principle on the basis of a sufficient number of comparative measurements spread over the period of unattended operation (see annex A).

NOTES

3 It is not possible to determine the standard deviation of an AMS under repeatable working conditions because

— commercially available calibration gas mixtures containing sulfur dioxide do not have all the properties of actual waste gas and do not cover all possible influences;

— the mass concentration of sulfur dioxide in waste gas usually varies with time;

— it is not possible to maintain the properties of a waste gas present in the waste gas flue when it is transferred into a vessel.

4 The reason that the integral performance is defined as a measure of the working accuracy, is that it contains, in addition to random errors, all the effects of interfering substances, changes in temperature and power line as well as zero drifts and span drifts. It also includes the standard deviation of the ISO manual method or the verified AMS using a different principle of detection, which can be determined separately and eliminated if necessary. Furthermore, it includes the effects, for the different methods, of a different response time to variations in the composition of the waste gas.

The integral performance defined in this subclause is an upper limiting value for the AMS. Relevant systematic errors of the measured values of the ISO manual method, or the verified AMS using a different principle of detection, have to be known and taken into account.

3.8 chimney: Stack or final exit duct on a stationary process used for the dispersion of residual process gases.

3.9 mass concentration: The concentration of a substance in an emission, expressed in milligrams per cubic metre or grams per cubic metre.