

This document is a preview generated by EVS

Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion Chromatography - Method for water and waste water

Water quality - Determination of dissolved Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba²⁺ using ion Chromatography - Method for water and waste water

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 14911:2000 sisaldab Euroopa standardi EN ISO 14911:1999 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 11.01.2000 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 14911:2000 consists of the English text of the European standard EN ISO 14911:1999.</p> <p>This document is endorsed on 11.01.2000 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p>Käsitlusala: This standard specifies a method for the determination of the dissolved cations Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba⁺ in water (e.g. drinking water, surface water, waste water). An appropriate pretreatment of the sample (e.g. dilution) and the application of a conductivity detector (CD) make the working ranges given in table 1 feasible. The applicability of the method for waste water samples should be proved in each case.</p>	<p>Scope: This standard specifies a method for the determination of the dissolved cations Li⁺, Na⁺, NH₄⁺, K⁺, Mn²⁺, Ca²⁺, Mg²⁺, Sr²⁺ and Ba⁺ in water (e.g. drinking water, surface water, waste water). An appropriate pretreatment of the sample (e.g. dilution) and the application of a conductivity detector (CD) make the working ranges given in table 1 feasible. The applicability of the method for waste water samples should be proved in each case.</p>
---	---

ICS 13.060.01

Võtmesõnad: calcium, cations, chemical analysis, determination of content, lithium, manganese, quality, water, water pollution, water tests

ICS 13.060.01

English version

Water quality

Determination of dissolved Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} ,
 Ca^{2+} , Mg^{2+} , Sr^{2+} and Ba^{2+} using ion chromatography –
Method for water and waste water
(ISO 14911 : 1998)

Qualité de l'eau – Dosage, par chromatographie ionique, des ions Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} , Ca^{2+} , Mg^{2+} , Sr^{2+} et Ba^{2+} dissous – Méthode applicable pour l'eau et les eaux résiduaires (ISO 14911 : 1998)

Wasserbeschaffenheit – Bestimmung der gelösten Kationen Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} , Ca^{2+} , Mg^{2+} , Sr^{2+} und Ba^{2+} mittels Ionenchromatographie – Verfahren für Wasser und Abwasser (ISO 14911 : 1998)

This European Standard was approved by CEN on 1999-07-07.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Foreword

International Standard

ISO 14911 : 1998 Water quality – Determination of dissolved Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} , Ca^{2+} , Mg^{2+} , Sr^{2+} and Ba^{2+} using ion chromatography – Method for water and waste water,

which was prepared by ISO/TC 147 'Water quality' of the International Organization for Standardization, has been adopted by Technical Committee CEN/TC 230 'Water analysis', the Secretariat of which is held by DIN, as a European Standard.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, and conflicting national standards withdrawn, by February 2000 at the latest.

In accordance with the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard:

Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 14911 : 1998 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to international publications are listed in Annex ZA (normative).

Introduction

The essential minimum requirements of an ion chromatographic system applied within the scope of this International Standard are given in clause 5.

The diversity of the appropriate and suitable assemblies and the procedural steps depending on them permit a general description only.

Further information on the analytical technique is given in the normative references (see clause 2) and the bibliography.

1 Scope

This International Standard specifies a method for the determination of the dissolved cations Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} , Ca^{2+} , Mg^{2+} , Sr^{2+} and Ba^{2+} in water (e.g. drinking water, surface water, waste water).

An appropriate pretreatment of the sample (e.g. dilution) and the application of a conductivity detector (CD) make the working ranges given in table 1 feasible.

The applicability of the method for waste water samples should be proved in each case.

Table 1 — Working ranges of the analytical method

Cation	Typical working range with 10 μl loop
	mg/l ¹⁾
Lithium	0,01 to 1
Sodium	0,1 to 10
Ammonium	0,1 to 10
Potassium	0,1 to 10
Manganese	0,5 to 50
Calcium	0,5 to 50
Magnesium	0,5 to 50
Strontium	0,5 to 50
Barium	1 to 100

1) The working range is limited by the ion-exchange capacity of the separator column. If necessary, the sample shall be diluted to meet the working range, or use a 100 μl loop for lower working ranges.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 5667-1:1980, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes.*

ISO 5667-2:1991, *Water quality — Sampling — Part 2: Guidance on sampling techniques.*

ISO 5667-3:1994, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples.*

ISO 6058:1984, *Water quality — Determination of calcium content — EDTA titrimetric method.*

ISO 6059:1984, *Water quality — Determination of the sum of calcium and magnesium — EDTA titrimetric method*

ISO 6333:1986, *Water quality — Determination of manganese — Formaldoxime spectrometric method.*

ISO 7980:1986, *Water quality — Determination of calcium and magnesium — Atomic absorption spectrometric method.*

ISO 8466-1:1990, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function.*

ISO 8466-2:1993, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 2: Calibration strategy for non-linear second order calibration functions.*

3 Interferences

3.1 Organics such as amino acids and aliphatic amines can interfere with the determination of inorganic cations.

3.2 If a strong complexing agent such as pyridine-2,6-dicarboxylic acid (PDA) is not present in the mobile phase, and if the suppressor technique is not used, cross-interferences from other cations like Zn^{2+} , Ni^{2+} , Cd^{2+} etc. can occur.

3.3 Cross-sensitivities with other cations, e.g. manganese, are dependent on the selectivity of the separator column used. If the quality requirements in clause 8 are not achieved, the sample shall be diluted.

3.4 Cross-sensitivities in the determination of NH_4^+ and Na^+ can occur when there are large differences in concentration.

3.5 Solid material and organic compounds (such as mineral oils, detergents and humic acids) shorten the lifetime of the separator column. They are therefore eliminated from the sample (9.3).

4 Principle

Liquid chromatographic separation of Li^+ , Na^+ , NH_4^+ , K^+ , Mn^{2+} , Ca^{2+} , Mg^{2+} , Sr^{2+} and Ba^{2+} by means of a separator column. A low-capacity cation exchanger is used as the stationary phase, and usually, aqueous solutions of mono- and dibasic acids as mobile phases (eluent, see 6.16).

These cations are detected by conductivity. It is essential that the eluents have a sufficiently low conductivity. For this reason, a conductivity detector (CD) is often combined with a suppressor device (e.g. an anion exchanger) which will reduce the conductivity of the eluent and transform the separated cations into their corresponding bases.

In the conductivity detection without chemical suppression, the difference between the ion equivalent conductivities is measured directly after the column. This difference should be as high as possible and the detector cell temperature should be stabilized within $\pm 0,1$ °C.

The concentration of the respective cations is determined by a calibration of the overall procedure. Particular cases can require calibration by means of standard addition (spiking).