

This document is a preview generated by EVS

## **Dental elevators - Part 6: Flohr elevators**

Dental elevators - Part 6: Flohr elevators

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 15087-6:2000 sisaldab Euroopa standardi EN ISO 15087-6:2000 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 13.10.2000 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 15087-6:2000 consists of the English text of the European standard EN ISO 15087-6:2000.</p> <p>This document is endorsed on 13.10.2000 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p><b>Käsitlusala:</b> This part of the standard specifies specific requirements including dimensions for Flohr dental elevators.</p>	<p><b>Scope:</b> This part of the standard specifies specific requirements including dimensions for Flohr dental elevators.</p>
---	---

ICS 11.060.20

Võtmesõnad:

ICS 11.060.20

English version

Dental elevators

Part 6: Flohr elevators  
(ISO 15087-6 : 2000)

Élévateurs dentaires – Partie 6:  
Élévateurs Flohr (ISO 15087-6 : 2000)

Zahnärztliche Wurzelheber – Teil 6:  
Flohr-Wurzelheber  
(ISO 15087-6 : 2000)

This European Standard was approved by CEN on 2000-04-30.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

**CEN**

European Committee for Standardization  
Comité Européen de Normalisation  
Europäisches Komitee für Normung

**Central Secretariat: rue de Stassart 36, B-1050 Brussels**

## Foreword

International Standard

ISO 15087-6 : 2000 Dental elevators – Part 6: Flohr elevators,

which was prepared by ISO/TC 106 'Dentistry' of the International Organization for Standardization, has been adopted by Technical Committee CEN/TC 55 'Dentistry', the Secretariat of which is held by DIN, as a European Standard.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, and conflicting national standards withdrawn, by November 2000 at the latest.

In accordance with the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard:

Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

## Endorsement notice

The text of the International Standard ISO 15087-6 : 2000 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to international publications are listed in Annex ZA (normative).

This document is a preview generated by EVS

## 1 Scope

This part of ISO 15087 gives requirements, including dimensions, for Flohr dental elevators.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 15087. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 15087 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 1942-3, *Dental vocabulary — Part 3: Dental instruments*.

ISO 15087 -1:1999, *Dental elevators — Part 1: General requirements*.

## 3 Terms and definitions

For the purposes of this part of ISO 15087, the terms and definitions given in ISO 1942-3 apply.

## 4 Design and dimensions

### 4.1 General

4.1.1 Flohr elevators shall conform to the general requirements given in ISO 15087-1.

4.1.2 Flohr elevators shall be grouped into the following types according to the shape of the working end:

- Type 1 = Straight working end (not used)
- Type 2 = Left working end
- Type 3 = Right working end

4.1.3 Flohr elevators shall be designed in accordance with Figure 1, and with dimensions in accordance with Table 1. Measurements of dimensions should be made as defined in Table 2.

NOTE Annex A of ISO 15087-1:1999 provides details of one method of measurement applicable to most types of dental hand instrument.

### 4.2 Length range

Flohr elevators shall have an overall length in the range between 145 mm and 155 mm.