

**Water quality - Determination of chromium(VI) - Photometric method for weakly contaminated water**

Water quality - Determination of chromium(VI) -  
Photometric method for weakly contaminated water

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 18412:2006 sisaldab Euroopa standardi EN ISO 18412:2006 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 20.09.2006 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 18412:2006 consists of the English text of the European standard EN ISO 18412:2006.</p> <p>This document is endorsed on 20.09.2006 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p><b>Käsitlusala:</b> This International Standard specifies a method for the determination of chromium(VI) in drinking water in mass concentrations between 2 µg/l and 50 µg/l. For the determination of higher concentrations, the sample is diluted prior to analysis.</p>	<p><b>Scope:</b> This International Standard specifies a method for the determination of chromium(VI) in drinking water in mass concentrations between 2 µg/l and 50 µg/l. For the determination of higher concentrations, the sample is diluted prior to analysis.</p>
---	---

ICS 13.060.50

Võtmesõnad:

ICS 13.060.50

English Version

## Water quality - Determination of chromium(VI) - Photometric method for weakly contaminated water (ISO 18412:2005)

Qualité de l'eau - Dosage du chrome(VI) - Méthode photométrique pour des eaux faiblement contaminées (ISO 18412:2005)

Wasserbeschaffenheit - Bestimmung von Chrom(VI) - Photometrisches Verfahren für gering belastetes Wasser (ISO 18412:2005)

This European Standard was approved by CEN on 3 August 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

## Foreword

The text of ISO 18412:2005 has been prepared by Technical Committee ISO/TC 147 "Water quality" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 18412:2006 by Technical Committee CEN/TC 230 "Water analysis", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2007, and conflicting national standards shall be withdrawn at the latest by February 2007.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

### Endorsement notice

The text of ISO 18412:2005 has been approved by CEN as EN ISO 18412:2006 without any modifications.

---

---

**Water quality — Determination of  
chromium(VI) — Photometric method  
for weakly contaminated water**

*Qualité de l'eau — Dosage du chrome(VI) — Méthode photométrique  
pour des eaux faiblement contaminées*



**PDF disclaimer**

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
Case postale 56 • CH-1211 Geneva 20  
Tel. + 41 22 749 01 11  
Fax + 41 22 749 09 47  
E-mail [copyright@iso.org](mailto:copyright@iso.org)  
Web [www.iso.org](http://www.iso.org)

Published in Switzerland

## Contents

Page

Foreword.....	iv
Introduction.....	v
1 Scope.....	1
2 Normative references.....	1
3 Interferences.....	1
4 Principle.....	2
5 Reagents.....	2
6 Apparatus.....	3
7 Sampling and sample pretreatment.....	3
8 Procedure.....	3
9 Evaluation.....	4
10 Expression of results.....	4
11 Test report.....	5
Annex A (informative) Precision data.....	6

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 18412 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

## Introduction

The user should be aware that particular problems could require the specification of additional conditions.

# Water quality — Determination of chromium(VI) — Photometric method for weakly contaminated water

**WARNING** — Persons using this International Standard should be familiar with normal laboratory practice. This International Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions. Care should be taken in handling potassium dichromate due to its carcinogenicity.

**IMPORTANT** — It is absolutely essential that tests conducted according to this standard be carried out by suitably trained staff.

## 1 Scope

This International Standard specifies a method for the determination of chromium(VI) in drinking water in mass concentrations between 2 µg/l and 50 µg/l. For the determination of higher concentrations, the sample is diluted prior to analysis. The method may also be applied to weakly polluted ground and surface water, provided the matrix does not contain interfering reducing agents. This method has not been verified for estuarine water and seawater, so the user is responsible for the validation of the method for these matrices. The photometric determination of chromium(VI) in waste water is carried out according to ISO 11083, *Water quality — Determination of chromium(VI) — Spectrometric method using 1,5-diphenylcarbazide*.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 3696, *Water for analytical laboratory use — Specification and test methods*

ISO 5667-1, *Water quality — Sampling — Part 1: Guidance on the design of sampling programmes*

ISO 5667-2, *Water quality — Sampling — Part 2: Guidance on sampling techniques*

ISO 5667-3, *Water quality — Sampling — Part 3: Guidance on the preservation and handling of samples*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 1: Statistical evaluation of the linear calibration function*

ISO 8466-2, *Water quality — Calibration and evaluation of analytical methods and estimation of performance characteristics — Part 2: Calibration strategy for non-linear second-order calibration functions*

## 3 Interferences

Reducing agents in the sample may lead to negative bias for the chromium(VI) concentration. Concentrations of sulfide up to 0,2 mg/l do not interfere.