

Acoustics - Reference zero for the calibration of audiometric equipment - Part 3: Reference equivalent threshold force levels for pure tones and bone vibrators

Acoustics - Reference zero for the calibration of audiometric equipment - Part 3: Reference equivalent threshold force levels for pure tones and bone vibrators

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>Käesolev Eesti standard EVS-EN ISO 389-3:1999 sisaldab Euroopa standardi EN ISO 389-3:1998 ingliskeelset teksti.</p> <p>Käesolev dokument on jõustatud 12.12.1999 ja selle kohta on avaldatud teade Eesti standardiorganisatsiooni ametlikus väljaandes.</p> <p>Standard on kättesaadav Eesti standardiorganisatsioonist.</p>	<p>This Estonian standard EVS-EN ISO 389-3:1999 consists of the English text of the European standard EN ISO 389-3:1998.</p> <p>This document is endorsed on 12.12.1999 with the notification being published in the official publication of the Estonian national standardisation organisation.</p> <p>The standard is available from Estonian standardisation organisation.</p>
--	---

<p>Käsitlusala: The standard specifies data applicable to the calibration of bone vibrators for pure-tone bone-conduction audiometry.</p>	<p>Scope: The standard specifies data applicable to the calibration of bone vibrators for pure-tone bone-conduction audiometry.</p>
--	--

ICS 13.140

Võtmesõnad: acoustics, audiometers, audiometry, auditory threshold, bone conduction, calibration, reference data

English version

Acoustics

Reference zero for the calibration of audiometric equipment

**Part 3: Reference equivalent threshold force levels for pure tones
and bone vibrators**

(ISO 389-3 : 1994, including Technical Corrigendum 1 : 1995)

Acoustique – Zéro de référence pour
l'étalonnage d'équipements audio-
métriques – Partie 3: Niveaux de
référence équivalents de force limi-
naire pour les vibrateurs à sons purs
et les ossivateurs
(ISO 389-3 : 1994, Rectificatif
Technique 1 : 1995 inclus)

Akustik – Standard-Bezugspegel für
die Kalibrierung audiometrischer
Geräte – Teil 3: Äquivalente Bezugs-
Schwellenkraftpegel für reine Töne
und Knochenleitungshörer
(ISO 389-3 : 1994, einschließlich
Technischer Korrektur 1 : 1995)

This European Standard was approved by CEN on 1998-12-04.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Foreword

International Standard

ISO 389-3 : 1994 Acoustics – Reference zero for the calibration of audiometric equipment – Part 3: Reference equivalent threshold force levels for pure tones and bone vibrators,

which was prepared by ISO/TC 43 'Acoustics' of the International Organization for Standardization, has been adopted by Technical Committee CEN/TC 211 'Acoustics', the Secretariat of which is held by DS, as a European Standard.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, and conflicting national standards withdrawn, by June 1999 at the latest.

In accordance with the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard:

Austria, Belgium, the Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, the Netherlands, Norway, Portugal, Spain, Sweden, Switzerland, and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 389-3 : 1994 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to international publications are listed in Annex ZA (normative).

This document is a preview generated by EVS

Introduction

Each part of ISO 389 specifies a specific reference zero for the calibration of audiometric equipment. ISO 389:1991 (to be re-issued as ISO 389-1) and ISO 389-2 are applicable to audiometric equipment for the transmission of pure tones by air conduction.

For clinical diagnostic and other audiometric purposes, it is often necessary to compare the measured hearing threshold levels of a person for sound transmitted to the inner ear by the air-conduction and bone-conduction pathways, respectively. Bone-conducted sound is provided for this purpose by an electromechanical vibrator applied to the mastoid prominence or to the forehead of the person under test.

The reference zero for the calibration of audiometric equipment for air conduction is defined in ISO 389:1991 and ISO 389-2 in terms of reference equivalent threshold sound pressure levels (RET SPLs), i.e. threshold sound pressure levels produced in a coupler, ear simulator or artificial ear of specified characteristics by supra-aural or insert earphones of various patterns, when excited electrically at a level corresponding to the threshold of hearing of young otologically normal persons. Similarly, this part of ISO 389 provides a reference zero for bone-conduction audiometry in terms of reference equivalent threshold force levels (RETFLs), i.e. the vibratory force levels produced by a bone vibrator on a specified mechanical coupler when the vibrator is excited electrically at a level corresponding to the threshold of hearing of young otologically normal persons. In some countries, the preferred location is the mastoid prominence; in other countries, the forehead location is used in addition to the mastoid prominence. Different RETFL values are valid for each of the two positions (see annex C).

For bone-conduction measurements, it is necessary to specify the static force of application of the vibrator to the skull of the test subject and to the mechanical coupler, as well as certain geometrical features of the vibrator tip. In addition, it is usually necessary to apply masking noise to the ear not under test, since excitation of the skull by the vibrator may be heard by that ear instead of (or in addition to) the ear intended for the test. An appropriate specification of the masking noise is, therefore, required as an adjunct to the reference equivalent threshold force levels, and such a specification is given in this part of ISO 389. Due to the so-called "occlusion effect" whereby the wearing of the transducer needed to provide the (air-conducted) masking noise causes a lowering of the bone-conduction threshold of hearing of the ear receiving the masking signal, it is necessary for the level of masking noise to be raised to cancel out the occlusion effect and provide adequate masking of the ear not under test. The specification of masking noise given in this part of ISO 389 is based on the procedures used in the experimental investigations from which the reference zero of this part of ISO 389 is derived.

Use of this reference zero to calibrate audiometers will ensure that measured bone-conduction hearing threshold levels of persons with unimpaired hearing or with hearing losses of purely sensorineural type (i.e. having unimpaired outer and middle ear function) will be compatible with the air-conduction hearing threshold levels of the same persons when using the reference zero of ISO 389:1991 or ISO 389-2, respectively. Although exact equivalence of air-conduction and bone-conduction thresholds for any individual in these classes cannot be expected, due to biological variability of sound transmission through the external and middle ear and through the skull, this part of ISO 389 will ensure that systematic deviations averaged over groups of such persons are reduced to a practical minimum.

This part of ISO 389 is based on an assessment of technical data provided by laboratories in three countries using methods of threshold testing which, in the respects described, were essentially uniform. Examination of the data showed that the experimental results were consistent. It has, therefore, been possible to standardize a reference zero by means of RETFL values which are to be used for all bone vibrators used in audiometry having similar characteristics to those used by the laboratories. The systematic uncertainties introduced by this deliberate simplification will be small in comparison to the usual step size of hearing level controls in clinical audiometers (5 dB).

1 Scope

This part of ISO 389 specifies the following data applicable to the calibration of bone vibrators for pure-tone bone-conduction audiometry.

- a) Reference equivalent threshold force levels (RETFLs), corresponding to the threshold of hearing of young otologically normal persons by bone-conduction audiometry. RETFL is the vibratory force level transmitted to a mechanical coupler of specified characteristics (see 5.3) by a vibrator when applied to the mechanical coupler under stated conditions of test and when energized at the voltage level corresponding to the normal threshold of hearing for location on the mastoid prominence.

NOTE 1 Interim values for the differences in reference equivalent threshold force levels between location on the forehead and mastoid are included for information in annex C.

- b) Essential characteristics of the bone vibrator and of its method of coupling to a person under test and to the mechanical coupler.
- c) Essential characteristics and datum level of the masking noise applied to the ear not under test.

1) To be re-issued as ISO 389-1.

Guidance on the practical application of this part of ISO 389 in the calibration of audiometers is given in annex B.

NOTE 2 Recommended procedures for carrying out bone-conduction audiometry are specified in ISO 8253-1.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this part of ISO 389. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this part of ISO 389 are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 389:1991, *Acoustics — Standard reference zero for the calibration of pure-tone air conduction audiometers.*¹⁾

ISO 389-2:1994, *Acoustics — Reference zero for the calibration of audiometric equipment — Part 2: Reference equivalent threshold sound pressure levels for pure tones and insert earphones.*

IEC 373:1990, *Mechanical coupler for measurements on bone vibrators.*