

ELEKTRILISED MEDITSIINISEADMED. OSA 2-68:
ERINÕUDED ELEKTRONKIIRENDITEL, KERGETE
IOONIDEGA JA RADIONUKLIIDALLIKAGA
VÄLISKIIRITUSRAVISEADMETEL KASUTATAVATE
RÖNTGENKUJUTISJUHTAVATE
KIIRITUSRAVISEADMETE ESMASELE OHUTUSELE JA
OLULISTELE TOIMIMISNÄITAJATELE

Medical electrical equipment - Part 2-68: Particular requirements for the basic safety and essential performance of X-ray-based image-guided radiotherapy equipment for use with electron accelerators, light ion beam therapy equipment and radionuclide beam therapy equipment

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN 60601-2-68:2015 sisaldab Euroopa standardi EN 60601-2-68:2015 ingliskeelset teksti.	This Estonian standard EVS-EN 60601-2-68:2015 consists of the English text of the European standard EN 60601-2-68:2015.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation.
Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 29.05.2015.	Date of Availability of the European standard is 29.05.2015.
Standard on kättesaadav Eesti Standardikeskusest.	The standard is available from the Estonian Centre for Standardisation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.60

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardikeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardikeskusega:

Aru 10, 10317 Tallinn, Eesti; koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation:

Aru 10, 10317 Tallinn, Estonia; homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

ICS 11.040.60

English Version

Medical electrical equipment - Part 2-68: Particular requirements for the basic safety and essential performance of X-ray-based image-guided radiotherapy equipment for use with electron accelerators, light ion beam therapy equipment and radionuclide beam therapy equipment
(IEC 60601-2-68:2014)

Appareils électromédicaux - Partie 2-68: Exigences particulières pour la sécurité de base et les performances essentielles des appareils de radiothérapie à rayonnement X assistée par imagerie médicale, destinés à être utilisés avec les accélérateurs d'électrons, les appareils de thérapie par faisceau d'ions légers et les appareils de thérapie par faisceau de radionucléides
(IEC 60601-2-68:2014)

Medizinische elektrische Geräte - Teil 2-68: Besondere Festlegungen für die Sicherheit einschließlich der wesentlichen Leistungsmerkmale von röntgenstrahlungsbasierten Geräten für die bildgesteuerte Strahlentherapie zur Verwendung mit Elektronenbeschleunigern, Leichtenion-Strahlentherapiesystemen und Radionuklid-Strahlentherapiesystemen
(IEC 60601-2-68:2014)

This European Standard was approved by CENELEC on 2014-10-09. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

The text of document 62C/595/FDIS, future edition 1 of IEC 60601-2-68 prepared by SC 62C "Equipment for radiotherapy, nuclear medicine and radiation dosimetry" of IEC/TC 62 "Electrical equipment in medical practice" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60601-2-68:2015.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2015-11-29
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2018-05-29

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For the relationship with EU Directive 93/42/EEC, see informative Annex ZZ, which is an integral part of this document.

Endorsement notice

The text of the International Standard IEC 60601-2-68:2014 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following note has to be added for the standard indicated:

IEC 60336:2005	NOTE	Harmonized as EN 60336:2005 (not modified).
IEC 60364-7-710:2002	NOTE	Harmonized as HD 60364-7-710:2012 (modified).
IEC 60522:1999	NOTE	Harmonized as EN 60522:1999 (not modified).
IEC 62220-1:2003	NOTE	Harmonized as EN 62220-1:2004 ¹⁾ (not modified).

¹⁾ Superseded by EN 62220-1-1:2015 (IEC 62220-1-1:2015): DOW = 2018-04-16.

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
--------------------	-------------	--------------	--------------	-------------

Annex ZA of EN 60601-1:2006 applies except as follows:

Amendment:

IEC 60601-1-3	2008	Medical electrical equipment - Part 1-3: General requirements for basic safety and essential performance - Collateral Standard: Radiation protection in diagnostic X-ray equipment	EN 60601-1-3 + corr. March	2008 2010
IEC 60601-1-6 +A1	2010 2013	Medical electrical equipment - Part 1-6: General requirements for basic safety and essential performance - Collateral standard: Usability	EN 60601-1-6 +A1	2010 2015

Addition:

IEC 60601-1 +A1	2005 2012	Medical electrical equipment - Part 1: General requirements for basic safety and essential performance	EN 60601-1 + corr. March + A1 + A1/AC +A12	2006 2010 2013 2014 2014
IEC 60601-2-1	2009	Medical electrical equipment - Part 2-1: Particular requirements for the basic safety and essential performance of electron accelerators in the range 1 MeV to 50 MeV	-	-
IEC 60601-2-4	2010	Medical electrical equipment - Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators	EN 60601-2-4	2011
IEC 60601-2-44 +A1	2009 2012	Medical electrical equipment - Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography	EN 60601-2-44 +A11 +A1	2009 2011 2012

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN/HD</u>	<u>Year</u>
IEC 60731	2011	Medical electrical equipment - Dosimeters with ionization chambers as used in radiotherapy	EN 60731	2012
IEC/TR 60788	2004	Medical electrical equipment - Glossary of defined terms	-	-
IEC 60976	2007	Medical electrical equipment - Medical electron accelerators - Functional performance characteristics	EN 60976	2007
IEC 61217	2011	Radiotherapy equipment - Coordinates, movements and scales	EN 61217	2012
IEC 61223-3-5	2004	Evaluation and routine testing in medical imaging departments - Part 3-5: Acceptance tests - Imaging performance of computed tomography X-ray equipment	EN 61223-3-5	2004
IEC 61262-7	1995	Medical electrical equipment - Characteristics of electro-optical X-ray image intensifiers - Part 7: Determination of the modulation transfer function	EN 61262-7	1995
IEC 62083	2009	Medical electrical equipment - Requirements for the safety of radiotherapy treatment planning systems	EN 62083	2009
IEC 62274	2005	Medical electrical equipment - Safety of radiotherapy record and verify systems	EN 62274	2005
IEC 62366	2007	Medical devices - Application of usability engineering to medical devices	-	-
IEC 62396-1	2012	Process management for avionics - Atmospheric radiation effects - Part 1: Accommodation of atmospheric radiation effects via single event effects within avionics electronic equipment	-	-
IEC 62563-1	2009	Medical electrical equipment - Medical image display systems - Part 1: Evaluation methods	EN 62563-1	2010

Annex ZZ (informative)

Coverage of Essential Requirements of EU Directives

This European Standard has been prepared under a mandate given to CENELEC by the European Commission and the European Free Trade Association, and within its scope the Standard covers all relevant essential requirements given in Annex I of EU Directive 93/42/EEC of 14 June 1993 concerning medical devices.

Compliance with this standard provides one means of conformity with the specified essential requirements of the Directive concerned.

WARNING: Other requirements and other EU Directives can be applied to the products falling within the scope of this standard.

This document is a preview generated by EVS

CONTENTS

FOREWORD.....	3
INTRODUCTION.....	5
201.1 Scope, object and related standards	7
201.2 Normative references.....	9
201.3 Terms and definitions.....	10
201.4 General requirements	18
201.5 General requirements for testing ME EQUIPMENT	19
201.6 Classification of me equipment and me systems	19
201.7 ME EQUIPMENT identification, marking and documents	19
201.8 Protection against electrical HAZARDS from ME EQUIPMENT	25
201.9 Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS	28
201.10 Protection against unwanted and excessive radiation HAZARDS	32
201.11 Protection against excessive temperatures and other HAZARDS	34
201.12 Accuracy of controls and instruments and protection against hazardous outputs	34
201.13 Hazardous situations and fault conditions for me equipment	34
201.14 PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS)	35
201.15 Construction of me equipment.....	35
201.16 ME SYSTEMS	35
201.17 Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS	35
201.101 Reference data for X-IGRT	36
201.102 X-IGRT IMAGING	40
201.103 IGRT analysis and correction	47
203 RADIATION protection in diagnostic X-RAY EQUIPMENT	51
206 Usability.....	52
Annex B (informative) Sequence of testing	54
Annex I (informative) ME SYSTEMS aspects.....	54
Annex AA (informative) Particular guidance and rationale.....	55
Annex BB (informative) Measuring $CTDI_{free\ air}$	57
Bibliography.....	58
Index of defined terms used in this standard	59
Figure 201.101 – PATIENT SUPPORT movements	53
Table 201.101 – Data required in the technical description	22
Table 201.102 – Clauses and subclauses in this particular standard that require the provision of information in the ACCOMPANYING DOCUMENTS, INSTRUCTIONS FOR USE and the technical description	23
Table 201.103 – Example test pattern for $CTDI_{free\ air}$ for kV.....	45

INTRODUCTION

Modern RADIOTHERAPY practices utilize information from various imaging modalities, acquired prior to initiating administration of the therapy, to plan the treatment. The imaging provides information about the location of the TARGET VOLUME and other anatomical features so that a treatment plan can be developed that provides an optimal dose distribution to have the best chance of achieving the intended effect of treatment while minimizing side effects.

However, difficulties arise when trying to administer the RADIATION, since TARGET VOLUMES/critical structures are constantly moving within the body. For example, in parts of the body moving with respiration, the TARGET VOLUMES/critical structures may change position or shape during the RADIATION BEAM delivery throughout any given fraction. Furthermore, a course of therapy may extend over many days, during which the TARGET VOLUME/PATIENT may shrink or grow and/or move. Hence, the exact location of the TARGET VOLUME/critical structures may change between the time of treatment planning imaging and the actual administration of a treatment.

IMAGE-GUIDED RADIOTHERAPY (IGRT) combines planar or volumetric imaging during the course of RADIOTHERAPY in order to adjust the treatment delivery based on the PATIENT anatomy and PATIENT position. This enables the OPERATOR and/or EXTERNAL BEAM EQUIPMENT (EBE) to adjust the RADIATION BEAM delivery based on the imaging information, such as the position of the TARGET VOLUME, critical organs and/or other reference features, to compensate for anatomical changes including internal organ motions and/or treatment setup uncertainties. The increased accuracy and precision achieved allows higher doses of RADIATION to be delivered to the TARGET VOLUME and a reduction in the margin of healthy cells affected by the RADIATION. This is often used in conjunction with other monitoring equipment.

This particular standard establishes requirements to be complied with by MANUFACTURERS in the design and construction of X-RAY IGRT EQUIPMENT (X-IGRT).

This particular standard covers safety aspects of kilovoltage (kV) and megavoltage (MV) X-ray imaging devices in a known geometrical relationship with an EXTERNAL BEAM EQUIPMENT such as an ELECTRON ACCELERATOR, medical light ion beam equipment or RADIONUCLIDE BEAM THERAPY EQUIPMENT, for the purpose of IGRT. It covers aspects of communication and relationships between the EXTERNAL BEAM EQUIPMENT and X-ray imaging devices, attached or not directly attached to but in the same RADIATION shielded area as, and dedicated for use only with, the EXTERNAL BEAM EQUIPMENT.

This particular standard applies to X-ray based IGRT equipment used in-room for IGRT purposes. This particular standard does not apply to standard CT scanners, which are not used for IGRT. However if a CT scanner is used in-room with a linear (electron) accelerator (linac) for IGRT then this particular standard applies.

When performing a HAZARD ANALYSIS, the MANUFACTURER should consider relevant diagnostic standards. For example, IMAGE DISPLAY DEVICE quality is specified in IEC documents in regards to diagnostic use (e.g. IEC 62563-1:2009, Ed. 1.0). However, since IGRT usage may or may not require such high requirements it is left to the MANUFACTURER to specify what is required for use with their X-IGRT EQUIPMENT.

This particular standard deals with the safety aspect of image acquisitions, image analysis, data transfer and treatment replanning or EBE/PATIENT repositioning.

This particular standard deals with equipment for REAL-TIME X-IGRT, ONLINE X-IGRT and OFFLINE X-IGRT.

X-IGRT EQUIPMENT is also related to the following current standards:

- IEC 62083, *Medical electrical equipment – Requirements for the safety of radiotherapy treatment planning systems*

- IEC 61217, *Radiotherapy equipment – Coordinates, movements and scales*
- IEC 62274, *Medical electrical equipment – Safety of radiotherapy record and verify systems*
- IEC 60976, *Medical electrical equipment – Medical electron accelerators – Functional performance characteristics*
- IEC TR 60977, *Medical electrical equipment – Medical electron accelerators – Guidelines for functional performance characteristics.*

This particular standard may give rise to amendments to some of the above standards.

This particular standard will focus on the safety aspects of the primary function of X-IGRT. It will not focus on emerging technologies within the field so as to not hinder progress, yet it will define a safe way of achieving X-IGRT.

This document is a preview generated by EVS

MEDICAL ELECTRICAL EQUIPMENT –

Part 2-68: Particular requirements for the basic safety and essential performance of X-ray-based image-guided radiotherapy equipment for use with electron accelerators, light ion beam therapy equipment and radionuclide beam therapy equipment

201.1 Scope, object and related standards

Clause 1 of the general standard¹ applies, except as follows:

201.1.1 Scope

Replacement:

This International Standard applies to the BASIC SAFETY and ESSENTIAL PERFORMANCE of X-ray based IMAGE-GUIDED RADIOTHERAPY equipment for use with EXTERNAL BEAM EQUIPMENT (EBE).

This particular standard covers safety aspects of kilovoltage (kV) and megavoltage (MV) X-ray imaging devices in a known geometrical relationship with EBE for the purpose of IGRT. It covers aspects of communication and relationships between the EXTERNAL BEAM EQUIPMENT and X-ray imaging devices, attached or not directly attached to, but in the same RADIATION shielded area as, and dedicated for use only with, the EXTERNAL BEAM EQUIPMENT.

This particular standard deals with equipment for REAL-TIME X-IGRT, ONLINE X-IGRT and OFFLINE X-IGRT. It covers procedures to reduce the risk of over-reliance on the X-IGRT EXTERNAL BEAM SYSTEM (X-IGRT EBS). For example the manufacturer will provide an interactive interface for user interaction with the correction suggested by the system.

If a clause or subclause is specifically intended to be applicable to X-IGRT EBE SYSTEMS the content of that clause or subclause will say so. If that is not the case, the clause or subclause applies only to X-IGRT EQUIPMENT.

This particular standard, with the inclusion of TYPE TESTS and SITE TESTS, applies respectively to the MANUFACTURER and some installation aspects of X-IGRT EBE SYSTEMS intended to be

- for NORMAL USE, operated under the authority of appropriately licensed or QUALIFIED PERSONS by OPERATORS having the required skills for a particular medical application, for particular specified clinical purposes, e.g. STATIONARY RADIOTHERAPY or MOVING BEAM RADIOTHERAPY,
- maintained in accordance with the recommendations given in the INSTRUCTIONS FOR USE,
- subject to regular quality assurance performance and calibration checks by a QUALIFIED PERSON.

NOTE In this particular standard, all references to installation refer to installation in the RESPONSIBLE ORGANIZATION'S premises

201.1.2 Object

Replacement:

¹ The general standard is IEC 60601-1:2005 + IEC 60601-1:2005/AMD1:2012, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*

The object of this particular standard is to establish particular BASIC SAFETY and ESSENTIAL PERFORMANCE requirements for X-IGRT EQUIPMENT and X-IGRT EBE SYSTEMS.

201.1.3 Collateral standards

Addition:

This particular standard refers to those applicable collateral standards that are listed in Clause 2 of the general standard and Clause 201.2 of this particular standard.

IEC60601-1-3 and IEC 60601-1-6 apply as modified in Clause 203 and Clause 206 respectively. IEC 60601-1-8, IEC 60601-1-9, IEC 60601-1-10 and IEC 60601-1-11 do not apply. All other published collateral standards in the IEC 60601-1 series apply as published.

Collateral standards published after the date of publication of this standard shall only apply subject to further amendment to this standard.

201.1.4 Particular standards

Replacement:

In the IEC 60601 series, particular standards may modify, replace or delete requirements contained in the general standard and collateral standards as appropriate for the particular ME EQUIPMENT under consideration, and may add other BASIC SAFETY and ESSENTIAL PERFORMANCE requirements.

A requirement of a particular standard takes priority over the general standard.

For brevity, IEC 60601-1 is referred to in this particular standard as the general standard. Collateral standards are referred to by their document number.

The numbering of clauses and subclauses of this particular standard corresponds to that of the general standard with the prefix "201" (e.g. 201.1 in this particular standard addresses the content of Clause 1 of the general standard) or applicable collateral standard with the prefix "20x", where x is the final digit(s) of the collateral standard document number (e.g. 202.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-2 collateral standard, 203.4 in this particular standard addresses the content of Clause 4 of the IEC 60601-1-3 collateral standard, etc.). The changes to the text of the general standard are specified by the use of the following words:

"Replacement" means that the clause or subclause of the general standard or applicable collateral standard is replaced completely by the text of this particular standard.

"Addition" means that the text of this particular standard is additional to the requirements of the general standard or applicable collateral standard.

"Amendment" means that the clause or subclause of the general standard or applicable collateral standard is amended as indicated by the text of this particular standard.

Subclauses, figures or tables which are additional to those of the general standard are numbered starting from 201.101. However, due to the fact that definitions in the general standard are numbered 3.1 through 3.139, additional definitions in this standard are numbered beginning from 201.3.201. Additional annexes are lettered AA, BB, etc., and additional items aa), bb), etc.

Subclauses, figures or tables which are additional to those of a collateral standard are numbered starting from 20x, where "x" is the number of the collateral standard, e.g. 202 for IEC 60601-1-2, 203 for IEC 60601-1-3, etc.

The term "this standard" is used to make reference to the general standard, any applicable collateral standards and this particular standard taken together.

Where there is no corresponding clause or subclause in this particular standard, the clause or subclause of the general standard or applicable collateral standard, although possibly not relevant, applies without modification; where it is intended that any part of the general standard or applicable collateral standard, although possibly relevant, is not to be applied, a statement to that effect is given in this particular standard.

201.2 Normative references

Clause 2 of the general standard applies, except as follows:

Amendment:

IEC 60601-1-3:2008, *Medical electrical equipment – Part 1-3: General requirements for basic safety and essential performance – Collateral Standard: Radiation protection in diagnostic X-ray equipment*

IEC 60601-1-6:2010, *Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral standard: Usability*
IEC 60601-1-6:2010/AMD1:2013

Addition:

IEC 60601-1:2005, *Medical electrical equipment – Part 1: General requirements for basic safety and essential performance*
IEC 60601-1:2005/AMD1:2012

IEC 60601-2-1:2009, *Medical electrical equipment – Part 2-1: Particular requirements for the basic safety and essential performance of electron accelerators in the range 1 MeV to 50 MeV*

IEC 60601-2-4:2010, *Medical electrical equipment – Part 2-4: Particular requirements for the basic safety and essential performance of cardiac defibrillators*

IEC 60601-2-44:2012, *Medical electrical equipment – Part 2-44: Particular requirements for the basic safety and essential performance of X-ray equipment for computed tomography*

IEC 60731:2011, *Medical electrical equipment – Dosimeters with ionization chambers as used in radiotherapy*

IEC/TR 60788:2004, *Medical electrical equipment – Glossary of defined terms*

IEC 60976:2007, *Medical electrical equipment – Medical electron accelerators – Functional performance characteristics*

IEC 61217:2011, *Radiotherapy equipment – Coordinates, movements and scales*

IEC 61223-3-5:2004, *Evaluation and routine testing in medical imaging departments – Part 3-5: Acceptance tests – Imaging performance of computed tomography X-ray equipment*

IEC 61262-7:1995, *Medical electrical equipment – Characteristics of electro-optical X-ray image intensifiers – Part 7: Determination of the modulation transfer function*

IEC 62083:2009, *Medical electrical equipment – Requirements for the safety of radiotherapy treatment planning systems*

IEC 62274:2005, *Medical electrical equipment – Safety of radiotherapy record and verify systems*

IEC 62366:2007, *Medical devices – Application of usability engineering to medical devices*

IEC 62396-1:2012, *Process management for avionics – Atmospheric radiation effects – Part 1: Accommodation of atmospheric radiation effects via single event effects within avionics electronic equipment*

IEC 62563-1:2009, *Medical electrical equipment – Medical image display systems – Part 1: Evaluation methods*

NOTE Informative references are listed in the bibliography beginning on page 58.

201.3 Terms and definitions

For the purposes of this document, the terms and definitions given in IEC 60601-2-1, IEC 60601-1:2005 + IEC 60601-1:2005 /AMD1:2012, and IEC/TR 60788:2004 apply, except as follows:

NOTE An index of defined terms is found at the end of the document.

Addition:

201.3.201 COMPUTED TOMOGRAPHY DOSE INDEX 100

$CTDI_{100}$

integral of the DOSE PROFILE representative of a single axial scan along a line perpendicular to the TOMOGRAPHIC PLANE divided by $N \times T$ according to the following:

for $N \times T$ less than or equal to 40 mm

$$CTDI_{100} = \int_{-50\text{ mm}}^{+50\text{ mm}} \frac{D(y)}{N \times T} dy$$

for $N \times T$ greater than 40 mm (all CT CONDITIONS OF OPERATION except collimation are kept the same for these measurements)

$$CTDI_{100} = \int_{-50\text{ mm}}^{+50\text{ mm}} \frac{D_{\text{Ref}}(y)}{(N \times T)_{\text{Ref}}} dz \times \frac{CTDI_{\text{free air}, N \times T}}{CTDI_{\text{free air}, \text{Ref}}}$$

where:

$D(y)$ is the DOSE PROFILE representative of a single axial scan along a line perpendicular to the TOMOGRAPHIC PLANE, where dose is reported as ABSORBED DOSE in air and is evaluated within a polymethylmethacrylate (PMMA) dosimetry PHANTOM (see 201.102.5.2);

$(N \times T)_{\text{Ref}}$ is a specific $N \times T$ of 20 mm or the largest $N \times T$ available not greater than 20 mm;

$D_{\text{Ref}}(y)$ is the DOSE PROFILE representative of a single axial scan along a line perpendicular to the TOMOGRAPHIC PLANE, where dose is reported as ABSORBED DOSE in air and is evaluated within a polymethylmethacrylate (PMMA) dosimetry PHANTOM (see 201.102.5.2) for $(N \times T)_{\text{Ref}}$;

$CTDI_{\text{free air}, N \times T}$ is the $CTDI_{\text{free air}}$ (201.3.202) for a specific value of $N \times T$;

$CTDI_{\text{free air}, \text{Ref}}$ is the $CTDI_{\text{free air}}$ (201.3.202) for $(N \times T)_{\text{Ref}}$;

N is the number of TOMOGRAPHIC SECTIONS produced in a single axial scan of the X-ray source;