

**Aktiivsed implanteeritavad meditsiiniseadmed. Osa
2-2: Erinõuded tahhüarütmia raviks mõeldud
aktiivsetele siirdatavatele meditsiiniseadmetele
(sealhulgas siirdatavatele defibrillaatoritele)**

Active implantable medical devices -- Part 2-2:
Particular requirements for active implantable medical
devices intended to treat tachyarrhythmia (includes
implantable defibrillators)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

Käesolev Eesti standard EVS-EN 45502-2-2:2008 sisaldab Euroopa standardi EN 45502-2-2:2008 ingliskeelset teksti.

Standard on kinnitatud Eesti Standardikeskuse 28.04.2008 käskkirjaga ja jõustub sellekohase teate avaldamisel EVS Teatajas.

Euroopa standardimisorganisatsioonide poolt rahvuslikele liikmetele Euroopa standardi teksti kättesaadavaks tegemise kuupäev on 31.03.2008.

Standard on kättesaadav Eesti standardiorganisatsioonist.

This Estonian standard EVS-EN 45502-2-2:2008 consists of the English text of the European standard EN 45502-2-2:2008.

This standard is ratified with the order of Estonian Centre for Standardisation dated 28.04.2008 and is endorsed with the notification published in the official bulletin of the Estonian national standardisation organisation.

Date of Availability of the European standard text 31.03.2008.

The standard is available from Estonian standardisation organisation.

ICS 11.040.40

Võtmesõnad:

Standardite reprodutseerimis- ja levitamiseõigus kuulub Eesti Standardikeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonilisse süsteemi või edastamine ükskõik millises vormis või millisel teel on keelatud ilma Eesti Standardikeskuse poolt antud kirjaliku loata.

Kui Teil on küsimusi standardite autorikaitse kohta, palun võtke ühendust Eesti Standardikeskusega:
Aru 10 Tallinn 10317 Eesti; www.evs.ee; Telefon: 605 5050; E-post: info@evs.ee

English version

**Active implantable medical devices -
Part 2-2: Particular requirements for active implantable medical devices
intended to treat tachyarrhythmia
(includes implantable defibrillators)**

Dispositifs médicaux implantables actifs -
Partie 2-2: Exigences particulières
pour les dispositifs médicaux implantables
actifs destinés au traitement des
tachyarythmies
(y compris les défibrillateurs implantables)

Aktive implantierbare Medizingeräte -
Teil 2-2: Besondere Festlegungen
für aktive implantierbare medizinische
Produkte zur Behandlung von
Tachyarrhythmie
(einschließlich implantierbaren
Defibrillatoren)

This European Standard was approved by CENELEC on 2007-05-01. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the Central Secretariat has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Cyprus, the Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

CENELEC

European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

Central Secretariat: rue de Stassart 35, B - 1050 Brussels

Foreword

This European Standard was prepared by the CEN/CENELEC Joint Working Group on ACTIVE IMPLANTABLE MEDICAL DEVICES (CEN/CLC JWG AIMD).

The text of the draft was submitted to the formal vote and was approved by CENELEC as EN 45502-2-2 on 2007-05-01.

This European Standard, together with EN 45502-2-1:2003, supersedes EN 50061:1988 + A1:1995 (+ corrigendum October 1995).

The following dates were fixed:

- latest date by which the EN has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2008-10-01
- latest date by which the national standards conflicting with the EN have to be withdrawn (dow) 2010-05-01

The requirements of this Particular Standard supplement or modify those of EN 45502-1:1997, active implantable medical devices – Part 1: General requirements for safety, marking and information to be provided by the manufacturer.

This European Standard has been prepared by the CEN/CENELEC Joint Working Group on Active Implantable Medical Devices (CEN/CLC JWG AIMD). Members of the Joint Working Group were nominated by one of the members of either CEN or CENELEC.

This European Standard has been prepared under mandates given to CEN and CENELEC by the Commission of the European Communities and the European Free Trade Association, and supports the essential requirements of Directive 90/385/EEC Council Directive of June 1990 on the approximation of the laws of the Member states relating to active implantable medical devices.

Although both this Particular Standard and the Directive deal with the same products, the structure and purpose of the two documents are different. Annex AA of this Particular Standard correlates the requirements of the Directive with the subclauses of EN 45502-1:1997 and of this Particular Standard. Annex BB provides references in the other direction, from this European Standard to the Directive. Annex CC is a rationale providing some further explanation of the subclauses of this Particular Standard.

Contents

1	Scope.....	7
2	Normative references	7
3	Definitions	8
4	Symbols and abbreviations (optional)	11
5	General requirements for non-implantable parts	11
6	Measurement of IMPLANTABLE PULSE GENERATOR and LEAD characteristics	12
7	General arrangement of the packaging	14
8	General MARKINGS for ACTIVE IMPLANTABLE MEDICAL DEVICES	14
9	MARKINGS on the SALES PACKAGING	15
10	Construction of the SALES PACKAGING	16
11	MARKINGS on the STERILE PACK.....	16
12	Construction of the NON-REUSABLE PACK.....	16
13	MARKINGS on the ACTIVE IMPLANTABLE MEDICAL DEVICE	17
14	Protection from unintended biological effects being caused by the ACTIVE IMPLANTABLE MEDICAL DEVICE	17
15	Protection from HARM to the patient or user caused by external physical features of the ACTIVE IMPLANTABLE MEDICAL DEVICE	18
16	Protection from HARM to the patient caused by electricity	18
17	Protection from HARM to the patient caused by heat.....	21
18	Protection from ionizing radiation released or emitted from the ACTIVE IMPLANTABLE MEDICAL DEVICE	21
19	Protection from unintended effects caused by the device.....	21
20	Protection of the device from damage caused by external defibrillators	23
21	Protection of the device from changes caused by high power electrical fields applied directly to the patient.....	27
22	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from changes caused by miscellaneous medical treatments	28
23	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from mechanical forces.....	28
24	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by electrostatic discharge.....	33
25	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from damage caused by atmospheric pressure changes	33
26	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from changes caused by temperature changes.....	34
27	Protection of the ACTIVE IMPLANTABLE MEDICAL DEVICE from electromagnetic non-ionizing radiation	34
28	Accompanying documentation	56
	Annex AA (informative) Table of cross-references from 90/385/EEC to EN 45502-2-2	62
	Annex BB (informative) Relationship between the subclauses of EN 45502-2-2 and the essential requirements of 90/385/EEC listed in Annex AA.....	70
	Annex CC (informative) Notes on EN 45502-2-2	72
	Annex DD (informative) Code for describing modes of IMPLANTABLE PULSE GENERATORS	93
	Annex EE (normative) Interface circuits	95

Annex FF (informative) Selection of capacitor C_x	99
Annex GG (normative) Calibration of injection network (Figure EE.105); Test signal of inhibition generator (Figure GG.101)	100
Annex HH (informative) Defined terms	102
Bibliography.....	104
Figure 101 – Measurement of CD PULSE characteristics	12
Figure 102 – Measurement of ICD OUTPUT VOLTAGE	13
Figure 103 – Test set-up for measurement of electrical neutrality	19
Figure 104 – Band pass filter for a.c. leakage current measurement	20
Figure 105 – Test set up for measurement of a.c. leakage current	20
Figure 106 – Test set-up for checking internal DEFIBRILLATION protection	22
Figure 107 – Damped sinus DEFIBRILLATION waveform.....	24
Figure 108 – RCL circuit for generating a damped sinus DEFIBRILLATION waveform.....	24
Figure 109 – Timing sequence used for DEFIBRILLATION test	25
Figure 110 – Resistor network for Test 1 and Test 2	25
Figure 111 – Test setup for truncated exponential DEFIBRILLATION waveform	26
Figure 112 – Biphasic DEFIBRILLATION waveform for Test 2.....	26
Figure 113 – Test set up for proof of protection from high frequency currents caused by surgical equipment	28
Figure 114 – Conductor current integrity test fixture.....	30
Figure 115 – Conductor flex test fixture	31
Figure 116 – Connector flex test fixture	32
Figure 117 – Test signal 2	35
Figure 118 – Test set-up for measurement of induced current flow	36
Figure 119 - Common mode connection to multichannel bipolar PULSE generators	36
Figure 120 – Differential mode connection to multichannel bipolar PULSE generators	37
Figure 121 – Common mode connection for CARDIOVERSION/DEFIBRILLATION TERMINALS	38
Figure 122 – Differential mode connection for CARDIOVERSION/DEFIBRILLATION TERMINALS	38
Figure 123 – Test set-up to check for induced malfunction.....	40
Figure 124 – Common mode connection to multichannel bipolar PULSE generators	41
Figure 125 – Differential mode connection to multichannel bipolar PULSE generators	41
Figure 126 – Test set-up to check for induced malfunction due to voltages induced on CARDIOVERSION/DEFIBRILLATION LEADS.....	42
Figure 127 – Common mode connection for CARDIOVERSION/DEFIBRILLATION TERMINALS	42
Figure 128 – Differential mode connection for CARDIOVERSION/DEFIBRILLATION TERMINALS	43
Figure 129 – Test signal for frequencies between 10 MHz and 450 MHz.....	43
Figure 130 – Test set-up to check for induced malfunction at high frequency	44
Figure 131 – Connection of the PULSE generator	45
Figure 132 – Test set-up to characterize performance while subject to interference	46
Figure 133 – Modulation 1	48
Figure 134 – Modulation 2 (alternative).....	49
Figure 135 – Test signal for frequencies between 150 kHz and 10 MHz.....	51
Figure 136 – Test set-up to check for malfunction at high frequency	53
Figure 137 – Connection of the PULSE generator	54
Figure 138 – Loop configuration for varying magnetic field test	55
Figure CC.101 – Safe exposure times for adipose/skeletal muscle tissue at various temperatures	76
Figure CC.102 – Average number of full shocks per year.....	78

Figure CC.103 – Reference test coil	83
Figure CC.104 – Amplitude at point D and C of the Tissue Interface in Annex EE (C_x selected for 5 kHz corner frequency)	87
Figure CC.105 – Test signal amplitudes for 27.3 to 27.5.3	88
Figure CC.106 – Expected induced current during test according to 27.3.....	90
Figure EE.101 – Tissue equivalent interface circuit for current measurements	95
Figure EE.102 – Tissue equivalent interface circuit to check for malfunction due to voltage induced on the sense/pace LEADS	96
Figure EE.103 – Tissue equivalent interface circuit to check for malfunction due to voltages induced on the CD LEADS.....	97
Figure EE.104 – Low pass filter used to attenuate the 500 kHz component of test signal	97
Figure EE.105 – Injection network	98
Figure FF.101 – Test to check for spurious low frequency noise and to determine the value of C_x	99
Figure GG.101 – Form of signal from a test signal generator used for determination of SENSITIVITY (SENSING THRESHOLD) and for inhibition	101
Table 101 – Connection sequence	23
Table 102 – Component values for resistor network shown in Figure 110	25
Table 103 – Timing parameters of test signal for Test 2.....	26
Table 104 – Spurious injection current limits for sense/pace TERMINALS	39
Table 105 – Spurious injection current limits for CARDIOVERSION/DEFIBRILLATION TERMINALS ...	39
Table 106 – Peak-to-peak amplitudes V_{pp} in the range 16,6 Hz to 10 MHz	39
Table 107 – Peak-to-peak amplitudes V_{pp} in the range 16,6 Hz to 150 kHz (Common Mode Test)	48
Table 108 – Peak-to-peak amplitudes V_{pp} in the range 150 kHz to 10 MHz	50
Table 109 – Sinusoidally modulated magnetic field strengths.....	56
Table DD.101 – Basic defibrillator code scheme	93
Table DD.102 – Examples of the defibrillator code	94
Table EE.101 – Component values for Figure EE.101 (A)	95
Table EE.102 – Component values for Figure EE.101 (B)	95
Table EE.103 – Component values for Figure EE.102.....	96
Table EE.104 – Component values for Figure EE.103.....	97
Table EE.105 – Component values for Figure EE.104.....	98
Table EE.106 – Component values for Figure EE.105.....	98
Table GG.101 – Calibration signal amplitude	101

Introduction

This Particular Standard specifies particular requirements for IMPLANTABLE CARDIOVERTER DEFIBRILLATORS and the functions of ACTIVE IMPLANTABLE MEDICAL DEVICES intended to treat tachyarrhythmia, to provide basic assurance of safety for both patients and users.

An external defibrillator is a MEDICAL DEVICE used, in the emergency setting, to deliver a high-energy shock to the heart, by means of ELECTRODES applied to the external chest wall, in patients suffering ventricular fibrillation (a rapid, disorganized and potentially lethal heart rhythm abnormality), to restore normal heart action. External defibrillators may also be used, in emergency or elective settings, to terminate other ventricular or atrial tachyarrhythmias by delivery of a high-energy shock, synchronised to the intrinsic cardiac rhythm, a procedure known as CARDIOVERSION. In patients known to be at risk of such arrhythmias, due to the occurrence of previous episodes or the presence of specific pre-disposing cardiac conditions, an IMPLANTABLE CARDIOVERTER DEFIBRILLATOR may be implanted to perform similar functions. The implantable device, which is much smaller than an external defibrillator, is contained within a sealed, encapsulating enclosure. It generates high voltage PULSES from an enclosed, miniature, electrical battery. The PULSES are transmitted to the heart by means of implanted, insulated conductors with ELECTRODES (LEADS) The IMPLANTABLE CARDIOVERTER DEFIBRILLATOR may also incorporate other sensing and pacing functions, such as rate support for bradycardia and ANTITACHYCARDIA PACING (ATP) to terminate certain tachyarrhythmias without the need of a high-energy shock. The defibrillator may be adjusted non-invasively by means of an electronic device, known as a programmer.

This Particular Standard is relevant to all parts of ACTIVE IMPLANTABLE MEDICAL DEVICES intended to treat tachyarrhythmia other than pacing functions to control bradyarrhythmia. Typical examples are IMPLANTABLE PULSE GENERATORS, LEADS, ADAPTORS, ACCESSORIES, programmers and the related software. (Bradyarrhythmia pacing functions are dealt with in EN 45502-2-1).

The requirements of this Particular Standard supplement or modify those of EN 45502-1:1997, *Active Implantable Medical Devices – Part 1: General requirements for safety, marking and for the information to be provided by the manufacturer*, hereinafter referred to as the General Standard. The requirements of this particular standard take priority over those of the General Standard.

Figures or tables that are additional to those of the General Standard are numbered starting from 101; additional annexes are lettered AA, BB, etc.

Annex DD describes a coding system that may be used to designate tachyarrhythmia therapy modes. Annex EE defines the tissue equivalent interface circuits and low pass filter required for some compliance tests. Annex FF describes a method for selecting the filter capacitor used in the tissue equivalent interface circuits defined by Annex EE. Annex GG defines the method of calibrating the injection network defined by Annex EE. All annexes except Annex EE and Annex GG are informative.

1 Scope

This Part 2-2 of EN 45502 specifies requirements that are applicable to IMPLANTABLE CARDIOVERTER DEFIBRILLATORS and the functions of ACTIVE IMPLANTABLE MEDICAL DEVICES intended to treat tachyarrhythmia.

The tests that are specified in EN 45502 are type tests and are to be carried out on samples of a device to show compliance.

This part of EN 45502 is also applicable to some non-implantable parts and accessories of the devices (see Note 1).

The characteristics of the IMPLANTABLE PULSE GENERATOR or LEAD shall be determined by either the appropriate method detailed in this particular standard or by any other method demonstrated to have accuracy equal to, or better than, the method specified. In the case of dispute, the method detailed in this Particular Standard shall apply.

Any aspect of an ACTIVE IMPLANTABLE MEDICAL DEVICE intended to treat bradyarrhythmias is covered by EN 45502-2-1 *Particular requirements for active implantable medical devices intended to treat bradyarrhythmia (Pacemakers)*.

NOTE 1 The device that is commonly referred to as an ACTIVE IMPLANTABLE MEDICAL DEVICE may in fact be a single device, a combination of devices, or a combination of a device or devices and one or more accessories. Not all of these parts are required to be either partially or totally implantable, but there is a need to specify some requirements of non-implantable parts and accessories if they could affect the safety or performance of the implantable device.

NOTE 2 The terminology used in this European Standard is intended to be consistent with the terminology of Directive 90/385/EEC.

NOTE 3 In this European Standard, terms printed in small capital letters are used as defined in Clause 3. Where a defined term is used as a qualifier in another term, it is not printed in small capital letters unless the concept thus qualified is also defined.

NOTE 4 Particular requirements for congestive heart failure devices are under consideration. These types of devices are not covered by this standard.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

This clause of the General Standard applies except as follows:

Additional references:

<u>Publication</u>	<u>Year</u>	<u>Title</u>
EN 980		Graphical symbols for use in the labelling of medical devices
EN 28601		Data elements and interchange formats – Information interchange – Representation of dates and times (ISO 8601 + technical corrigendum 1)
EN 45502-1		Active implantable medical devices – Part 1: General requirements for safety, marking and information to be provided by the manufacturer
EN 45502-2-1		Active implantable medical devices – Part 2-1: Particular requirements for active implantable medical devices intended to treat bradyarrhythmia (cardiac pacemakers)

EN 60068-2-27	Environmental testing – Part 2-27: Tests – Test Ea and guidance: Shock (IEC 60068–2–27)
EN 60068-2-47	Environmental testing – Part 2-47: Tests – Mounting of specimens for vibration, impact and similar dynamic tests (IEC 60068-2-47)
EN 60068-2-64	Environmental testing – Part 2: Test methods – Test Fh: Vibration, broad-band random (digital control) and guidance (IEC 60068-2-64)
IEC 60878	Graphical symbols for electrical equipment in medical practice
ISO 5841-3 + corr. 1	Implants for surgery - Cardiac pacemakers - Part 3: Low-profile connectors (IS-1) for implantable pacemakers
ISO 11318	Cardiac defibrillators – Connector assembly DF-1 for implantable defibrillators – Dimensions and test requirements
ANSI/AAMI PC69	Active implantable medical devices – Electromagnetic compatibility – EMC test protocols for implantable cardiac pacemakers and implantable cardioverter defibrillators

3 Definitions

This clause of the General Standard applies except as follows:

Additional definitions:

3.3.1

adaptor

special connector used between an otherwise incompatible IMPLANTABLE PULSE GENERATOR and a LEAD

3.3.2

implantable cardioverter defibrillator (ICD)

ACTIVE IMPLANTABLE MEDICAL DEVICE comprising an IMPLANTABLE PULSE GENERATOR and LEAD(S) that is intended to detect and correct tachycardias and fibrillation by application of CARDIOVERSION/-DEFIBRILLATION PULSE(S) to the heart

3.3.3

implantable pulse generator (IPG)

part of the ACTIVE IMPLANTABLE MEDICAL DEVICE, including the power supply and electronic circuit that produces an electrical output

NOTE For purposes of this Particular Standard, the term IMPLANTABLE PULSE GENERATOR describes any ACTIVE IMPLANTABLE MEDICAL DEVICE that incorporates functions intended to treat tachyarrhythmias.

3.3.4

sensitivity (sensing threshold)

minimum signal required to control consistently the function of the IMPLANTABLE PULSE GENERATOR [see 6.1.5]

3.3.5

sensor

special part of an IMPLANTABLE PULSE GENERATOR that is designed to detect signals for the purpose of RATE MODULATION or other control purposes

3.5.1

electrode

electrically conducting part (usually the termination of a LEAD), which is designed to form an interface with body tissue or body fluid