

**TERVISHOIUTOODETE STERILISEERIMINE.
MIKROBIOLOOGILISED MEETODID. OSA 1: MIKROOBSE
POPULATSIOONI MÄÄRAMINE TOODETEL**

**Sterilization of health care products - Microbiological
methods - Part 1: Determination of a population of
microorganisms on products (ISO 11737-1:2018 +
ISO 11737-1:2018/Amd 1:2021)**

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 11737-1:2018+A1:2021 sisaldab Euroopa standardi EN ISO 11737-1:2018 ja selle muudatuse A1:2021 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 11737-1:2018+A1:2021 consists of the English text of the European standard EN ISO 11737-1:2018 and its amendment A1:2021.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas. Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 31.01.2018, muudatus A1 16.06.2021.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation. Date of Availability of the European standard is 31.01.2018, for A1 16.06.2021.
Muudatusega A1 lisatud või muudetud teksti algus ja lõpp on tekstis tähistatud sümbolitega A_1 A_1 . Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The start and finish of text introduced or altered by amendment A1 is indicated in the text by tags A_1 A_1 . The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 07.100.10; 11.080.01

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 11737-1 + A1

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2018, June 2021

ICS 07.100.10; 11.080.01

Supersedes EN ISO 11737-1:2006

English Version

Sterilization of health care products - Microbiological methods - Part 1: Determination of a population of microorganisms on products (ISO 11737-1:2018 + ISO 11737-1:2018/Amd 1:2021)

Stérilisation des produits de santé - Méthodes microbiologiques - Partie 1: Détermination d'une population de micro-organismes sur des produits (ISO 11737-1:2018 + ISO 11737-1:2018/Amd 1:2021)

Sterilisation von Produkten für die Gesundheitsfürsorge - Mikrobiologische Verfahren - Teil 1: Bestimmung der Population von Mikroorganismen auf Produkten (ISO 11737-1:2018 + ISO 11737-1:2018/Amd 1:2021)

This European Standard was approved by CEN on 6 December 2017. Amendment A1 was approved by CEN on 28 September 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard and its amendment the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard and its Amendment A1 exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 11737-1:2018) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2018, and conflicting national standards shall be withdrawn at the latest by July 2018.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11737-1:2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive(s), see informative Annex ZA, ZB, and ZC, which are integral parts of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the edition of the referenced document (including any amendments) listed below applies. For dated references, only the edition cited applies. However, for any use of this standard within the meaning of Annex ZA, ZB or ZC, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this should be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard as listed below.

NOTE 1 The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table – Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 10012	EN ISO 10012:2003	ISO 10012:2003
ISO 13485	EN ISO 13485:2016	ISO 13485:2016
ISO 15189	EN ISO 15189:2012	ISO 15189:2012
ISO/IEC 17025	EN ISO/IEC 17025:2017	ISO/IEC 17025:2017

NOTE 2 Many of the standards normatively referred to by ISO 11737-1 are undated. These referred standards also include normative references themselves to other dated and undated standards. For undated normative references, it should always be assumed that the latest edition applies.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11737-1:2018 has been approved by CEN as EN ISO 11737-1:2018 without any modification.

This document is a preview generated by EVS

A1) Amendment A1 European foreword

This document (EN ISO 11737-1:2018/A1:2021) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 204 "Sterilization of medical devices" the secretariat of which is held by BSI.

This Amendment to the European Standard EN ISO 11737-1:2018 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2021, and conflicting national standards shall be withdrawn at the latest by December 2021.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s) and Regulations, see informative Annex ZA and ZB, which are an integral part of this document.

The following referenced documents are indispensable for the application of this document. For undated references, the edition of the referenced document (including any amendments) listed below applies. For dated references, only the edition cited applies. However, for any use of this standard within the meaning of Annex ZA or ZB, the user should always check that any referenced document has not been superseded and that its relevant contents can still be considered the generally acknowledged state-of-art.

When an IEC or ISO standard is referred to in the ISO standard text, this should be understood as a normative reference to the corresponding EN standard, if available, and otherwise to the dated version of the ISO or IEC standard as listed below.

NOTE The way in which these referenced documents are cited in normative requirements determines the extent (in whole or in part) to which they apply.

Table – Correlation between normative references and dated EN and ISO standards

Normative references as listed in Clause 2 of the ISO standard	Equivalent dated standard	
	EN	ISO
ISO 10012	EN ISO 10012:2003	ISO 10012:2003
ISO 13485	EN ISO 13485:2016	ISO 13485:2016
ISO 15189	EN ISO 15189:2012	ISO 15189:2012
ISO/IEC 17025	EN ISO/IEC 17025:2017	ISO/IEC 17025:2017

NOTE The standards normatively referred to by ISO 11737-1:2018/Amd 1:2021 are undated. These referred standards also include normative references to other dated and undated standards. For undated normative references, it should always be assumed that the latest edition applies.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11737-1:2018/Amd1:2021 has been approved by CEN as EN ISO 11737-1:2018/A1:2021 without any modification. ^(A1)

This document is a preview generated by EVS

This document is a preview generated by EVS

Contents

Foreword	iv
[A1] Amendment A1 foreword [A1]	vi
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 [A1] General [A1]	5
5 Selection of products	5
5.1 General.....	5
5.2 Sample item portion (SIP)	5
6 Methods of determination and microbial characterization of bioburden	6
6.1 Determination of bioburden.....	6
6.1.1 Selection of an appropriate method.....	6
6.1.2 Neutralization of inhibitory substances	6
6.1.3 Removal of microorganisms.....	6
6.1.4 Culturing of microorganisms	7
6.1.5 Enumeration of microorganisms.....	7
6.2 Microbial characterization of bioburden	7
7 Validation of the method for determining bioburden	7
7.1 General.....	7
7.2 Validation	8
8 Routine determination of bioburden and interpretation of data	8
8.1 General.....	8
8.2 Limits of detection and plate counting.....	8
8.3 Microbial characterization	8
8.4 Bioburden data for extent of treatment.....	8
8.5 Bioburden spikes.....	8
8.6 Bioburden levels.....	8
8.7 Data analysis	8
8.8 Statistical methods	9
9 Maintenance of the method for determining bioburden	9
9.1 Changes to the product and/or manufacturing process	9
9.2 Changes to the method for determining bioburden	9
9.3 Requalification of the method for determining bioburden	9
Annex A (informative) Guidance on the determination of a population of microorganisms on products	10
Annex B (informative) Guidance on methods to determine bioburden	27
Annex C (informative) Validation of bioburden recovery efficiency	37
Annex D (informative) Typical assignment of responsibilities	45
[A1] Annex ZA (informative) Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered [A1]	47
[A1] Annex ZB (informative) Relationship between this European standard and the General Safety and Performance Requirements of Regulation (EU) 2017/746 aimed to be covered [A1]	50
Bibliography	52

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

This third edition cancels and replaces the second edition (ISO 11737-1:2006), which has been technically revised. It also incorporates the Technical Corrigendum ISO 11737-1:2006/Cor.1:2007.

The main changes compared to the previous edition are as follows:

- the term “bioburden spikes” has been introduced as a normal and consistent part of the bioburden, and examples of data have been provided;
- clarification has been added that package testing is not typically done except when it is an integral part of the product;
- more information has been provided on the most probable number (MPN) technique and its applications;
- details have been provided on ways to improve limit of detection (LOD) and correct use of the data;
- some discussion has been deleted of statistical methods for the evaluation of bioburden data where information was not typical or not required;
- a table has been added with criteria for selection of a bioburden recovery efficiency approach, the use of the correction factor (CF) has been explained, and the bioburden recovery efficiency value of < 50 % mentioned for technique modifications has been eliminated;
- more information has been provided on the application and performance of a bioburden method suitability test;

- a section has been added to detail rules for direct plate counts, estimated counts and counts beyond the ideal range;
- a table has been added to clarify where typical responsibilities reside for the manufacturer or the laboratory;
- the focus on a risk-based approach has been increased, including the purpose for which bioburden data will be used.

A list of all parts in the ISO 11737 series can be found on the ISO website.

A1 Amendment A1 foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html. **A1**

Introduction

A sterile health care product is one that is free of viable microorganisms. International Standards that specify requirements for the validation and routine control of sterilization processes require, when it is necessary to supply a sterile health care product, that adventitious microbiological contamination of a health care product prior to sterilization be minimized. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize health care products can generally best be described by an exponential relationship between the numbers of microorganisms surviving and the extent of treatment with the sterilizing agent. Inevitably, this means there is always a finite probability that a microorganism can survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the microorganisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization processing cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product item.

Generic requirements of the quality management system for design and development, production, installation and servicing are given in ISO 9001 and particular requirements for quality management systems for medical device production are given in ISO 13485. The standards for quality management systems recognize that, for certain processes used in manufacturing, the effectiveness of the process cannot be fully verified by subsequent inspection and testing of the product. Sterilization is an example of such a process. For this reason, sterilization processes are validated for use, the performance of the sterilization process is monitored routinely and the equipment is maintained.

International Standards specifying procedures for the validation and routine control of the processes used for the sterilization of health care products have been prepared (see, for example, ISO 14937, ISO 11135, the ISO 11137 series, the ISO 17665 series and ISO 14160). However, it is important to be aware that exposure to a properly validated and accurately controlled sterilization process is not the only factor associated with the provision of assurance that the product is sterile and, in this respect, suitable for its intended use. Furthermore, for the effective validation and routine control of a sterilization process, it is important to be aware of the microbiological challenge that is presented in the process, in terms of number, characteristics and properties of microorganisms.

The term “bioburden” is used to describe the population of viable microorganisms present on or in a product and/or a sterile barrier system. A knowledge of bioburden can be used in a number of situations as part of the following:

- validation and requalification of sterilization processes;
- routine monitoring for control of manufacturing processes;
- monitoring of raw materials, components or packaging;
- assessment of the efficiency of cleaning processes;
- an overall environmental monitoring programme.

Bioburden is the sum of the microbial contributions from a number of sources, including raw materials, manufacturing of components, assembly processes, manufacturing environment, assembly/manufacturing aids (e.g. compressed gases, water, lubricants), cleaning processes and packaging of finished products. To control bioburden, attention should be given to the microbiological status of these sources.

It is not possible to enumerate bioburden exactly and, in practice, a determination of bioburden is made using a defined method. Definition of a single method for use in determining bioburden in all situations is not practicable because of the wide variety of designs and materials of construction of health care products. Nor is it possible to define a single technique to be used in all situations for the removal of microorganisms in preparation for enumeration. Furthermore, the selection of culture conditions for enumeration of microorganisms will be influenced by the types of microorganism likely to be present on or in health care products.

This document specifies the requirements to be met for the determination of bioburden. In addition, it gives guidance in the annexes to provide explanations and methods that are deemed suitable to conform with the requirements. Methods other than those given in the guidance may be used, if they are effective in achieving conformity with the requirements of this document.

Sterilization of health care products — Microbiological methods —

Part 1:

Determination of a population of microorganisms on products

1 Scope

This document specifies requirements and provides guidance on the enumeration and microbial characterization of the population of viable microorganisms on or in a health care product, component, raw material or package.

NOTE 1 The nature and extent of microbial characterization is dependent on the intended use of bioburden data.

NOTE 2 See Annex A for guidance on Clauses 1 to 9.

This document does not apply to the enumeration or identification of viral, prion or protozoan contaminants. This includes the removal and detection of the causative agents of spongiform encephalopathies, such as scrapie, bovine spongiform encephalopathy and Creutzfeldt-Jakob disease.

NOTE 3 Guidance on inactivating viruses and prions can be found in ISO 22442-3, ICH Q5A(R1) and ISO 13022.

This document does not apply to the microbiological monitoring of the environment in which health care products are manufactured.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10012, *Measurement management systems — Requirements for measurement processes and measuring equipment*

ISO 13485, *Medical devices — Quality management systems — Requirements for regulatory purposes*

ISO 15189, *Medical laboratories — Requirements for quality and competence*

ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— IEC Electropedia: available at <http://www.electropedia.org/>

— ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

batch

defined quantity of a *product* (3.16) intended or purported to be uniform in character and quality, which has been produced during a defined cycle of manufacture