

TOORNAFTA, NAFTASAADUSED JA SAMAVÄÄRSED
TOOTED. TIHEDUSE MÄÄRAMINE. LABORI
TIHEDUSMÕÕTUR OSTSILLEERUVA U-TORU
SENSORIGA

Crude petroleum, petroleum products and related
products - Determination of density - Laboratory
density meter with an oscillating U tube sensor (ISO
12185:2024)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>See Eesti standard EVS-EN ISO 12185:2024 sisaldab Euroopa standardi EN ISO 12185:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 27.03.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 12185:2024 consists of the English text of the European standard EN ISO 12185:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 27.03.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 75.180.30

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 12185

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2024

ICS 75.180.30

Supersedes EN ISO 12185:1996

English Version

Crude petroleum, petroleum products and related products - Determination of density - Laboratory density meter with an oscillating U tube sensor (ISO 12185:2024)

Pétroles bruts, produits pétroliers et produits connexes - Détermination de la masse volumique - Appareil de masse volumique de laboratoire à capteur à tube en U oscillant (ISO 12185:2024)

Rohöl und Mineralölerzeugnisse - Bestimmung der Dichte - U-Rohr-Oszillationsverfahren (ISO 12185:2024)

This European Standard was approved by CEN on 23 February 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 12185:2024) has been prepared by Technical Committee ISO/TC 28 "Petroleum and related products, fuels and lubricants from natural or synthetic sources" in collaboration with Technical Committee CEN/TC 19 "Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2024, and conflicting national standards shall be withdrawn at the latest by September 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 12185:1996.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 12185:2024 has been approved by CEN as EN ISO 12185:2024 without any modification.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Principle	3
5 Apparatus	3
5.1 Density meter.....	3
5.2 Homogenizer.....	3
5.3 Constant-temperature bath.....	3
6 Reagents and materials	3
6.1 Flushing solvent.....	3
6.2 Adjustment liquids.....	4
7 Sampling	4
8 Sample preparation	5
9 Apparatus preparation	5
9.1 Test temperature.....	5
9.2 Cell cleaning.....	5
9.3 Meter verification and adjustment.....	6
9.4 Meter calibration.....	6
9.5 Quality control checks.....	6
10 Test procedure	6
11 Calculation	7
12 Test report	7
13 Precision	7
13.1 Repeatability, <i>r</i>	7
13.2 Reproducibility, <i>R</i>	8
Annex A (informative) Meter calibration	9
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 28, *Petroleum and related products, fuels and lubricants from natural or synthetic sources*, Subcommittee SC 2, *Measurement of petroleum and related products*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 19, *Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 12185:1996), which has been technically revised. It also incorporates the Technical Corrigendum ISO 12185:1996/Cor 1:2001.

The main changes are as follows:

- definitions have been added in [Clause 3](#);
- a quality control (QC) check has been added in [9.5](#).

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

The first edition of this standard (ISO 12185:1996) was written at a time when there were relatively few models of density meter with an oscillating U-tube sensor on the market.

There are now a considerable number of different manufacturers and models of laboratory density meter available worldwide, many of which use different methodologies or algorithms to cope with the effect of viscosity on displayed density.

This document therefore encompasses a wider range of instruments than those covered in the first edition and gives guidance and requirements for accurate density analyses, such as apparatus and apparatus preparation (see [Clauses 5](#) and [9, Annex A](#)).

This document is a preview generated by EVS

Crude petroleum, petroleum products and related products — Determination of density — Laboratory density meter with an oscillating U-tube sensor

1 Scope

This document specifies a method for the determination, using an oscillating U-tube density meter, of the density of crude petroleum and related products within the range 600 kg/m^3 to $1\,100 \text{ kg/m}^3$, which can be handled as single-phase liquids at the test temperature and pressure.

This document is applicable to liquids of any vapour pressure as long as suitable precautions are taken to ensure that they remain in single phase. Loss of light components leads to changes in density during both the sample handling and the density determination.

This method is not intended for use with in-line density meters.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 91, *Petroleum and related products — Temperature and pressure volume correction factors (petroleum measurement tables) and standard reference conditions*

ISO 3015, *Petroleum and related products from natural or synthetic sources — Determination of cloud point*

ISO 3016, *Petroleum and related products from natural or synthetic sources — Determination of pour point*

ISO 3170, *Petroleum liquids — Manual sampling*

ISO 3171, *Petroleum liquids — Automatic pipeline sampling*

IP 389: *Determination of wax appearance temperature (WAT) of middle distillate fuels by differential thermal analysis (DTA) or differential scanning calorimetry (DSC)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>

— IEC Electropedia: available at <https://www.electropedia.org/>

3.1 density

ρ

mass of liquid per unit volume at a specified temperature

Note 1 to entry: This is usually the mass of liquid expressed in kilograms, divided by its volume, expressed in cubic metres. The unit of measurement can be displayed as either $\text{kg}\cdot\text{m}^{-3}$ or kg/m^3 . When quoting liquid density, the temperature at which it has been measured shall also be quoted (e.g. $840,0 \text{ kg/m}^3$ at $20,1 \text{ }^\circ\text{C}$).