

IN VITRO DIAGNOSTIKAMEDITSIINISEADMED. TOOTJA  
ANTAV TEAVE (ETIKETTIMINE). OSA 2: IN VITRO  
DIAGNOSTIKA REAGENDID PROFESSIONAALSEKS  
KASUTUSEKS

In vitro diagnostic medical devices - Information  
supplied by the manufacturer (labelling) - Part 2: In  
vitro diagnostic reagents for professional use (ISO  
18113-2:2022)

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

<p>See Eesti standard EVS-EN ISO 18113-2:2024 sisaldab Euroopa standardi EN ISO 18113-2:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 12.06.2024.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 18113-2:2024 consists of the English text of the European standard EN ISO 18113-2:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 12.06.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 11.100.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht [www.evs.ee](http://www.evs.ee); telefon 605 5050; e-post [info@evs.ee](mailto:info@evs.ee)

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage [www.evs.ee](http://www.evs.ee); phone +372 605 5050; e-mail [info@evs.ee](mailto:info@evs.ee)

EUROPEAN STANDARD

EN ISO 18113-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2024

ICS 11.100.10

Supersedes EN ISO 18113-2:2011

English Version

**In vitro diagnostic medical devices - Information supplied  
by the manufacturer (labelling) - Part 2: In vitro diagnostic  
reagents for professional use (ISO 18113-2:2022)**

Dispositifs médicaux de diagnostic in vitro -  
Informations fournies par le fabricant (étiquetage) -  
Partie 2: Réactifs de diagnostic in vitro à usage  
professionnel (ISO 18113-2:2022)

In-vitro-Diagnostika - Bereitstellung von  
Informationen durch den Hersteller - Teil 2: In-vitro-  
diagnostische Reagenzien für den Gebrauch durch  
Fachpersonal (ISO 18113-2:2022)

This European Standard was approved by CEN on 2 October 2022.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## European foreword

This document (EN ISO 18113-2:2024) has been prepared by Technical Committee ISO/TC 212 "Medical laboratories and in vitro diagnostic systems" in collaboration with Technical Committee CEN/TC 140 "In vitro diagnostic medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2024, and conflicting national standards shall be withdrawn at the latest by June 2027.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 18113-2:2011.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

## Endorsement notice

The text of ISO 18113-2:2022 has been approved by CEN as EN ISO 18113-2:2024 without any modification.

# Contents

	Page
<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General</b> .....	<b>1</b>
4.1 Essential requirements.....	1
4.2 Identification of kit components.....	2
<b>5 Content of the outer container label</b> .....	<b>2</b>
5.1 Manufacturer.....	2
5.2 Identification of the in vitro diagnostic (IVD) reagent.....	2
5.2.1 IVD reagent name.....	2
5.2.2 Batch code/lot number.....	2
5.2.3 Unique device identifier (UDI).....	2
5.3 Contents.....	3
5.4 Intended use/Intended purpose.....	3
5.5 In vitro diagnostic use.....	3
5.6 Storage, transport, and handling conditions.....	3
5.7 Expiry date.....	3
5.8 Warnings and precautions.....	4
<b>6 Content of the immediate container label</b> .....	<b>4</b>
6.1 General provisions.....	4
6.1.1 Single container.....	4
6.1.2 Small label.....	4
6.2 Manufacturer.....	4
6.3 Identification of the IVD reagent.....	4
6.3.1 IVD reagent or component name.....	4
6.3.2 Batch code/lot number.....	4
6.3.3 Unique device identifier (UDI).....	5
6.4 Contents.....	5
6.5 In vitro diagnostic use.....	5
6.6 Storage and handling conditions.....	5
6.7 Expiry date.....	5
6.8 Warnings and precautions.....	5
<b>7 Content of the instructions for use</b> .....	<b>5</b>
7.1 Manufacturer.....	5
7.2 Identification of the IVD reagent.....	6
7.3 Intended use/intended purpose.....	6
7.4 Principles of the examination method.....	6
7.5 Traceability of values assigned to calibrators and trueness-control materials.....	7
7.6 Components.....	7
7.7 Additional required equipment and/or materials.....	7
7.8 Reagent preparation.....	7
7.9 Storage and shelf life after first opening.....	8
7.10 Warnings and precautions and/or measures to be taken and limitations of use regarding the device.....	8
7.11 Primary sample collection, handling, and storage.....	8
7.12 Examination procedure.....	9
7.13 Control procedure.....	9
7.14 Calculation of examination results.....	9
7.15 Interpretation of results.....	9
7.16 Performance characteristics.....	9

7.16.1	Analytical performance characteristics .....	9
7.16.2	Clinical performance characteristics.....	10
7.16.3	Measuring interval.....	10
7.17	Biological reference intervals .....	10
7.18	Limitations of the examination procedure .....	10
7.19	Literature references.....	10
7.20	Document control .....	10
<b>Bibliography</b>	.....	<b>11</b>

This document is a preview generated by EVS

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 212, *Clinical laboratory testing and in vitro diagnostic test systems*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 140, *In vitro diagnostic medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 18113-2:2009), which has been technically revised.

The main changes are as follows:

- Added Information pertaining to (unique device identifier-device identifier) UDI;
- Updated with examples to reference European Union and other regulations;
- Added additional detail for clarification;
- Updated the Bibliography.

A list of all parts in the ISO 18113 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

## Introduction

Manufacturers of in vitro diagnostic (IVD) reagents for professional use, supply users with information to enable the safe use and the expected performance of their devices. The type and level of detail varies according to the intended uses and country-specific regulations.

The International Medical Devices Regulators Forum (IMDRF) encourages convergence of the evolution of regulatory systems for medical devices at the global level. Eliminating differences among regulatory jurisdictions can allow patients earlier access to new technologies and treatments. This document provides a basis for harmonization of labelling requirements for IVD reagents for professional use.

This document is concerned solely with information supplied with IVD reagents, calibrators and control materials intended for professional use. It is intended to be used in conjunction with ISO 18113-1, which contains the general requirements for information supplied by the manufacturer and definitions of general labelling concepts.

This document is intended to support the essential labelling requirements of all the IMDRF<sup>[8]</sup> partners, as well as other countries that have or plan to enact labelling regulations for IVD medical devices.

For IVD reagents, calibrators and/or control materials that are intended to be used as a system with an instrument provided by the same manufacturer, this document is also intended to be used together with ISO 18113-1 and ISO 18113-3.

# In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) —

## Part 2:

## In vitro diagnostic reagents for professional use

### 1 Scope

This document specifies requirements for information supplied by the manufacturer of in vitro diagnostic (IVD) reagents, calibrators and controls intended for professional use.

This document can also be applicable to accessories.

This document is applicable to the labels for outer and immediate containers and to the instructions for use.

This document does not apply to:

- a) IVD instruments or equipment;
- b) IVD reagents for self-testing.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 8601-1, *Date and time — Representations for information interchange — Part 1: Basic rules*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 15223-1, *Medical devices — Symbols to be used with information to be supplied by the manufacturer — Part 1: General requirements*

ISO 18113-1, *In vitro diagnostic medical devices — Information supplied by the manufacturer (labelling) — Part 1: Terms, definitions and general requirements*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18113-1 apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

### 4 General

#### 4.1 Essential requirements

The requirements of ISO 18113-1 apply.