

This document is a preview generated by EVS

Ophthalmic instruments - Fundamental requirements and test methods - Part 2: Light hazard protection (ISO 15004-2:2024)

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

<p>See Eesti standard EVS-EN ISO 15004-2:2024 sisaldab Euroopa standardi EN ISO 15004-2:2024 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 11.12.2024.</p> <p>Standard on kättesaadav Eesti Standardimis-ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 15004-2:2024 consists of the English text of the European standard EN ISO 15004-2:2024.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 11.12.2024.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
---	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.040.70

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis-ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis-ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation

No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation:

Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 15004-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2024

ICS 11.040.70

Supersedes EN ISO 15004-2:2007

English Version

Ophthalmic instruments - Fundamental requirements and test methods - Part 2: Light hazard protection (ISO 15004-2:2024)

Instruments ophtalmiques - Exigences fondamentales et méthodes d'essai - Partie 2: Protection contre les dangers de la lumière (ISO 15004-2:2024)

Ophthalmische Instrumente - Grundlegende Anforderungen und Prüfverfahren - Teil 2: Schutz gegen Gefährdung durch Licht (ISO 15004-2:2024)

This European Standard was approved by CEN on 18 November 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 15004-2:2024) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 15004-2:2007.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 15004-2:2024 has been approved by CEN as EN ISO 15004-2:2024 without any modification.

Contents

Page

Foreword	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Classification principles	7
4.1 General.....	7
4.2 Group 1 instruments.....	9
4.3 Group 2 instruments.....	9
5 Requirements for classification	9
5.1 General.....	9
5.2 Requirements for classification as a Group 1 instrument.....	9
5.3 Requirements for classification as a Group 2 instrument.....	10
5.4 Exposure limits for Group 1 classification.....	10
5.4.1 Continuous wave instruments.....	10
5.4.2 Pulsed instruments.....	10
5.4.3 Time-limited instruments.....	11
5.4.4 Dose limited instruments.....	11
5.4.5 Scanning instruments.....	11
5.4.6 Instruments with multiple sources.....	12
5.4.7 Instruments for long-term repetitive daily use.....	12
5.5 Recommended maximum exposure (RME) values for Group 2 instruments.....	12
5.5.1 Continuous wave instruments.....	12
5.5.2 Pulsed instruments.....	12
5.5.3 Time-limited instruments.....	13
5.5.4 Scanning instruments.....	13
5.5.5 Instruments with multiple sources.....	13
5.5.6 Instruments for long-term repetitive daily use.....	14
5.6 Exposure limits and maximum recommended exposure values.....	14
6 Test methods	19
6.1 General.....	19
6.2 Classification of instruments into Group 1 or Group 2.....	19
6.3 Spectral measurements.....	20
6.4 Group 2 instruments: determination of time and number of pulses to reach recommended maximum exposure.....	20
6.4.1 Determination of time t_{\max} to reach the recommended maximum exposure for weighted corneal and lenticular ultraviolet radiation radiant exposure, H_{S-CL}	20
6.4.2 Determination of time t_{\max} to reach the recommended maximum exposure for photochemical aphakic retinal radiant exposure.....	20
6.4.3 Determination of the number of pulses necessary to reach the recommended maximum exposure for photochemical aphakic retinal exposure, n_{\max} (for pulsed instruments).....	21
7 Information supplied by the manufacturer	21
7.1 Information provided on request.....	21
7.2 Accompanying documents.....	21
7.2.1 General.....	21
7.2.2 Dose-limited instruments.....	21
7.2.3 Cautionary statements.....	21
7.3 Marking.....	25
7.3.1 Radiation output.....	25
7.3.2 Consult accompanying documents.....	25
7.3.3 Safety information for optical radiation.....	25

Annex A (normative) Spectral weighting functions	26
Annex B (informative) Measurement instruments	33
Annex C (informative) Measurement methods for radiance/irradiance	37
Annex D (informative) Guidance on the direct measurement of irradiance	42
Annex E (informative) Examples of measurement methods for specific ophthalmic instruments	44
Annex F (informative) Safety signs	55
Bibliography	56

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 170, *Ophthalmic optics*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 15004-2:2007), which has been technically revised.

The main changes are as follows:

- The terms and definitions include dose limited instruments and time limited instruments;
- The safe exposure limits have been reorganized into 4 tables ([Tables 2 to 5](#)), and the associated measurement conditions have been reorganized into a companion table ([Table 6](#));
- A number of Group 1 exposure limits and Group 2 recommended maximum exposures (RMEs) have been updated to conform to recent research and relevant standards;
- The language has been clarified and simplified throughout the document, and a flowchart has been added as a guide to make the standard more accessible to first-time users;
- Clauses and associated exposure limits have been added for long-term repetitive exposures, such as may apply to extensive use of head-mounted displays by people with visual impairments;
- Provisions have been added to ensure that the exposure limits and RMEs applicable to specific devices are easily accessible to end users.

A list of all parts in the ISO 15004 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

Introduction

Ophthalmic instruments are classified into two groups to distinguish instruments that can present a potential hazard from those that cannot. The two groups are named Group 1 and Group 2. They are defined as follows:

Group 1 instruments: ophthalmic instruments for which no potential light hazard exists when used as intended (see [Clause 4](#)).

Group 2 instruments: ophthalmic instruments for which a potential light hazard exists (see [Clause 4](#)).

Limits and guidelines for optical radiation exposure of the eye during ophthalmic examination can differ from those of non-ophthalmic applications. They can be more restrictive because of pupils dilation or retinal image stabilization, or less restrictive based on benefit/risk ratios. Furthermore, interruptions of exposure during surgical procedures mitigate the risk.

All group 2 instruments pose a potential risk of injury at the upper emission values of the instrument. This is true for both photochemical (where time is critical) and thermal, where transmission and absorption values can vary. Clinical judgement of individual susceptibility is also required.

NOTE 1 The basic limits and guidelines in this document are based on the International Commission on Non-Ionizing Radiation Protection (ICNIRP) guidelines for human exposure to optical radiation. They also cover cases where pupils are dilated, or the image is stabilized on the retina during ophthalmic examinations.

NOTE 2 This document provides exposure limits for ocular tissues. The exposures can be calculated based upon the measured instrument emissions.

The flow chart in [Figure 1](#) at the beginning of [Clause 4](#) provides guidance on how to apply this document to any device to be tested or designed for light hazards conformity.

Ophthalmic instruments — Fundamental requirements and test methods —

Part 2: Light hazard protection

1 Scope

This document specifies fundamental requirements for optical radiation safety for ophthalmic instruments and is applicable to all ophthalmic instruments that direct optical radiation into or at the eye. It is also applicable to all new and emerging ophthalmic instruments that direct optical radiation into or at the eye, as well as to those portions of therapeutic or surgical systems that direct optical radiation into or at the eye for diagnostic, illumination, measurement, imaging or alignment purposes.

NOTE For the purpose of this document, optical radiation relates to the wavelength range of 250 nm to 2 500 nm.

This document does not apply to therapeutic radiation. However, in the case of the treatment beams of therapeutic devices, when conducting risk assessments for non-target tissues, the limits given in this document may be applied to those parts of the treatment beam that strike non-target tissue.

Where vertical (instrument-specific) International Standards contain specific light hazard requirements different from those given in ISO 15004-2, then those in the vertical International Standard take precedence.

This document classifies ophthalmic instruments into either Group 1 or Group 2 to distinguish instruments that are non-hazardous from those that are potentially hazardous.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 15004-1, *Ophthalmic instruments — Fundamental requirements and test methods — Part 1: General requirements applicable to all ophthalmic instruments*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 15004-1 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

aperture

aperture stop

opening that defines the area over which average optical emission is measured

Note 1 to entry: For spectral irradiance measurements this opening is usually the entrance of a small integrating sphere placed in front of the radiometer/spectroradiometer entrance slit.