

**Piston-operated volumetric apparatus -  
Part 7: Alternative measurement procedures for the  
determination of volume (ISO 8655-7:2022 +  
ISO 8655-7:2022/Amd 1:2024)**

## EESTI STANDARDI EESSÕNA

## NATIONAL FOREWORD

See Eesti standard EVS-EN ISO 8655-7:2022+A1:2024 sisaldab Euroopa standardi EN ISO 8655-7:2022 ja selle muudatuse A1:2024 ingliskeelset teksti.	This Estonian standard EVS-EN ISO 8655-7:2022+A1:2024 consists of the English text of the European standard EN ISO 8655-7:2022 and its amendment A1:2024.
Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.  Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 11.05.2022, muudatus A1 18.12.2024.	This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.  Date of Availability of the European standard is 11.05.2022, for A1 18.12.2024.
Muudatusega A1 lisatud või muudetud teksti algus ja lõpp on tekstis tähistatud sümbolitega $\boxed{A1}$ $\langle A1 \rangle$ .  Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.	The start and finish of text introduced or altered by amendment A1 is indicated in the text by tags $\boxed{A1}$ $\langle A1 \rangle$ .  The standard is available from the Estonian Centre for Standardisation and Accreditation.

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile [standardiosakond@evs.ee](mailto:standardiosakond@evs.ee).

ICS 17.060

<p><b>Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele</b></p> <p>Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.</p> <p>Kui Teil on küsimusi standardite autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht <a href="http://www.evs.ee">www.evs.ee</a>; telefon 605 5050; e-post <a href="mailto:info@evs.ee">info@evs.ee</a></p> <p><b>The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation</b></p> <p>No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.</p> <p>If you have any questions about standards copyright protection, please contact the Estonian Centre for Standardisation and Accreditation: Homepage <a href="http://www.evs.ee">www.evs.ee</a>; phone +372 605 5050; e-mail <a href="mailto:info@evs.ee">info@evs.ee</a></p>
--

EUROPEAN STANDARD

EN ISO 8655-7 + A1

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2022, December 2024

ICS 17.060

Supersedes EN ISO 8655-7:2005, EN ISO 8655-7:2005/AC:2009

English Version

**Piston-operated volumetric apparatus - Part 7: Alternative measurement procedures for the determination of volume (ISO 8655-7:2022 + ISO 8655-7:2022/Amd 1:2024)**

Appareils volumétriques à piston - Partie 7: Modes opératoires de mesure alternatifs pour la détermination de volumes (ISO 8655-7:2022 + ISO 8655-7:2022/Amd 1:2024)

Volumenmessgeräte mit Hubkolben - Teil 7: Alternatives Prüfverfahren zur Bestimmung des Volumens (ISO 8655-7:2022 + ISO 8655-7:2022/Amd 1:2024)

This European Standard was approved by CEN on 13 February 2022. Amendment A1 was approved by CEN on 17 December 2024.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard and its amendment the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard and its Amendment A1 exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

## European foreword

This document (EN ISO 8655-7:2022) has been prepared by Technical Committee ISO/TC 48 "Laboratory equipment" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2022, and conflicting national standards shall be withdrawn at the latest by November 2022.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 8655-7:2005, EN ISO 8655-7:2005/AC:2009.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Endorsement notice

The text of ISO 8655-7:2022 has been approved by CEN as EN ISO 8655-7:2022 without any modification.

## **A1** Amendment A1 European foreword

This document (EN ISO 8655-7:2022/A1:2024) has been prepared by Technical Committee ISO/TC 48 "Laboratory equipment" in collaboration with Technical Committee CEN/TC 332 "Laboratory equipment" the secretariat of which is held by DIN.

This Amendment to the European Standard EN ISO 8655-7:2022 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2025, and conflicting national standards shall be withdrawn at the latest by June 2025.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

### **Endorsement notice**

The text of ISO 8655-7:2022/Amd 1:2024 has been approved by CEN as EN ISO 8655-7:2022/A1:2024 without any modification. **A1**

This document is a preview generated by EVS

<b>Contents</b>		<b>Page</b>
<b>Foreword</b> .....		<b>v</b>
<b>[A1] Amendment A1 Foreword [A1]</b> .....		<b>vii</b>
<b>Introduction</b> .....		<b>viii</b>
<b>1</b>	<b>Scope</b> .....	<b>1</b>
<b>2</b>	<b>Normative references</b> .....	<b>1</b>
<b>3</b>	<b>Terms and definitions</b> .....	<b>2</b>
<b>4</b>	<b>General requirements</b> .....	<b>2</b>
	4.1 Metrological confirmation.....	2
	4.2 Uncertainty of measurement.....	2
	4.3 Operator qualification.....	2
<b>5</b>	<b>Performance requirements</b> .....	<b>3</b>
	5.1 Performance tolerances.....	3
	5.2 Operator impact.....	3
<b>6</b>	<b>Test conditions</b> .....	<b>3</b>
	6.1 General.....	3
	6.2 Test equipment.....	3
	6.3 Test room, environmental conditions.....	3
	6.4 Test volumes.....	4
	6.4.1 Fixed volume POVA.....	4
	6.4.2 Adjustable volume POVA.....	4
	6.5 Number of measurements per test volume.....	5
	6.6 Test liquids.....	5
<b>7</b>	<b>Evaluation</b> .....	<b>5</b>
	7.1 Mean volume.....	5
	7.2 Systematic error of measurement.....	5
	7.3 Random error of measurement.....	6
<b>8</b>	<b>Test methods</b> .....	<b>6</b>
	8.1 General.....	6
	8.2 Gravimetric method.....	7
	8.3 Dual-dye ratiometric photometric method.....	7
	8.4 Single dye photometric method.....	8
	8.5 Hybrid photometric/gravimetric method for multichannel POVA.....	8
	8.6 Titration method.....	8
	8.7 Batch testing.....	8
<b>9</b>	<b>Dispense procedures</b> .....	<b>8</b>
	9.1 General.....	8
	9.2 Preparation.....	9
	9.3 Single-channel air displacement pipettes (in accordance with ISO 8655-2).....	9
	9.3.1 General.....	9
	9.3.2 Test cycle.....	9
	9.4 Multi-channel pipettes (in accordance with ISO 8655-2).....	10
	9.5 Positive displacement pipettes (in accordance with ISO 8655-2).....	11
	9.6 Burettes (in accordance with ISO 8655-3).....	11
	9.7 Dilutors (in accordance with ISO 8655-4).....	12
	9.7.1 General.....	12

9.7.2	Test cycle.....	12
9.8	Dispensers (in accordance with ISO 8655-5) .....	13
9.9	Syringes (in accordance with ISO 8655-9) .....	13
9.9.1	General.....	13
9.9.2	Test cycle.....	13
<b>10</b>	<b>Reporting of results.....</b>	<b>14</b>
<b>Annex A</b> (normative)	<b>Gravimetric procedure.....</b>	<b>16</b>
<b>Annex B</b> (normative)	<b>Dual-dye ratiometric photometric procedure .....</b>	<b>21</b>
<b>Annex C</b> (normative)	<b>Single dye photometric procedure .....</b>	<b>28</b>
<b>Annex D</b> (normative)	<b>Photometric/gravimetric hybrid procedure .....</b>	<b>32</b>
<b>Annex E</b> (normative)	<b>Titrimetric procedure.....</b>	<b>41</b>
<b>Annex F</b> (normative)	<b>Conversion of liquid mass to volume.....</b>	<b>45</b>
<b>Bibliography</b> .....		<b>48</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 332, *Laboratory equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 8655-7:2005), which has been technically revised. It also incorporates the Technical Corrigendum ISO 8655-7:2005/Cor.1:2008.

The main changes are as follows:

- a gravimetric test method was added (see 8.2);
- a photometric/gravimetric hybrid test method was added (see 8.5);
- a batch testing method was added (see 8.7);
- measurement procedures for all methods are given in normative Annexes A to E;
- standard dispense procedures for POVA described in ISO 8655-2, ISO 8655-3, ISO 8655-4, ISO 8655-5, and ISO 8655-9 were added (see Clause 9);
- requirements for operator qualification have been added (see 4.3);
- requirements for testing of multi-channel POVA is described in more detail, with specific procedures given for these apparatus (see 8.5, and Annex D);
- Annexes A, B, and C of the first edition have been deleted and replaced.

A list of all parts in the ISO 8655 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html).

This document is a preview generated by EVS

## **A1** Amendment A1 Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at [www.iso.org/patents](http://www.iso.org/patents). ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see [www.iso.org/iso/foreword.html](http://www.iso.org/iso/foreword.html).

This document was prepared by Technical Committee ISO/TC 48, *Laboratory equipment*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 332, *Laboratory equipment*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 8655 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at [www.iso.org/members.html](http://www.iso.org/members.html). **A1**

## Introduction

The ISO 8655 series addresses the needs of:

- manufacturers, as a basis for quality control including, where appropriate, the issuance of manufacturer's declarations;
- calibration laboratories, test houses, users of the equipment and other bodies as a basis for independent calibration, testing, verification, and routine tests.

The tests specified in the ISO 8655 series are intended to be carried out by trained personnel.

# Piston-operated volumetric apparatus —

## Part 7:

# Alternative measurement procedures for the determination of volume

## 1 Scope

This document specifies alternative measurement procedures for the determination of volume of piston-operated volumetric apparatus.

The procedures are applicable to complete systems comprising the basic apparatus and all parts selected for use with the apparatus, disposable or reusable, involved in the measurement by delivery process (Ex). Methods described in this document are suitable for various maximum nominal volumes of piston-operated volumetric apparatus. It is the responsibility of the user to select the appropriate method.

## 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1042, *Laboratory glassware — One-mark volumetric flasks*

ISO 2859-1, *Sampling procedures for inspection by attributes — Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

ISO 3951-1, *Sampling procedures for inspection by variables — Part 1: Specification for single sampling plans indexed by acceptance quality limit (AQL) for lot-by-lot inspection for a single quality characteristic and a single AQL*

ISO 8655-1, *Piston-operated volumetric apparatus — Part 1: Terminology, general requirements and user recommendations*

ISO 8655-2, *Piston-operated volumetric apparatus — Part 2: Pipettes*

ISO 8655-3, *Piston-operated volumetric apparatus — Part 3: Burettes*

ISO 8655-4, *Piston-operated volumetric apparatus — Part 4: Dilutors*

ISO 8655-5, *Piston-operated volumetric apparatus — Part 5: Dispensers*

ISO 8655-6, *Piston-operated volumetric apparatus — Part 6: Gravimetric reference measurement procedure for the determination of volume*

ISO 8655-8, *Piston-operated volumetric apparatus — Part 8: Photometric reference measurement procedure for the determination of volume*

ISO 8655-9, *Piston-operated volumetric apparatus — Part 9: Manually operated precision laboratory syringes*

ISO/IEC Guide 2, *Standardization and related activities — General vocabulary*

ISO/IEC Guide 99, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 8655-1, ISO/IEC Guide 2, ISO/IEC Guide 99 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

#### 3.1 acceptance quality limit AQL

<acceptance sampling> worst tolerable quality level

Note 1 to entry: This concept only applies when a sampling scheme with rules for switching and for discontinuation, such as in ISO 2859-1, ISO 3951-1 or ISO 3951-5 is used.

Note 2 to entry: Although individual lots with quality as bad as the acceptance quality limit may be accepted with fairly high probability, the designation of an acceptance quality limit does not suggest that this is a desirable quality level. Sampling schemes found in International Standards such as ISO 2859-1, ISO 3951-1 or ISO 3951-5, with their rules for switching and for discontinuation of sampling inspection, are designed to encourage suppliers to have process averages consistently better than the AQL.

### 4 General requirements

#### 4.1 Metrological confirmation

Metrological confirmation of all POVA shall be performed on a regular basis to ensure the apparatus conforms to requirements for its intended use. The requirements of the methods and procedures described in this document are suitable to be used in the metrological confirmation of POVA. For calibrations and testing, no less than ten replicate measurements per selected volume shall be performed and the measurement procedures in this document shall be validated by comparison to one of the reference measurement procedures described in ISO 8655-6 or ISO 8655-8.

#### 4.2 Uncertainty of measurement

When performing calibrations (ISO 8655-1:2022, 6.4) according to measurement procedures described in this document, the expanded measurement uncertainty of the mean delivered volume for each selected volume shall be estimated and reported (see Clause 10 (m)).

When performing testing (ISO 8655-1:2022, 6.4) or routine tests (ISO 8655-1:2022, 6.5), it is optional to estimate and report the expanded measurement uncertainty.

NOTE For further information on uncertainty for the photometric and gravimetric methods, refer to ISO/TR 16153<sup>[1]</sup> and ISO/TR 20461<sup>[2]</sup> respectively.

#### 4.3 Operator qualification

An operator who uses POVA for volumetric transfers, performs metrological confirmation or routine tests of POVA shall be adequately trained on the use of the type of POVA under test. Operator training and competence should be documented.