

Environmental solid matrices - Determination of polychlorinated biphenyls (PCB) by gas chromatography - mass selective detection (GC-MS) or electron-capture detection (GC-ECD) (ISO 18475:2023)

EESTI STANDARDI EESSÕNA

NATIONAL FOREWORD

<p>See Eesti standard EVS-EN ISO 18475:2025 sisaldab Euroopa standardi EN ISO 18475:2025 ingliskeelset teksti.</p> <p>Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas.</p> <p>Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 16.07.2025.</p> <p>Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest.</p>	<p>This Estonian standard EVS-EN ISO 18475:2025 consists of the English text of the European standard EN ISO 18475:2025.</p> <p>This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation.</p> <p>Date of Availability of the European standard is 16.07.2025.</p> <p>The standard is available from the Estonian Centre for Standardisation and Accreditation.</p>
--	---

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 13.080.10

Standardite reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele. Andmete paljundamine, taastekitamine, kopeerimine, salvestamine elektroonsesse süsteemi või edastamine ükskõik millises vormis või millisel teel ilma Eesti Standardimis- ja Akrediteerimiskeskuse kirjaliku loata on keelatud.

Kui Teil on küsimusi standardite autorikaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega: Koduleht www.evs.ee; telefon 605 5050; e-post info@evs.ee

The right to reproduce and distribute standards belongs to the Estonian Centre for Standardisation and Accreditation. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying, without a written permission from the Estonian Centre for Standardisation and Accreditation.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 18475

NORME EUROPÉENNE

EUROPÄISCHE NORM

July 2025

ICS 13.080.10

Supersedes EN 17322:2020

English Version

Environmental solid matrices - Determination of polychlorinated biphenyls (PCB) by gas chromatography - mass selective detection (GC-MS) or electron-capture detection (GC-ECD) (ISO 18475:2023)

Matrices solides environnementales - Dosage des polychlorobiphényles (PCB) par chromatographie en phase gazeuse-spectrométrie de masse (CG-SM) ou chromatographie en phase gazeuse avec détection par capture d'électrons (CG-ECD) (ISO 18475:2023)

Feststoffe in der Umwelt - Bestimmung von polychlorierten Biphenylen (PCB) mittels Gaschromatographie und massenspektrometrischer Detektion (GC-MS) oder Elektronen-Einfang-Detektion (GC-ECD) (ISO 18475:2023)

This European Standard was approved by CEN on 13 July 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

The text of ISO 18475:2023 has been prepared by Technical Committee ISO/TC 190 "Soil quality" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 18475:2025 by Technical Committee CEN/TC 444 "Environmental characterization of solid matrices" the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2026, and conflicting national standards shall be withdrawn at the latest by January 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 17322:2020.

Any feedback and questions on this document should be directed to the users' national standards body. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 18475:2023 has been approved by CEN as EN ISO 18475:2025 without any modification.

Contents

Page

Foreword	v
Introduction	vii
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	3
5 Interferences	4
5.1 Interference with sampling and extraction.....	4
5.2 Interference with GC.....	4
6 Safety remarks	5
7 Reagents	5
7.1 General.....	5
7.2 Reagents for extraction.....	5
7.3 Reagents for clean-up.....	5
7.3.1 Clean-up A using aluminium oxide.....	5
7.3.2 Clean-up B using silica gel 60 for column chromatography.....	6
7.3.3 Clean-up C using gel permeation chromatography (GPC).....	6
7.3.4 Clean-up D using Florisil®.....	6
7.3.5 Clean-up E using silica H ₂ SO ₄ /silica NaOH.....	7
7.3.6 Clean-up F using benzenesulfonic acid/sulfuric acid.....	7
7.3.7 Clean-up G using TBA sulfite reagent.....	7
7.3.8 Clean-up H using pyrogenic copper.....	7
7.3.9 Clean-up I using silica/silver nitrate.....	8
7.4 Gas chromatographic analysis.....	8
7.5 Standards.....	9
7.5.1 General.....	9
7.5.2 Calibration standards.....	9
7.5.3 Internal and injection standards.....	9
7.6 Preparation of standard solutions.....	10
7.6.1 Preparation of calibration standard solutions of PCBs.....	10
7.6.2 Preparation of internal standard solution.....	11
7.6.3 Preparation of injection standard solution.....	11
8 Apparatus	11
8.1 Extraction and clean-up procedures.....	11
8.2 Gas chromatograph.....	12
9 Sample storage and preservation	12
9.1 Sample storage.....	12
9.2 Sample pre-treatment.....	13
10 Procedure	13
10.1 Blank test.....	13
10.2 Extraction.....	13
10.2.1 General.....	13
10.2.2 Extraction procedure 1: Samples using acetone/petroleum ether or hexane-like solvent and agitation or sonication.....	14
10.2.3 Extraction procedure 2: Samples using Soxhlet or pressurized liquid extraction.....	15
10.2.4 Extraction procedure 3: Samples using acetone/petroleum ether or hexane-like solvent/sodium chloride and agitation.....	15
10.3 Concentration.....	16
10.4 Clean-up of the extract.....	16

10.4.1	General	16
10.4.2	Clean-up A – Aluminium oxide	17
10.4.3	Clean-up B – Silica gel	17
10.4.4	Clean-up C – Gel permeation chromatography	17
10.4.5	Clean-up D – Florisil®	18
10.4.6	Clean-up E – Silica H ₂ SO ₄ /silica NaOH	18
10.4.7	Clean-up F – Benzenesulfonic acid/sulfuric acid	18
10.4.8	Clean-up G – TBA sulfite reagent	19
10.4.9	Clean-up H – Clean-up using pyrogenic copper to remove elemental sulfur and some other organic sulfur compounds	19
10.4.10	Clean up I – AgNO ₃ /silica	19
10.5	Addition of the injection standard	19
10.6	Gas chromatographic analysis (GC)	19
10.6.1	General	19
10.6.2	Setting the gas chromatograph	19
10.7	Mass spectrometry (MS)	20
10.7.1	Mass spectrometric conditions	20
10.7.2	Calibration of the method using an internal standard	20
10.7.3	Measurement	21
10.7.4	Identification	22
10.7.5	Check on method performance	22
10.7.6	Calculation	23
10.8	Electron capture detection (ECD)	24
10.8.1	General	24
10.8.2	ECD conditions	24
10.8.3	Calibration of the method using internal standards	24
10.8.4	Measurement	24
10.8.5	Identification	24
10.8.6	Check on ECD method performance	24
10.8.7	Calculation	25
11	Performance characteristics	25
12	Precision	25
13	Test report	26
Annex A (informative) Repeatability and reproducibility data		27
Annex B (informative) Examples for gas chromatographic conditions and retention times of PCBs		32
Annex C (informative) Calculation method for the estimation of total PCB content		33
Bibliography		40

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by the European Committee for Standardization (CEN) (as EN 17322:2020) and was adopted, without modification other than those given below, by Technical Committee ISO/TC 190, *Soil quality*, Subcommittee SC 3, *Chemical and physical characterization*.

- change of EN ISO 5667-15 reference to ISO 5667-15 reference;
- change of EN ISO 16720 reference to ISO 16720 reference;
- change of EN ISO 22892 reference to ISO 22892 reference;
- change of EN ISO 5667-13 reference to ISO 5667-13 reference;
- change of EN ISO 6468 reference to ISO 6468 reference;
- uniform spelling of sulfate and sulfite;
- editorially revised.

This first edition cancels and replaces ISO 10382:2002 and ISO 13876:2013, which have been technically revised.

The main changes are as follows:

- deletion of OCP analysis (this document specifies methods for quantitative determination of polychlorinated biphenyls);
- addition of GC-MS as a detection method;
- extension of the scope to sludge, sediment, treated biowaste and waste;

- addition of modern extraction techniques and commonly used methods with optimized extraction time, proven clean-up methods and state of the art quantification methods;
- update of normative references.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

This document is a preview generated by EVS

Introduction

Polychlorinated biphenyls (PCB) have been widely used as additives in industrial applications where chemical stability has been required. This stability on the other hand creates environmental problems when PCB are eventually released into the environment. Since some of these PCB compounds are highly toxic, their presence in the environment (air, water, soil, sediment and waste) is regularly monitored and controlled. At present determination of PCB is carried out in these matrices in most of the routine laboratories following the preceding steps for sampling, pre-treatment, extraction and clean-up, by measurement of specific PCB by means of gas chromatography in combination with mass spectrometric detection (GC-MS) or gas chromatography with electron capture detector (GC-ECD).

This document was developed by merging of EN 16167:2018, initially elaborated as a CEN Technical Specification in the European project 'HORIZONTAL' and validated by CEN/TC 400 with the support of BAM, with EN 15308, published by CEN/TC 292.

Considering the different matrices and possible interfering compounds, this document does not contain one single possible way of working. Several choices are possible, in particular relating to clean-up. Detection with both MS-detection and ECD-detection is possible. Two different extraction procedures are described and 9 clean-up procedures. The use of internal and injection standards is described in order to have an internal check on choice of the extraction and clean-up procedure. The method is as far as possible in agreement with the method described for PAH (EN 16181:2018 and EN 15527:2008). It has been tested for ruggedness.

This document is applicable and validated for several types of matrices as indicated in [Table 1](#) (see also [Annex A](#) for the results of the validation).

Table 1 — Matrices for which this document is applicable and validated

Matrix	Materials used for validation
Soil	Sandy soil Mix of soil from the vicinity of Berlin, Germany and PCB-free German reference soil
Sludge	Mix of municipal waste water treatment plant sludge from North Rhine Westphalia, Germany
Biowaste	Mix of compost from the vicinity of Berlin, Germany and sludge from North Rhine Westphalia, Germany
Waste	Contaminated soil, building debris, waste wood, sealant waste, electronic waste, shredder light fraction, cable shredder waste

WARNING — Persons using this document should be familiar with usual laboratory practice. This document does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this document be carried out by suitably trained staff.

Environmental solid matrices — Determination of polychlorinated biphenyls (PCB) by gas chromatography - mass selective detection (GC-MS) or electron-capture detection (GC-ECD)

1 Scope

This document specifies methods for quantitative determination of seven selected polychlorinated biphenyls (PCB28, PCB52, PCB101, PCB118, PCB138, PCB153 and PCB180) in soil, sludge, sediment, treated biowaste, and waste using GC-MS and GC-ECD (see [Table 2](#)).

Table 2 — Target analytes of this document

	Target analyte	CAS-RN ^a
PCB28	2,4,4'-trichlorobiphenyl	7012-37-5
PCB52	2,2',5,5'-tetrachlorobiphenyl	35693-99-3
PCB101	2,2',4,5,5'-pentachlorobiphenyl	37680-73-2
PCB118	2,3',4,4',5-pentachlorobiphenyl	31508-00-6
PCB138	2,2',3,4,4',5'-hexachlorobiphenyl	35065-28-2
PCB153	2,2',4,4',5,5'-hexachlorobiphenyl	35065-27-1
PCB180	2,2',3,4,4',5,5'-heptachlorobiphenyl	35065-29-3

^a CAS-RN Chemical Abstracts Service Registry Number.

The limit of detection depends on the determinants, the equipment used, the quality of chemicals used for the extraction of the sample and the clean-up of the extract.

Under the conditions specified in this document, lower limit of application from 1 µg/kg (expressed as dry matter) for soils, sludge and biowaste to 10 µg/kg (expressed as dry matter) for solid waste can be achieved. For some specific samples the limit of 10 µg/kg cannot be reached.

Sludge, waste and treated biowaste may differ in properties, as well as in the expected contamination levels of PCB and presence of interfering substances. These differences make it impossible to describe one general procedure. This document contains decision tables based on the properties of the sample and the extraction and clean-up procedure to be used.

NOTE The analysis of PCB in insulating liquids, petroleum products, used oils and aqueous samples is referred to in EN 61619, EN 12766-1 and ISO 6468 respectively.

The method can be applied to the analysis of other PCB congeners not specified in the scope, provided suitability is proven by proper in-house validation experiments.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5667-15, *Water quality — Sampling — Part 15: Guidance on the preservation and handling of sludge and sediment samples*

ISO 8466-1, *Water quality — Calibration and evaluation of analytical methods — Part 1: Linear calibration function*

ISO 16720, *Soil quality — Pretreatment of samples by freeze-drying for subsequent analysis*

ISO 18512, *Soil quality — Guidance on long and short term storage of soil samples*

ISO 22892, *Soil quality — Guidelines for the identification of target compounds by gas chromatography and mass spectrometry*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1 critical pair

pair of congeners that shall be separated to a predefined degree (e.g. $R = 0,5$) to ensure chromatographic separation meets minimum quality criteria

$$R = 2 \times \frac{\Delta t}{Y_a + Y_b} \quad (x) \quad (1)$$

where

R resolution

Δt difference in retention times of the two peaks a and b in seconds (s)

Y_a peak width at the base of peak a in seconds (s)

Y_b peak width at the base of peak b in seconds (s)