

**MEDITSIINISEADMETE BIOLOOGILINE HINDAMINE.
OSA 17: MEDITSIINISEADME OSADE TOKSIKOLOOGILISE
RISKI HINDAMINE**

**Biological evaluation of medical devices -
Part 17: Toxicological risk assessment of medical device
constituents (ISO 10993-17:2023 +
ISO 10993-17:2023/Amd 1:2025)**

EESTI STANDARDI EESSÕNA**NATIONAL FOREWORD**

| | |
|---|---|
| See Eesti standard EVS-EN ISO 10993-17:2023+A1:2025 sisaldab Euroopa standardi EN ISO 10993-17:2023 ja selle muudatuse A1:2025 ingliskeelset teksti. | This Estonian standard EVS-EN ISO 10993-17:2023+A1:2025 consists of the English text of the European standard EN ISO 10993-17:2023 and its amendment A1:2025. |
| Standard on jõustunud sellekohase teate avaldamisega EVS Teatajas. Euroopa standardimisorganisatsioonid on teinud Euroopa standardi rahvuslikele liikmetele kättesaadavaks 22.11.2023, muudatus A1 24.12.2025. | This standard has been endorsed with a notification published in the official bulletin of the Estonian Centre for Standardisation and Accreditation. Date of Availability of the European standard is 22.11.2023, for A1 24.12.2025. |
| Muudatusega A1 lisatud või muudetud teksti algus ja lõpp on tekstis tähistatud sümbolitega A1 A1 . | The start and finish of text introduced or altered by amendment A1 is indicated in the text by tags A1 A1 . |
| Standard on kättesaadav Eesti Standardimis- ja Akrediteerimiskeskusest. | The standard is available from the Estonian Centre for Standardisation and Accreditation. |

Tagasisidet standardi sisu kohta on võimalik edastada, kasutades EVS-i veebilehel asuvat tagasiside vormi või saates e-kirja meiliaadressile standardiosakond@evs.ee.

ICS 11.100.20

Standardite ja standardilaadsete dokumentide reprodutseerimise ja levitamise õigus kuulub Eesti Standardimis- ja Akrediteerimiskeskusele

Eesti standardid ja standardilaadsed dokumendid on Eesti Standardimis- ja Akrediteerimiskeskuse intellektuaalomand ning neid kasutatakse litsentsi alusel dokumentide kasutuslepingu tingimuste kohaselt.

Ilma Eesti Standardimis- ja Akrediteerimiskeskuse eelneva kirjaliku loata on keelatud standardite ja standardilaadsete dokumentide täielik või osaline reprodutseerimine, levitamine, muutmise või kasutamine mis tahes kujul ja viisil - sealhulgas kopeerimise, skaneerimise, salvestamise või jagamise teel digiplatvormidel (k.a masinõppe ja tehisintellekti rakendustes). Loata kasutamine väljaspool litsentsi tingimusi käsitletakse õigusrikkumisena.

Kui Teil on küsimusi standardite ja standardilaadsete dokumentide autoriõiguse kaitse kohta, võtke palun ühendust Eesti Standardimis- ja Akrediteerimiskeskusega:

Veebileht www.evs.ee; telefon +372 6055050; e-post info@evs.ee

The right to reproduce and distribute standards and standard-like documents belongs to the Estonian Centre for Standardisation and Accreditation

Estonian standards and standard-like documents are the intellectual property of the Estonian Centre for Standardisation and Accreditation and are made available under license in accordance with the terms and conditions of the document use agreement.

Without the prior written permission of the Estonian Centre for Standardisation and Accreditation, the full or partial reproduction, distribution, modification, or use of standards and standard-like documents in any form or by any means - including photocopying, scanning, storing, or sharing via digital platforms (incl. in machine learning and artificial intelligence applications) - is strictly prohibited. Any unauthorized use beyond the scope of the granted license is prohibited and may result in legal action.

If you have any questions about copyright, please contact Estonian Centre for Standardisation and Accreditation: Homepage www.evs.ee; phone +372 605 5050; e-mail info@evs.ee

EUROPEAN STANDARD

EN ISO 10993-17 + A1

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2023, December 2025

ICS 11.100.20

Supersedes EN ISO 10993-17:2009

English Version

Biological evaluation of medical devices - Part 17:
Toxicological risk assessment of medical device
constituents (ISO 10993-17:2023 + ISO 10993-
17:2023/Amd 1:2025)

Évaluation biologique des dispositifs médicaux - Partie
17: Appréciation du risque toxicologique des
constituants des dispositifs médicaux (ISO 10993-
17:2023 + ISO 10993-17:2023/Amd 1:2025)

Biologische Beurteilung von Medizinprodukten - Teil
17: Toxikologische Risikobewertung von
Medizinproduktbestandteilen (ISO 10993-17:2023 +
ISO 10993-17:2023/Amd 1:2025)

This European Standard was approved by CEN on 2 July 2023. Amendment A1 was approved by CEN on 2 November 2025.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard and its amendment the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard and its Amendment A1 exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels

European foreword

This document (EN ISO 10993-17:2023) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2024, and conflicting national standards shall be withdrawn at the latest by May 2024.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 10993-17:2009.

This document has been prepared under a Standardization Request given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s) / Regulation(s).

For the relationship with EU Directive(s) / Regulation(s), see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10993-17:2023 has been approved by CEN as EN ISO 10993-17:2023 without any modification.

A1 Amendment A1 European foreword

This document (EN ISO 10993-17:2023/A1:2025) has been prepared by Technical Committee ISO/TC 194 "Biological and clinical evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biological and clinical evaluation of medical devices" the secretariat of which is held by DIN.

This Amendment to the European Standard EN ISO 10993-17:2023 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2026, and conflicting national standards shall be withdrawn at the latest by June 2026.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a standardization request addressed to CEN by the European Commission. The Standing Committee of the EFTA States subsequently approves these requests for its Member States.

For the relationship with EU Legislation, see informative Annex ZA, which is an integral part of this document.

Any feedback and questions on this document should be directed to the users' national standards body/national committee. A complete listing of these bodies can be found on the CEN website.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Türkiye and the United Kingdom.

Endorsement notice

The text of ISO 10993-17:2023/Amd 1:2025 has been approved by CEN as EN ISO 10993-17:2023/A1:2025 without any modification. **A1**

This document is a preview generated by EVS

| Contents | Page |
|---|-------------|
| Foreword | v |
| Amendment A1 Foreword | vii |
| Introduction | viii |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 2 |
| 4 Abbreviated terms and symbols | 8 |
| 5 Toxicological risk assessment within the biological evaluation process | 9 |
| 5.1 General..... | 9 |
| 5.1.1 Risk assessment principles..... | 9 |
| 5.1.2 Hazard identification..... | 10 |
| 5.1.3 Risk estimation..... | 10 |
| 5.2 Toxicological risk assessment process..... | 12 |
| 6 Constituent specific toxicological information | 14 |
| 6.1 General..... | 14 |
| 6.2 Identification of hazardous constituents..... | 15 |
| 6.2.1 General..... | 15 |
| 6.2.2 Application of the toxicological screening limit..... | 16 |
| 6.2.3 Identification of human carcinogens or suspected human carcinogens..... | 17 |
| 6.2.4 Selection of the point of departure..... | 17 |
| 7 Tolerable contact level, tolerable intake and threshold of toxicological concern | 18 |
| 7.1 Derivation of TCL and TI..... | 18 |
| 7.2 Application of TTC..... | 19 |
| 8 Exposure dose estimation | 19 |
| 9 Margin of safety | 20 |
| 9.1 General..... | 20 |
| 9.2 Calculating the margin of safety..... | 21 |
| 9.2.1 General..... | 21 |
| 9.2.2 Combining MoS values to address additivity of harm..... | 23 |
| 10 Toxicological risk acceptance criteria | 23 |
| 10.1 General..... | 23 |
| 10.2 Further risk analysis or risk evaluation or risk control..... | 24 |
| 11 Reporting requirements | 24 |
| Annex A (normative) Evaluation of toxicological data quality when selecting a point of departure | 25 |
| Annex B (normative) Derivation of toxicological screening limits | 26 |
| Annex C (normative) Derivation of constituent TI or TCL for select endpoints | 32 |
| Annex D (informative) Typical assumptions for biological parameters | 40 |
| Annex E (normative) Estimation of an exposure dose | 43 |

| | |
|---|-----------|
| Annex F (informative) Reporting of toxicological risk assessment information | 52 |
| Annex ZA (informative) Relationship between this European Standard the General Safety and Performance Requirements of Regulation (EU) 2017/745 aimed to be covered | 58 |
| Bibliography | 63 |

This document is a preview generated by EVS

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO document should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 194, *Biological and clinical evaluation of medical devices*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 206, *Biocompatibility of medical and dental materials and devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 10993-17:2002), which has been technically revised.

The main changes are as follows:

- the title has been changed;
- the scope has been revised and a new statement on its applicability has been added;
- the following terms have been removed: allowable limit, benefit factor, concomitant exposure factor, health benefit, health hazard, health risk, health risk analysis, leachable substance, multiple exposure, physiologically based pharmacokinetic modelling, proportional exposure factor, repeated use, simultaneous use, TCL modifying factor, tolerable exposure, and tolerable risk, utilization factor;

- the following terms have been added: *analogue* (3.1), *benchmark dose low* (3.2), *carcinogen* (3.3), *constituent* (3.4), *dose-response* (3.6), *exposure dose* (3.7), *harmful dose* (3.9), *human carcinogen* (3.10), *identified constituent* (3.11), *irritation* (3.12), *margin of safety* (3.14), *point of departure* (3.19), *release kinetics* (3.20), *slope factor* (3.21), *suspected human carcinogen* (3.22), *systemic toxicity* (3.23), *threshold of toxicological concern* (3.24), *total quantity* (3.27), *toxicological risk*, (3.28), *toxicological risk assessment* (3.29), *toxicological screening limit* (3.30) and *worst-case estimated exposure dose* (3.32);
- the following clauses have been removed: former Clause 4 on the general principles for establishing allowable limits, former Clause 5 on the establishment of tolerable intake for specific leachable substances, former Clause 6 on the calculation of tolerable exposure, former Clause 7 on the feasibility evaluation, former Clause 8 on benefit evaluation, and former Clause 9 on allowable limits;
- the following clauses have been added: Clause 4 on abbreviated terms and symbols, Clause 5 on toxicological risk assessment within the biological evaluation process, Clause 6 on constituent toxicological information, Clause 7 on the tolerable contact level, tolerable intake and the threshold of toxicological concern, Clause 8 on the exposure dose estimation, and Clause 9 on margin of safety;
- former Annex A has been moved to Annex D;
- Annex B and Annex C have been deleted;
- the following annexes have added: Annex A on evaluating toxicological data quality when selecting a POD, Annex B on derivation of toxicological screening limits, Annex C on deriving constituent TI or TCL for select endpoints, Annex E on estimating an exposure dose, and Annex F on reporting toxicological risk assessment information.

A list of all parts in the ISO 10993 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

A1 Amendment A1 Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

ISO draws attention to the possibility that the implementation of this document may involve the use of (a) patent(s). ISO takes no position concerning the evidence, validity or applicability of any claimed patent rights in respect thereof. As of the date of publication of this document, ISO had not received notice of (a) patent(s) which may be required to implement this document. However, implementers are cautioned that this may not represent the latest information, which may be obtained from the patent database available at www.iso.org/patents. ISO shall not be held responsible for identifying any or all such patent rights.

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 194, *Biological and clinical evaluation of medical devices*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 206, *Biological and clinical evaluation of medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

A list of all parts in the ISO 10993 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html. **A1**

Introduction

A medical device or material that has direct or indirect contact with the patient's body or the user's body is expected to perform its intended use while being free from unacceptable risks, including biological and toxicological risks. For this reason, medical devices are typically subject to a biological evaluation within a risk management process to assess their safety. The ISO 10993 series specifies a process through which the manufacturer of a medical device can identify biological hazards associated with the medical device, estimate and evaluate the risks associated with these hazards, control these risks, and monitor the effectiveness of the controls throughout the life cycle of the medical device.

ISO 10993-1, in line with ISO 14971, facilitates a common understanding of biological evaluation within a risk management process. ISO 10993-18 includes methods for identifying and quantifying hazardous medical device constituents so that their toxicological risk can be evaluated. Furthermore, ISO 10993-18 specifies when to consider conducting a toxicological risk assessment per this document.

This document specifies requirements for a toxicological risk assessment process for specific medical device constituents that is used within the biological evaluation process specified by ISO 10993-1 and Clause 1. For example, the biological risk analysis of a medical device includes obtaining constituent information as described in ISO 10993-1:2018, 6.2 and ISO 10993-18. The extent to which constituent information is needed depends on what is known about the material formulation, manufacturing process (i.e. processing aid chemicals, process steps, etc.), what nonclinical or clinical information exist, and on the nature and duration of body contact with the medical device. This toxicological risk assessment process is based on the principle that the biological evaluation and risk assessment process is most efficient and effective when the minimum information necessary is used to assess if exposure to a harmful dose of any medical device constituent can occur. The process, requirements, criteria and methods specified in this document are intended to yield the following information, which is useful in the overall biological risk assessment of the final product:

- whether constituents present in, on or extracted from the medical device are at a quantity that can be a potential source of harm to health;
- derivation of a tolerable intake or tolerable contact level, for a constituent over a specified time period, on the basis of body mass or surface area, that is considered to be without appreciable harm to health;
- a worst-case estimated exposure dose for each constituent and subsequent toxicological risk estimation;
- a toxicological risk estimate based on the tolerable intake or tolerable contact level, and on the worst-case estimated exposure dose for each constituent.

This document is intended for use by toxicologists or other knowledgeable and experienced professionals, appropriately qualified by training and experience, capable of making informed decisions based upon scientific data and a knowledge of medical devices.

Lastly, this latest revision of this document is more extensive than the previous edition as it clarifies when a toxicological risk assessment is recommended, how to calculate the worst-case estimated exposure dose of a constituent and when the probability of occurrence of harm to health should be addressed by other means (e.g. frequency based dose-response (if available), probabilistic dose-response, or biological testing).

Biological evaluation of medical devices —

Part 17:

Toxicological risk assessment of medical device constituents

1 Scope

This document specifies the process and requirements for the toxicological risk assessment of medical device constituents. The methods and criteria used to assess whether exposure to a constituent is without appreciable harm are also specified. The toxicological risk assessment can be part of the biological evaluation of the final product, as described in ISO 10993-1.

The process described in this document applies to chemical characterization information obtained in line with ISO 10993-18. When a toxicological risk assessment of either the compositional information or analytical chemistry data (e.g. extractable data or leachable data) are required to determine whether the toxicological risks related to the constituents are negligible or tolerable.

The process described in this document is not intended to apply to circumstances where the toxicological risk has been estimated by other means, such as:

- constituents, excluding cohort of concern or excluded chemicals, that are present in or extracted from a medical device at an amount representative of patient exposure below a relevant, toxicologically-based reporting threshold (see applicable requirements in ISO 10993-18:2020, Annex E and ISO/TS 21726);
- a new or changed medical device for which chemical or biological equivalence has been established with an existing biocompatible or clinically established medical device (see applicable requirements in ISO 10993-18:2020, Annex C).

The process described in this document is also not applicable to:

- medical device constituents that do not contact the body (e.g. in vitro diagnostics);
- biological risks associated with physical interactions of the medical device with the body (i.e. application of mechanical forces, energy or surface morphology, etc.), provided that the chemical exposure is not changed;
- active pharmaceutical ingredients of device-drug combination products or biologic components of device-biologic combination products as additional regulatory considerations can apply;
- exposure to a particular constituent that arises from sources other than the device, such as food, water or air.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10993-1:2018, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 10993-18:2020 [A1] ¹ [A1], *Biological evaluation of medical devices — Part 18: Chemical characterization of medical device materials within a risk management process*

ISO/TS 21726:2019, *Biological evaluation of medical devices — Application of the threshold of toxicological concern (TTC) for assessing biocompatibility of medical device constituents*

ISO 14971:2019, *Medical devices — Application of risk management to medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 10993-1 and the following apply.

ISO and IEC maintain terminology databases for use in standardization at the following addresses:

- ISO Online browsing platform: available at <https://www.iso.org/obp>
- IEC Electropedia: available at <https://www.electropedia.org/>

3.1

analogue

substance with similar molecular, physical, chemical or toxicological properties

3.2

benchmark dose low

BMD_L

lower one-sided confidence limit of a dose derived from *dose-response* (3.6) modelling that is associated with a specified change (e.g. 5 % or 10 %) in the dose-response relationship

Note 1 to entry: A specified change of 5 % is applied when a reported harm applies to individual animals. A specified change of 10 % is applied when a reported harm applies to a fraction of animals in a population.

[SOURCE: EPA 2012^[2]]

3.3

carcinogen

constituent (3.4) that causes cancer in humans or experimental animals as determined by valid experimental or observational evidence

Note 1 to entry: Carcinogens are either genotoxic carcinogens or non-genotoxic carcinogens. A genotoxic carcinogen is a constituent capable of causing cancer by a mechanism that involves direct alteration of the genetic material of target cells, as a key event at an early stage in tumour development. A non-genotoxic carcinogen is a constituent capable of producing cancer by a mechanism where direct gene damage is not the key event in tumour development (C.3.1).

[SOURCE: International Agency for Research on Cancer^[3]]

3.4

constituent

chemical that is present in or on the finished medical device or its materials of construction

¹ As amended by ISO 10993-18:2020/Amd 1:2022.